






• Labeling



ECG, Respiration and SpO2 Transmitter		
MODEL LX-8300M		IPX5/IPX8
VOLTAGE LR6 x 2		  
EQUIPMENT TYPE	INT	
[SN]		
TYPE LX-8300M	ID FA	US
FCC ID DV8LX8300M		ANSI/AAMI ES60601-1 (2005) + AAMI 1 (2012)
		MEDICAL - PATIENT-MONITORING EQUIPMENT
Fukuda Denshi Co., Ltd.		
3-38-4 Hongo, Bunkyo-ku, Tokyo, Japan		MADE IN JAPAN

The size of the product is small, and it is difficult to secure a space on the product to describe the full text of the description requirement of § 95.2393.

Since the notation on the product is not realistic, the note of the description requirement is described in the user manual.