Appendix A. WIFI6E and PD System Verification of Result

Tissue Verification

The measuring results for tissue simulating liquid are shown as below.

Note:

The dielectric properties of the tissue simulating liquid have been measured within 24 hours before the SAR testing and within ± 10 % of the target values. Liquid temperature during the SAR testing has kept within ± 2 °C.

System Validation

The SAR measurement system was validated according to procedures in KDB 865664 D01. The validation status in tabulated summary is as below.

System Verification

Note:

Comparing to the reference SAR value provided by SPEAG in dipole calibration certificate, the deviation of system check results is within its specification of 10 %. The result indicates the system check can meet the variation criterion and the plots please refer to Appendix AA-1 of this report.

| | System Validation & System Verification | | | | | | | | | | | | | | | | | | | | | | | |
|---|---|--------------------|------------------------|---------------------|----------------------|---------------------------------|----------------------------------|----------------------------------|-----------------------------------|----------------------|--------------------|-------------------|--------------------|----------------|------|---------------|--------------------|------------------------------|------------------------------|--------------------------------|------------------|---------------|--------------|------------|
| Р | lot No. | Frequency (MHz) | Liquid Temp. (℃) | Conductivity (σ) | Permittivity (εr) | Targeted Conductivity (σ) | Targeted Permittivity (ɛr) | Deviation Conductivity (σ) | Deviation Permittivity (εr) | Sensitivity Range | Probe Linearity | Probe Isotropy | Modulation Type | Duty Factor | PAR | Date | Frequency (MHz) | Targeted 1g SAR (W/kg) | Measured 1g SAR (W/kg) | Normalized 1g SAR (W/kg) | Deviation (%) | Dipole S/N | Probe S/N | DAE S/N |
| | S01 | 6500 | 23.3 | 6.184 | 33.72 | 6.07 | 34.5 | 1.88 | -2.26 | PASS | PASS | PASS | OFDM | N/A | PASS | Apr. 09, 2021 | 6500 | 285.00 | 30.7 | 307.00 | 7.72 | 1008 | 7555 | 1589 |

| Date | Frequency (MHz) | Liquid Temp. (°C) | Conductivity (σ) | Permittivity (ε _r) | Conductivity Target (σ) | Permittivity Target (ε _r) | Delta (σ) (%) | Delta (ε _r) (%) | Band | Channel |
|---------------|--------------------|----------------------|---------------------|-----------------------------------|----------------------------|--|------------------|--------------------------------|--------|---------|
| Apr. 09, 2021 | 6025 | 23.3 | 5.588 | 34.389 | 5.509 | 33.870 | 1.42 | 1.44 | UNII-5 | 15 |
| Apr. 09, 2021 | 6185 | 23.3 | 5.679 | 34.366 | 5.698 | 34.878 | -0.37 | -1.53 | UNII-5 | 47 |
| Apr. 09, 2021 | 6345 | 23.3 | 6.195 | 33.925 | 6.001 | 34.566 | 3.25 | -1.95 | UNII-5 | 79 |
| Apr. 09, 2021 | 6505 | 23.3 | 6.171 | 33.718 | 6.075 | 34.494 | 1.50 | -2.27 | UNII-6 | 111 |
| Apr. 09, 2021 | 6665 | 23.3 | 6.327 | 33.080 | 6.261 | 34.302 | 1.07 | -3.56 | UNII-7 | 143 |
| Apr. 09, 2021 | 6825 | 23.3 | 6.561 | 33.145 | 6.447 | 34.110 | 1.72 | -2.80 | UNII-7 | 175 |
| Apr. 09, 2021 | 6985 | 23.3 | 6.592 | 32.745 | 6.632 | 33.918 | -0.57 | -3.41 | UNII-8 | 207 |

| Test Date | Frequency [GHz] | mmWave Probe S/N | Verification Source S/N | Averaging Area [cm ²] | Distance [mm] | Target Power Density [W/m ²] | Measured Power Density [W/m ²] | Deviation [%] | |
|---------------|--------------------|---------------------|----------------------------|---|------------------|--|--|------------------|--|
| Apr. 09, 2021 | 30 | 9454 | 1016 | 4 | 10.0 | 37.0 | 36.7 | -0.81% | |