

Client **C&C Taiwan (Auden)**

CALIBRATION CERTIFICATE



Object(s) **ES3DV2 - SN: 3023**
 Calibration procedure(s) **QA CAL-01 v2
Calibration procedure for dosimetric E-field probes**
 Calibration date: **September 23, 2003**
 Condition of the calibrated item **In Tolerance (according to the specific calibration document)**

This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.

All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.

Calibration Equipment used (M&TE critical for calibration)

| Model Type | ID # | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration |
|-----------------------------------|----------------|---|------------------------|
| Power meter EPM E4419B | GB41293874 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 |
| Power sensor E4412A | MY41495277 | 2-Apr-03 (METAS, No 252-0250) | Apr-04 |
| Reference 20 dB Attenuator | SN: 5086 (20b) | 3-Apr-03 (METAS No. 251-0340) | Apr-04 |
| Fluke Process Calibrator Type 702 | SN: 6295803 | 8-Sep-03 (Sintrel SCS No. E-030020) | Sep-04 |
| Power sensor HP 8481A | MY41092180 | 18-Sep-02 (Agilent, No. 20020918) | In house check: Oct 03 |
| RF generator HP 8684C | US3642U01700 | 4-Aug-99 (SPEAG, in house check Aug-02) | In house check: Aug-05 |
| Network Analyzer HP 8753E | US37390585 | 18-Oct-01 (Agilent, No. 24BR1033101) | In house check: Oct 03 |

| | Name | Function | Signature |
|----------------|---------------|---------------------|---|
| Calibrated by: | Katja Pokovic | Laboratory Director |  |
| Approved by: | Niels Kuster | Quality Manager |  |

Date issued: October 5, 2003

This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.

Probe ES3DV2

SN:3023

Manufactured: April 15, 2003
Last calibration: September 23, 2003

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

DASY - Parameters of Probe: ES3DV2 SN:3023

Sensitivity in Free Space

| | |
|-------|---|
| NormX | 0.85 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormY | 0.94 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormZ | 1.01 $\mu\text{V}/(\text{V}/\text{m})^2$ |

Diode Compression

| | | |
|-------|-----------|----|
| DCP X | 96 | mV |
| DCP Y | 96 | mV |
| DCP Z | 96 | mV |

Sensitivity in Tissue Simulating Liquid

Head **900 MHz** $\epsilon_r = 41.5 \pm 5\%$ $\sigma = 0.97 \pm 5\% \text{ mho/m}$
 Valid for f=800-1000 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

| | | |
|---------|------------------------------|-------------------|
| ConvF X | 6.0 $\pm 9.5\%$ (k=2) | Boundary effect: |
| ConvF Y | 6.0 $\pm 9.5\%$ (k=2) | Alpha 0.33 |
| ConvF Z | 6.0 $\pm 9.5\%$ (k=2) | Depth 1.66 |

Head **1800 MHz** $\epsilon_r = 40.0 \pm 5\%$ $\sigma = 1.40 \pm 5\% \text{ mho/m}$
 Valid for f=1710-1910 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

| | | |
|---------|------------------------------|-------------------|
| ConvF X | 4.9 $\pm 9.5\%$ (k=2) | Boundary effect: |
| ConvF Y | 4.9 $\pm 9.5\%$ (k=2) | Alpha 0.23 |
| ConvF Z | 4.9 $\pm 9.5\%$ (k=2) | Depth 2.54 |

Boundary Effect

Head **900 MHz** Typical SAR gradient: 5 % per mm

| | | | |
|--|--|-------------|-------------|
| Probe Tip to Boundary | | 1 mm | 2 mm |
| SAR _{be} [%] Without Correction Algorithm | | 5.8 | 2.8 |
| SAR _{be} [%] With Correction Algorithm | | 0.1 | 0.3 |

Head **1800 MHz** Typical SAR gradient: 10 % per mm


| | | | |
|--|--|-------------|-------------|
| Probe Tip to Boundary | | 1 mm | 2 mm |
| SAR _{be} [%] Without Correction Algorithm | | 7.7 | 4.7 |
| SAR _{be} [%] With Correction Algorithm | | 0.1 | 0.3 |

Sensor Offset


| | | |
|----------------------------|------------|----|
| Probe Tip to Sensor Center | 2.0 | mm |
|----------------------------|------------|----|

Receiving Pattern (ϕ , $\theta = 0^\circ$)


f = 30 MHz, TEM cell ifi110




f = 100 MHz, TEM cell ifi110





f = 300 MHz, TEM cell ifi110




f = 900 MHz, TEM cell ifi110






Isotropy Error (ϕ), $\theta = 0^\circ$





Frequency Response of E-Field


(TEM-Cell:ifi1110, Waveguide R22)



Dynamic Range f(SAR_{brain}) (Waveguide R22)



Conversion Factor Assessment



Head 900 MHz $\epsilon_r = 41.5 \pm 5\%$ $\sigma = 0.97 \pm 5\%$ mho/m

Valid for f=800-1000 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X


| | | |
|---------|------------------------------|-------------------|
| ConvF X | 6.0 $\pm 9.5\%$ (k=2) | Boundary effect: |
| ConvF Y | 6.0 $\pm 9.5\%$ (k=2) | Alpha 0.33 |
| ConvF Z | 6.0 $\pm 9.5\%$ (k=2) | Depth 1.66 |

Head 1800 MHz $\epsilon_r = 40.0 \pm 5\%$ $\sigma = 1.40 \pm 5\%$ mho/m

Valid for f=1710-1910 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

| | | |
|---------|------------------------------|-------------------|
| ConvF X | 4.9 $\pm 9.5\%$ (k=2) | Boundary effect: |
| ConvF Y | 4.9 $\pm 9.5\%$ (k=2) | Alpha 0.23 |
| ConvF Z | 4.9 $\pm 9.5\%$ (k=2) | Depth 2.54 |

Conversion Factor Assessment



Body **900 MHz** $\epsilon_r = 55.0 \pm 5\%$ $\sigma = 1.05 \pm 5\%$ mho/m

Valid for f=800-1000 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C


| | | |
|---------|------------------|-------------------|
| ConvF X | 6.0 ± 9.5% (k=2) | Boundary effect: |
| ConvF Y | 6.0 ± 9.5% (k=2) | Alpha 0.43 |
| ConvF Z | 6.0 ± 9.5% (k=2) | Depth 1.44 |

Body **1800 MHz** $\epsilon_r = 53.3 \pm 5\%$ $\sigma = 1.52 \pm 5\%$ mho/m

Valid for f=1710-1910 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

| | | |
|---------|------------------|-------------------|
| ConvF X | 4.5 ± 9.5% (k=2) | Boundary effect: |
| ConvF Y | 4.5 ± 9.5% (k=2) | Alpha 0.26 |
| ConvF Z | 4.5 ± 9.5% (k=2) | Depth 2.61 |

Conversion Factor Assessment



Head 2450 MHz $\epsilon_r = 39.2 \pm 5\%$ $\sigma = 1.80 \pm 5\%$ mho/m

Valid for f=2400-2500 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X


| | | | |
|---------|-----------------------------|------------------|-------------|
| ConvF X | 4.4 \pm 9.5% (k=2) | Boundary effect: | |
| ConvF Y | 4.4 \pm 9.5% (k=2) | Alpha | 0.38 |
| ConvF Z | 4.4 \pm 9.5% (k=2) | Depth | 1.66 |

Body 2450 MHz $\epsilon_r = 52.7 \pm 5\%$ $\sigma = 1.95 \pm 5\%$ mho/m

Valid for f=2400-2500 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

| | | | |
|---------|-----------------------------|------------------|-------------|
| ConvF X | 4.1 \pm 9.5% (k=2) | Boundary effect: | |
| ConvF Y | 4.1 \pm 9.5% (k=2) | Alpha | 0.35 |
| ConvF Z | 4.1 \pm 9.5% (k=2) | Depth | 1.94 |

Conversion Factor Assessment



Head 5200 MHz $\epsilon_r = 36.0 \pm 5\%$ $\sigma = 4.66 \pm 5\%$ mho/m

Valid for f=4940-5460 MHz with Head Tissue Simulating Liquid according to OET65-SuppC


| | | | |
|---------|--------------------------------|------------------|-------------|
| ConvF X | 2.70 $\pm 16.6\%$ (k=2) | Boundary effect: | |
| ConvF Y | 2.70 $\pm 16.6\%$ (k=2) | Alpha | 0.75 |
| ConvF Z | 2.70 $\pm 16.6\%$ (k=2) | Depth | 1.45 |

Body 5200 MHz $\epsilon_r = 49.0 \pm 5\%$ $\sigma = 5.30 \pm 5\%$ mho/m

Valid for f=4940-5460 MHz with Body Tissue Simulating Liquid according to OET65-SuppC

| | | | |
|---------|--------------------------------|------------------|-------------|
| ConvF X | 1.82 $\pm 16.6\%$ (k=2) | Boundary effect: | |
| ConvF Y | 1.82 $\pm 16.6\%$ (k=2) | Alpha | 0.90 |
| ConvF Z | 1.82 $\pm 16.6\%$ (k=2) | Depth | 1.70 |

Conversion Factor Assessment



Head 5800 MHz $\epsilon_r = 35.3 \pm 5\%$ $\sigma = 5.27 \pm 5\%$ mho/m

Valid for f=4940-5460 MHz with Head Tissue Simulating Liquid according to OET65-SuppC

| | | | |
|---------|--------------------------------|------------------|-------------|
| ConvF X | 2.40 $\pm 16.6\%$ (k=2) | Boundary effect: | |
| ConvF Y | 2.40 $\pm 16.6\%$ (k=2) | Alpha | 0.89 |
| ConvF Z | 2.40 $\pm 16.6\%$ (k=2) | Depth | 1.30 |

Body 5800 MHz $\epsilon_r = 48.2 \pm 5\%$ $\sigma = 6.0 \pm 5\%$ mho/m

Valid for f=4940-5460 MHz with Body Tissue Simulating Liquid according to OET65-SuppC

| | | | |
|---------|--------------------------------|------------------|-------------|
| ConvF X | 1.50 $\pm 16.6\%$ (k=2) | Boundary effect: | |
| ConvF Y | 1.50 $\pm 16.6\%$ (k=2) | Alpha | 1.01 |
| ConvF Z | 1.50 $\pm 16.6\%$ (k=2) | Depth | 1.85 |

Deviation from Isotropy in HSL

Error (θ, ϕ), $f = 900$ MHz

