

From: drew.rosenberg@itron.com  
Sent: Thursday, September 02, 2004 8:22 AM  
To: info@ckccertification.com  
Subject: Tune-up procedure

There is no tune-up procedure for this device. The frequency and amplitude of the transmitter cannot be changed by the end user. Routine calibration is performed whenever the device is returned for service.