



6700 Washington Avenue S
Eden Prairie, Minnesota 55344
T: 1.800.328.8602

Hear better. Live better.

Starkey.com

30 April 2024

Federal Communications Commission
45 L Street, NE
Washington, DC 20554
Attention: Application Examiner

Reference:	Certification:	Parts 15.209 and 15.247 Intentional Radiator
	Applicant:	Starkey Laboratories, Inc.
	Equipment:	Low Power Device (2400-2483.5 MHz) Low Power Device (10.281 MHz)
	Model #:	Edge AI RIC RT Hearing Aid FCC ID: EOA-24EDGRICRT

Dear Sir/Madam:

Enclosed please find an FCC EOA-24EDGRICRT certification application. The device is a wireless hearing aid intended for use by the hearing impaired. The wireless features are used during the initial fitting of a hearing aid to program the hearing aid and used during normal usage to communicate control information and digital audio to and from any properly configured Bluetooth Low Energy compliant accessory and between a binaural pair of Edge AI hearing aids. They contain a Bluetooth Low Energy radio being certified under Part 15.247 and a Near Frequency Magnetic Induction radio being certified under Part 15.209. Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

A handwritten signature in black ink, appearing to read "William J. Mitchell", written over a light blue horizontal line.

William J. Mitchell, PE
Principal Wireless and Safety Regulatory Engineer
952-918-5683
bill_mitchell@starkey.com