



6700 Washington Avenue S  
Eden Prairie, Minnesota 55344  
T: 1.800.328.8602

Hear better. Live better.

[Starkey.com](http://Starkey.com)

21 November 2023

Federal Communications Commission  
45 L Street, NE  
Washington, DC 20554

Subject: FCC ID labeling on EOA-24GENCICLPO and EOA-24GENCICRPO wireless hearing aids

Re: Response to Inquiry to FCC (Tracking Number 775017)

To whom it may concern,

Consistent to the direction provided on a previous similar product via your July 12, 2010 Response to Inquiry to FCC (Tracking Number 775017) concerning approval by the FCC for alternate labeling to our hearing aid under CFR 2.925(f), we are placing the FCC ID label for the EOA-24GENCICLPO and EOA-24GENCICRPO wireless hearing aids in the User Manual and on the outer package.

Consistent to the direction provided on the earlier product, we are including the following documents with our submission for certification of Starkey's wireless hearing aids having an FCC ID of EOA-24GENCICLPO or EOA-24GENCICRPO.

- The FCC Response to Inquiry to FCC (Tracking Number 775017)
- Supporting documents similar to those associated with that Inquiry
  - Figure 1 below showing curved nature, relative size, and available printed area and FCC ID of EOA-24GENCICLPO or EOA-24GENCICRPO hearing aids.
  - The relevant page from the User Manual showing the alternate FCC ID marking location and other Part 15 required wording.

If you have any questions or comments, please do not hesitate to contact me.

Sincerely,

William J. Mitchell, PE  
Principal Wireless and Safety Regulatory Engineer  
952-918-5683  
[bill\\_mitchell@starkey.com](mailto:bill_mitchell@starkey.com)



Figure 1 (scale in inches and mm)