



Hear better. Live better.

6700 Washington Avenue S  
Eden Prairie, Minnesota 55344  
T: 1.800.328.8602

[Starkey.com](http://Starkey.com)

April 13, 2023

Federal Communications Commission  
45 L Street, NE  
Washington, DC 20554  
Attention: Application Examiner

Reference:	Certification:	Part 15.247 Intentional Radiator
	Applicant:	Starkey Laboratories, Inc.
	Equipment:	Low Power Device (2400-2483.5 MHz)
	Model #:	Genesis AI Custom CIC Hearing Aid
		FCC ID: EOA-24GENCICZN

Dear Sir/Madam:

Enclosed please find an FCC EOA-24GENCICZN certification application. The device is a wireless hearing aid intended for use by the hearing impaired. The wireless features are used during the initial fitting of a hearing aid to program the hearing aid and used during normal usage to communicate control information and digital audio to and from any properly configured Bluetooth Low Energy compliant accessory and between a binaural pair of Genesis hearing aids. Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

A handwritten signature in black ink, appearing to read "William J. Mitchell".

William J. Mitchell, PE  
Principal Wireless and Safety Regulatory Engineer  
952-918-5683  
[bill\\_mitchell@starkey.com](mailto:bill_mitchell@starkey.com)