



6700 Washington Avenue S  
Eden Prairie, Minnesota 55344  
T: 1.800.328.8602

Hear better. Live better.

[Starkey.com](http://Starkey.com)

August 31, 2022

Federal Communications Commission  
45 L Street, NE  
Washington, DC 20554  
Attention: Application Examiner

Reference: Certification: Part 15.247 Intentional Radiator  
Applicant: Starkey Laboratories, Inc.  
Equipment: Low Power Device (2400-2483.5 MHz)  
Model #: Genesis AI ITE R custom hearing aid (right ear)  
FCC ID: EOA-24GENITERCH

Dear Sir/Madam:

Enclosed please find an FCC EOA-24GENITERCH certification application. The device is a wireless hearing aid intended for use by the hearing impaired. The wireless features are used during the initial fitting of a hearing aid to program the hearing aid and used during normal usage to communicate control information and digital audio to and from any properly configured Bluetooth Low Energy compliant accessory and between a binaural pair of Genesis hearing aids. Note that this is the right hearing aid. The corresponding left hearing aid is certified with FCC ID: EOA-24GENITELCH. Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

A handwritten signature in black ink that reads "William J. Mitchell".

William J. Mitchell, PE  
Principal Wireless and Safety Regulatory Engineer  
952-918-5683  
[bill\\_mitchell@starkey.com](mailto:bill_mitchell@starkey.com)