



6600 Washington Avenue S.  
Eden Prairie, Minnesota 55344  
T: 1.800.328.8602

[www.starkey.com](http://www.starkey.com)

September 23, 2019

Federal Communications Commission  
445 12th Street SW  
Washington, DC 20554  
Attention: Application Examiner

Reference: Certification: Parts 15.209 and 15.247 Intentional Radiator  
Applicant: Starkey Laboratories, Inc.  
Equipment: Low Power Device (2400-2483.5 MHz)  
Low Power Device (10.281 MHz)  
Model #: Livio RIC 312 Hearing Aid  
FCC ID: EOA-24LIVIOR312

Dear Sir/Madam:

Enclosed please find an FCC EOA-24LIVIOR312 Class 2 permissive change application. The device is a wireless hearing aid intended for use by the hearing impaired. The wireless features are used during the initial fitting of a hearing aid to program the hearing aid, and used during normal usage to communicate control information and digital audio to and from any properly configured Bluetooth Low Energy compliant accessory and between a binaural pair of Livio hearing aids.

No changes have been made to the device hardware; the change is a software change enabling a new transmission mode with the 2 MB/s Bluetooth Low Energy protocol as defined in version 5.0 of the Bluetooth Core Specification that has a 1 MHz occupied bandwidth.

Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

A handwritten signature in black ink that reads "William J. Mitchell". The signature is written in a cursive style with a large, prominent initial "W".

William J. Mitchell  
Principal Regulatory Affairs Engineer  
952-918-5683  
[bill\\_mitchell@starkey.com](mailto:bill_mitchell@starkey.com)