



6600 Washington Avenue S.
Eden Prairie, Minnesota 55344
T: 1.800.328.8602

www.starkey.com

May 11, 2016

Federal Communications Commission
445 12th Street SW
Washington, DC 20554

Attention: Application Examiner

Reference: Certification: Part 15.247 Intentional Radiator
Applicant: Starkey Laboratories, Inc.
Equipment: Digital Transmission System (902-928 MHz)
Model #: 400
FCC ID: EOA-CM

Dear Sir/Madam:

Enclosed please find an application for a Class II permissive change for Part 15.247 Certification of Starkey Laboratories Model 400. This device is an accessory for wireless hearing aids intended for use by the hearing impaired. The wireless features are used during normal usage to receive digital audio from a built-in microphone and stream it to the wireless hearing aids.

The device has not been changed in any way since its original FCC approval on January 20, 2016. The purpose of this Class II permissive change is add the results of RF Exposure (SAR) evaluation testing, since the device is now considered to be a body worn device subject to portable device RF exposure conditions.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

A handwritten signature in black ink that reads "William J. Mitchell". The signature is written in a cursive style.

William J. Mitchell, PE
Wireless Regulatory Engineer
952-918-5683
bill_mitchell@starkey.com