



6600 Washington Avenue S.  
Eden Prairie, Minnesota 55344  
T: 1.800.328.8602

[www.starkey.com](http://www.starkey.com)

September 2, 2016  
Federal Communications Commission  
445 12<sup>th</sup> Street SW  
Washington, DC 20554

Subject: FCC ID labeling on EOA-EXPSTANDARD wireless hearing aid

Re: Response to Inquiry to FCC (Tracking Number 775017)

To whom it may concern,

Consistent to the direction provided on a previous similar product via your July 12, 2010 Response to Inquiry to FCC (Tracking Number 775017) concerning approval by the FCC for alternate labeling to our hearing aid under CFR 2.925(f), we are placing the FCC ID label for the EOA-EXPSTANDARD wireless hearing aid in the User Manual.

Consistent to the direction provided on the earlier product, we are including the following documents with our submission for Certification of Starkey's wireless hearing aid having an FCC ID of EOA-EXPSTANDARD.

- The FCC Response to Inquiry to FCC (Tracking Number 775017)
- Supporting documents similar to those associated with that Inquiry
- Hearing aid label control drawing showing curved nature, relative size, and available printed area and FCC ID of EOA-EXPSTANDARD hearing aid
- The relevant page from the User Manual showing the alternate FCC ID marking location and other Part 15 required wording.

If you have any questions or comments, please do not hesitate to contact me.

Sincerely,

A handwritten signature in black ink that reads "William J. Mitchell". The signature is written in a cursive, flowing style.

William J. Mitchell, PE  
Wireless Regulatory Engineer  
952-918-5683  
[bill\\_mitchell@starkey.com](mailto:bill_mitchell@starkey.com)