



6600 Washington Avenue S.
Eden Prairie, Minnesota 55344
T: 1.800.328.8602

www.starkey.com

October 12, 2017

Federal Communications Commission
445 12th Street SW
Washington, DC 20554

Attention: Application Examiner
Reference: Certification: Part 15.249 Intentional Radiator
Applicant: Starkey Laboratories, Inc. DBA Starkey Hearing Technologies
Equipment: Low Power Device (902-928 MHz)
Model: Wireless hearing aid
FCC ID: EOA-MUSEPOWER13

Dear Sir/Madam:

Enclosed please find an FCC EOA-MUSEPOWER13 certification application. This device is a wireless hearing aid by use by the hearing impaired. The wireless features are used during the initial fitting of a hearing aid to program the hearing aid, during normal usage to receive control information from a remote control, and during normal usage to receive digital audio from a TV accessory.

This application covers hearing aids with the Muse P+ BTE 13 and Muse iQ P+ BTE 13 model names. Both models are electrically and physically identical to each other, the only difference is the product marketing name.

Further details and supporting documents are attached.

Should you have any questions, please feel free to contact the undersigned.

Sincerely,

A handwritten signature in black ink that reads "William J. Mitchell". The signature is written in a cursive, flowing style.

William J. Mitchell, PE
Wireless Regulatory Engineer
952-918-5683
bill_mitchell@starkey.com