

July 16, 2009

Federal Communications Commission
Office of Engineering and Technology
7435 Oakland Mills Road
Columbia, Maryland 21046

Re: Boston Scientific requests Class II permissive change of a previously certified device


Dear Sir or Madam:

Guidant Corporation, a wholly owned subsidiary of Boston Scientific Corporation doing business as Boston Scientific Cardiac Rhythm Management, located at 4100 Hamline Avenue North, St. Paul, MN 55112-5798, is requesting a Class II permissive change to the previously approved equipment specified by FCC ID: ESCCRM312005.

The antenna has been replaced by a new design which improves bandwidth performance within the assigned ISM band of 902 to 928 MHz. A review of both test reports (original and new design) determined that the new antenna will not degrade the radio performance characteristics previously communicated to and accepted by the Commission under equipment authorization FCC ID: ESCCRM312005.

In support of the Class II change, a radiated emissions test report is included with the application, along with photos of the new antenna design.

Sincerely,



James Kippola
Electrical Engineer
Boston Scientific Corporation
Cardiac Rhythm Management
Tel: (651) 581-3635
Fax: (651) 582-2830
Email: james.kippola@bsci.com



John Rymkiewicz
Fellow Electrical Engineer
Boston Scientific Corporation
Cardiac Rhythm Management
Tel: (651) 582-6869
Fax: (651) 582-2830
Email: john.rymkiewicz@bsci.com