



WTD 13.4.9

FCC Compliance Opinion

Information about the Applicant	
Grantee (Company Name)	Boston Scientific Corporation
FCC Contact Person	James Kippola
Address	4100 Hamline Avenue North
City, State, Zip	St. Paul, MN 55112-5798
Job Number	BSTN0429
Model	Model 6290
FCC ID	ESCCRM628013
Agent	Peter Musto
Approval Type	Original
Equipment Class	DXT Part 15, Low Power Transceiver, Rx Verified
Rule Part	15.249

Overview

Model 6280

LATITUDE™ Communicator is an in-home patient monitor. The Model 6280 Communicator provides radio frequency (RF) telemetry in the 902 – 928 MHz ISM band for use with radio frequency (RF) enabled BSC implantable pulse generators.

Confidentiality requested for allowed files. Request is properly addressed and referenced.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

Signature

A handwritten signature in blue ink that reads "Dave Tolman".

Dave Tolman TCB Committee
9/25/2013