

# FCC Compliance Opinion



Information about the Applicant	
Grantee (Company Name)	Boston Scientific Corporation
FCC Contact Person	James Kippola
Address	4100 Hamline Avenue N.
City, State, Zip	St. Paul, MN 55112-5798
Job Number	BSTN0527
Model	6290
FCC ID	ESCCRM629013
Agent	Peter Musto
Approval Type	Class II Permissive Change
Equipment Class	TNB Licensed Non-broadcast Station Transmitter
Rule Part	95I

## Overview

This application is for a class 2 permissive change to the Boston Scientific Model 6290 Latitude™ Communicator to include firmware updates to add support for Boston Scientific Corporation's Subcutaneous Implantable Cardioverter Defibrillator device. The existing hardware and functionality with other Boston Scientific devices remains unchanged. The model 6290 is a master device which is paired with a patient implant in order to receive information and download instructions. The device performs a Listen Before Talk function prior to initiating a communication event.

Confidentiality requested for allowed files. Request is properly addressed and referenced.

## Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

**Signature**

Dave Tolman, TCB Committee  
2/3/2015