

TÜV AMERICA INC - PRODUCT SERVICE
19333 WILD MOUNTAIN ROAD
TAYLORS FALLS, MINNESOTA 55084-1786

ENGINEERING STATEMENT

For Certification of

Guidant Corporation

Implantable Defibrillator

Model No: CONTAK RENEWAL[®] 3 RF

FCC ID: ESCCRMH21004


FRN: 0009-7001-70

I am an Electromagnetic Compatibility Engineer for TÜV AMERICA INC - PRODUCT SERVICE, Taylors Falls, Minnesota. My education and experience are a matter of record with the Federal Communications Commission.

TÜV AMERICA INC - PRODUCT SERVICE has been authorized by Randy Nuernberg, Vice President of Product Engineering, of Guidant Corporation, to make Certification measurements and act as the authorized representative in the preparation of applications for registration of the above mentioned equipment. TÜV AMERICA INC - PRODUCT SERVICE is on file with the FCC under Paragraph 2.948 of the FCC Rules. The transmitter is to be used in Models H210, H215, H217, H219, H250, H255, H257 and H259 defibrillators – these are identical to the unit tested except some have higher energy defibrillation output or use a different type of battery. The RF portion is used to reprogram the implanted defibrillator and to retrieve data from the implant.

Test data required by the FCC for Certification is included in this report.

It is submitted that the above equipment meets the FCC Part 15 requirements and Certification is hereby requested.



Joel T. Schneider
Senior Engineer
Minnesota Wild River Lab
TÜV America
Product Service
Tel: 651 583 3322
Fax: 651 638 0298
jschneider@tuvam.com
Dated: 06 January 2004