



# element

**Boston Scientific Corporation**

**Bluetooth Low Energy**

**Insertable Cardiac Monitor (ICM)**


**LUX-Dx**

**M301**

**FCC 15.247:2019**

**Bluetooth LE Radio**

**Report # BSTN0975**



NVLAP LAB CODE: 200881-0



*This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the U.S. Government. This Report shall not be reproduced, except in full without written approval of the laboratory.*

# CERTIFICATE OF TEST

**Last Date of Test: November 7, 2019**  
**Boston Scientific Corporation**  
**EUT: Bluetooth Low Energy ICM**

## Radio Equipment Testing

### Standards

Specification	Method
FCC 15.247:2019	ANSI C63.10:2013, KDB 558074


### Results

Method Clause	Test Description	Applied	Results	Comments
6.2	Powerline Conducted Emissions	No	N/A	Not required for a battery powered EUT.
6.5, 6.6, 11.12.1, 11.13.2	Spurious Radiated Emissions	Yes	Pass	
11.6	Duty Cycle	Yes	N/A	
11.8.2	Occupied Bandwidth	Yes	Pass	
11.9.1.1	Output Power	Yes	Pass	
11.9.1.1	Equivalent Isotropic Radiated Power	Yes	Pass	
11.10.2	Power Spectral Density	Yes	Pass	
11.11	Band Edge Compliance	Yes	Pass	
11.11	Spurious Conducted Emissions	Yes	Pass	

### Deviations From Test Standards

None

### Approved By:



Matt Nuernberg, Operations Manager

*Product compliance is the responsibility of the client; therefore, the tests and equipment modes of operation represented in this report were agreed upon by the client, prior to testing. The results of this test pertain only to the sample(s) tested. The specific description is noted in each of the individual sections of the test report supporting this certificate of test. This report reflects only those tests from the referenced standards shown in the certificate of test. It does not include inspection or verification of labels, identification, marking or user information. As indicated in the Statement of Work sent with the quotation, Element's standard process is to always use the latest published version of the test methods even when earlier versions are cited in the test specification. Issuance of a purchase order was de facto acceptance of this approach. Otherwise, the client would have advised Element in writing of the specific version of the test methods they wanted applied to the subject testing.*

# REVISION HISTORY



Revision Number	Description	Date (yyyy-mm-dd)	Page Number
00	None		

# ACCREDITATIONS AND AUTHORIZATIONS



---

## United States

**FCC** - Designated by the FCC as a Telecommunications Certification Body (TCB). Certification chambers, Open Area Test Sites, and conducted measurement facilities are listed with the FCC.

**A2LA** - Accredited by A2LA to ISO / IEC 17065 as a product certifier. This allows Element to certify transmitters to FCC and IC specifications.

**NVLAP** - Each laboratory is accredited by NVLAP to ISO 17025

---

## Canada

**ISED** - Recognized by Innovation, Science and Economic Development Canada as a Certification Body (CB) and as a CAB for the acceptance of test data.

---

## European Union

**European Commission** – Within Element, we have a EU Notified Body validated for the EMCD and RED Directives.

---

## Australia/New Zealand

**ACMA** - Recognized by ACMA as a CAB for the acceptance of test data.

---

## Korea

**MSIT / RRA** - Recognized by KCC's RRA as a CAB for the acceptance of test data.

---

## Japan

**VCCI** - Associate Member of the VCCI. Conducted and radiated measurement facilities are registered.

---

## Taiwan

**BSMI** – Recognized by BSMI as a CAB for the acceptance of test data.

**NCC** - Recognized by NCC as a CAB for the acceptance of test data.

---

## Singapore

**IDA** – Recognized by IDA as a CAB for the acceptance of test data.

---

## Israel

**MOC** – Recognized by MOC as a CAB for the acceptance of test data.

---

## Hong Kong

**OFCA** – Recognized by OFCA as a CAB for the acceptance of test data.

---

## Vietnam

**MIC** – Recognized by MIC as a CAB for the acceptance of test data.

---

## SCOPE


For details on the Scopes of our Accreditations, please visit:

<https://www.nwemc.com/emc-testing-accreditations>

# FACILITIES



<b>California</b> Labs OC01-17 41 Tesla Irvine, CA 92618 (949) 861-8918	<b>Minnesota</b> Labs MN01-10 9349 W Broadway Ave. Brooklyn Park, MN 55445 (612)-638-5136	<b>Oregon</b> Labs EV01-12 6775 NE Evergreen Pkwy #400 Hillsboro, OR 97124 (503) 844-4066	<b>Texas</b> Labs TX01-09 3801 E Plano Pkwy Plano, TX 75074 (469) 304-5255	<b>Washington</b> Labs NC01-05 19201 120 <sup>th</sup> Ave NE Bothell, WA 98011 (425)984-6600
<b>NVLAP</b>				
NVLAP Lab Code: 200676-0	NVLAP Lab Code: 200881-0	NVLAP Lab Code: 200630-0	NVLAP Lab Code:201049-0	NVLAP Lab Code: 200629-0
<b>Innovation, Science and Economic Development Canada</b>				
2834B-1, 2834B-3	2834E-1, 2834E-3	2834D-1	2834G-1	2834F-1
<b>BSMI</b>				
SL2-IN-E-1154R	SL2-IN-E-1152R	SL2-IN-E-1017	SL2-IN-E-1158R	SL2-IN-E-1153R
<b>VCCI</b>				
A-0029	A-0109	A-0108	A-0201	A-0110
<b>Recognized Phase I CAB for ISED, ACMA, BSMI, IDA, KCC/RRA, MIC, MOC, NCC, OFCA</b>				
US0158	US0175	US0017	US0191	US0157



# MEASUREMENT UNCERTAINTY



## Measurement Uncertainty

When a measurement is made, the result will be different from the true or theoretically correct value. The difference is the result of tolerances in the measurement system that cannot be completely eliminated. To the extent that technology allows us, it has been our aim to minimize this error. Measurement uncertainty is a statistical expression of measurement error qualified by a probability distribution.


A measurement uncertainty estimation has been performed for each test per our internal quality document QM205.4.6. The estimation is used to compare the measured result with its "true" or theoretically correct value. The expanded measurement uncertainty (K=2) can be found included as part of the applicable test description page. Our measurement data meets or exceeds the measurement uncertainty requirements of the applicable specification; therefore, the test data can be compared directly to the specification limit to determine compliance. The calculations for estimating measurement uncertainty are based upon ETSI TR 100 028 (or CISPR 16-4-2 as applicable), and are available upon request.

The following table represents the Measurement Uncertainty (MU) budgets for each of the tests that may be contained in this report.


<b>Test</b>	<b>+ MU</b>	<b>- MU</b>
Frequency Accuracy	0.0007%	-0.0007%
Amplitude Accuracy (dB)	1.2 dB	-1.2 dB
Conducted Power (dB)	1.2 dB	-1.2 dB
Radiated Power via Substitution (dB)	0.7 dB	-0.7 dB
Temperature (degrees C)	0.7°C	-0.7°C
Humidity (% RH)	2.5% RH	-2.5% RH
Voltage (AC)	1.0%	-1.0%
Voltage (DC)	0.7%	-0.7%
Field Strength (dB)	5.2 dB	-5.2 dB
AC Powerline Conducted Emissions (dB)	2.4 dB	-2.4 dB

# Test Setup Block Diagrams


## Antenna Port Conducted Measurements



## Near Field Test Fixture Measurements



## Spurious Radiated Emissions



# PRODUCT DESCRIPTION



## Client and Equipment Under Test (EUT) Information

<b>Company Name:</b>	Boston Scientific Corporation
<b>Address:</b>	4100 Hamline Avenue North
<b>City, State, Zip:</b>	St. Paul, MN 55112-5798
<b>Test Requested By:</b>	Daniel Landherr
<b>EUT:</b>	Bluetooth Low Energy ICM
<b>First Date of Test:</b>	October 31, 2019
<b>Last Date of Test:</b>	November 7, 2019
<b>Receipt Date of Samples:</b>	October 31, 2019
<b>Equipment Design Stage:</b>	Production
<b>Equipment Condition:</b>	No Damage
<b>Purchase Authorization:</b>	Verified

## Information Provided by the Party Requesting the Test

<b>Functional Description of the EUT:</b>
Implantable cardiac monitor with BLE
<b>Testing Objective:</b>
To demonstrate compliance of the Bluetooth low energy radio to FCC 15.247 requirements.



# CONFIGURATIONS



## Configuration BSTN0975- 1

Software/Firmware Running during test	
Description	Version
Bootloader	1.0.0.0

EUT			
Description	Manufacturer	Model/Part Number	Serial Number
Bluetooth Low Energy ICM (Hybrid)	Boston Scientific Corporation	M301	93075980

Remote Equipment Outside of Test Setup Boundary			
Description	Manufacturer	Model/Part Number	Serial Number
DC Power Supply	Agilent	U8002A	TPZ
Multimeter	Fluke	114	MMU

Cables					
Cable Type	Shield	Length (m)	Ferrite	Connection 1	Connection 2
DC Leads (x2)	No	0.5 m	No	DC Power Supply	Bluetooth Low Energy ICM (Hybrid)
DC Leads (x2)	No	0.5 m	No	DC Power Supply	Multimeter
AC Cable	No	1.8 m	No	AC Mains	DC Power Supply

## Configuration BSTN0975- 4

EUT			
Description	Manufacturer	Model/Part Number	Serial Number
Bluetooth Low Energy ICM	Boston Scientific Corporation	M301	109681

# MODIFICATIONS



## Equipment Modifications

Item	Date	Test	Modification	Note	Disposition of EUT
1	2019-10-31	Occupied Bandwidth	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Element following the test.
2	2019-10-31	Output Power	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Element following the test.
3	2019-10-31	Equivalent Isotropic Radiated Power	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Element following the test.
4	2019-10-31	Power Spectral Density	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Element following the test.
5	2019-10-31	Band Edge Compliance	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Element following the test.
6	2019-10-31	Spurious Conducted Emissions	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	EUT remained at Element following the test.
7	2019-11-07	Spurious Radiated Emissions	Tested as delivered to Test Station.	No EMI suppression devices were added or modified during this test.	Scheduled testing was completed.

# DUTY CYCLE



## TEST DESCRIPTION

---

The Duty Cycle (x) were measured for each of the EUT operating modes. The measurements were made using a zero span on the spectrum analyzer to see the pulses in the time domain. The transmit power was set to its default maximum. A direct connection was made between the RF output of the EUT and a spectrum analyzer. Attenuation and a DC block were used

The duty cycle was calculated by dividing the transmission pulse duration (T) by the total period of a single on and total off time.

The EUT operates at 100% Duty Cycle.

# SPURIOUS RADIATED EMISSIONS



PSA-ESCI 2019.05.10

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data. The test data represents the configuration / operating mode/ model that produced the highest emission levels as compared to the specification limit.

## MODES OF OPERATION

Tx on Low channel (2402 MHz), Mid channel (2442 MHz), and High channel (2480 MHz); BLE

## POWER SETTINGS INVESTIGATED

3.0 VDC via Battery

## CONFIGURATIONS INVESTIGATED

BSTN0975 - 4

## FREQUENCY RANGE INVESTIGATED

Start Frequency	30 MHz	Stop Frequency	26500 MHz
-----------------	--------	----------------	-----------

## SAMPLE CALCULATIONS

Radiated Emissions: Field Strength = Measured Level + Antenna Factor + Cable Factor - Amplifier Gain + Distance Adjustment Factor + External Attenuation

## TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Interval
Filter - High Pass	Micro-Tronics	HPM50111	LFN	12-Sep-2019	12 mo
Filter - Low Pass	Micro-Tronics	LPM50004	LFK	17-Sep-2019	12 mo
Attenuator	Fairview Microwave	SA18E-20	TWZ	17-Sep-2019	12 mo
Attenuator	Fairview Microwave	SA18E-10	TYA	17-Sep-2019	12 mo
Amplifier - Pre-Amplifier	Miteq	AM-1616-1000	AVO	18-Oct-2019	12 mo
Cable	ESM Cable Corp.	Bilog Cables	MNH	18-Oct-2019	12 mo
Antenna - Biconilog	Teseq	CBL 6141B	AYD	25-Jan-2018	24 mo
Amplifier - Pre-Amplifier	Miteq	JSD4-18002600-26-8P	APU	11-Sep-2019	12 mo
Cable	ESM Cable Corp.	TTBJ141 KMKM-72	MNP	11-Sep-2019	12 mo
Antenna - Standard Gain	ETS Lindgren	3160-09	AHG	NCR	0 mo
Amplifier - Pre-Amplifier	Miteq	AMF-6F-12001800-30-10P	AVW	8-Feb-2019	12 mo
Antenna - Standard Gain	ETS Lindgren	3160-08	AIQ	NCR	0 mo
Amplifier - Pre-Amplifier	Miteq	AMF-6F-08001200-30-10P	AVV	8-Feb-2019	12 mo
Cable	ESM Cable Corp.	Standard Gain Horn Cables	MNJ	8-Mar-2019	12 mo
Antenna - Standard Gain	ETS Lindgren	3160-07	AXP	NCR	0 mo
Analyzer - Spectrum Analyzer	Keysight	N9010A (EXA)	AFQ	13-Dec-2018	12 mo
Amplifier - Pre-Amplifier	Miteq	AMF-3D-00100800-32-13P	AVT	8-Feb-2019	12 mo
Cable	ESM Cable Corp.	Double Ridge Guide Horn Cables	MNI	17-Sep-2019	12 mo
Antenna - Double Ridge	ETS-Lindgren	3115	AJQ	16-Jan-2019	24 mo

## TEST DESCRIPTION

The highest gain antenna of each type to be used with the EUT was tested. The EUT was configured for the required transmit frequencies and the modes as showed in the data sheets.

For each configuration, the spectrum was scanned throughout the specified range as part of the exploratory investigation of the emissions. These "pre-scans" are not included in the report. Final measurements on individual emissions were then made and included in this test report.

The individual emissions from the EUT were maximized by rotating the EUT on a turntable, adjusting the position of the EUT and EUT antenna in three orthogonal axis if required, and adjusting the measurement antenna height and polarization (per ANSI C63.10). A preamp and high pass filter (and notch filter) were used for this test in order to provide sufficient measurement sensitivity.

Measurements were made with the required detectors and annotated on the data for each individual point using the following annotation:

QP = Quasi-Peak Detector  
PK = Peak Detector  
AV = RMS Detector

Measurements were made to satisfy the specific requirements of the test specification for out of band emissions as well as the restricted band requirements.

If there are no detectable emissions above the noise floor, the data included may show noise floor measurements for reference only.

Measurements at the edges of the allowable band may be presented in an alternative method as provided for in the ANSI C63.10 Marker-Delta method. This method involves performing an in-band fundamental measurement followed by a screen capture of the fundamental and out-of-band emission using reduced measurement instrumentation bandwidths. The amplitude delta measured on this screen capture is applied to the fundamental emission value to show the out-of-band emission level as applied to the limit.

Where the radio test software does not provide for a duty cycle at continuous transmit conditions (> 98%) and the RMS (power average) measurements were made across the on and off times of the EUT transmissions, a duty cycle correction is added to the measurements using the formula of  $10 \cdot \text{LOG}(dc)$ .

# SPURIOUS RADIATED EMISSIONS




EmR5 2019.08.15.1

PSA-ESCI 2019.05.10

<b>Work Order:</b>	BSTN0975	<b>Date:</b>	7-Nov-2019	
<b>Project:</b>	None	<b>Temperature:</b>	21.5 °C	
<b>Job Site:</b>	MN05	<b>Humidity:</b>	22.4% RH	
<b>Serial Number:</b>	109681	<b>Barometric Pres.:</b>	1037 mbar	
<b>Tested by:</b>	Andrew Rogstad			
<b>EUT:</b>	Bluetooth Low Energy ICM			
<b>Configuration:</b>	4			
<b>Customer:</b>	Boston Scientific Corporation			
<b>Attendees:</b>	Dan Landherr			
<b>EUT Power:</b>	3.0 VDC via Battery			
<b>Operating Mode:</b>	Tx on Low channel (2402 MHz), Mid channel (2442 MHz), and High channel (2480 MHz); BLE			
<b>Deviations:</b>	None			
<b>Comments:</b>	The EUT used a 100% duty cycle during testing. See data comments for EUT orientation and Tx channel.			

<b>Test Specifications</b>	<b>Test Method</b>
FCC 15.247:2019	ANSI C63.10:2013

<b>Run #</b>	12	<b>Test Distance (m)</b>	3	<b>Antenna Height(s)</b>	1 to 4(m)	<b>Results</b>	Pass
--------------	----	--------------------------	---	--------------------------	-----------	----------------	------



Freq (MHz)	Amplitude (dBuV)	Factor (dB)	Antenna Height (meters)	Azimuth (degrees)	Test Distance (meters)	External Attenuation (dB)	Polarity/Transducer Type	Detector	Distance Adjustment (dB)	Adjusted (dBuV/m)	Spec. Limit (dBuV/m)	Compared to Spec. (dB)	Comments
4883.892	42.0	4.7	3.9	100.0	3.0	0.0	Vert	AV	0.0	46.7	54.0	-7.3	EUT vert, Mid ch.
4883.958	41.0	4.7	1.5	173.0	3.0	0.0	Horz	AV	0.0	45.7	54.0	-8.3	EUT horz, Mid ch.
7326.247	30.6	13.5	3.2	292.0	3.0	0.0	Vert	AV	0.0	44.1	54.0	-9.9	EUT vert, Mid ch.
7325.560	30.5	13.5	3.0	199.9	3.0	0.0	Horz	AV	0.0	44.0	54.0	-10.0	EUT horz, Mid ch.
7440.727	30.7	13.2	1.5	354.0	3.0	0.0	Vert	AV	0.0	43.9	54.0	-10.1	EUT vert, High ch.
7440.817	30.7	13.2	1.5	359.0	3.0	0.0	Horz	AV	0.0	43.9	54.0	-10.1	EUT horz, High ch.
4804.058	39.3	4.6	1.7	164.9	3.0	0.0	Horz	AV	0.0	43.9	54.0	-10.1	EUT horz, Low ch.
4959.858	38.8	4.8	1.0	123.0	3.0	0.0	Horz	AV	0.0	43.6	54.0	-10.4	EUT horz, High ch.
4803.917	38.9	4.6	2.2	38.0	3.0	0.0	Vert	AV	0.0	43.5	54.0	-10.5	EUT vert, Low ch.
4883.875	38.6	4.7	1.1	274.0	3.0	0.0	Vert	AV	0.0	43.3	54.0	-10.7	EUT on side, Mid ch.
4960.017	36.4	4.8	1.5	318.1	3.0	0.0	Vert	AV	0.0	41.2	54.0	-12.8	EUT vert, High ch.
4884.117	35.3	4.7	3.7	88.1	3.0	0.0	Vert	AV	0.0	40.0	54.0	-14.0	EUT horz, Mid ch.
2483.647	33.1	-3.8	1.5	321.0	3.0	10.0	Horz	AV	0.0	39.3	54.0	-14.7	EUT horz, High ch.
2483.507	33.0	-3.8	1.5	232.9	3.0	10.0	Vert	AV	0.0	39.2	54.0	-14.8	EUT vert, High ch.
2483.810	32.8	-3.8	1.5	228.0	3.0	10.0	Horz	AV	0.0	39.0	54.0	-15.0	EUT on side, High ch.
2389.630	32.6	-3.6	2.2	168.9	3.0	10.0	Vert	AV	0.0	39.0	54.0	-15.0	EUT vert, Low ch.
4883.917	34.2	4.7	1.5	171.0	3.0	0.0	Horz	AV	0.0	38.9	54.0	-15.1	EUT on side, Mid ch.

Freq (MHz)	Amplitude (dBuV)	Factor (dB)	Antenna Height (meters)	Azimuth (degrees)	Test Distance (meters)	External Attenuation (dB)	Polarity/ Transducer Type	Detector	Distance Adjustment (dB)	Adjusted (dBuV/m)	Spec. Limit (dBuV/m)	Compared to Spec. (dB)	Comments
2484.297	32.7	-3.8	1.5	252.0	3.0	10.0	Vert	AV	0.0	38.9	54.0	-15.1	EUT horz, High ch.
2483.627	32.7	-3.8	1.5	275.9	3.0	10.0	Vert	AV	0.0	38.9	54.0	-15.1	EUT on side, High ch.
2388.533	32.5	-3.6	1.5	306.1	3.0	10.0	Horz	AV	0.0	38.9	54.0	-15.1	EUT horz, Low ch.
2484.773	32.6	-3.8	1.1	229.9	3.0	10.0	Horz	AV	0.0	38.8	54.0	-15.2	EUT vert, High ch.
4884.067	33.0	4.7	1.5	300.9	3.0	0.0	Horz	AV	0.0	37.7	54.0	-16.3	EUT vert, Mid ch.
7325.993	42.2	13.5	3.2	292.0	3.0	0.0	Vert	PK	0.0	55.7	74.0	-18.3	EUT vert, Mid ch.
7439.073	42.5	13.2	1.5	359.0	3.0	0.0	Horz	PK	0.0	55.7	74.0	-18.3	EUT horz, High ch.
7326.907	42.0	13.5	3.0	199.9	3.0	0.0	Horz	PK	0.0	55.5	74.0	-18.5	EUT horz, Mid ch.
7440.013	42.1	13.2	1.5	354.0	3.0	0.0	Vert	PK	0.0	55.3	74.0	-18.7	EUT vert, High ch.
12400.790	29.0	5.4	1.5	88.0	3.0	0.0	Vert	AV	0.0	34.4	54.0	-19.6	EUT vert, High ch.
12400.670	28.9	5.4	1.5	299.0	3.0	0.0	Horz	AV	0.0	34.3	54.0	-19.7	EUT horz, High ch.
4884.533	48.0	4.7	3.9	100.0	3.0	0.0	Vert	PK	0.0	52.7	74.0	-21.3	EUT vert, Mid ch.
4884.350	47.3	4.7	1.5	173.0	3.0	0.0	Horz	PK	0.0	52.0	74.0	-22.0	EUT horz, Mid ch.
4803.792	46.9	4.6	1.7	164.9	3.0	0.0	Horz	PK	0.0	51.5	74.0	-22.5	EUT horz, Low ch.
4803.758	46.6	4.6	2.2	38.0	3.0	0.0	Vert	PK	0.0	51.2	74.0	-22.8	EUT vert, Low ch.
2389.807	44.6	-3.6	2.2	168.9	3.0	10.0	Vert	PK	0.0	51.0	74.0	-23.0	EUT vert, Low ch.
4883.217	46.2	4.7	1.1	274.0	3.0	0.0	Vert	PK	0.0	50.9	74.0	-23.1	EUT on side, Mid ch.
4959.333	46.0	4.8	1.0	123.0	3.0	0.0	Horz	PK	0.0	50.8	74.0	-23.2	EUT horz, High ch.
2388.300	44.4	-3.6	1.5	306.1	3.0	10.0	Horz	PK	0.0	50.8	74.0	-23.2	EUT horz, Low ch.
2483.537	44.5	-3.8	1.5	321.0	3.0	10.0	Horz	PK	0.0	50.7	74.0	-23.3	EUT horz, High ch.
2483.767	44.2	-3.8	1.5	232.9	3.0	10.0	Vert	PK	0.0	50.4	74.0	-23.6	EUT vert, High ch.
2483.713	44.1	-3.8	1.5	252.0	3.0	10.0	Vert	PK	0.0	50.3	74.0	-23.7	EUT horz, High ch.
2484.897	43.8	-3.8	1.5	275.9	3.0	10.0	Vert	PK	0.0	50.0	74.0	-24.0	EUT on side, High ch.
2483.977	43.8	-3.8	1.1	229.9	3.0	10.0	Horz	PK	0.0	50.0	74.0	-24.0	EUT vert, High ch.
2483.680	43.7	-3.8	1.5	228.0	3.0	10.0	Horz	PK	0.0	49.9	74.0	-24.1	EUT on side, High ch.
4959.475	45.0	4.8	1.5	318.1	3.0	0.0	Vert	PK	0.0	49.8	74.0	-24.2	EUT vert, High ch.
4883.958	44.9	4.7	1.5	171.0	3.0	0.0	Horz	PK	0.0	49.6	74.0	-24.4	EUT on side, Mid ch.
4883.375	44.8	4.7	3.7	88.1	3.0	0.0	Vert	PK	0.0	49.5	74.0	-24.5	EUT horz, Mid ch.
12399.980	29.1	-0.6	1.5	185.9	3.0	0.0	Vert	AV	0.0	28.5	54.0	-25.5	EUT vert, High ch.
4882.733	43.7	4.7	1.5	300.9	3.0	0.0	Horz	PK	0.0	48.4	74.0	-25.6	EUT vert, Mid ch.
12399.530	28.9	-0.6	2.8	218.9	3.0	0.0	Horz	AV	0.0	28.3	54.0	-25.7	EUT horz, High ch.
12208.930	29.6	-1.7	1.5	27.0	3.0	0.0	Horz	AV	0.0	27.9	54.0	-26.1	EUT horz, Mid ch.
12209.260	29.6	-1.7	1.5	357.1	3.0	0.0	Vert	AV	0.0	27.9	54.0	-26.1	EUT vert, Mid ch.
12009.510	29.6	-2.2	3.1	225.9	3.0	0.0	Vert	AV	0.0	27.4	54.0	-26.6	EUT vert, Low ch.
12010.970	29.6	-2.2	2.7	288.1	3.0	0.0	Horz	AV	0.0	27.4	54.0	-26.6	EUT horz, Low ch.
12400.370	40.6	5.4	1.5	299.0	3.0	0.0	Horz	PK	0.0	46.0	74.0	-28.0	EUT horz, High ch.
12400.500	40.5	5.4	1.5	88.0	3.0	0.0	Vert	PK	0.0	45.9	74.0	-28.1	EUT vert, High ch.
12399.680	41.0	-0.6	1.5	185.9	3.0	0.0	Vert	PK	0.0	40.4	74.0	-33.6	EUT vert, High ch.
12399.030	40.5	-0.6	2.8	218.9	3.0	0.0	Horz	PK	0.0	39.9	74.0	-34.1	EUT horz, High ch.
12210.170	41.1	-1.7	1.5	357.1	3.0	0.0	Vert	PK	0.0	39.4	74.0	-34.6	EUT vert, Mid ch.
12207.830	41.1	-1.8	1.5	27.0	3.0	0.0	Horz	PK	0.0	39.3	74.0	-34.7	EUT horz, Mid ch.
12009.560	41.4	-2.2	3.1	225.9	3.0	0.0	Vert	PK	0.0	39.2	74.0	-34.8	EUT vert, Low ch.
12009.120	40.9	-2.2	2.7	288.1	3.0	0.0	Horz	PK	0.0	38.7	74.0	-35.3	EUT horz, Low ch.

# OCCUPIED BANDWIDTH



XMit 2019.09.05

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

## TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Cal. Due
Meter - Multimeter	Fluke	114	MMU	18-Jul-17	18-Jul-20
Power Supply - DC	Agilent	U8002A	TPZ	NCR	NCR
Generator - Signal	Agilent	N5173B	TIW	5-Jul-17	5-Jul-20
Analyzer - Spectrum Analyzer	Keysight	N9010A	AFN	1-May-19	1-May-20
Cable	ESM Cable Corp.	TTBJ141-KMKM-72	MNU	11-Apr-19	11-Apr-20
Block - DC	Fairview Microwave	SD3379	AMI	6-Aug-19	6-Aug-20
Attenuator	S.M. Electronics	SA26B-20	RFW	13-Feb-19	13-Feb-20

## TEST DESCRIPTION

The measurement was made using a direct connection between the RF output of the EUT and a spectrum analyzer. The EUT was set to the channels and modes listed in the datasheet.

The 6dB occupied bandwidth was measured using 100 kHz resolution bandwidth and 300 kHz video bandwidth. The 99.0% occupied bandwidth was also measured at the same time which can be needed during Output Power depending on the applicable method.






# OCCUPIED BANDWIDTH




TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK Low Channel, 2402 MHz						
				Value	Limit	Result
					(≥)	
				754.596 kHz	500 kHz	Pass



BLE/GFSK Mid Channel, 2442 MHz						
				Value	Limit	Result
					(≥)	
				750.87 kHz	500 kHz	Pass




# OCCUPIED BANDWIDTH



TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK High Channel, 2480 MHz				Value	Limit	Result
				(≥)		
				745.883 kHz	500 kHz	Pass



# OUTPUT POWER



XMI 2019.09.05

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

## TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Cal. Due
Meter - Multimeter	Fluke	114	MMU	18-Jul-17	18-Jul-20
Power Supply - DC	Agilent	U8002A	TPZ	NCR	NCR
Generator - Signal	Agilent	N5173B	TIW	5-Jul-17	5-Jul-20
Analyzer - Spectrum Analyzer	Keysight	N9010A	AFN	1-May-19	1-May-20
Cable	ESM Cable Corp.	TTBJ141-KMKM-72	MNU	11-Apr-19	11-Apr-20
Block - DC	Fairview Microwave	SD3379	AMI	6-Aug-19	6-Aug-20
Attenuator	S.M. Electronics	SA26B-20	RFW	13-Feb-19	13-Feb-20

## TEST DESCRIPTION

The measurement was made using a direct connection between the RF output of the EUT and a spectrum analyzer. The transmit frequency was set to the required channels in each band. The transmit power was set to its default maximum.

Prior to measuring peak transmit power the DTS bandwidth (B) was measured.

The method found in ANSI C63.10:2013 Section 11.9.1.1 was used because the RBW on the analyzer was greater than the DTS Bandwidth of the radio.

# OUTPUT POWER



TbTfx 2019.08.30.0 XM 2019.09.05


EUT: Bluetooth Low Energy ICM		Work Order: BSTN0975	
Serial Number: 93075980		Date: 31-Oct-19	
Customer: Boston Scientific Corporation		Temperature: 21.6 °C	
Attendees: Dan Landherr		Humidity: 26.1% RH	
Project: None		Barometric Pres.: 1022 mbar	
Tested by: Andrew Rogstad		Job Site: MN08	
Power: 3.0 VDC			
TEST SPECIFICATIONS		Test Method	
FCC 15.247:2019		ANSI C63.10:2013	
COMMENTS			
Reference level offset includes measurement cable, DC block, and 20 dB attenuator.			
DEVIATIONS FROM TEST STANDARD			
None			
Configuration #	1	Signature <i>Andrew Rogstad</i>	
		Out Pwr (dBm)	Limit (dBm) Result
		-0.874	30 Pass
		-1.142	30 Pass
		-1.32	30 Pass
BLE/GFSK Low Channel, 2402 MHz			
BLE/GFSK Mid Channel, 2442 MHz			
BLE/GFSK High Channel, 2480 MHz			

# OUTPUT POWER




TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK Low Channel, 2402 MHz						
				Out Pwr (dBm)	Limit (dBm)	Result
				-0.874	30	Pass



BLE/GFSK Mid Channel, 2442 MHz						
				Out Pwr (dBm)	Limit (dBm)	Result
				-1.142	30	Pass




# OUTPUT POWER



TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK High Channel, 2480 MHz						
	Out Pwr (dBm)	Limit (dBm)	Result			
	-1.32	30	Pass			



# EQUIVALENT ISOTROPIC RADIATED POWER (EIRP)



XMit 2019.09.05

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

## TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Cal. Due
Meter - Multimeter	Fluke	114	MMU	18-Jul-17	18-Jul-20
Power Supply - DC	Agilent	U8002A	TPZ	NCR	NCR
Generator - Signal	Agilent	N5173B	TIW	5-Jul-17	5-Jul-20
Analyzer - Spectrum Analyzer	Keysight	N9010A	AFN	1-May-19	1-May-20
Cable	ESM Cable Corp.	TTBJ141-KMKM-72	MNU	11-Apr-19	11-Apr-20
Block - DC	Fairview Microwave	SD3379	AMI	6-Aug-19	6-Aug-20
Attenuator	S.M. Electronics	SA26B-20	RFW	13-Feb-19	13-Feb-20

## TEST DESCRIPTION

The measurement was made using a direct connection between the RF output of the EUT and a spectrum analyzer. The peak output power was measured with the EUT set to low, medium and high transmit frequencies. The EUT was transmitting in a no hop mode at the data rate(s) listed in the datasheet.

The method found in ANSI C63.10:2013 Section 7.8.5 was used for a FHSS radio.

The antenna gain was added to the conducted output power value to calculate the EIRP.



# EQUIVALENT ISOTROPIC RADIATED POWER (EIRP)



TstTx 2019.08.30.0 XMt 2019.09.05

EUT: Bluetooth Low Energy ICM		Work Order: BSTN0975				
Serial Number: 93075980		Date: 31-Oct-19				
Customer: Boston Scientific Corporation		Temperature: 21.6 °C				
Attendees: Dan Landherr		Humidity: 25.9% RH				
Project: None		Barometric Pres.: 1022 mbar				
Tested by: Andrew Rogstad		Power: 3.0 VDC				
		Job Site: MN08				
TEST SPECIFICATIONS		Test Method				
FCC 15.247:2019		ANSI C63.10:2013				
COMMENTS						
Reference level offset includes measurement cable, DC block, and 20 dB attenuator.						
DEVIATIONS FROM TEST STANDARD						
None						
Configuration #	1	Signature <i>Andrew Rogstad</i>				
		Out Pwr (dBm)	Antenna Gain (dBi)	EIRP (dBm)	EIRP Limit (dBm)	Result
		-0.874	-2.7	-3.574	36	Pass
		-1.142	-2.7	-3.842	36	Pass
		-1.32	-2.7	-4.02	36	Pass


BLE/GFSK Low Channel, 2402 MHz  
 BLE/GFSK Mid Channel, 2442 MHz  
 BLE/GFSK High Channel, 2480 MHz

# EQUIVALENT ISOTROPIC RADIATED POWER (EIRP)




TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK Low Channel, 2402 MHz						
Out Pwr (dBm)	Antenna Gain (dBi)	EIRP (dBm)	EIRP Limit (dBm)	Result		
-0.874	-2.7	-3.574	36	Pass		



BLE/GFSK Mid Channel, 2442 MHz						
Out Pwr (dBm)	Antenna Gain (dBi)	EIRP (dBm)	EIRP Limit (dBm)	Result		
-1.142	-2.7	-3.842	36	Pass		





# POWER SPECTRAL DENSITY



XMI 2019.09.05

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

## TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Cal. Due
Meter - Multimeter	Fluke	114	MMU	18-Jul-17	18-Jul-20
Power Supply - DC	Agilent	U8002A	TPZ	NCR	NCR
Generator - Signal	Agilent	N5173B	TIW	5-Jul-17	5-Jul-20
Analyzer - Spectrum Analyzer	Keysight	N9010A	AFN	1-May-19	1-May-20
Cable	ESM Cable Corp.	TTBJ141-KMKM-72	MNU	11-Apr-19	11-Apr-20
Block - DC	Fairview Microwave	SD3379	AMI	6-Aug-19	6-Aug-20
Attenuator	S.M. Electronics	SA26B-20	RFW	13-Feb-19	13-Feb-20

## TEST DESCRIPTION

The measurement was made using a direct connection between the RF output of the EUT and a spectrum analyzer. The maximum power spectral density measurements was measured using the channels and modes as called out on the following data sheets.

Per the procedure outlined in ANSI C63.10 the peak power spectral density was measured in a 3 kHz RBW.

# POWER SPECTRAL DENSITY



TbTx 2019.08.30.0 XMI 2019.09.05


EUT: Bluetooth Low Energy ICM		Work Order: BSTN0975	
Serial Number: 93075980		Date: 31-Oct-19	
Customer: Boston Scientific Corporation		Temperature: 21.7 °C	
Attendees: Dan Landherr		Humidity: 26.2% RH	
Project: None		Barometric Pres.: 1022 mbar	
Tested by: Andrew Rogstad		Job Site: MN08	
Power: 3.0 VDC			
TEST SPECIFICATIONS		Test Method	
FCC 15.247:2019		ANSI C63.10:2013	
COMMENTS			
Reference level offset includes measurement cable, DC block, and 20 dB attenuator.			
DEVIATIONS FROM TEST STANDARD			
None			
Configuration #	1	Signature <i>Andrew Rogstad</i>	
		Value dBm/3kHz	Limit < dBm/3kHz
BLE/GFSK Low Channel, 2402 MHz		-17.132	8
BLE/GFSK Mid Channel, 2442 MHz		-17.745	8
BLE/GFSK High Channel, 2480 MHz		-17.115	8
			Results
			Pass
			Pass
			Pass

# POWER SPECTRAL DENSITY




TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK Low Channel, 2402 MHz						
	Value	Limit	Results			
	dBm/3kHz	< dBm/3kHz				
	-17.132	8	Pass			



BLE/GFSK Mid Channel, 2442 MHz						
	Value	Limit	Results			
	dBm/3kHz	< dBm/3kHz				
	-17.745	8	Pass			




# POWER SPECTRAL DENSITY



TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK High Channel, 2480 MHz			
	Value	Limit	Results
	dBm/3kHz	< dBm/3kHz	
	-17.115	8	Pass



# BAND EDGE COMPLIANCE



XMIT 2019.09.05

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

## TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Cal. Due
Meter - Multimeter	Fluke	114	MMU	18-Jul-17	18-Jul-20
Power Supply - DC	Agilent	U8002A	TPZ	NCR	NCR
Generator - Signal	Agilent	N5173B	TIW	5-Jul-17	5-Jul-20
Analyzer - Spectrum Analyzer	Keysight	N9010A	AFN	1-May-19	1-May-20
Cable	ESM Cable Corp.	TTBJ141-KMKM-72	MNU	11-Apr-19	11-Apr-20
Block - DC	Fairview Microwave	SD3379	AMI	6-Aug-19	6-Aug-20
Attenuator	S.M. Electronics	SA26B-20	RFW	13-Feb-19	13-Feb-20

## TEST DESCRIPTION

The measurement was made using a direct connection between the RF output of the EUT and a spectrum analyzer. The spurious RF conducted emissions at the edges of the authorized bands were measured with the EUT set to low and high transmit frequencies in each available band. The channels closest to the band edges were selected. The EUT was transmitting at the data rate(s) listed in the datasheet.

The spectrum was scanned below the lower band edge and above the higher band edge.



# BAND EDGE COMPLIANCE



TbTx 2019.08.30.0 XMi 2019.09.05


EUT: Bluetooth Low Energy ICM		Work Order: BSTN0975	
Serial Number: 93075980		Date: 31-Oct-19	
Customer: Boston Scientific Corporation		Temperature: 21.6 °C	
Attendees: Dan Landherr		Humidity: 25.9% RH	
Project: None		Barometric Pres.: 1022 mbar	
Tested by: Andrew Rogstad		Job Site: MN08	
Power: 3.0 VDC			
TEST SPECIFICATIONS		Test Method	
FCC 15.247:2019		ANSI C63.10:2013	
COMMENTS			
Reference level offset includes measurement cable, DC block, and 20 dB attenuator.			
DEVIATIONS FROM TEST STANDARD			
None			
Configuration #	1	Signature	<i>Andrew Rogstad</i>
		Value (dBc)	Limit ≤ (dBc) Result
		BLE/GFSK Low Channel, 2402 MHz	-47.87 -20 Pass
		BLE/GFSK High Channel, 2480 MHz	-52.56 -20 Pass

# BAND EDGE COMPLIANCE




TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK Low Channel, 2402 MHz						
				Value (dBc)	Limit ≤ (dBc)	Result
				-47.87	-20	Pass



BLE/GFSK High Channel, 2480 MHz						
				Value (dBc)	Limit ≤ (dBc)	Result
				-52.56	-20	Pass



# SPURIOUS CONDUCTED EMISSIONS



XMit 2019.09.05

Testing was performed using the mode(s) of operation and configuration(s) noted within the report. The individuals and/or the organization requesting the test provided the modes, configurations and settings used to complete the evaluation. The actual test parameters are specified in the test data, this includes items such as investigated frequency range (scanned) and test levels. The testing methods and performance specifications, as well as the test site used for the evaluation are indicated in the test data.

## TEST EQUIPMENT

Description	Manufacturer	Model	ID	Last Cal.	Cal. Due
Meter - Multimeter	Fluke	114	MMU	18-Jul-17	18-Jul-20
Power Supply - DC	Agilent	U8002A	TPZ	NCR	NCR
Generator - Signal	Agilent	N5173B	TIW	5-Jul-17	5-Jul-20
Analyzer - Spectrum Analyzer	Keysight	N9010A	AFN	1-May-19	1-May-20
Cable	ESM Cable Corp.	TTBJ141-KMKM-72	MNU	11-Apr-19	11-Apr-20
Block - DC	Fairview Microwave	SD3379	AMI	6-Aug-19	6-Aug-20
Attenuator	S.M. Electronics	SA26B-20	RFW	13-Feb-19	13-Feb-20

## TEST DESCRIPTION

The measurement was made using a direct connection between the RF output of the EUT and a spectrum analyzer. The spurious RF conducted emissions were measured with the EUT set to low, medium and high transmit frequencies. The EUT was transmitting at the data rate(s) listed in the datasheet. For each transmit frequency, the spectrum was scanned throughout the specified frequency range.

# SPURIOUS CONDUCTED EMISSIONS



TstTx 2019.08.30.0 XMI 2019.09.05

EUT: Bluetooth Low Energy ICM		Work Order: BSTN0975	
Serial Number: 93075980		Date: 31-Oct-19	
Customer: Boston Scientific Corporation		Temperature: 21.7 °C	
Attendees: Dan Landherr		Humidity: 26.1% RH	
Project: None		Barometric Pres.: 1022 mbar	
Tested by: Andrew Rogstad		Power: 3.0 VDC	
		Job Site: MN08	
TEST SPECIFICATIONS		Test Method	
FCC 15.247:2019		ANSI C63.10:2013	
COMMENTS			
Reference level offset includes measurement cable, DC block, and 20 dB attenuator.			
DEVIATIONS FROM TEST STANDARD			
None			
Configuration #	1	Signature <i>Andrew Rogstad</i>	


	Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result
BLE/GFSK Low Channel, 2402 MHz	Fundamental	2402.19	N/A	N/A	N/A
BLE/GFSK Low Channel, 2402 MHz	30 MHz - 12.5 GHz	4804.25	-50.13	-20	Pass
BLE/GFSK Low Channel, 2402 MHz	12.5 GHz - 25 GHz	23907.34	-49.48	-20	Pass
BLE/GFSK Mid Channel, 2442 MHz	Fundamental	2442.19	N/A	N/A	N/A
BLE/GFSK Mid Channel, 2442 MHz	30 MHz - 12.5 GHz	4884.94	-50.77	-20	Pass
BLE/GFSK Mid Channel, 2442 MHz	12.5 GHz - 25 GHz	24084.36	-49.32	-20	Pass
BLE/GFSK High Channel, 2480 MHz	Fundamental	2480.2	N/A	N/A	N/A
BLE/GFSK High Channel, 2480 MHz	30 MHz - 12.5 GHz	4961.06	-50.26	-20	Pass
BLE/GFSK High Channel, 2480 MHz	12.5 GHz - 25 GHz	23731.84	-48.81	-20	Pass

# SPURIOUS CONDUCTED EMISSIONS




TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK Low Channel, 2402 MHz					
Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result	
Fundamental	2402.19	N/A	N/A	N/A	



BLE/GFSK Low Channel, 2402 MHz					
Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result	
30 MHz - 12.5 GHz	4804.25	-50.13	-20	Pass	




# SPURIOUS CONDUCTED EMISSIONS




TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK Low Channel, 2402 MHz					
Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result	
12.5 GHz - 25 GHz	23907.34	-49.48	-20	Pass	



BLE/GFSK Mid Channel, 2442 MHz					
Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result	
Fundamental	2442.19	N/A	N/A	N/A	




# SPURIOUS CONDUCTED EMISSIONS




TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK Mid Channel, 2442 MHz				
Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result
30 MHz - 12.5 GHz	4884.94	-50.77	-20	Pass



BLE/GFSK Mid Channel, 2442 MHz				
Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result
12.5 GHz - 25 GHz	24084.36	-49.32	-20	Pass




# SPURIOUS CONDUCTED EMISSIONS




TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK High Channel, 2480 MHz					
Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result	
Fundamental	2480.2	N/A	N/A	N/A	



BLE/GFSK High Channel, 2480 MHz					
Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result	
30 MHz - 12.5 GHz	4961.06	-50.26	-20	Pass	





# SPURIOUS CONDUCTED EMISSIONS



TbTx 2019.08.30.0 XMI 2019.09.05

BLE/GFSK High Channel, 2480 MHz				
Frequency Range	Measured Freq (MHz)	Max Value (dBc)	Limit ≤ (dBc)	Result
12.5 GHz - 25 GHz	23731.84	-48.81	-20	Pass

