

Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/4/2007 5:18:16 PM

# Body\_GSM850 Ch189\_Keypad Up with 1.5cm Gap\_20070204\_EDGE8\_PC528

## DUT: 710211

Communication System: GSM850; Frequency: 836.4 MHz;Duty Cycle: 1:8.3 Medium: MSL\_850 Medium parameters used : f = 836.4 MHz;  $\sigma = 0.972$  mho/m;  $\epsilon_r = 56.2$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 21.6 °C; Liquid Temperature : 21.4 °C


## DASY4 Configuration:

- Probe: ET3DV6 - SN1788; ConvF(6.33, 6.33, 6.33); Calibrated: 9/19/2006

- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)
- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-B; Type: QD 000 P40 C; Serial: TP-1383
- Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch189/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.086 mW/g

Ch189/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 3.79 V/m; Power Drift = -0.132 dB Peak SAR (extrapolated) = 0.113 W/kg SAR(1 g) = 0.081 mW/g; SAR(10 g) = 0.058 mW/g Maximum value of SAR (measured) = 0.086 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/4/2007 5:49:37 PM

# Body\_GSM850 Ch189\_Keypad Up with 1.5cm Gap\_20070204\_EDGE10\_PC528

### DUT: 710211


Communication System: GSM850; Frequency: 836.4 MHz;Duty Cycle: 1:4 Medium: MSL\_850 Medium parameters used : f = 836.4 MHz;  $\sigma = 0.972$  mho/m;  $\epsilon_r = 56.2$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 21.8 °C; Liquid Temperature : 21.4 °C

#### DASY4 Configuration:

- Probe: ET3DV6 SN1788; ConvF(6.33, 6.33, 6.33); Calibrated: 9/19/2006
- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)
- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-B; Type: QD 000 P40 C; Serial: TP-1383
- Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch189/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.096 mW/g

Ch189/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 4.23 V/m; Power Drift = -0.172 dB Peak SAR (extrapolated) = 0.125 W/kg SAR(1 g) = 0.093 mW/g; SAR(10 g) = 0.067 mW/g Maximum value of SAR (measured) = 0.099 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/4/2007 6:11:41 PM

# Body\_GSM850 Ch189\_Keypad Up with 1.5cm Gap\_20070204\_EDGE12\_PC528

## DUT: 710211


Communication System: GSM850; Frequency: 836.4 MHz;Duty Cycle: 1:2 Medium: MSL\_850 Medium parameters used : f = 836.4 MHz;  $\sigma = 0.972$  mho/m;  $\epsilon_r = 56.2$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 21.6 °C; Liquid Temperature : 21.4 °C

DASY4 Configuration:

- Probe: ET3DV6 SN1788; ConvF(6.33, 6.33, 6.33); Calibrated: 9/19/2006
- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)
- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-B; Type: QD 000 P40 C; Serial: TP-1383
- Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch189/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.069 mW/g

Ch189/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 3.49 V/m; Power Drift = -0.132 dB Peak SAR (extrapolated) = 0.088 W/kg SAR(1 g) = 0.068 mW/g; SAR(10 g) = 0.049 mW/g Maximum value of SAR (measured) = 0.072 mW/g





Date/Time: 2/5/2007 11:57:01 AM

# Body PCS Ch661 Keypad Up with 1.5cm Gap 20070205 GPRS8 PC528

DUT: 710211

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3 Medium: MSL\_1900 Medium parameters used: f = 1880 MHz;  $\sigma = 1.52 \text{ mho/m}$ ;  $\epsilon_r = 53$ ;  $\rho = 1000 \text{ kg/m}^3$ Ambient Temperature : 21.5 °C; Liquid Temperature : 18.9 °C


DASY4 Configuration: - Probe: ET3DV6 - SN1788; ConvF(4.67, 4.67, 4.67); Calibrated: 9/19/2006

- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch661/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.132 mW/g

Ch661/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 7.23 V/m; Power Drift = -0.069 dB Peak SAR (extrapolated) = 0.175 W/kg SAR(1 g) = 0.121 mW/g; SAR(10 g) = 0.079 mW/g Maximum value of SAR (measured) = 0.128 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/5/2007 5:31:33 PM

# Body PCS Ch512 Keypad Up with 1.5cm Gap 20070205 GPRS10 PC528

DUT: 710211

Communication System: PCS 1900; Frequency: 1850.2 MHz; Duty Cycle: 1:4 Medium: MSL\_1900 Medium parameters used : f = 1850.2 MHz;  $\sigma = 1.48 \text{ mho/m}$ ;  $\epsilon_r = 53.2$ ;  $\rho = 1000 \text{ kg/m}^3$ Ambient Temperature : 21.8 °C; Liquid Temperature : 18.9 °C

DASY4 Configuration:

- Probe: ET3DV6 - SN1788; ConvF(4.67, 4.67, 4.67); Calibrated: 9/19/2006


- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch512/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.173 mW/g

Ch512/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 8.40 V/m; Power Drift = -0.046 dB Peak SAR (extrapolated) = 0.233 W/kg SAR(1 g) = 0.161 mW/g; SAR(10 g) = 0.106 mW/g Maximum value of SAR (measured) = 0.172 mW/g

Ch512/Zoom Scan (5x5x7)/Cube 1: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 8.40 V/m; Power Drift = -0.046 dB Peak SAR (extrapolated) = 0.230 W/kg SAR(1 g) = 0.153 mW/g; SAR(10 g) = 0.101 mW/g Maximum value of SAR (measured) = 0.167 mW/g





Date/Time: 2/5/2007 1:27:57 PM

# Body PCS Ch661 Keypad Up with 1.5cm Gap 20070205 GPRS12 PC528

DUT: 710211

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:2 Medium: MSL\_1900 Medium parameters used: f = 1880 MHz;  $\sigma = 1.52 \text{ mho/m}$ ;  $\epsilon_r = 53$ ;  $\rho = 1000 \text{ kg/m}^3$ Ambient Temperature : 21.8 °C; Liquid Temperature : 18.9 °C


DASY4 Configuration: - Probe: ET3DV6 - SN1788; ConvF(4.67, 4.67, 4.67); Calibrated: 9/19/2006

- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch661/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.130 mW/g

Ch661/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 7.00 V/m; Power Drift = -0.042 dB Peak SAR (extrapolated) = 0.176 W/kg SAR(1 g) = 0.120 mW/g; SAR(10 g) = 0.079 mW/g Maximum value of SAR (measured) = 0.129 mW/g





Date/Time: 2/5/2007 3:37:01 PM

# Body PCS Ch661 Keypad Up with 1.5cm Gap 20070205 EDGE8 PC528

DUT: 710211

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3 Medium: MSL\_1900 Medium parameters used: f = 1880 MHz;  $\sigma = 1.52 \text{ mho/m}$ ;  $\epsilon_r = 53$ ;  $\rho = 1000 \text{ kg/m}^3$ Ambient Temperature : 21.7 °C; Liquid Temperature : 18.9 °C

DASY4 Configuration:


- Probe: ET3DV6 - SN1788; ConvF(4.67, 4.67, 4.67); Calibrated: 9/19/2006

- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch661/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.039 mW/g

Ch661/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 3.80 V/m; Power Drift = -0.152 dB Peak SAR (extrapolated) = 0.052 W/kg SAR(1 g) = 0.036 mW/g; SAR(10 g) = 0.023 mW/g Maximum value of SAR (measured) = 0.038 mW/g





Date/Time: 2/5/2007 3:03:49 PM

# Body PCS Ch661 Keypad Up with 1.5cm Gap 20070205 EDGE10 PC528

DUT: 710211

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:4 Medium: MSL\_1900 Medium parameters used: f = 1880 MHz;  $\sigma = 1.52 \text{ mho/m}$ ;  $\epsilon_r = 53$ ;  $\rho = 1000 \text{ kg/m}^3$ Ambient Temperature : 21.6 °C; Liquid Temperature : 18.9 °C

DASY4 Configuration:

- Probe: ET3DV6 - SN1788; ConvF(4.67, 4.67, 4.67); Calibrated: 9/19/2006


- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch661/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.046 mW/g

Ch661/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 4.16 V/m; Power Drift = -0.150 dB Peak SAR (extrapolated) = 0.065 W/kg SAR(1 g) = 0.043 mW/g; SAR(10 g) = 0.028 mW/g Maximum value of SAR (measured) = 0.047 mW/g

Ch661/Zoom Scan (5x5x7)/Cube 1: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 4.16 V/m; Power Drift = -0.150 dB Peak SAR (extrapolated) = 0.100 W/kg SAR(1 g) = 0.040 mW/g; SAR(10 g) = 0.025 mW/g Maximum value of SAR (measured) = 0.050 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/5/2007 4:06:24 PM

# Body PCS Ch661 Keypad Up with 1.5cm Gap 20070205 EDGE12 PC528

DUT: 710211

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:2 Medium: MSL\_1900 Medium parameters used: f = 1880 MHz;  $\sigma = 1.52 \text{ mho/m}$ ;  $\epsilon_r = 53$ ;  $\rho = 1000 \text{ kg/m}^3$ Ambient Temperature : 21.7 °C; Liquid Temperature : 18.9 °C


DASY4 Configuration: - Probe: ET3DV6 - SN1788; ConvF(4.67, 4.67, 4.67); Calibrated: 9/19/2006

- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch661/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.039 mW/g

Ch661/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 3.74 V/m; Power Drift = -0.110 dB Peak SAR (extrapolated) = 0.048 W/kg SAR(1 g) = 0.033 mW/g; SAR(10 g) = 0.022 mW/g Maximum value of SAR (measured) = 0.036 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/4/2007 3:30:58 PM

# Body GSM850 Ch189 Keypad Down with 1.5cm Gap 20070204 GPRS8 PC528

DUT: 710211

Communication System: GSM850; Frequency: 836.4 MHz; Duty Cycle: 1:8.3 Medium: MSL\_850 Medium parameters used : f = 836.4 MHz;  $\sigma = 0.972$  mho/m;  $\epsilon_r = 56.2$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 21.1 °C; Liquid Temperature : 21.4 °C


DASY4 Configuration: - Probe: ET3DV6 - SN1788; ConvF(6.33, 6.33, 6.33); Calibrated: 9/19/2006

- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-B; Type: QD 000 P40 C; Serial: TP-1383
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch189/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.262 mW/g

Ch189/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 7.69 V/m; Power Drift = -0.156 dB Peak SAR (extrapolated) = 0.318 W/kg SAR(1 g) = 0.246 mW/g; SAR(10 g) = 0.179 mW/g Maximum value of SAR (measured) = 0.260 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/5/2007 11:02:10 AM

# Body PCS Ch661 Keypad Down with 1.5cm Gap 20070205 GPRS8 PC528

DUT: 710211

Communication System: PCS 1900; Frequency: 1880 MHz; Duty Cycle: 1:8.3 Medium: MSL\_1900 Medium parameters used: f = 1880 MHz;  $\sigma = 1.52 \text{ mho/m}$ ;  $\epsilon_r = 53$ ;  $\rho = 1000 \text{ kg/m}^3$ Ambient Temperature : 21.4 °C; Liquid Temperature : 18.9 °C

DASY4 Configuration:

- Probe: ET3DV6 - SN1788; ConvF(4.67, 4.67, 4.67); Calibrated: 9/19/2006


- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch661/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.079 mW/g

Ch661/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 3.08 V/m; Power Drift = -0.062 dB Peak SAR (extrapolated) = 0.118 W/kg SAR(1 g) = 0.072 mW/g; SAR(10 g) = 0.043 mW/g Maximum value of SAR (measured) = 0.078 mW/g

Ch661/Zoom Scan (5x5x7)/Cube 1: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 3.08 V/m; Power Drift = -0.062 dB Peak SAR (extrapolated) = 0.094 W/kg SAR(1 g) = 0.063 mW/g; SAR(10 g) = 0.041 mW/gMaximum value of SAR (measured) = 0.068 mW/g





Date/Time: 2/4/2007 7:20:15 PM

# Body GSM850 Ch251 Keypad Up with 1.5cm Gap 20070204 GPRS10 BT On PC528

DUT: 710211

Communication System: GSM850; Frequency: 848.8 MHz; Duty Cycle: 1:4 Medium: MSL\_850 Medium parameters used: f = 849 MHz;  $\sigma = 0.984$  mho/m;  $\epsilon_r = 56.1$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 21.9 °C; Liquid Temperature : 21.4 °C


DASY4 Configuration: - Probe: ET3DV6 - SN1788; ConvF(6.33, 6.33, 6.33); Calibrated: 9/19/2006

- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-B; Type: QD 000 P40 C; Serial: TP-1383
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch251/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.435 mW/g

Ch251/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 8.72 V/m; Power Drift = -0.027 dB Peak SAR (extrapolated) = 0.519 W/kg SAR(1 g) = 0.409 mW/g; SAR(10 g) = 0.299 mW/g Maximum value of SAR (measured) = 0.435 mW/g





Date/Time: 2/5/2007 6:59:36 PM

# Body PCS Ch512 Keypad Up with 1.5cm Gap 20070205 GPRS10 Bluetooth On PC528

DUT: 710211

Communication System: PCS 1900; Frequency: 1850.2 MHz; Duty Cycle: 1:4 Medium: MSL 1900 Medium parameters used : f = 1850.2 MHz;  $\sigma = 1.48$  mho/m;  $\varepsilon_c = 53.2$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 21.6 °C; Liquid Temperature : 18.9 °C

DASY4 Configuration:

- Probe: ET3DV6 - SN1788; ConvF(4.67, 4.67, 4.67); Calibrated: 9/19/2006


- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch512/Area Scan (71x191x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.161 mW/g

Ch512/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 8.38 V/m; Power Drift = -0.019 dB Peak SAR (extrapolated) = 0.218 W/kg SAR(1 g) = 0.151 mW/g; SAR(10 g) = 0.100 mW/g Maximum value of SAR (measured) = 0.162 mW/g

Ch512/Zoom Scan (5x5x7)/Cube 1: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 8.38 V/m; Power Drift = -0.019 dB Peak SAR (extrapolated) = 0.136 W/kg SAR(1 g) = 0.096 mW/g; SAR(10 g) = 0.064 mW/g Maximum value of SAR (measured) = 0.102 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/4/2007 8:00:53 PM

# Body GSM850 Ch251 Keypad Up with 1.5cm Gap 20070204 GPRS10 PC529

DUT: 710211

Communication System: GSM850; Frequency: 848.8 MHz; Duty Cycle: 1:4 Medium: MSL\_850 Medium parameters used: f = 849 MHz;  $\sigma = 0.984$  mho/m;  $\epsilon_r = 56.1$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 21.9 °C; Liquid Temperature : 21.4 °C


DASY4 Configuration: - Probe: ET3DV6 - SN1788; ConvF(6.33, 6.33, 6.33); Calibrated: 9/19/2006

- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-B; Type: QD 000 P40 C; Serial: TP-1383
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch251/Area Scan (71x171x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.519 mW/g

Ch251/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 11.9 V/m; Power Drift = -0.098 dB Peak SAR (extrapolated) = 0.647 W/kg SAR(1 g) = 0.489 mW/g; SAR(10 g) = 0.348 mW/g Maximum value of SAR (measured) = 0.518 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/5/2007 7:56:43 PM

# Body PCS Ch512 Keypad Up with 1.5cm Gap GPRS10 PC529

DUT: 710211

Communication System: PCS 1900; Frequency: 1850.2 MHz; Duty Cycle: 1:4 Medium: MSL\_1900 Medium parameters used : f = 1850.2 MHz;  $\sigma = 1.48 \text{ mho/m}$ ;  $\epsilon_r = 53.2$ ;  $\rho = 1000 \text{ kg/m}^3$ Ambient Temperature : 21.5 °C; Liquid Temperature : 18.9 °C

DASY4 Configuration:

- Probe: ET3DV6 - SN1788; ConvF(4.67, 4.67, 4.67); Calibrated: 9/19/2006


- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)

- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
   Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch512/Area Scan (71x171x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.182 mW/g

Ch512/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 8.93 V/m; Power Drift = -0.119 dB Peak SAR (extrapolated) = 0.256 W/kg SAR(1 g) = 0.171 mW/g; SAR(10 g) = 0.114 mW/g Maximum value of SAR (measured) = 0.183 mW/g

Ch512/Zoom Scan (5x5x7)/Cube 1: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 8.93 V/m; Power Drift = -0.119 dB Peak SAR (extrapolated) = 0.160 W/kg SAR(1 g) = 0.113 mW/g; SAR(10 g) = 0.076 mW/g Maximum value of SAR (measured) = 0.120 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/4/2007 11:20:15 AM

# Left Tilted\_GSM850 Ch251\_20070204\_Bluetooth On\_PC529\_2D

## DUT: 710211

Communication System: GSM850; Frequency: 848.8 MHz;Duty Cycle: 1:8.3 Medium: HSL\_850 Medium parameters used: f = 849 MHz;  $\sigma = 0.912$  mho/m;  $e_r = 42.9$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 22.4 °C; Liquid Temperature : 20.9 °C


DASY4 Configuration:

- Probe: ET3DV6 - SN1788; ConvF(6.6, 6.6, 6.6); Calibrated: 9/19/2006

- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)
- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-B; Type: QD 000 P40 C; Serial: TP-1383
- Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch251/Area Scan (71x161x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 1.25 mW/g

Ch251/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 28.3 V/m; Power Drift = -0.184 dB Peak SAR (extrapolated) = 1.63 W/kg SAR(1 g) = 1.15 mW/g; SAR(10 g) = 0.775 mW/g Maximum value of SAR (measured) = 1.25 mW/g





Test Laboratory: Sporton International Inc. SAR Testing Lab

Date/Time: 2/4/2007 2:42:50 PM

Left Tilted\_PCS Ch512\_20070204\_Bluetooth On\_PC529\_2D

DUT: 710211


Communication System: PCS; Frequency: 1850.2 MHz;Duty Cycle: 1:8.3 Medium: HSL\_1900 Medium parameters used : f = 1850.2 MHz;  $\sigma = 1.43$  mho/m;  $\varepsilon_r = 41.6$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 22.6 °C; Liquid Temperature : 20.3 °C

# DASY4 Configuration:

- Probe: ET3DV6 SN1788; ConvF(5.3, 5.3, 5.3); Calibrated: 9/19/2006
- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)
- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-A; Type: QD 000 P40 C; Serial: TP-1303
- Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch512/Area Scan (71x161x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.696 mW/g

Ch512/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 16.9 V/m; Power Drift = -0.149 dB Peak SAR (extrapolated) = 1.05 W/kg SAR(1 g) = 0.632 mW/g; SAR(10 g) = 0.368 mW/g Maximum value of SAR (measured) = 0.690 mW/g





Date/Time: 2/4/2007 8:00:53 PM

# Body\_GSM850 Ch251\_Keypad Up with 1.5cm Gap\_20070204\_GPRS10\_PC529\_2D

#### DUT: 710211


Communication System: GSM850; Frequency: 848.8 MHz;Duty Cycle: 1:4 Medium: MSL\_850 Medium parameters used: f = 849 MHz;  $\sigma = 0.984$  mho/m;  $\epsilon_r = 56.1$ ;  $\rho = 1000$  kg/m<sup>3</sup> Ambient Temperature : 21.9 °C; Liquid Temperature : 21.4 °C

#### DASY4 Configuration:


- Probe: ET3DV6 SN1788; ConvF(6.33, 6.33, 6.33); Calibrated: 9/19/2006
- Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)
- Electronics: DAE3 Sn577; Calibrated: 11/21/2006
- Phantom: SAM-B; Type: QD 000 P40 C; Serial: TP-1383
- Measurement SW: DASY4, V4.7 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 171

Ch251/Area Scan (71x171x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 0.519 mW/g

Ch251/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 11.9 V/m; Power Drift = -0.098 dB Peak SAR (extrapolated) = 0.647 W/kg SAR(1 g) = 0.489 mW/g; SAR(10 g) = 0.348 mW/g Maximum value of SAR (measured) = 0.518 mW/g









Schmid & Partner Engineering AG <sub>Zeughausstrasse</sub> 43, 8004 Zuric	y of h, Switzerland		chweizerischer Kalibrierdien: ervice suisse d'étalonnage ervízio svizzero di taratura wiss Calibration Service
Accredited by the Swiss Federal C The Swiss Accreditation Service Multilateral Agreement for the re	e is one of the signatorie	s to the EA	.: SCS 108
Client Sporton (Aude	n)	Certificate No: D	835V2-499_Mar06
CALIBRATION C	CERTIFICATE		
Object	D835V2 - SN: 49	9	
Calibration procedure(s)	QA CAL-05.v6 Calibration proce	dure for dipole validation kits	
Calibration date	March 15, 2006		
Condition of the calibrated item	In Tolerance		
The measurements and the unce	rtainties with confidence p	onal standards, which realize the physical units of robability are given on the following pages and arr by facility: environment temperature $(22 \pm 3)^{\circ}$ C and	e part of the certificate.
The measurements and the unce	rtainties with confidence p ded in the closed laborato	robability are given on the following pages and are	e part of the certificate.
The measurements and the unce All calibrations have been conduct	rtainties with confidence p ded in the closed laborato	robability are given on the following pages and are	e part of the certificate.
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&1 Primary Standards Power meter EPM-442A	rtainties with confidence p ded in the closed laborato FE official for calibration) ID # GB37480704	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&1 Primary Standards Power meter EPM-442A Power sensor HP 8481A	tainties with confidence p ated in the closed laborato FE official for calibration) 10 # GB37480704 US37292783	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator	tainties with confidence p ted in the closed laborato IE offical for calibration) ID # GB37480704 US37292783 SN: 5086 (20g)	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No. 251-00498)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&1 Primary Standards Power meter EPM-442A Power sensor HP 8481A	rtainties with confidence p sted in the closed laborato FE official for calibration) 10 # GB37480704 US37292783	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8-481A Reference 20 dB Attenuator Reference 10 dB Attenuator	rtainties with confidence p ded in the closed laborato IE oftical for calibration) ID # GB37480704 US37292783 SN: 5066 (20g) SN: 5047.2 (10r)	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6	rtainties with confidence p ted in the closed laborato TE offical for calibration) ID # GB37480704 US37292783 SN: 5066 (20g) SN: 5047.2 (10r) SN 1507	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 28-Oct-05 (SPEAG, No. ET3-1507_Oct05)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06 Oct-08
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A	rtainties with confidence p ted in the closed laborato TE official for calibration) ID # GB37480704 US37292783 SN: 5066 (20g) SN: 5067 (20g) SN: 5067 (20g) SN: 5047.2 (10r) SN 1507 SN 601 ID # MY41092317	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 12-Oct-05 (SPEAG, No. ET3-1507, Oct05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (In house) 18-Oct-02 (SPEAG, In house check Oct-05)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check. Oct-07
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator Agilent E4421B	rtainties with confidence p ted in the closed laborato IE critical for calibration) ID # GB37480704 US37292783 SN: 5066 (20g) SN: 5066 (20g) SN: 5066 (20g) SN: 5066 (20g) SN: 5066 (20g) SN: 5066 (20g) SN: 5067 2 (10r) SN: 507 SN 601 ID # MY41092317 MY41000675	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Cemificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 28-Oct-05 (SPEAG, No. ET3-1507, Oct05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (In house) 18-Oct-02 (SPEAG, In house check Oct-05) 11-May-05 (SPEAG, In house check Nov-05)	e part of the certificate. d humidity < 70%. Scheduled Celibration Oct-06 Oct-06 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Oct-07 In house check: Nov-07
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A	rtainties with confidence p ted in the closed laborato TE official for calibration) ID # GB37480704 US37292783 SN: 5066 (20g) SN: 5067 (20g) SN: 5067 (20g) SN: 5047.2 (10r) SN 1507 SN 601 ID # MY41092317	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 12-Oct-05 (SPEAG, No. ET3-1507, Oct05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (In house) 18-Oct-02 (SPEAG, In house check Oct-05)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Oct-07
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator Agilent E4421B	rtainties with confidence p ted in the closed laborato IE critical for calibration) ID # GB37480704 US37292783 SN: 5066 (20g) SN: 5066 (20g) SN: 5066 (20g) SN: 5066 (20g) SN: 5066 (20g) SN: 5066 (20g) SN: 5067 2 (10r) SN: 507 SN 601 ID # MY41092317 MY41000675	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Cemificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 28-Oct-05 (SPEAG, No. ET3-1507, Oct05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (In house) 18-Oct-02 (SPEAG, In house check Oct-05) 11-May-05 (SPEAG, In house check Nov-05)	e part of the certificate. d humidity < 70%. Scheduled Celibration Oct-06 Oct-06 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Oct-07 In house check: Nov-07
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator Agilent E4421B	rtainties with confidence p ted in the closed laborato ID # GB37480704 US37292783 SN: 5066 (20g) SN: 5066 (20g) SN: 5067 (20g) SN: 5047.2 (10r) SN 1507 SN 601 ID # MY41092317 MY41000575 US37390585 S4205	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No. 251-00498) 28-Oct-05 (SPEAS, No. 251-00498) 28-Oct-05 (SPEAS, No. 251-00498) 28-Oct-05 (SPEAS, No. 251-00498) 28-Oct-05 (SPEAS, No. 251-00498) 16-Dec-05 (SPEAS, No. DAE4-601_Dec05) Check Date (in house) 18-Oct-02 (SPEAS, in house check Oct-05) 11-May-05 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Oct-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Oct-07 In house check: Nov-07 In house check: Nov-06 Signature
The measurements and the uncer All calibrations have been conduct Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator Agilent E4421B Network Analyzer HP 8753E	rtainties with confidence p ted in the closed laborato FE official for calibration) ID # GB37480704 US37292783 SN: 5086 (20g) SN: 5086 (20g) SN: 5086 (20g) SN: 5087 (20g) SN: 507 SN 601 ID # MY41092317 MY41092317 MY41090575 US37390585 S4208 Name	robability are given on the following pages and arr y facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No. 251-00498) 28-Oct-05 (SPEAG, No. 251-00498) 28-Oct-05 (SPEAG, No. DAE4-601_Dec05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-05) 11-May-05 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05) Function	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Nov-07 In house check: Nov-07
The measurements and the unce All calibrations have been conduc Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator Agilent E4421B Network Analyzer HP 8753E - Calibrated by:	rtainties with confidence p ted in the closed laborato FE official for calibration) ID # GB37480704 US37292783 SN: 5066 (20g) SN: 5047.2 (10r) SN 507 SN 601 ID # MY41092317 MY41092317 MY41000575 US37390585 \$4205 Name Judith Müller	robability are given on the following pages and arr ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No. 251-00498) 28-Oct-05 (SPEAG, No. ET3-1507 Oct05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-05) 11-May-05 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05) Function Laboratory Technician	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Oct-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Oct-07 In house check: Nov-07 In house check: Nov-06 Signature



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



SWISS

S

С

S

Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

#### Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

#### Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

#### Additional Documentation:

d) DASY4 System Handbook

#### Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- · Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole . positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

Certificate No: D835V2-499\_Mar06

Page 2 of 9



#### Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY4	V4.7
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V4.9	
Distance Dipole Center - TSL	15 mm	' with Spacer
Area Scan resolution	dx, dy = 15 mm	
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

#### Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.1 ± 6 %	0.94mho/m ± 6 %
Head TSL temperature during test	(22.2 ± 0.2) °C		

## SAR result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	condition	
SAR measured	250 mW input power	2.35 mW / g
SAR normalized	normalized to 1W	9.40 mW / g
SAR for nominal Head TSL parameters 1	normalized to 1W	9.24 mW / g ± 17.0 % (k=2)
i i		
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.53 mW / g
SAR normalized	normalized to 1W	6.12 mW / g
SAR for nominal Head TSL parameters 1	normalized to 1W	6.07 mW / g ± 16.5 % (k=2)

<sup>1</sup> Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

Certificate No: D835V2-499\_Mar06

Page 3 of 9



#### **Body TSL parameters**

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	55.2	0.97 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	56.8±6%	0.98 mho/m ± 6 %
Body TSL temperature during test	(21.4 ± 0.2) °C		

2.8

#### SAR result with Body TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL	condition	
SAR measured	250 mW input power	2.45 mW/g
SAR normalized	normalized to 1W	9.80 mW / g
SAR for nominal Body TSL parameters 2	normalized to 1W	9.91 mW / g ± 17.0 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL	condition	
	condition 250 mW input power	1.62 mW / g
SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL SAR measured SAR normalized		1.62 mW / g 6.48 mW / g

<sup>2</sup> Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

Certificate No: D835V2-499\_Mar06

Page 4 of 9



52.1 Ω - 2.9 jΩ - 29.1 dB

# Appendix Antenna Parameters with Head TSL Impedance, transformed to feed point Return Loss

#### Antenna Parameters with Body TSL

Impedance, transformed to feed point	47.9 Ω - 5.1 jΩ
Return Loss	- 24.9 dB

#### General Antenna Parameters and Design

Electrical Delay (one direction)	1.391ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### Additional EUT Data

Manufactured by	SPEAG
Manufactured on	July 10, 2003

Certificate No: D835V2-499\_Mar06


Page 5 of 9



# DASY4 Validation Report for Head TSL Date/Time: 15.03.2006 12:51:44 Test Laboratory: SPEAG, Zurich, Switzerland DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN:499 Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1 Medium: HSL U10 BB; Medium parameters used: f = 835 MHz; $\sigma = 0.942$ mho/m; $\epsilon_r = 42.1$ ; $\rho = 1000$ kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY4 (High Precision Assessment) DASY4 Configuration: Probe: ET3DV6 - SN1507 (HF); ConvF(6.09, 6.09, 6.09); Calibrated: 28.10.2005 Sensor-Surface: 4mm (Mechanical Surface Detection) Electronics: DAE4 Sn601; Calibrated: 15.12.2005 Phantom: Flat Phantom 4.9L; Type: QD000P49AA; ; Measurement SW: DASY4, V4.7 Build 14; Postprocessing SW: SEMCAD, V1.8 Build 165 Pin = 250 mW; d = 10 mm/Area Scan (71x81x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 2.54 mW/g Pin = 250 mW; d = 10 mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 53.7 V/m; Power Drift = -0.008 dB Peak SAR (extrapolated) = 3:53 W/kg SAR(1 g) = 2.35 mW/g; SAR(10 g) = 1.53 mW/g Maximum value of SAR (measured) = 2.55 mW/g dB 0.000 -5.00 -10.0 15.0 20.0 -25.0 $0 \, dB = 2.55 \, mW/g$ Page 6 of 9 Certificate No: D835V2-499\_Mar06



## Impedance Measurement Plot for Head TSL






# DASY4 Validation Report for Body TSL Date/Time: 14.03.2006 12:37:15 Test Laboratory: SPEAG, Zurich, Switzerland DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN:499 Communication System: CW; Frequency: 835 MHz;Duty Cycle: 1:1 Medium: MSL U10; Medium parameters used: f = 835 MHz; $\sigma$ = 0.972 mho/m; $\epsilon_i$ = 56.9; $\rho$ = 1000 kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY4 (High Precision Assessment) DASY4 Configuration: Probe: ET3DV6 - SN1507 (HF); ConvF(5.84, 5.84, 5.84); Calibrated: 28.10.2005 Sensor-Surface: 4mm (Mechanical Surface Detection) · Electronics: DAE4 Sn601; Calibrated: 15.12.2005 Phantom: Flat Phantom 4.9L; Type: QD000P49AA; ; Measurement SW: DASY4, V4.7 Build 14: Postprocessing SW: SEMCAD, V1.8 Build 165 ٠ Pin = 250 mW; d = 10 mm/Area Scan (71x81x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 2.63 mW/g Pin = 250 mW; d = 10 mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 53.3 V/m; Power Drift = 0.026 dB Peak SAR (extrapolated) = 3:51 W/kg SAR(1 g) = 2.45 mW/g; SAR(10 g) = 1.62 mW/g Maximum value of SAR (measured) = 2.65 mW/g dB 0.000 -5.00 -10.0 -15.0 -20.0 25.0

Certificate No: D835V2-499\_Mar06

Page 8 of 9

 $0 \, dB = 2.65 \, mW/g$ 








Calibration Laborator Schmid & Partner Engineering AG Reughausstrasse 43, 8004 Zuric			chweizerischer Kalibrierdienst ervice suisse d'étalonnage nrvizio avizzero di taratura wiss Calibration Service
Accredited by the Swiss Federal C			: SCS 108
The Swiss Accreditation Service Multilateral Agreement for the re			
Client Sporton (Aude	n)	Certificate No: D	1900V2-5d041_Mar06
CALIBRATION C	ERTIFICATE		
	B40001/0 011 5	10.14	
Object	D1900V2 - SN: 5	d041	
Calibration procedure(s)	QA CAL-05.v6		
	Calibration proce	dure for dipole validation kits	
Calibration date:	March 21, 2006		
Gard and Folde.	March 21, 2000		
Condition of the calibrated item	In Tolerance		
	ents the traceability to nati	onal standards, which realize the physical units of robability are given on the following pages and are	
The measurements and the unce All calibrations have been conduct	ents the traceability to nati intainties with confidence p cted in the closed laborator		e part of the certificate.
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&1)	ents the traceability to nati intainties with confidence p cted in the closed laborator	robability are given on the following pages and are	e part of the certificate.
The measurements and the unce All celibrations have been conduct Celibration Equipment used (M&1 Primary Standards Power meter EPM-442A	ents the traceability to nati rtainties with confidence p cted in the closed laborator TE critical for calibration) ID # GB37480704	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C an <u>Cal Date (Calibrated by, Certificate No.)</u> 04-Oct-05 (METAS, No. 251-00516)	e part of the cartificate. d humildity < 70%. Scheduled Calibration Oct-06
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&1 Primary Standards Power meter EPM-442A Power sensor HP 8481A	ents the traceability to nati rtainties with confidence p ctad in the closed laborator TE critical for calibration) ID # GB37480704 US37292783	robability are given on the following pages and are ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516)	e part of the cartificate. d humildity < 70%. Scheduled Calibration Oct-06 Oct-06
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&T Primary Standards Power meter EPM-442A Power sensor HP 8461A Reference 20 dB Attenuator	ents the traceability to nati rtainties with confidence p cted in the closed laborator TE critical for calibration) ID # GB37480704 US37292783 SN: 5086 (20g)	robability are given on the following pages and are ry facility: environment temperature (22 ± 3)*C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No 251-00498)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&) Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Reference 10 dB Attenuator	ents the traceability to nati rtainties with confidence p cted in the closed laborator TE critical for calibration) ID # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 (10r)	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496)	e part of the cartificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06 Aug-06
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&1 Primary Standards Power meter EPM-442A Power sensor HF 6461A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6	ents the traceability to nati rtainties with confidence p cted in the closed laborator TE critical for calibration) ID # GB37480704 US37292783 SN: 5086 (20g)	robability are given on the following pages and are ry facility: environment temperature (22 ± 3)*C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No 251-00498)	e part of the certificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&) Primary Standards Power meter EPM-442A Power sensor HIP 8481A Reference 20 dB Attenuator Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards	ens the traceability to nati rtainties with confidence p cted in the closed laborator TE critical for calibration) ID # GB37480704 US37292783 SN: 5086 (20g) SN: 5087 2 (10r) SN: 1507	robability are given on the following pages and arr ty facility: environment temperature (22 ± 3)°C and <u>Cal Date (Calibrated by, Certificate No.)</u> 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No. 251-00496) 11-Aug-05 (METAS, No. 251-00496) 28-Oct-05 (SPEAG, No. ET3-1507_Oct05)	e part of the cartificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06 Oct-08
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&) Primary Standards Power meter: EPM-442A Power sensor HP 8461A Reference 20 dB Attenuator Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A	ents the traceability to nati rtainties with confidence p cted in the closed laborator TE critical for calibration) ID # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 (10r) SN: 1507 SN: 1507 SN: 601 ID # MY41092317	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (SPEAG, No. DAE4-601_De005) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-05)	e part of the cartificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Oct-07
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&1 Primary Standards Power meter EPM-442A Power sensor HF 24814 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference 10 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HF 2481A RF generator Agilent E4421B	ents the traceability to nati rtainties with confidence p ctad in the closed laborator TE critical for calibration) ID # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 (10r) SN: 5047.2 (10r) SN: 601 ID # MY41092317 MY41092517	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C an <u>Cal Date (Calibrated by, Certificate No.)</u> 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00516) 11-Aug-05 (METAS, No. 251-00496) 11-Aug-05 (METAS, No. 251-00496) 28-Oct-05 (SPEAG, No. ET3-1507_Oct05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) <u>Check Date (in house)</u> 18-Oct-02 (SPEAG, in house check Oct-05) 11-May-05 (SPEAG, in house check Nov-05)	e part of the cartificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Nov-07
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&1 Primary Standards Power meter EPM-442A Power sensor HE 6461A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference 10 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 6481A RF generator Apilent E4421B	ents the traceability to nati rtainties with confidence p cted in the closed laborator TE critical for calibration) ID # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 (10r) SN: 1507 SN: 1507 SN: 601 ID # MY41092317	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (SPEAG, No. DAE4-601_De005) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-05)	e part of the cartificate. d humidity < 70%. Scheduled Calibration Oct-06 Oct-06 Aug-06 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Oct-07
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&) Primary Standards Power meter EPM-442A Power sensor HIP 8481A Reference 20 dB Attenuator Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards	ents the traceability to nati rtainties with confidence p ctad in the closed laborator TE critical for calibration) ID # G837480704 US37292783 SN: 5086 (20g) SN: 5086 (20g) SN: 5047.2 (10r) SN: 1507 SN: 601 ID # MY41092317 MY41092317 MY41092317 MY4109235 S4206	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C and Cal Date (Galibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No 251-00498) 11-Aug-05 (METAS, No 251-00498) 28-Oct-05 (SPEAG, No. D351-00498) 28-Oct-05 (SPEAG, No. DAE4-601_Dec05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (in house) 18-Oct-02 (SPEAG, in house check Not-05) 18-Oct-01 (SPEAG, in house check Not-05)	e part of the cartificate. d humildity < 70%. Scheduled Calibration Oct-06 Oct-08 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Nov-07 In house check: Nov-07 In house check: Nov-06
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&) Primary Standards Power meter EPM-442A Power sensor HP 8461A Reference 20 dB Attenuator Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator Apilent E4421B Notwork Analyzer HP 8763E	ents the traceability to nati rtainties with confidence p cted in the closed laborator ID # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 (10r) SN: 1507 SN: 1507 SN: 1507 SN: 1507 SN: 601 ID # MY41092317 MY41092317 MY41092317 MY41092317 MY41092317 MY41092317 MY41092317 MY41092317 MY41092317	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (SPEAG, No. DAE4-601_Dec05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-05) 11-May-05 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05) Function	e part of the cartificate. d humidity < 70%. Scheduled Calibration Oct:06 Oct:06 Aug:06 Oct:08 Dec-06 Scheduled Check In house check: Nov-07 In house check: Nov-06 Signature
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&I Primary Standards Power meter EPM-442A Power sensor HP 8461A Reference 20 dB Attenuator Reference 20 dB Attenuator Reference 10 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator Agilent E4421B Notwork Analyzer HP 8763E	ents the traceability to nati rtainties with confidence p ctad in the closed laborator TE critical for calibration) ID # G837480704 US37292783 SN: 5086 (20g) SN: 5086 (20g) SN: 5047.2 (10r) SN: 1507 SN: 601 ID # MY41092317 MY41092317 MY41092317 MY4109235 S4206	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C and Cal Date (Galibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00498) 11-Aug-05 (METAS, No 251-00498) 11-Aug-05 (METAS, No 251-00498) 28-Oct-05 (SPEAG, No. D351-00498) 28-Oct-05 (SPEAG, No. DAE4-601_Dec05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (in house) 18-Oct-02 (SPEAG, in house check Not-05) 18-Oct-01 (SPEAG, in house check Not-05)	e part of the cartificate. d humidity < 70%. Scheduled Calibration Oct:06 Oct:06 Aug:06 Oct:08 Dec-06 Scheduled Check In house check: Nov-07 In house check: Nov-06 Signature
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&1 Primary Standards Power meter EPM-442A Power sensor HE 6461A Reference 20 dB Attenuator Reference 10 dB Attenuator Reference 10 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 6481A RF generator Apilent E4421B	ents the traceability to nati rtainties with confidence p cted in the closed laborator ID # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 (10r) SN: 1507 SN: 1507 SN: 1507 SN: 1507 SN: 601 ID # MY41092317 MY41092317 MY41092317 MY41092317 MY41092317 MY41092317 MY41092317 MY41092317 MY41092317	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (METAS, No 251-00496) 11-Aug-05 (SPEAG, No. DAE4-601_Dec05) 15-Dec-05 (SPEAG, No. DAE4-601_Dec05) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-05) 11-May-05 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05) Function	e part of the cartificate. d humildity < 70%. Scheduled Calibration Oct-06 Oct-08 Aug-06 Aug-06 Oct-08 Dec-06 Scheduled Check In house check: Nov-07 In house check: Nov-07 In house check: Nov-06
The measurements and the unce All calibrations have been conduct Calibration Equipment used (M&1 Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 10 dB Attenuator Reference 20 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator Apilent E4421B Notwork Analyzet HP 8763E	ents the traceability to nati rtainties with confidence p cted in the closed laborator TE critical for calibration) ID # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 (10r) SN: 5047.2 (10r) SN: 601 ID # MY41092317 MY41092317 MY41092317 MY41090675 US37390585 S4206 Namo Judith Müller	robability are given on the following pages and an ry facility: environment temperature (22 ± 3)°C and Cal Date (Calibrated by, Certificate No.) 04-Oct-05 (METAS, No. 251-00516) 04-Oct-05 (METAS, No. 251-00496) 11-Aug-05 (METAS, No. 251-00496) 12-Aug-05 (METAS, No. 251-00498) 28-Oct-05 (SPEAG, No. DAE4-001_Dec05) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-05) 11-May-05 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05) I8-Oct-01 (SPEAG, in house check Nov-05) Function Laboratory Technician	e part of the cartificate. d humidity < 70%. Scheduled Calibration Oct:06 Oct:06 Aug:06 Oct:08 Dec-06 Scheduled Check In house check: Nov-07 In house check: Nov-06 Signature







#### Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY4	V4.7
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Area Scan resolution	dx, dy = 15 mm	
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1900 MHz ± 1 MHz	

#### Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.4 ± 6 %	1.42 mho/m ± 6 %
Head TSL temperature during test	(21.5 ± 0.2) °C		

## SAR result with Head TSL

SAR averaged over $1 \text{ cm}^3$ (1 g) of Head TSL	condition	
SAR measured	250 mW input power	9.75 mW / g
SAR normalized	normalized to 1W	39.0 mW / g
SAR for nominal Head TSL parameters 1	normalized to 1W	38.4 mW / g ± 17.0 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.17 mW / g
SAR normalized	normalized to 1W	20.7 mW / g
SAR for nominal Head TSL parameters 1	normalized to 1W	20.5 mW / g ± 16.5 % (k=2)

<sup>1</sup> Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

Certificate No: D1900V2-5d041\_Mar06

Page 3 of 9



## Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	53.3	1.52 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	54.7 ± 6 %	1.54 mho/m ± 6 %
Body TSL temperature during test	(21.6 ± 0.2) °C		

## SAR result with Body TSL

SAR averaged over 1 cm3 (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	10.2 mW / g
SAR normalized	normalized to 1W	40.8 mW / g
SAR for nominal Body TSL parameters 2	normalized to 1W	41.1 mW/g±17.0 % (k=2)

SAR averaged over 10 $\mbox{cm}^3$ (10 g) of Body TSL	condition	
SAR measured	250 mW input power	5.40 mW / g
SAR normalized	normalized to 1W	21.6 mW / g
SAR for nominal Body TSL parameters <sup>2</sup>	normalized to 1W	21.8 mW / g ± 16.5 % (k=2)


<sup>2</sup> Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

Certificate No: D1900V2-5d041\_Mar06 Page 4 of 9



pendix	
tenna Parameters with Head TSL	
Impedance, transformed to feed point	53.5 Ω + 5.1 jΩ
Return Loss	- 24.8 dB
tenna Parameters with Body TSL	
Impedance, transformed to feed point	47.9 Ω + 6.3 jΩ
Return Loss	- 23.4 dB
neral Antenna Parameters and Design	
Electrical Delay (one direction)	1.200 ns
r long term use with 100W radiated power, only a slight v	warming of the dipole near the feedpoint can be measured.
ditional EUT Data Manufactured by	SPEAG
Manufactured on	July 4, 2003






Certificate No: D1900V2-5d041\_Mar06


Page 6 of 9






Impedance Measurement Plot for Head TSL












Calibration Laborator Schmid & Partner Engineering AG Reughausstrasse 43, 8004 Zurio			hweizerischer Kalibrierdienst rvice suisse d'étalonnage rvizio svizzero di taratura viss Calibration Service	
Accredited by the Swiss Federal	Office of Metrology and Ar	coreditation Accreditation No.	: SCS 108	
The Swiss Accreditation Servic Multilateral Agreement for the r		n certificates		
Client Sporton (Aude	en)	Certificate No: E	T3-1788_Sep06	
CALIBRATION	CERTIFICAT	Ender All Street Street		
Object	ET3DV6 - SN:17	788		
Calibration procedure(s)	QA CAL-01.v5 Calibration proc	edure for dosimetric E-field probes		
Calibration date:	September 19, 2	2006	a provide the second	
	Vala Brankis su use ve		and the survey many processes	
Condition of the calibrated item	In Tolerance			
Condition of the calibrated item This calibration certificate docum The measurements and the unco	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ory facility: environment temperature $(22 \pm 3)$ °C and	e part of the certificate.	
Condition of the calibrated item This calibration certificate docum The measurements and the unco	In Tolerance	tional standards, which mailize the physical units of probability are given on the following pages and an	e part of the certificate.	
Condition of the calibrated item This calibration certificate docum The measurements and the unio All calibrations have been concu Calibration Equipment used (M& Primary Standards	In Tolerance	tional standards, which realize the physical units o probability are given on the following pages and an ory facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.)	a part of the certificate. 5 humidity < 70%, Scheduled Calibration	
Condition of the calibrated item This calibration certificate docum The measurements and the uno All calibrations have been concu Calibration Equipment used (M& Primary Standards Power meter E44198	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ory facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557)	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07	
Condition of the calibrated item This calibration certificate docum The measurements and the uno All calibrations have been concu Calibration Equipment used (M& Primary Standards Power meter E44198 Power sonsor E4412A	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ory facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5 Apr-06 (METAS, No. 251-00557) 5 Apr-06 (METAS, No. 251-00557)	a part of the certificate. d humidity < 70%, Scheduled Calibration Apr-07 Apr-07	
Condition of the calibrated item This calibration certificate docum The measurements and the uno All calibrations have been concu Calibration Equipment used (M& Primary Standards Power meter E44198	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ory facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557)	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07	
Condition of the calibrated item This calibration certificate docum The measurements and the unic All calibrations have been concu Calibration Equipment used (M& Primary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ory facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557)	a part of the certificate. d humidity < 70%, Scheduled Calibration Apr-07 Apr-07 Apr-07	
Condition of the calibrated item This calibration certificate docum The measurements and the uno All calibrations have been concu Calibration Equipment used (M& Primary Standards Power sensor E4412A Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 30 dB Attenuator	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ory facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00559) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00559) 10-Aug-06 (METAS, No. 217-00593)	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07 Apr-07 Aug-07 Aug-07 Aug-07 Aug-07	
Condition of the calibrated item This calibration certificate docum The measurements and the unco- All calibrations have been concu- Calibration Equipment used (M& Primary Standards Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 30 dB Attenuator	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ony facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5: Apr-06 (METAS, No. 251:-00557) 5: Apr-06 (METAS, No. 251:-00557) 10: Aug-06 (METAS, No. 251:-00557) 10: Aug-06 (METAS, No. 217:-00592) 4: Apr-06 (METAS, No. 217:-00593) 2: Jan-06 (SPEAG, No. 253:-3013_Jan06)	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07 Apr-07 Apr-07 Apr-07 Aug-07 Aug-07 Jan-07	
Condition of the calibrated item This calibration certificate docum The measurements and the uno All calibrations have been concu Calibration Equipment used (M& Primary Standards Power sonsor E44128 Power sonsor E4412A Reference 30 B Attenuator Reference 20 B Attenuator	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ory facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00559) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00559) 10-Aug-06 (METAS, No. 217-00593)	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07 Apr-07 Aug-07 Aug-07 Aug-07 Aug-07	
Condition of the calibrated item This calibration certificate docum The measurements and the unco- All calibrations have been concu- Calibration Equipment used (M& Primary Standards Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 30 dB Attenuator Reference 30 dB Attenuator Reference 30 dB Attenuator Reference Probe ES3DV2	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ony facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5: Apr-06 (METAS, No. 251:-00557) 5: Apr-06 (METAS, No. 251:-00557) 10: Aug-06 (METAS, No. 251:-00557) 10: Aug-06 (METAS, No. 217:-00592) 4: Apr-06 (METAS, No. 217:-00593) 2: Jan-06 (SPEAG, No. 253:-3013_Jan06)	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07 Apr-07 Apr-07 Apr-07 Aug-07 Aug-07 Jan-07	
Condition of the calibrated item This calibration certificate docum The measurements and the unco- All calibrations have been concu- Calibration Equipment used (M& Primary Standards Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 3 dB Attenuator Reference 30 dB Attenuator	In Tolerance           tents the traceebility to na entainties with confidence plate           intee in the closed laborate           TE-critical for calibration)           ID #           GB41293874           MY41498087           SN: S5054 (3c)           SN: S5054 (3c)           SN: S5129 (3ob)           SN: 3013           SN: 654           ID #           US3642U01700	tional standards, which realize the physical units of probability are given on the following pages and an ony facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. 253-3013_Jam06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-99 (SPEAG, in house check Nov-05)	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07 Apr-07 Apr-07 Aug-07 Jun-07 Scheduled Check In house check: Nov-07	
Condition of the calibrated item This calibration certificate docum The measurements and the uno All calibrations have been concu Calibration Equipment used (M& Primary Standards Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 30 dB Attenuator	In Tolerance	tional standards, which realize the physical units of probability are given on the following pages and an ony facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5 Apr-06 (METAS, No. 251-00557) 5 Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4 Apr-06 (METAS, No. 217-00592) 10-Aug-08 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. ES3-3013_Jan06) 21-Jan-06 (SPEAG, No. DAE4-654_Jun06) 21-Jan-06 (SPEAG, No. DAE4-654_Jun06)	a part of the certificate. d humidity < 70%, Scheduled Calibration Apr-07 Apr-07 Apr-07 Aug-07 Aug-07 Aug-07 Jun-07 Scheduled Check	
Condition of the calibrated item This calibration certificate docum The measurements and the union All calibrations have been concul Calibration Equipment used (M8 Primary Standards Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 9 Probe ES3DV2 DAE4 Soconcary Standards RF generator HP 8648C Network Analyzer HP 8763E	In Tolerance           tents the traceshilty to na entainties with confidence of intee in the closed laborate TEI-entical for calibration)           ID # (B&41293874 MY4149807 SN: 55054 (3c) SN: 55054 (3c) SN: 55086 (20b) SN: 55129 (30b) SN: 3013 SN: 654           ID # US3642U01700 US37390586           Name	tional standards, which realize the physical units of probability are given on the following pages and an ony facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5 Apr-06 (METAS, No. 251-00557) 5 Apr-06 (METAS, No. 251-00557) 10 Aug-06 (METAS, No. 251-00557) 10 Aug-06 (METAS, No. 251-00558) 10 Aug-06 (METAS, No. 217-00592) 4 Apr-06 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. ES3-3013_Jan06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-98 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05) Function	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07 Apr-07 Apr-07 Aug-07 Jun-07 Scheduled Check In house check: Nov-07	
Condition of the calibrated item This calibration certificate docum The measurements and the unco- All calibrations have been concu- Calibration Equipment used (M& Primary Standards Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 3 dB Attenuator Reference 30 dB Attenuator	In Tolerance           tents the traceebility to na ortainties with confidence plants           inter in the closed laborate           TE-entical for calibration)           ID #           GB41293874           MY41495277           MY41495277           MY41498087           SN: S5054 (3c)           SN: S5054 (3c)           SN: S5120 (30b)           SN: 3013           SN: 654           ID #           US3842U01700           US3842U01700           US37390585	tional standards, which realize the physical units of probability are given on the following pages and an ory facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5 Apr-06 (METAS, No. 251-00557) 5 Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00559) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00593) 2-Jun-06 (SPEAG, No. ES3-9013,Jun06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-99 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05)	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07 Apr-07 Aug-07 Aug-07 Aug-07 Jun-07 Jun-07 Scheduled Check In house check: Nov-07 In house check: Nov-08	
Condition of the calibrated item This calibration certificate docum The measurements and the union All calibrations have been concul Calibration Equipment used (M8 Primary Standards Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 9 Probe ES3DV2 DAE4 Soconcary Standards RF generator HP 8648C Network Analyzer HP 8763E	In Tolerance           tents the traceshilty to na entainties with confidence of intee in the closed laborate TEI-entical for calibration)           ID # (B&41293874 MY4149807 SN: 55054 (3c) SN: 55054 (3c) SN: 55086 (20b) SN: 55129 (30b) SN: 3013 SN: 654           ID # US3642U01700 US37390586           Name	tional standards, which realize the physical units of probability are given on the following pages and an ony facility: environment temperature (22 ± 3)°C an Cal Date (Calibrated by, Certificate No.) 5 Apr-06 (METAS, No. 251-00557) 5 Apr-06 (METAS, No. 251-00557) 10 Aug-06 (METAS, No. 251-00557) 10 Aug-06 (METAS, No. 251-00558) 10 Aug-06 (METAS, No. 217-00592) 4 Apr-06 (METAS, No. 217-00593) 2-Jan-06 (SPEAG, No. ES3-3013_Jan06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-98 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Nov-05) Function	a part of the certificate. d humidity < 70%. Scheduled Calibration Apr-07 Apr-07 Aug-07 Aug-07 Aug-07 Jun-07 Jun-07 Scheduled Check In house check: Nov-07 In house check: Nov-08	







ET3DV6 SN:1788

September 19, 2006

### Probe ET3DV6

### SN:1788

Manufactured: Last calibrated: Recalibrated: May 28, 2003 September 30, 2004 September 19, 2006

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)


Certificate No: ET3-1788\_Sep06

Page 3 of 9




ET3DV6 SN:1788			5	September 19, 2006	
DASY - Par	ameters of P	robe: ET3	DV6 SN:1	788	
Sensitivity in Fre	ee Space <sup>A</sup>		Diode (	Compression <sup>B</sup>	
NormX NormY NormZ	<b>1.73</b> ± 10.1% <b>1.67</b> ± 10.1% <b>1.70</b> ± 10.1%	$\mu$ V/(V/m) <sup>2</sup> $\mu$ V/(V/m) <sup>2</sup> $\mu$ V/(V/m) <sup>2</sup>	DCP X DCP Y DCP Z	95 mV 101 mV 93 mV	
	ssue Simulating Li				
Please see Page 8.			50111 40(015)		
Boundary Effect	t				
TSL	000 MHz Typical SA	AR gradient: 5 %	per mm		
Sensor Cente SAR <sub>be</sub> [%] SAR <sub>be</sub> [%]		lgorithm	3.7 mm 4 7.9 0.1	<b>1.7 mm</b> 4.3 0.3	
TSL 18	810 MHz Typical SA	AR gradient: 10 %	per mm		
Sensor Cente SAR <sub>be</sub> [%] SAR <sub>be</sub> [%]	er to Phantom Surface Di Without Correction A With Correction Algo	Igorithm	<b>3.7 mm</b> 4 11.8 0.2	<b>1.7 mm</b> 7.0 0.4	
Sensor Offset					
Probe Tip to	Sensor Center		2.7 mm		
	rtainty of measureme tiplied by the coverag				
	X,Y,Z do not affect the E <sup>2</sup> -field to		-		
<sup>9</sup> Numerical linearization pa	rameter: uncertainty not require	əd.			






# FCC/IC SAR Test Report

#### Test Report No 🔅 FA710211-1-2-01



FCC/IC SAR Test Report

Test Report No FA710211-1-2-01




Certificate No: ET3-1788\_Sep06


Page 7 of 9

# FCC/IC SAR Test Report

#### Test Report No 🔅 FA710211-1-2-01









Calibration Laboratory Schmid & Partner Engineering AG eughausstrasse 43, 8004 Zurich,		Hac MRA RAC MRA RAC MRA S	Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
Accredited by the Swiss Federal Off The Swiss Accreditation Service I	is one of the signatori	les to the EA	No.: SCS 108
Aultilateral Agreement for the rec	94 <u>7</u> - 2017 - 2		DAE3-577_Nov06
CALIBRATION CI	ERTIFICAT	E	
Object	DAE3 - SD 000	D03 AA - SN: 577	100000000000
Calibration procedure(s)	QA CAL-06.v12 Calibration proc	edure for the data acquisition elect	ronics (DAE)
Calibration date:	November 21, 2	2006	
		ational standards, which realize the physical units probability are given on the following pages and	
This calibration certificate documen The measurements and the uncerta	ats the traceability to na ainties with confidence ad in the closed laborat	probability are given on the following pages and ory facility: environment temperature (22 $\pm$ 3)°C	are part of the certificate.
This calibration certificate documen The measurements and the uncert All calibrations have been conducte	ats the traceability to na ainties with confidence ad in the closed laborat	probability are given on the following pages and ory facility: environment temperature (22 $\pm$ 3)°C	are part of the certificate.
This calibration certificate documer The measurements and the uncert All calibrations have been conducte Calibration Equipment used (M&TE	ats the traceability to na ainties with confidence ad in the closed laborat critical for calibration)	probability are given on the following pages and ory facility: environment temperature (22 $\pm$ 3) $^\circ {\mathbb C}$	are part of the certificate. and humidity < 70%.
This calibration certificate document The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Fluke Process Calibrator Type 702 Keithley Multimeter Type 2001	ats the traceability to na ainties with confidence ad in the closed laborat critical for calibration) ID # SN: 6295803 SN: 0810278	probability are given on the following pages and ony facility: environment temperature (22 ± 3)°C Cal Date (Calibrated by, Certificate No.) 13-Oct-06 (Elcal AG, No: 5492) 03-Oct-06 (Elcal AG, No: 5478)	are part of the certificate. and humidity < 70%. Scheduled Calibration Oct-07 Oct-07
This calibration certificate documen The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Fluke Process Calibrator Type 702	ats the traceability to na ainties with confidence ad in the closed laborat critical for calibration) ID # SN: 6295803 SN: 0810278 ID #	probability are given on the following pages and ory facility: environment temperature (22 ± 3)°C Cal Date (Calibrated by, Certificate No.) 13-Oct-06 (Elcal AG, No: 5492)	are part of the certificate. and humidity < 70%. Scheduled Calibration Oct-07
This calibration certificate document The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Fluke Process Calibrator Type 702 Keithley Multimeter Type 2001 Secondary Standards	ats the traceability to na ainties with confidence ad in the closed laborat critical for calibration) ID # SN: 6295803 SN: 0810278 ID #	probability are given on the following pages and ory facility: environment temperature (22 ± 3)°C Cal Date (Calibrated by, Certificate No.) 13-Oct-06 (Elcal AG, No: 5492) 03-Oct-06 (Elcal AG, No: 5478) Check Date (in house)	are part of the certificate. and humidity < 70%. Scheduled Calibration Oct-07 Oct-07 Scheduled Check
This calibration certificate document The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Fluke Process Calibrator Type 702 Keithley Multimeter Type 2001 Secondary Standards	ID # SN: 6295803 SN: 6810278	probability are given on the following pages and ory facility: environment temperature (22 ± 3)°C Cal Date (Calibrated by, Certificate No.) 13-Oct-06 (Elcal AG, No: 5492) 03-Oct-06 (Elcal AG, No: 5478) Check Date (in house) 12 16 Jun-06 (SPEAG, in house check)	are part of the certificate. and humidity < 70%. Scheduled Calibration Oct-07 Oct-07 Scheduled Check In house check Jun-07
This calibration certificate document The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Fluke Process Calibrator Type 702 Keithley Multimeter Type 2001 Secondary Standards	ats the traceability to na ainties with confidence ad in the closed laborat critical for calibration) ID # SN: 6295803 SN: 0810278 ID #	probability are given on the following pages and ory facility: environment temperature (22 ± 3)°C Cal Date (Calibrated by, Certificate No.) 13-Oct-06 (Elcal AG, No: 5492) 03-Oct-06 (Elcal AG, No: 5478) Check Date (in house)	are part of the certificate. and humidity < 70%. Scheduled Calibration Oct-07 Oct-07 Scheduled Check
This calibration certificate document The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Fluke Process Calibrator Type 702 Keithley Multimeter Type 2001 Secondary Standards Calibrator Box V1.1	Ats the traceability to na ainties with confidence ad in the closed laborat critical for calibration) ID # SN: 6295803 SN: 0810276 ID # SE UMS 006 AB 100 Name	probability are given on the following pages and ony facility: environment temperature (22 ± 3)°C Cal Date (Calibrated by, Certificate No.) 13-Oct-06 (Elcal AG, No: 5492) 03-Oct-06 (Elcal AG, No: 5478) Check Date (in house) D2 15-Jun-06 (SPEAG, in house check) Function	are part of the certificate. and humidity < 70%. Scheduled Calibration Oct-07 Oct-07 Scheduled Check In house check Jun-07



Calibration Laboratory of SWISS Schweizerischer Kalibrierdienst S Schmid & Partner Service suisse d'étalonnage C Engineering AG Servizio svizzero di taratura S Zeughausstrasse 43, 8004 Zurich, Switzerland BRA Swiss Calibration Service Accredited by the Swiss Federal Office of Metrology and Accreditation Accreditation No.: SCS 108 The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates Glossary DAE data acquisition electronics information used in DASY system to align probe sensor X to the robot Connector angle coordinate system. Methods Applied and Interpretation of Parameters DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range. · Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required. The following parameters contain technical information as a result from the performance test and require no uncertainty. DC Voltage Measurement Linearity: Verification of the Linearity at +10% and -10% of the ۰ nominal calibration voltage. Influence of offset voltage is included in this measurement. Common mode sensitivity: Influence of a positive or negative common mode voltage on the differential measurement. Channel separation: Influence of a voltage on the neighbor channels not subject to an input . voltage. AD Converter Values with inputs shorted: Values on the internal AD converter . corresponding to zero input voltage Input Offset Measurement: Output voltage and statistical results over a large number of zero voltage measurements.

- Input Offset Current: Typical value for information; Maximum channel input offset current, not considering the input resistance.
- Input resistance: DAE input resistance at the connector, during internal auto-zeroing and during measurement.
- Low Battery Alarm Voltage: Typical value for information. Below this voltage, a battery alarm signal is generated.
- Power consumption: Typical value for information. Supply currents in various operating modes.

Certificate No: DAE3-577 Nov06

Page 2 of 5

©2007 SPORTON International Inc. SAR Testing Lab

This report shall not be reproduced except in full, without the written approval of Sporton.



#### DC Voltage Measurement

A/D - Converter Reso	olution nominal			
High Range:	1LSB =	6.1µV,	full range =	-100+300 mV
Low Range:	1LSB =	61nV ,	full range =	-1+3mV
DASY measurement	parameters: Auto	Zero Time: 3	3 sec; Measuring t	time: 3 sec

Calibration Factors	х	Y	Z
High Range	404.355 ± 0.1% (k=2)	$403.806 \pm 0.1\% \ (k{=}2)$	404.276 ± 0.1% (k=2)
Low Range	3.92854 ± 0.7% (k=2)	$3.93862 \pm 0.7\%$ (k=2)	$3.93591 \pm 0.7\%$ (k=2)

#### **Connector Angle**

Certificate No: DAE3-577\_Nov06

Page 3 of 5



#### Appendix

#### 1. DC Voltage Linearity

High Range		Input (µV)	Reading (µV)	Error (%)
Channel X	+ Input	200000	199999.5	0.00
Channel X	+ Input	20000	20005.87	0.03
Channel X	- Input	20000	-19998.71	-0.01
Channel Y	+ Input	200000	200000	0.00
Channel Y	+ Input	20000	20004.22	0.02
Channel Y	- Input	20000	-20003.23	0.02
Channel Z	+ Input	200000	200000.6	0.00
Channel Z	+ Input	20000	20005.24	0.03
Channel Z	- Input	20000	-20001.80	0.01

Low Range	Input (μV)	Reading (µV)	Error (%)
Channel X + Input	2000	1999.9	0.00
Channel X + Input	200	200.27	0.13
Channel X - Input	200	-200.73	0.36
Channel Y + Input	2000	2000.1	0.00
Channel Y + Input	200	199.22	-0.39
Channel Y - Input	200	-200.86	0.43
Channel Z + Input	2000	1999.9	0.00
Channel Z + Input	200	199.28	-0.36
Channel Z - Input	200	-200.94	0.47

#### 2. Common mode sensitivity

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Common mode Input Voltage (mV)	High Range Average Reading (μV)	Low Range Average Reading (µV)
200	14.24	12.49
- 200	-12.13	-12.92
200	-6.51	-7.06
- 200	6.05	5.81
200	1.09	0.86
- 200	-2.86	-2.63
	Input Voltage (mV) 200 - 200 200 - 200 200 200 200 200 200 200 200 200 200	Input Voltage (mV)         Average Reading (μV)           200         14.24           - 200         -12.13           200         -6.51           - 200         6.05           200         1.09

#### 3. Channel separation

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X (µV)	Channel Y (μV)	Channel Z (µV)
Channel X	200	-	2.51	0.09
Channel Y	200	0.43	-	3.37
Channel Z	200	-0.55	0.96	-

Certificate No: DAE3-577\_Nov06

Page 4 of 5



#### 4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	15970	16306
Channel Y	15851	16305
Channel Z	16208	17068

#### 5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec Input 10M $\Omega$ 

	Average (µV)	min. Offset (μV)	max. Offset (μV)	Std. Deviation (µV)
Channel X	-0.51	-1.55	0.47	0.50
Channel Y	-2.06	-4.32	-0.65	0.60
Channel Z	-1.63	-2.56	-0.15	0.35

#### 6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

#### 7. Input Resistance

	Zeroing (MOhm)	Measuring (MOhm)
Channel X	0.2000	199.8
Channel Y	0.2000	200.7
Channel Z	0.2000	199.8

#### 8. Low Battery Alarm Voltage (verified during pre-test)

Typical values	Alarm Level (VDC)	
Supply (+ Vcc)	+7.9	
Supply (- Vcc)	-7.6	

#### 9. Power Consumption (verified during pre test)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.0	+6	+14
Supply (- Vcc)	-0.01	-8	-9

Certificate No: DAE3-577\_Nov06 Page 5 of 5