

Appendix 9

pages 1

VERIFY WITH DIPOLE VALIDATION KIT

Validation Kit 1800 MHz, SN: 207, d = 10 mm, f = 1800 MHz, Pin = 250 mW, Generic Flat Phantom, Measured 2000-2-29/NMP

$\sigma = 1.66$ [mho/m] $\epsilon_r = 41.9$ $\rho = 1.00$ [g/cm³]

Coarse Grid Dx = 20.0 Dy = 20.0 Dz = 5.0 [mm]

SAR [mW/g] Max: 9.18

SAR (1g): 9.98 [mW/g] SAR (10g): 5.04 [mW/g]

