

Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Calibration Certificate

Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1504

Place of Calibration:

Zurich

Date of Calibration:

July 26, 2002

Calibration Interval:

12 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

U. Vella

Approved by:

Adrian Katja

Probe ET3DV6

SN:1504

| | |
|-------------------|------------------|
| Manufactured: | October 24, 1999 |
| Last calibration: | January 10, 2002 |
| Recalibrated: | July 26, 2002 |

Calibrated for System DASY3

DASY3 - Parameters of Probe: ET3DV6 SN:1504

Sensitivity in Free Space

| | |
|-------|---|
| NormX | 2.02 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormY | 1.78 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormZ | 1.73 $\mu\text{V}/(\text{V}/\text{m})^2$ |

Diode Compression

| | | |
|-------|-----------|----|
| DCP X | 95 | mV |
| DCP Y | 95 | mV |
| DCP Z | 95 | mV |

Sensitivity in Tissue Simulating Liquid

| | | | |
|---------|------------------------------|-----------------------------|-------------------------------|
| Head | 835 MHz | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.90 \pm 5\%$ mho/m |
| Head | 900 MHz | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.97 \pm 5\%$ mho/m |
| ConvF X | 6.5 $\pm 9.5\%$ (k=2) | Boundary effect: | |
| ConvF Y | 6.5 $\pm 9.5\%$ (k=2) | Alpha | 0.39 |
| ConvF Z | 6.5 $\pm 9.5\%$ (k=2) | Depth | 2.42 |
| Head | 1880 MHz | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\%$ mho/m |
| Head | 1800 MHz | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\%$ mho/m |
| ConvF X | 5.4 $\pm 9.5\%$ (k=2) | Boundary effect: | |
| ConvF Y | 5.4 $\pm 9.5\%$ (k=2) | Alpha | 0.53 |
| ConvF Z | 5.4 $\pm 9.5\%$ (k=2) | Depth | 2.44 |

Boundary Effect



| | | | |
|------|--|--|-------------|
| Head | 835 MHz | Typical SAR gradient: 5 % per mm | |
| | Probe Tip to Boundary | 1 mm | 2 mm |
| | SAR _{be} [%] Without Correction Algorithm | 9.6 | 5.3 |
| | SAR _{be} [%] With Correction Algorithm | 0.3 | 0.5 |
| Head | 1880 MHz | Typical SAR gradient: 10 % per mm | |
| | Probe Tip to Boundary | 1 mm | 2 mm |
| | SAR _{be} [%] Without Correction Algorithm | 13.0 | 8.5 |
| | SAR _{be} [%] With Correction Algorithm | 0.2 | 0.2 |

Sensor Offset



| | | |
|----------------------------|---------------------------------|----|
| Probe Tip to Sensor Center | 2.7 | mm |
| Optical Surface Detection | 1.4 \pm 0.2 | mm |

Receiving Pattern (ϕ), $\theta = 0^\circ$



f = 30 MHz, TEM cell ifi110





f = 100 MHz, TEM cell ifi110





f = 300 MHz, TEM cell ifi110




f = 900 MHz, TEM cell ifi110






Isotropy Error (ϕ), $\theta = 0^\circ$





Frequency Response of E-Field



(TEM-Cell:ifi110, Waveguide R22)



Dynamic Range f(SAR_{brain}) (Waveguide R22)





Conversion Factor Assessment



| | | | |
|------|----------|------------------------------|-------------------------------|
| Head | 835 MHz | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.90 \pm 5\%$ mho/m |
| Head | 900 MHz | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.97 \pm 5\%$ mho/m |
| | ConvF X | 6.5 $\pm 9.5\%$ (k=2) | Boundary effect: |
| | ConvF Y | 6.5 $\pm 9.5\%$ (k=2) | Alpha 0.39 |
| | ConvF Z | 6.5 $\pm 9.5\%$ (k=2) | Depth 2.42 |
| | | | |
| Head | 1880 MHz | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\%$ mho/m |
| Head | 1800 MHz | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\%$ mho/m |
| | ConvF X | 5.4 $\pm 9.5\%$ (k=2) | Boundary effect: |
| | ConvF Y | 5.4 $\pm 9.5\%$ (k=2) | Alpha 0.53 |
| | ConvF Z | 5.4 $\pm 9.5\%$ (k=2) | Depth 2.44 |


Conversion Factor Assessment



| | | | |
|-------------|-----------------|------------------------------|-------------------------------|
| Body | 835 MHz | $\epsilon_r = 55.2 \pm 5\%$ | $\sigma = 0.97 \pm 5\%$ mho/m |
| Body | 900 MHz | $\epsilon_r = 55.0 \pm 5\%$ | $\sigma = 1.05 \pm 5\%$ mho/m |
| | ConvF X | 6.5 $\pm 9.5\%$ (k=2) | Boundary effect: |
| | ConvF Y | 6.5 $\pm 9.5\%$ (k=2) | Alpha 0.42 |
| | ConvF Z | 6.5 $\pm 9.5\%$ (k=2) | Depth 2.38 |
| | | | |
| Body | 1880 MHz | $\epsilon_r = 53.3 \pm 5\%$ | $\sigma = 1.52 \pm 5\%$ mho/m |
| Body | 1800 MHz | $\epsilon_r = 53.3 \pm 5\%$ | $\sigma = 1.52 \pm 5\%$ mho/m |
| | ConvF X | 5.0 $\pm 9.5\%$ (k=2) | Boundary effect: |
| | ConvF Y | 5.0 $\pm 9.5\%$ (k=2) | Alpha 0.74 |
| | ConvF Z | 5.0 $\pm 9.5\%$ (k=2) | Depth 2.06 |

Deviation from Isotropy in HSL

Error (θ, ϕ), $f = 900$ MHz



Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Calibration Certificate

Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1505

Place of Calibration:

Zurich

Date of Calibration:

September 7, 2002

Calibration Interval:

12 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

N. Vetter

Approved by:

Alexander Kaya

Probe ET3DV6

SN:1505

| | |
|-------------------|-------------------|
| Manufactured: | October 24, 1999 |
| Last calibration: | May 22, 2002 |
| Repaired: | August 29, 2002 |
| Recalibrated: | September 7, 2002 |

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

DASY - Parameters of Probe: ET3DV6 SN:1505

Sensitivity in Free Space

| | |
|-------|---|
| NormX | 1.67 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormY | 1.69 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormZ | 1.68 $\mu\text{V}/(\text{V}/\text{m})^2$ |

Diode Compression

| | | |
|-------|-----------|----|
| DCP X | 94 | mV |
| DCP Y | 94 | mV |
| DCP Z | 94 | mV |

Sensitivity in Tissue Simulating Liquid

| | | | |
|---------|------------------------------|-----------------------------|-------------------------------|
| Head | 835 MHz | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.90 \pm 5\%$ mho/m |
| Head | 900 MHz | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.97 \pm 5\%$ mho/m |
| ConvF X | 7.0 $\pm 9.5\%$ (k=2) | Boundary effect: | |
| ConvF Y | 7.0 $\pm 9.5\%$ (k=2) | Alpha | 0.21 |
| ConvF Z | 7.0 $\pm 9.5\%$ (k=2) | Depth | 3.66 |
| Head | 1880 MHz | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\%$ mho/m |
| Head | 1800 MHz | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\%$ mho/m |
| ConvF X | 5.6 $\pm 9.5\%$ (k=2) | Boundary effect: | |
| ConvF Y | 5.6 $\pm 9.5\%$ (k=2) | Alpha | 0.43 |
| ConvF Z | 5.6 $\pm 9.5\%$ (k=2) | Depth | 2.63 |

Boundary Effect



| | | | |
|--|-----------------|-----------------------------------|-------------|
| Head | 835 MHz | Typical SAR gradient: 5 % per mm | |
| Probe Tip to Boundary | | 1 mm | 2 mm |
| SAR _{be} [%] Without Correction Algorithm | | 9.2 | 5.6 |
| SAR _{be} [%] With Correction Algorithm | | 0.6 | 0.6 |
| Head | 1880 MHz | Typical SAR gradient: 10 % per mm | |
| Probe Tip to Boundary | | 1 mm | 2 mm |
| SAR _{be} [%] Without Correction Algorithm | | 11.4 | 8.1 |
| SAR _{be} [%] With Correction Algorithm | | 0.3 | 0.3 |

Sensor Offset



| | | |
|----------------------------|---------------------------------|----|
| Probe Tip to Sensor Center | 2.7 | mm |
| Optical Surface Detection | 1.4 \pm 0.2 | mm |

Receiving Pattern (ϕ), $\theta = 0^\circ$



f = 30 MHz, TEM cell ifi110





f = 100 MHz, TEM cell ifi110





f = 300 MHz, TEM cell ifi110




f = 900 MHz, TEM cell ifi110






Isotropy Error (ϕ), $\theta = 0^\circ$





Frequency Response of E-Field


(TEM-Cell:ifi110, Waveguide R22)



Dynamic Range $f(\text{SAR}_{\text{brain}})$ (Waveguide R22)





Conversion Factor Assessment



| | | | |
|-------------|-----------------|------------------------------|-------------------------------|
| Head | 835 MHz | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.90 \pm 5\%$ mho/m |
| Head | 900 MHz | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.97 \pm 5\%$ mho/m |
| | ConvF X | 7.0 $\pm 9.5\%$ (k=2) | Boundary effect: |
| | ConvF Y | 7.0 $\pm 9.5\%$ (k=2) | Alpha 0.21 |
| | ConvF Z | 7.0 $\pm 9.5\%$ (k=2) | Depth 3.66 |
| | | | |
| Head | 1880 MHz | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\%$ mho/m |
| Head | 1800 MHz | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\%$ mho/m |
| | ConvF X | 5.6 $\pm 9.5\%$ (k=2) | Boundary effect: |
| | ConvF Y | 5.6 $\pm 9.5\%$ (k=2) | Alpha 0.43 |
| | ConvF Z | 5.6 $\pm 9.5\%$ (k=2) | Depth 2.63 |


Conversion Factor Assessment



| | | | |
|-------------|-----------------|------------------------------|-------------------------------|
| Body | 835 MHz | $\epsilon_r = 55.2 \pm 5\%$ | $\sigma = 0.97 \pm 5\%$ mho/m |
| Body | 900 MHz | $\epsilon_r = 55.0 \pm 5\%$ | $\sigma = 1.05 \pm 5\%$ mho/m |
| | ConvF X | 6.7 $\pm 9.5\%$ (k=2) | Boundary effect: |
| | ConvF Y | 6.7 $\pm 9.5\%$ (k=2) | Alpha 0.30 |
| | ConvF Z | 6.7 $\pm 9.5\%$ (k=2) | Depth 2.69 |
| | | | |
| Body | 1880 MHz | $\epsilon_r = 53.3 \pm 5\%$ | $\sigma = 1.52 \pm 5\%$ mho/m |
| Body | 1800 MHz | $\epsilon_r = 53.3 \pm 5\%$ | $\sigma = 1.52 \pm 5\%$ mho/m |
| | ConvF X | 5.0 $\pm 9.5\%$ (k=2) | Boundary effect: |
| | ConvF Y | 5.0 $\pm 9.5\%$ (k=2) | Alpha 0.58 |
| | ConvF Z | 5.0 $\pm 9.5\%$ (k=2) | Depth 2.32 |

Deviation from Isotropy in HSL

Error (θ, ϕ), $f = 900$ MHz



Client **Nokia Inc. Texas**

CALIBRATION CERTIFICATE

Object(s) **D835V2 - SN:486**

Calibration procedure(s) **QA CAL-05.v2
Calibration procedure for dipole validation kits**

Calibration date: **May 26, 2003**



Condition of the calibrated item **In Tolerance (according to the specific calibration document)**

This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.

All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.

Calibration Equipment used (M&TE critical for calibration)

| Model Type | ID # | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration |
|---------------------------|------------|---|------------------------|
| RF generator R&S SML-03 | 100698 | 27-Mar-2002 (R&S, No. 20-92389) | In house check: Mar-05 |
| Power sensor HP 8481A | MY41092317 | 18-Oct-02 (Agilent, No. 20021018) | Oct-04 |
| Power sensor HP 8481A | US37292783 | 30-Oct-02 (METAS, No. 252-0236) | Oct-03 |
| Power meter EPM E442 | GB37480704 | 30-Oct-02 (METAS, No. 252-0236) | Oct-03 |
| Network Analyzer HP 8753E | US38432426 | 3-May-00 (Agilent, No. 8702K064602) | In house check: May 03 |

| | Name | Function | Signature |
|----------------|----------------|---------------------|---|
| Calibrated by: | Judith Mueller | Technician |  |
| Approved by: | Katja Pokovic | Laboratory Director |  |

Date issued: May 26, 2003

This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.

DASY

Dipole Validation Kit

Type: D835V2

Serial: 486

Manufactured: May 19, 2003

Calibrated: May 26, 2003

1. Measurement Conditions

The measurements were performed in the flat section of the SAM twin phantom filled with head simulating solution of the following electrical parameters at 835 MHz:

| | | |
|------------------------|-------------------|-----------|
| Relative Dielectricity | 42.8 | $\pm 5\%$ |
| Conductivity | 0.89 mho/m | $\pm 5\%$ |

The DASY4 System with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.7 at 835 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 7x7x7 fine cube was chosen for cube integration.

The dipole input power (forward power) was 250 mW $\pm 3\%$. The results are normalized to 1W input power.

2. SAR Measurement with DASY4 System

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

averaged over 1 cm³ (1 g) of tissue: **9.80 mW/g** $\pm 16.8\%$ (k=2)¹

averaged over 10 cm³ (10 g) of tissue: **6.40 mW/g** $\pm 16.2\%$ (k=2)¹

¹ validation uncertainty

Date/Time: 05/26/03 17:23:08

Test Laboratory: SPEAG, Zurich, Switzerland
 File Name: SN486_SN1507_HSL835_260503.da4

DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN486
Program: Dipole Calibration

Communication System: CW-835; Frequency: 835 MHz; Duty Cycle: 1:1

Medium: HSL 835 MHz ($\sigma = 0.89$ mho/m, $\epsilon_r = 42.8$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(6.7, 6.7, 6.7); Calibrated: 1/18/2003
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 - SN411; Calibrated: 1/16/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006
- Measurement SW: DASY4, V4.1 Build 47; Postprocessing SW: SEMCAD, V1.6 Build 115

Pin = 250 mW; d = 15 mm/Area Scan (81x81x1): Measurement grid: dx=15mm, dy=15mm

Reference Value = 56.8 V/m

Power Drift = -0.004 dB

Maximum value of SAR = 2.61 mW/g

Pin = 250 mW; d = 15 mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm



Peak SAR (extrapolated) = 3.56 W/kg

SAR(1 g) = 2.45 mW/g; SAR(10 g) = 1.6 mW/g

Reference Value = 56.8 V/m

Power Drift = -0.004 dB

Maximum value of SAR = 2.61 mW/g




0 dB = 2.61mW/g

26 May 2003 14:28:05


[CH1] S11 1 U FS 1: 50.514 Ω -2.8906 Ω 65.939 pF 835.000 000 MHz

PRm
Cor
Avg
15
↑



CH2 S11 LOG 4 dB/REF -20 dB 1: -38.690 dB 835.000 000 MHz

PRm
Cor
↑



CENTER 835.000 000 MHz

SPAN 400.000 000 MHz

Calibration Certificate

835 MHz System Validation Dipole

Type:

D835V2

Serial Number:

455

Place of Calibration:

Zurich

Date of Calibration:

July 16, 2002

Calibration Interval:

24 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

N. Vetter

Approved by:

Alconic Klatza

DASY

Dipole Validation Kit

Type: D835V2

Serial: 455

Manufactured: January 31, 2002
Calibrated: July 16, 2002

1. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with head simulating solution of the following electrical parameters at 835 MHz:

| | | |
|------------------------|-------------------|-----------|
| Relative Dielectricity | 42.5 | $\pm 5\%$ |
| Conductivity | 0.90 mho/m | $\pm 5\%$ |

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.6 at 835 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW $\pm 3\%$. The results are normalized to 1W input power.

2.1. SAR Measurement with DASY3 System

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the worst-case extrapolation are:

averaged over 1 cm³ (1 g) of tissue: **9.84 mW/g**

averaged over 10 cm³ (10 g) of tissue: **6.32 mW/g**

2.2 SAR Measurement with DASY4 System

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

averaged over 1 cm³ (1 g) of tissue: **9.20 mW/g**

averaged over 10 cm³ (10 g) of tissue: **6.08 mW/g**

3. Dipole Impedance and Return Loss

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

| | | |
|----------------------|-----------------|---------------------------------------|
| Electrical delay: | 1.375 ns | (one direction) |
| Transmission factor: | 0.992 | (voltage transmission, one direction) |

The dipole was positioned at the flat phantom sections according to section 1 and the distance holder was in place during impedance measurements.

| | |
|---------------------------------|--------------------------------|
| Feedpoint impedance at 835 MHz: | $\text{Re}\{Z\} = 49.6 \Omega$ |
| | $\text{Im}\{Z\} = -1.8 \Omega$ |
| Return Loss at 835 MHz | -34.7 dB |

4. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with body simulating solution of the following electrical parameters at 835 MHz:

| | | |
|------------------------|-------------------|-----------|
| Relative Dielectricity | 55.3 | $\pm 5\%$ |
| Conductivity | 0.95 mho/m | $\pm 5\%$ |

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.2 at 835 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW $\pm 3\%$. The results are normalized to 1W input power.

5.1. SAR Measurement with DASY3 System

Standard SAR-measurements were performed according to the measurement conditions described in section 4. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the worst-case extrapolation are:

averaged over 1 cm³ (1 g) of tissue: **10.1 mW/g**

averaged over 10 cm³ (10 g) of tissue: **6.60 mW/g**

5.2 SAR Measurement with DASY4 System

Standard SAR-measurements were performed according to the measurement conditions described in section 4. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

averaged over 1 cm³ (1 g) of tissue: **9.24 mW/g**

averaged over 10 cm³ (10 g) of tissue: **6.20 mW/g**

6. Dipole Impedance and Return Loss

The dipole was positioned at the flat phantom sections according to section 4 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 835 MHz: **Re{Z} = 45.6 Ω**

Im {Z} = -4.3 Ω

Return Loss at 835 MHz **-23.7 dB**

4. Handling

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

5. Design

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

6. Power Test

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Validation Dipole D835V2 SN455, d = 15 mm

Frequency: 835 MHz; Antenna Input Power: 250 [mW]



SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0

Probe: ET3DV6 - SN1507; ConvF(6.60,6.60) at 835 MHz; IEEE1528 835 MHz: $\sigma = 0.90 \text{ mho/m}$, $\epsilon_r = 42.5$, $\rho = 1.00 \text{ g/cm}^3$

Cubes (2): Peak: 3.84 mW/g \pm 0.02 dB, SAR (1g): 2.46 mW/g \pm 0.02 dB, SAR (10g): 1.58 mW/g \pm 0.01 dB, (Worst-case extrapolation)



Penetration depth: 12.1 (11.1, 13.5) [mm]

Powerdrift: 0.00 dB



Validation Dipole D835V2 SN455, d = 15 mm

Frequency: 835 MHz; Antenna Input Power: 250 [mW]
SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0
Probe: ET3DV6 - SN1507; ConvF(6.60,6.60,6.60) at 835 MHz; IEEE1528 835 MHz: $\sigma = 0.90$ mho/m $\epsilon_r = 42.5$ $\rho = 1.00$ g/cm³
Cubes (2): Peak: 3.40 mW/g ± 0.02 dB, SAR (1g): 2.30 mW/g ± 0.02 dB, SAR (10g): 1.52 mW/g ± 0.01 dB, (Advanced extrapolation)
Penetration depth: 13.1 (12.8, 13.6) [mm]
Powerdrift: 0.00 dB




De1

PRM

Cor
Avg
16


↑



CH2 S11 LOG 5 dB/REF 0 dB 1:-34.736 dB 835.000 000 MHz

PRM
Cor

↑



START 635.000 000 MHz

STOP 1 835.000 000 MHz

Validation Dipole D835V2 SN455, d = 15 mm

Frequency: 835 MHz; Antenna Input Power: 250 [mW]

SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0



Probe: ET3DV6 - SNI507; ConvF(6.20,6.20,6.20) at 835 MHz; IEEE1528 835 MHz: $\sigma = 0.95$ mho/m $\epsilon_r = 55.3$ $\rho = 1.00$ g/cm³

Cubes (2): Peak: 3.91 mW/g ± 0.01 dB, SAR (1g): 2.53 mW/g ± 0.01 dB, SAR (10g): 1.65 mW/g ± 0.01 dB, (Worst-case extrapolation)

Penetration depth: 12.7 (11.6, 14.2) [mm]



Powerdrift: 0.01 dB

SAR_{Tot} [mW/g]



Validation Dipole D835V2 SN455, d = 15 mm

Frequency: 835 MHz; Antenna Input Power: 250 [mW]
SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0
Probe: ET3DV6 - SN1507; ConvF(6.20,6.20,6.20) at 835 MHz; IEEE1528 835 MHz: $\sigma = 0.95 \text{ mho/m}$, $\epsilon_r = 55.3$, $\rho = 1.00 \text{ g/cm}^3$
Cubes (2): Peak: 3.30 mW/g \pm 0.01 dB, SAR (1g): 2.31 mW/g \pm 0.01 dB, SAR (10g): 1.55 mW/g \pm 0.01 dB, (Advanced extrapolation)
Penetration depth: 14.3 (14.2, 14.5) [mm]
Powerdrift: 0.01 dB




Del

PRM

Cor
Avg
16


↑



CH2 S11 LOG 5 dB/REF 0 dB 1:-23.713 dB 835.000 000 MHz

PRM
Cor

↑



START 635.000 000 MHz

STOP 1 835.000 000 MHz