


Duty cycle

Agilent 15:48:47 Mar 7, 2007

R L



Total "ON TIME" = 17.5 ms + 10.5ms = 28 ms

$$DCCF = 20 * \log(28/100)$$

$$DCCF = - 11.05 \text{ dB}$$

Therefore, 10.5dB factor provided by client was used.