



Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Glossary

DAE	data acquisition electronics
Connector angle	information used in DASY system to align probe sensor X to the robot coordinate system.

Methods Applied and Interpretation of Parameters

- *DC Voltage Measurement:* Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- *Connector angle:* The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.
 - *DC Voltage Measurement Linearity:* Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
 - *Common mode sensitivity:* Influence of a positive or negative common mode voltage on the differential measurement.
 - *Channel separation:* Influence of a voltage on the neighbor channels not subject to an input voltage.
 - *AD Converter Values with inputs shorted:* Values on the internal AD converter corresponding to zero input voltage
 - *Input Offset Measurement:* Output voltage and statistical results over a large number of zero voltage measurements.
 - *Input Offset Current:* Typical value for information; Maximum channel input offset current, not considering the input resistance.
 - *Input resistance:* Typical value for information: DAE input resistance at the connector, during internal auto-zeroing and during measurement.
 - *Low Battery Alarm Voltage:* Typical value for information. Below this voltage, a battery alarm signal is generated.
 - *Power consumption:* Typical value for information. Supply currents in various operating modes.

DC Voltage Measurement

A/D - Converter Resolution nominal

High Range: 1LSB = $6.1\mu V$, full range = $-100...+300 mV$

Low Range: 1LSB = $61nV$, full range = $-1.....+3mV$

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Calibration Factors	X	Y	Z
High Range	$405.212 \pm 0.02\% (k=2)$	$405.495 \pm 0.02\% (k=2)$	$404.992 \pm 0.02\% (k=2)$
Low Range	$4.00132 \pm 1.50\% (k=2)$	$4.00055 \pm 1.50\% (k=2)$	$4.01173 \pm 1.50\% (k=2)$

Connector Angle

Connector Angle to be used in DASY system	$316.5^\circ \pm 1^\circ$
---	---------------------------

Appendix (Additional assessments outside the scope of SCS0108)

1. DC Voltage Linearity

High Range		Reading (μ V)	Difference (μ V)	Error (%)
Channel X	+ Input	200030.95	-10.22	-0.01
Channel X	+ Input	20006.14	0.69	0.00
Channel X	- Input	-20003.24	2.19	-0.01
Channel Y	+ Input	200030.42	-10.80	-0.01
Channel Y	+ Input	20003.36	-1.94	-0.01
Channel Y	- Input	-20006.62	-1.11	0.01
Channel Z	+ Input	200032.70	-0.21	-0.00
Channel Z	+ Input	20002.68	-2.63	-0.01
Channel Z	- Input	-20006.13	-0.53	0.00

Low Range		Reading (μ V)	Difference (μ V)	Error (%)
Channel X	+ Input	2001.20	-0.10	-0.01
Channel X	+ Input	200.69	-0.63	-0.31
Channel X	- Input	-198.00	0.79	-0.40
Channel Y	+ Input	2001.03	-0.13	-0.01
Channel Y	+ Input	200.67	-0.54	-0.27
Channel Y	- Input	-199.46	-0.59	0.30
Channel Z	+ Input	2001.23	0.11	0.01
Channel Z	+ Input	200.07	-1.08	-0.54
Channel Z	- Input	-199.93	-0.98	0.49

2. Common mode sensitivity

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Common mode Input Voltage (mV)	High Range Average Reading (μ V)	Low Range Average Reading (μ V)
Channel X	200	0.85	-1.33
	-200	2.39	0.51
Channel Y	200	20.36	19.83
	-200	-23.04	-22.89
Channel Z	200	-19.87	-20.86
	-200	19.46	19.34

3. Channel separation

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X (μ V)	Channel Y (μ V)	Channel Z (μ V)
Channel X	200	-	2.15	-3.05
Channel Y	200	6.25	-	4.13
Channel Z	200	7.88	3.65	-

4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	16193	15359
Channel Y	15873	13390
Channel Z	16219	17704

5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input 10MΩ

	Average (μ V)	min. Offset (μ V)	max. Offset (μ V)	Std. Deviation (μ V)
Channel X	0.28	-0.85	1.34	0.49
Channel Y	-1.28	-2.53	0.12	0.49
Channel Z	-0.63	-2.60	0.79	0.52

6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+6	+14
Supply (- Vcc)	-0.01	-8	-9