

# FCC SAR Test Report

APPLICANT	:	Acer Incorporated
EQUIPMENT	:	Tablet Computer
BRAND NAME	:	Acer
MODEL NAME	:	A100
FCC ID	:	HLZTMDMA100
STANDARD	:	FCC 47 CFR Part 2 (2.1093)
		IEEE C95.1-1991
		IEEE 1528-2003
		FCC OET Bulletin 65 Supplement C (Edition 01-01)

The product was received on Mar. 23, 2011 and completely tested on May. 25, 2011. We, SPORTON INTERNATIONAL INC., would like to declare that the tested sample has been evaluated in accordance with the procedures and shown the compliance with the applicable technical standards.

The test results in this report apply exclusively to the tested model / sample. Without written approval of SPORTON INTERNATIONAL INC., the test report shall not be reproduced except in full.

**Reviewed by:** 

Jones Tsai / Manager



# SPORTON INTERNATIONAL INC.

No. 52, Hwa Ya 1<sup>st</sup> Rd., Hwa Ya Technology Park, Kwei-Shan Hsiang, Tao Yuan Hsien, Taiwan, R.O.C.

**SPORTON INTERNATIONAL INC.** TEL : 886-3-327-3456 FAX : 886-3-328-4978 FCC ID : HLZTMDMA100

Page Number	: 1 of 31		
Report Issued Date	: Jun. 03, 2011		
Report Version	: Rev. 02		



# Table of Contents

Re۱		listory	
1.		nent of Compliance	
2.	Admir	nistration Data	5
	2.1	Testing Laboratory	5
	2.2	Applicant	
	2.3	Manufacturer	
	2.4	Application Details	
3.	Gener	al Information	
	3.1	Description of Device Under Test (DUT)	6
	3.2	Product Photos	
	3.3	Applied Standards	
	3.4	Device Category and SAR Limits	7
	3.5	Test Conditions	
		3.5.1 Ambient Condition	
		3.5.2 Test Configuration	
4.	Speci	fic Absorption Rate (SAR)	8
	4.1	Introduction	
	4.2	SAR Definition	
5.	SAR N	leasurement System	
	5.1	E-Field Probe	
		5.1.1 E-Field Probe Specification	
		5.1.2 E-Field Probe Calibration	.11
	5.2	Data Acquisition Electronics (DAE)	
	5.3	Robot	.11
	5.4	Measurement Server	
	5.5	Phantom	.13
	5.6	Device Holder	.14
	5.7	Data Storage and Evaluation	.16
		5.7.1 Data Storage	.16
		5.7.2 Data Evaluation	
	5.8	Test Equipment List	
6.	Tissu	e Simulating Liquids	.19
7.	Uncer	tainty Assessment	.21
8.	SAR N	leasurement Evaluation	.24
	8.1	Purpose of System Performance check	
	8.2	System Setup	.24
	8.3	Validation Results	.25
9.	DUT T	esting Position	.26
10.	Measu	Irement Procedures	
	10.1	Spatial Peak SAR Evaluation	
	10.2	Area & Zoom Scan Procedures	.28
	10.3	Volume Scan Procedures	
	10.4	SAR Averaged Methods	.28
	10.5	Power Drift Monitoring	.28
11.	SAR 1	est Results	
	11.1	SAR Test Setup	
	11.2	Conducted Average Power (Unit: dBm)	
	11.3	Test Records for Mini-Tablet SAR Test	.30

Appendix A. Plots of System Performance Check Appendix B. Plots of SAR Measurement Appendix C. DASY Calibration Certificate Appendix D. Product Photos Appendix E. Test Setup Photos

**SPORTON INTERNATIONAL INC.** TEL : 886-3-327-3456 FAX : 886-3-328-4978 FCC ID : HLZTMDMA100



# **Revision History**

REPORT NO.	VERSION	DESCRIPTION	ISSUED DATE
FA132346	Rev. 01	Initial issue of report	May 26, 2011
FA132346	Rev. 02	Revised descriptions at section 11	Jun. 03, 2011



# 1. Statement of Compliance

The maximum results of Specific Absorption Rate (SAR) found during testing for **Acer Incorporated Tablet Computer Acer A100** are as follows (with expanded uncertainty 21.4 % for 300 MHz to 3 GHz, and 25.6% for 3 GHz to 6 GHz).

Band	Position	SAR <sub>1g</sub> (W/kg)	
802.11b/g/n	Body	0.796	
Bluetooth	Body	N/A	

This device is in compliance with Specific Absorption Rate (SAR) for general population/uncontrolled exposure limits (1.6 W/kg) specified in FCC 47 CFR part 2 (2.1093) and ANSI/IEEE C95.1-1991, and had been tested in accordance with the measurement methods and procedures specified in IEEE 1528-2003 and FCC OET Bulletin 65 Supplement C (Edition 01-01).



# 2. Administration Data

# 2.1 Testing Laboratory

Test Site	SPORTON INTERNATIONAL INC.
Test Site Location	No. 52, Hwa Ya 1 <sup>st</sup> Rd., Hwa Ya Technology Park, Kwei-Shan Hsiang, Tao Yuan Hsien, Taiwan, R.O.C. TEL: +886-3-327-3456 FAX: +886-3-328-4978

# 2.2 Applicant

Company Name	Acer Incorporated
	8F., No.88, Sec. 1, Xintai 5th. Rd., Xizhi Dist., New Taipei City 22181, Taiwan (R.O.C)

# 2.3 Manufacturer

Company Name	<ol> <li>(1) Compal Electronics, Inc</li> <li>(2) Compal Electronics Technology (Kunshan) Co., Ltd.</li> <li>(3) Compal Information (Kunshan) Co., Ltd.</li> <li>(4) Compal Information Technology (Kunshan) Co., Ltd.</li> <li>(5) Compalead Eletrônica Do Brasil Indústria E Comércio Ltda</li> <li>(6) Compal (Vietnam) Co., Ltd.</li> </ol>	
Address	<ul> <li>(1)No. 581, Ruiguang Rd., Neihu District, Taipei City 11492, Taiwan</li> <li>(2)No. 25, Third Avenue, A Zone, Kunshan Comprehensive Free Trade Zone, Kunshan, Jiangsu, China</li> <li>(3)No. 15, Third Avenue, A Zone, Kunshan Comprehensive Free Trade Zone, Kunshan, Jiangsu, China</li> <li>(4) No. 58, First Avenue, A Zone, Kunshan Comprehensive Free Trade Zone, Kunshan, Jiangsu, China</li> <li>(5)Rua Kanebo 175, Galpões C1, C2, C3, C4, C5 C6 E C12, Bairro Distrito Industrial Jundiaí Business Park, Cep 13213-090, Jundiaí - São Paulo, Brasil</li> <li>(6)Ba Thien Industrial Zone, Ba Hien Commune, Binh Xuyen County, Vinh Phuc Province, Vietnam</li> </ul>	

# 2.4 Application Details

Date of Receipt of Application	Mar. 23, 2011
Date of Start during the Test	May. 25, 2011
Date of End during the Test	May. 25, 2011



# 3. General Information

# 3.1 Description of Device Under Test (DUT)

Product Feature & Specification			
DUT Type	Tablet Computer		
Brand Name	Acer		
Model Name	A100		
FCC ID	HLZTMDMA100		
Sample 1	EUT with LP DDR2 Main and eMMC Main		
Sample 2	EUT with LP DDR2 2nd and eMMC 2nd		
Sample 3	EUT with LP DDR2 3rd and eMMC Main		
Tx Frequency	802.11b/g/n : 2400 MHz ~ 2483.5 MHz Bluetooth : 2400 MHz ~ 2483.5 MHz		
Rx Frequency	802.11b/g/n : 2400 MHz ~ 2483.5 MHz Bluetooth : 2400 MHz ~ 2483.5 MHz		
Maximum Output Power to Antenna	802.11b : 14.86 dBm 802.11g : 13.73 dBm 802.11n (2.4GHz) : 11.13 dBm (BW 20MHz)		
Antenna Type	PIFA Antenna		
HW Version	LA7251PR10 Rev. 1.0		
SW Version	NA		
Type of Modulation	802.11b : DSSS (BPSK / QPSK / CCK) 802.11g/n : OFDM (BPSK / QPSK / 16QAM / 64QAM) Bluetooth (1Mbps) : GFSK Bluetooth EDR (2Mbps) : π/4-DQPSK		
	Bluetooth EDR (3Mbps) : 8-DPSK		
DUT Stage	Identical Prototype		

**Remark:** The above DUT's information was declared by manufacturer. Please refer to the specifications or user's manual for more detailed description.



# 3.2 Product Photos

Please refer to Appendix D.

# 3.3 Applied Standards

The Specific Absorption Rate (SAR) testing specification, method and procedure for this device is in accordance with the following standards:

- FCC 47 CFR Part 2 (2.1093)
- IEEE C95.1-1991
- IEEE 1528-2003
- FCC OET Bulletin 65 Supplement C (Edition 01-01)
- FCC KDB 248227 D01 v01r02
- FCC KDB 447498 D01 v04
- FCC KDB 616217 D03 v01

# 3.4 Device Category and SAR Limits

This device belongs to portable device category because its radiating structure is allowed to be used within 20 centimeters of the body of the user. Limit for General Population/Uncontrolled exposure should be applied for this device, it is 1.6 W/kg as averaged over any 1 gram of tissue.

### 3.5 Test Conditions

#### 3.5.1 Ambient Condition

Ambient Temperature	<b>20 to 24</b> °C
Humidity	< 60 %

#### 3.5.2 Test Configuration

For WLAN SAR testing, WLAN engineering testing software installed on the DUT can provide continuous transmitting RF signal. This RF signal utilized in SAR measurement has almost 100% duty cycle and its crest factor is 1.



# 4. Specific Absorption Rate (SAR)

### 4.1 Introduction

SAR is related to the rate at which energy is absorbed per unit mass in an object exposed to a radio field. The SAR distribution in a biological body is complicated and is usually carried out by experimental techniques or numerical modeling. The standard recommends limits for two tiers of groups, occupational/controlled and general population/uncontrolled, based on a person's awareness and ability to exercise control over his or her exposure. In general, occupational/controlled exposure limits are higher than the limits for general population/uncontrolled.

# 4.2 SAR Definition

The SAR definition is the time derivative (rate) of the incremental energy (dW) absorbed by (dissipated in) an incremental mass (dm) contained in a volume element (dv) of a given density ( $\rho$ ). The equation description is as below:

$$SAR = \frac{d}{dt} \left( \frac{dW}{dm} \right) = \frac{d}{dt} \left( \frac{dW}{\rho dv} \right)$$

SAR is expressed in units of Watts per kilogram (W/kg)

SAR measurement can be either related to the temperature elevation in tissue by

$$SAR = C\left(\frac{\delta T}{\delta t}\right)$$

Where: C is the specific heat capacity,  $\delta T$  is the temperature rise and  $\delta t$  is the exposure duration, or related to the electrical field in the tissue by


$$SAR = \frac{\sigma |E|^2}{\rho}$$

Where:  $\zeta$  is the conductivity of the tissue,  $\rho$  is the mass density of the tissue and E is the RMS electrical field strength.

However for evaluating SAR of low power transmitter, electrical field measurement is typically applied.



# 5. SAR Measurement System



### Fig 5.1 SPEAG DASY4 or DASY5 System Configurations

The DASY4 or DASY5 system for performance compliance tests is illustrated above graphically. This system consists of the following items:

- > A standard high precision 6-axis robot with controller, a teach pendant and software
- A data acquisition electronic (DAE) attached to the robot arm extension
- A dosimetric probe equipped with an optical surface detector system
- > The electro-optical converter (ECO) performs the conversion between optical and electrical signals
- A measurement server performs the time critical tasks such as signal filtering, control of the robot operation and fast movement interrupts.
- > A probe alignment unit which improves the accuracy of the probe positioning
- > A computer operating Windows XP
- DASY4 or DASY5 software
- Remove control with teach pendant and additional circuitry for robot safety such as warming lamps, etc.
- > The SAM twin phantom
- A device holder
- > Tissue simulating liquid
- > Dipole for evaluating the proper functioning of the system

Some of the components are described in details in the following sub-sections.



# 5.1 E-Field Probe

The SAR measurement is conducted with the dosimetric probe (manufactured by SPEAG). The probe is specially designed and calibrated for use in liquid with high permittivity. The dosimetric probe has special calibration in liquid at different frequency. This probe has a built in optical surface detection system to prevent from collision with phantom.

### 5.1.1 E-Field Probe Specification

<et3dv6></et3dv6>				_
Construction	Symmetrical design with triangular core Built-in optical fiber for surface detection system. Built-in shielding against static charges. PEEK enclosure material (resistant to organic solvents, e.g., DGBE)			
Frequency	10 MHz to 3 GHz; Linearity: ± 0.2 dB			
Directivity	± 0.2 dB in HSL (rotation around probe axis) ± 0.4 dB in HSL (rotation normal to probe axis)		2006 View	
Dynamic Range	5 μW/g to 100 mW/g; Linearity: ± 0.2 dB			
Dimensions	Overall length: 330 mm (Tip: 16 mm) Tip diameter: 6.8 mm (Body: 12 mm) Distance from probe tip to dipole centers: 2.7 mm	Fig 5	.2 Photo	of ET3DV6

#### <EX3DV4 Probe>

<ex3dv4 probe=""></ex3dv4>			
Construction	Symmetrical design with triangular core Built-in shielding against static charges PEEK enclosure material (resistant to organic solvents, e.g., DGBE)		<b>.</b>
Frequency	10 MHz to 6 GHz; Linearity: ± 0.2 dB		
Directivity	± 0.3 dB in HSL (rotation around probe axis) ± 0.5 dB in tissue material (rotation normal to probe axis)		2014
Dynamic Range	10 μW/g to 100 mW/g; Linearity: ± 0.2 dB (noise: typically < 1 μW/g)		
Dimensions	Overall length: 330 mm (Tip: 20 mm) Tip diameter: 2.5 mm (Body: 12 mm) Typical distance from probe tip to dipole centers: 1 mm		
		Fig 5.3	Photo of EX3DV4



#### 5.1.2 E-Field Probe Calibration

Each probe needs to be calibrated according to a dosimetric assessment procedure with accuracy better than  $\pm$  10%. The spherical isotropy shall be evaluated and within  $\pm$  0.25 dB. The sensitivity parameters (NormX, NormY, and NormZ), the diode compression parameter (DCP) and the conversion factor (ConvF) of the probe are tested. The calibration data can be referred to appendix C of this report.

### 5.2 Data Acquisition Electronics (DAE)

The data acquisition electronics (DAE) consists of a highly sensitive electrometer-grade preamplifier with auto-zeroing, a channel and gain-switching multiplexer, a fast 16 bit AD-converter and a command decoder and control logic unit. Transmission to the measurement server is accomplished through an optical downlink for data and status information as well as an optical uplink for commands and the clock.

The input impedance of the DAE is 200 MOhm; the inputs are symmetrical and floating. Common mode rejection is above 80 dB.



Fig 5.4 Photo of DAE

### 5.3<u>Robot</u>

The SPEAG DASY system uses the high precision robots (DASY4: RX90BL; DASY5: TX90XL) type from Stäubli SA (France). For the 6-axis controller system, the robot controller version (DASY4: CS7MB; DASY5: CS8c) from Stäubli is used. The Stäubli robot series have many features that are important for our application:

- ➢ High precision (repeatability ±0.035 mm)
- High reliability (industrial design)
- Jerk-free straight movements
- > Low ELF interference (the closed metallic construction shields against motor control fields)



**SPORTON INTERNATIONAL INC.** TEL : 886-3-327-3456 FAX : 886-3-328-4978 FCC ID : HLZTMDMA100 Page Number : 11 of 31

Report Issued Date : Jun. 03, 2011 Report Version : Rev. 02



### 5.4 Measurement Server

The measurement server is based on a PC/104 CPU board with CPU (DASY4: 166 MHz, Intel Pentium; DASY5: 400 MHz, Intel Celeron), chipdisk (DASY4: 32 MB; DASY5: 128 MB), RAM (DASY4: 64 MB, DASY5: 128 MB). The necessary circuits for communication with the DAE electronic box, as well as the 16 bit AD converter system for optical detection and digital I/O interface are contained on the DASY I/O board, which is directly connected to the PC/104 bus of the CPU board.

The measurement server performs all the real-time data evaluation for field measurements and surface detection, controls robot movements and handles safety operations.





Fig 5.7 Photo of Server for DASY4

Fig 5.8 Photo of Server for DASY5



# 5.5<u>Phantom</u>

#### <SAM Twin Phantom>

Shell Thickness	2 ± 0.2 mm;	
	Center ear point: 6 ± 0.2 mm	
Filling Volume	Approx. 25 liters	The second secon
Dimensions	Length: 1000 mm; Width: 500 mm;	
	Height: adjustable feet	Υ Y
Measurement Areas	Left Hand, Right Hand, Flat Phantom	
		Contraction of the second seco
		Fig 5.9 Photo of SAM Phantom

The bottom plate contains three pair of bolts for locking the device holder. The device holder positions are adjusted to the standard measurement positions in the three sections. A white cover is provided to tap the phantom during off-periods to prevent water evaporation and changes in the liquid parameters. On the phantom top, three reference markers are provided to identify the phantom position with respect to the robot.

#### <ELI4 Phantom>

Shell Thickness	2 ± 0.2 mm (sagging: <1%)	
Filling Volume	Approx. 30 liters	
Dimensions	Major ellipse axis: 600 mm Minor axis: 400 mm	Fig 5.10 Photo of ELI4 Phantom

The ELI4 phantom is intended for compliance testing of handheld and body-mounted wireless devices in the frequency range of 30 MHz to 6 GHz. ELI4 is fully compatible with standard and all known tissue simulating liquids.



### 5.6 Device Holder

#### <Device Holder for SAM Twin Phantom>

The SAR in the phantom is approximately inversely proportional to the square of the distance between the source and the liquid surface. For a source at 5 mm distance, a positioning uncertainty of  $\pm$  0.5 mm would produce a SAR uncertainty of  $\pm$  20 %. Accurate device positioning is therefore crucial for accurate and repeatable measurements. The positions in which the devices must be measured are defined by the standards.

The DASY device holder is designed to cope with different positions given in the standard. It has two scales for the device rotation (with respect to the body axis) and the device inclination (with respect to the line between the ear reference points). The rotation center for both scales is the ear reference point (EPR). Thus the device needs no repositioning when changing the angles.

The DASY device holder is constructed of low-loss POM material having the following dielectric parameters: relative permittivity  $\varepsilon$  = 3 and loss tangent  $\delta$  = 0.02. The amount of dielectric material has been reduced in the closest vicinity of the device, since measurements have suggested that the influence of the clamp on the test results could thus be lowered.



Fig 5.11 Device Holder



#### <Laptop Extension Kit>

The extension is lightweight and made of POM, acrylic glass and foam. It fits easily on the upper part of the mounting device in place of the phone positioned. The extension is fully compatible with the SAM Twin and ELI phantoms.




Fig 5.12 Laptop Extension Kit



### 5.7 Data Storage and Evaluation

#### 5.7.1 Data Storage

The DASY software stores the assessed data from the data acquisition electronics as raw data (in microvolt readings from the probe sensors), together with all the necessary software parameters for the data evaluation (probe calibration data, liquid parameters and device frequency and modulation data) in measurement files. The post-processing software evaluates the desired unit and format for output each time the data is visualized or exported. This allows verification of the complete software setup even after the measurement and allows correction of erroneous parameter settings. For example, if a measurement has been performed with an incorrect crest factor parameter in the device setup, the parameter can be corrected afterwards and the data can be reevaluated.

The measured data can be visualized or exported in different units or formats, depending on the selected probe type (e.g., [V/m], [A/m], [mW/g]). Some of these units are not available in certain situations or give meaningless results, e.g., a SAR-output in a non-lose media, will always be zero. Raw data can also be exported to perform the evaluation with other software packages.

#### 5.7.2 Data Evaluation

The DASY post-processing software (SEMCAD) automatically executes the following procedures to calculate the field units from the microvolt readings at the probe connector. The parameters used in the evaluation are stored in the configuration modules of the software :

Probe parameters :	- Sensitivity	Norm <sub>i</sub> , a <sub>i0</sub> , a <sub>i1</sub> , a <sub>i2</sub>
	<ul> <li>Conversion factor</li> </ul>	ConvFi
	<ul> <li>Diode compression point</li> </ul>	dcp <sub>i</sub>
Device parameters :	- Frequency	f
	- Crest factor	cf
Media parameters :	<ul> <li>Conductivity</li> </ul>	ζ
	- Density	ρ

These parameters must be set correctly in the software. They can be found in the component documents or they can be imported into the software from the configuration files issued for the DASY components. In the direct measuring mode of the multi-meter option, the parameters of the actual system setup are used. In the scan visualization and export modes, the parameters stored in the corresponding document files are used.

The first step of the evaluation is a linearization of the filtered input signal to account for the compression characteristics of the detector diode. The compensation depends on the input signal, the diode type and the DC-transmission factor from the diode to the evaluation electronics. If the exciting field is pulsed, the crest factor of the signal must be known to correctly compensate for peak power.



The formula for each channel can be given as :

$$V_i = U_i + U_i^2 \cdot \frac{cf}{dcp_i}$$

with  $V_i$  = compensated signal of channel i, (i = x, y, z)  $U_i$  = input signal of channel i, (i = x, y, z) cf = crest factor of exciting field (DASY parameter) dcp<sub>i</sub> = diode compression point (DASY parameter)

From the compensated input signals, the primary field data for each channel can be evaluated :

E-field Probes : 
$$\mathbf{E}_i = \sqrt{\frac{V_i}{Norm_i \cdot ConvF}}$$
  
H-field Probes :  $\mathbf{H}_i = \sqrt{V_i} \cdot \frac{\mathbf{a}_{i0} + \mathbf{a}_{i1}f + \mathbf{a}_{i2}f^2}{f}$ 

with  $V_i$  = compensated signal of channel i, (i = x, y, z) Norm<sub>i</sub> = sensor sensitivity of channel i, (i = x, y, z),  $\mu V/(V/m)^2$  for E-field Probes ConvF = sensitivity enhancement in solution  $a_{ij}$  = sensor sensitivity factors for H-field probes f = carrier frequency [GHz]  $E_i$  = electric field strength of channel i in V/m  $H_i$  = magnetic field strength of channel i in A/m

The RSS value of the field components gives the total field strength (Hermitian magnitude) :

$$\mathbf{E}_{\rm tot} = \sqrt{\mathbf{E}_{\rm x}^2 + \mathbf{E}_{\rm y}^2 + \mathbf{E}_{\rm z}^2}$$

The primary field data are used to calculate the derived field units.

$$SAR = E_{tot}^2 \cdot \frac{\sigma}{\rho \cdot 1000}$$

with SAR = local specific absorption rate in mW/g

 $E_{tot}$  = total field strength in V/m

 $\zeta$  = conductivity in [mho/m] or [Siemens/m]

 $\rho$  = equivalent tissue density in g/cm<sup>3</sup>

Note that the density is set to 1, to account for actual head tissue density rather than the density of the tissue simulating liquid.



# 5.8 Test Equipment List

Manufacturan	Name of Equipment	Type/Medal	Carial Number	Calibration	
Manufacturer	Name of Equipment	Type/Model	Serial Number	Last Cal.	Due Date
SPEAG	Dosimetric E-Field Probe	ES3DV3	3071	Jun. 22, 2010	Jun. 21, 2011
SPEAG	2450MHz System Validation Kit	D2450V2	735	Jun. 17, 2010	Jun. 16, 2012
SPEAG	Data Acquisition Electronics	DAE3	577	Jan. 13, 2011	Jan. 12, 2012
SPEAG	Data Acquisition Electronics	DAE4	778	Oct. 22, 2010	Oct. 21, 2011
SPEAG	Device Holder	N/A	N/A	NCR	NCR
SPEAG	SAM Phantom	QD 000 P40 C	TP-1303	NCR	NCR
SPEAG	SAM Phantom	QD 000 P40 C	TP-1383	NCR	NCR
SPEAG	SAM Phantom	QD 000 P40 C	TP-1446	NCR	NCR
SPEAG	SAM Phantom	QD 000 P40 C	TP-1478	NCR	NCR
SPEAG	ELI4 Phantom	QD 0VA 001 BB	1026	NCR	NCR
SPEAG	ELI4 Phantom	QD 0VA 001 BA	1029	NCR	NCR
Agilent	Network Analyzer	E5071C	MY46101588	Apr. 20, 2011	Apr. 19, 2012
Agilent	Wireless Communication Test Set	E5515C	MY48360820	Jan. 12, 2010	Jan. 11, 2012
Agilent	Wireless Communication Test Set	E5515C	GB46311322	Mar. 23, 2011	Mar. 22, 2013
R&S	Universal Radio Communication Tester	CMU200	114256	Feb. 08, 2010	Feb. 07, 2012
Agilent	Dielectric Probe Kit	85070D	US01440205	NCR	NCR
Agilent	Dual Directional Coupler	778D	50422	NCR	NCR
AR	Power Amplifier	5S1G4M2	0328767	NCR	NCR
R&S	Spectrum Analyzer	FSP40	100055	Jun. 11, 2010	Jun. 10, 2011

Table 5.1 Test Equipment List

Note: The calibration certificate of DASY can be referred to appendix C of this report.





# 6. Tissue Simulating Liquids

For the measurement of the field distribution inside the SAM phantom with DASY, the phantom must be filled with around 25 liters of homogeneous body tissue simulating liquid. For head SAR testing, the liquid height from the ear reference point (ERP) of the phantom to the liquid top surface is larger than 15 cm, which is shown in Fig. 6.1. For body SAR testing, the liquid height from the center of the flat phantom to the liquid top surface is larger than 15 cm, which is shown in Fig. 6.2.





Fig 6.1 Photo of Liquid Height for Head SAR

Fig 6.2 Photo of Liquid Height for Body SAR

Frequency	Water	Sugar	Cellulose	Salt	Preventol	DGBE	Conductivity	Permittivity	
(MHz)	(%)	(%)	(%)	(%)	(%)	(%)	(σ)	(ε <sub>r</sub> )	
	For Head								
835	40.3	57.9	0.2	1.4	0.2	0	0.90	41.5	
900	40.3	57.9	0.2	1.4	0.2	0	0.97	41.5	
1800, 1900, 2000	55.2	0	0	0.3	0	44.5	1.40	40.0	
2450	55.0	0	0	0	0	45.0	1.80	39.2	
				For Body					
835	50.8	48.2	0	0.9	0.1	0	0.97	55.2	
900	50.8	48.2	0	0.9	0.1	0	1.05	55.0	
1800, 1900, 2000	70.2	0	0	0.4	0	29.4	1.52	53.3	
2450	68.6	0	0	0	0	31.4	1.95	52.7	

The following table gives the recipes for tissue simulating liquid.

Table 6.1 Recipes of Tissue Simulating Liquid



Frequency (MHz)	Liquid Type	Conductivity (σ)	±5% Range	Permittivity (ε <sub>r</sub> )	±5% Range
2450	Head	1.80	1.71 ~ 1.89	39.2	37.2 ~ 41.2
5200	Head	4.66	4.43 ~ 4.89	36.0	34.2 ~ 37.8
5500	Head	4.96	4.71 ~ 5.21	35.6	33.8 ~ 37.4
5800	Head	5.27	5.01 ~ 5.53	35.3	33.5 ~ 37.1
2450	Body	1.95	1.85 ~ 2.05	52.7	50.1 ~ 55.3
5200	Body	5.30	5.04 ~ 5.57	49.0	46.6 ~ 51.5
5500	Body	5.65	5.37 ~ 5.93	48.6	46.2 ~ 51.0
5800	Body	6.00	5.70 ~ 6.30	48.2	45.8 ~ 50.6

The following table gives the targets for tissue simulating liquid.

Table 6.2 Targets of Tissue Simulating Liquid
---

The dielectric parameters of the liquids were verified prior to the SAR evaluation using an Agilent 85070D Dielectric Probe Kit and an Agilent Network Analyzer.

The following table shows the measuring results for simulating liquid.

Frequency	Liquid	Temperature	Conductivity	Permittivity	Measurement
(MHz)	Type	(℃)	(σ)	(ε <sub>r</sub> )	Date
2450	Body	21.5	1.97	53.8	May. 25, 2011

Table 6.3 Measuring Results for Simulating Liquid



# 7. Uncertainty Assessment

The component of uncertainly may generally be categorized according to the methods used to evaluate them. The evaluation of uncertainly by the statistical analysis of a series of observations is termed a Type An evaluation of uncertainty. The evaluation of uncertainty by means other than the statistical analysis of a series of observation is termed a Type B evaluation of uncertainty. Each component of uncertainty, however evaluated, is represented by an estimated standard deviation, termed standard uncertainty, which is determined by the positive square root of the estimated variance.

A Type A evaluation of standard uncertainty may be based on any valid statistical method for treating data. This includes calculating the standard deviation of the mean of a series of independent observations; using the method of least squares to fit a curve to the data in order to estimate the parameter of the curve and their standard deviations; or carrying out an analysis of variance in order to identify and quantify random effects in certain kinds of measurement.

A type B evaluation of standard uncertainty is typically based on scientific judgment using all of the relevant information available. These may include previous measurement data, experience and knowledge of the behavior and properties of relevant materials and instruments, manufacture's specification, data provided in calibration reports and uncertainties assigned to reference data taken from handbooks. Broadly speaking, the uncertainty is either obtained from an outdoor source or obtained from an assumed distribution, such as the normal distribution, rectangular or triangular distributions indicated in Table 7.1

Uncertainty Distributions	Normal	Rectangular	Triangular	U-Shape
Multi-plying Factor <sup>(a)</sup>	1/k <sup>(b)</sup>	1/√3	1/√6	1/√2

(a) standard uncertainty is determined as the product of the multiplying factor and the estimated range of variations in the measured quantity

(b)  $\kappa$  is the coverage factor

#### Table 7.1 Standard Uncertainty for Assumed Distribution

The combined standard uncertainty of the measurement result represents the estimated standard deviation of the result. It is obtained by combining the individual standard uncertainties of both Type A and Type B evaluation using the usual "root-sum-squares" (RSS) methods of combining standard deviations by taking the positive square root of the estimated variances.

Expanded uncertainty is a measure of uncertainty that defines an interval about the measurement result within which the measured value is confidently believed to lie. It is obtained by multiplying the combined standard uncertainty by a coverage factor. Typically, the coverage factor ranges from 2 to 3. Using a coverage factor allows the true value of a measured quantity to be specified with a defined probability within the specified uncertainty range. For purpose of this document, a coverage factor two is used, which corresponds to confidence interval of about 95 %. The DASY uncertainty Budget is showed in Table 7.2 and Table 7.3.



Error Description	Uncertainty Value (±%)	Probability Distribution	Divisor	Ci (1g)	Standard Uncertainty (1g)	
Measurement System	-	-	-	-		
Probe Calibration	5.5	Normal	1	1	± 5.5 %	
Axial Isotropy	4.7	Rectangular	√3	0.7	± 1.9 %	
Hemispherical Isotropy	9.6	Rectangular	√3	0.7	± 3.9 %	
Boundary Effects	1.0	Rectangular	√3	1	± 0.6 %	
Linearity	4.7	Rectangular	√3	1	± 2.7 %	
System Detection Limits	1.0	Rectangular	√3	1	± 0.6 %	
Readout Electronics	0.3	Normal	1	1	± 0.3 %	
Response Time	0.8	Rectangular	√3	1	± 0.5 %	
Integration Time	2.6	Rectangular	√3	1	± 1.5 %	
RF Ambient Noise	3.0	Rectangular	√3	1	± 1.7 %	
RF Ambient Reflections	3.0	Rectangular	√3	1	± 1.7 %	
Probe Positioner	0.4	Rectangular	√3	1	± 0.2 %	
Probe Positioning	2.9	Rectangular	√3	1	± 1.7 %	
Max. SAR Eval.	1.0	Rectangular	√3	1	± 0.6 %	
Test Sample Related						
Device Positioning	2.9	Normal	1	1	± 2.9 %	
Device Holder	3.6	Normal	1	1	± 3.6 %	
Power Drift	5.0	Rectangular	√3	1	± 2.9 %	
Phantom and Setup						
Phantom Uncertainty	4.0	Rectangular	√3	1	± 2.3 %	
Liquid Conductivity (Target)	5.0	Rectangular	√3	0.64	± 1.8 %	
Liquid Conductivity (Meas.)	2.5	Normal	1	0.64	± 1.6 %	
Liquid Permittivity (Target)	5.0	Rectangular	√3	0.6	± 1.7 %	
Liquid Permittivity (Meas.)	2.5	Normal	1	0.6	± 1.5 %	
Combined Standard Uncerta	inty				± 10.7 %	
Coverage Factor for 95 %					K = 2	
Expanded Uncertainty						

Table 7.2 Uncertainty Budget of DASY for frequency range 300 MHz to 3 GHz



Error Description	Uncertainty Value (±%)	Probability Distribution	Divisor	Ci (1g)	Standard Uncertainty (1g)	
Measurement System	•	•	-	•	•	
Probe Calibration	6.55	Normal	1	1	± 6.55 %	
Axial Isotropy	4.7	Rectangular	√3	0.7	± 1.9 %	
Hemispherical Isotropy	9.6	Rectangular	√3	0.7	± 3.9 %	
Boundary Effects	2.0	Rectangular	√3	1	± 1.2 %	
Linearity	4.7	Rectangular	√3	1	± 2.7 %	
System Detection Limits	1.0	Rectangular	√3	1	± 0.6 %	
Readout Electronics	0.3	Normal	1	1	± 0.3 %	
Response Time	0.8	Rectangular	√3	1	± 0.5 %	
Integration Time	2.6	Rectangular	√3	1	± 1.5 %	
RF Ambient Noise	3.0	Rectangular	√3	1	± 1.7 %	
RF Ambient Reflections	3.0	Rectangular	√3	1	± 1.7 %	
Probe Positioner	0.8	Rectangular	√3	1	± 0.5 %	
Probe Positioning	9.9	Rectangular	√3	1	± 5.7 %	
Max. SAR Eval.	4.0	Rectangular	√3	1	± 2.3 %	
Test Sample Related						
Device Positioning	2.9	Normal	1	1	± 2.9 %	
Device Holder	3.6	Normal	1	1	± 3.6 %	
Power Drift	5.0	Rectangular	√3	1	± 2.9 %	
Phantom and Setup						
Phantom Uncertainty	4.0	Rectangular	√3	1	± 2.3 %	
Liquid Conductivity (Target)	5.0	Rectangular	√3	0.43	± 1.8 %	
Liquid Conductivity (Meas.)	2.5	Normal	1	0.43	± 1.6 %	
Liquid Permittivity (Target)	5.0	Rectangular	√3	0.49	± 1.7 %	
Liquid Permittivity (Meas.)	2.5	Normal	1	0.49	± 1.5 %	
Combined Standard Uncerta	inty				± 12.8 %	
Coverage Factor for 95 %					K = 2	
Expanded Uncertainty						

Table 7.3 Uncertainty Budget of DASY for frequency range 3 GHz to 6 GHz



# 8. SAR Measurement Evaluation


Each DASY system is equipped with one or more system validation kits. These units, together with the predefined measurement procedures within the DASY software, enable the user to conduct the system performance check and system validation. System validation kit includes a dipole, tripod holder to fix it underneath the flat phantom and a corresponding distance holder.


# 8.1 Purpose of System Performance check

The system performance check verifies that the system operates within its specifications. System and operator errors can be detected and corrected. It is recommended that the system performance check be performed prior to any usage of the system in order to guarantee reproducible results. The system performance check uses normal SAR measurements in a simplified setup with a well characterized source. This setup was selected to give a high sensitivity to all parameters that might fail or vary over time. The system check does not intend to replace the calibration of the components, but indicates situations where the system uncertainty is exceeded due to drift or failure.

# 8.2 System Setup

In the simplified setup for system evaluation, the DUT is replaced by a calibrated dipole and the power source is replaced by a continuous wave that comes from a signal generator. The calibrated dipole must be placed beneath the flat phantom section of the SAM twin phantom with the correct distance holder. The distance holder should touch the phantom surface with a light pressure at the reference marking and be oriented parallel to the long side of the phantom. The equipment setup is shown below:





Page Number	: 24 of 31			
Report Issued Date	: Jun. 03, 2011			
Report Version	: Rev. 02			



- 1. Signal Generator
- 2. Amplifier
- 3. Directional Coupler
- 4. Power Meter
- 5. Calibrated Dipole

The output power on dipole port must be calibrated to 24 dBm (250 mW) before dipole is connected.



Fig 8.2 Photo of Dipole Setup

# 8.3 Validation Results

Comparing to the original SAR value provided by SPEAG, the validation data should be within its specification of 10 %. Table 8.1 shows the target SAR and measured SAR after normalized to 1W input power. The table below indicates the system performance check can meet the variation criterion and the plots can be referred to Appendix A of this report.

Measurement	Frequency	Targeted SAR <sub>1g</sub>	Measured SAR <sub>1g</sub>	Deviation	
Date	(MHz)	(W/kg)	(W/kg)	(%)	
May. 25, 2011	2450	53.00	12.60	50.40	

Table 8.1 Target and Measurement SAR after Normalized
---



# 9. DUT Testing Position

Please refer to Appendix E for the test setup photos.

# 10. Measurement Procedures

The measurement procedures are as follows:

- (a) Use engineering software to transmit RF power continuously (continuous Tx) in the highest power channel
- (b) Measure output power through RF cable and power meter
- (c) Place the DUT in the positions described in the last section
- (d) Set scan area, grid size and other setting on the DASY software
- (e) Taking data for the middle channel on each testing position
- (f) Find out the largest SAR result on these testing positions of each band
- (g) Measure SAR results for other channels in worst SAR testing position if the SAR of highest power channel is larger than 0.8 W/kg

According to the test standard, the recommended procedure for assessing the peak spatial-average SAR value consists of the following steps:

- (a) Power reference measurement
- (b) Area scan
- (c) Zoom scan
- (d) Power drift measurement

### 10.1 Spatial Peak SAR Evaluation

The procedure for spatial peak SAR evaluation has been implemented according to the test standard. It can be conducted for 1g and 10g, as well as for user-specific masses. The DASY software includes all numerical procedures necessary to evaluate the spatial peak SAR value.

The base for the evaluation is a "cube" measurement. The measured volume must include the 1g and 10g cubes with the highest averaged SAR values. For that purpose, the center of the measured volume is aligned to the interpolated peak SAR value of a previously performed area scan.

The entire evaluation of the spatial peak values is performed within the post-processing engine (SEMCAD). The system always gives the maximum values for the 1g and 10g cubes. The algorithm to find the cube with highest averaged SAR is divided into the following stages:

- (a) Extraction of the measured data (grid and values) from the Zoom Scan
- (b) Calculation of the SAR value at every measurement point based on all stored data (A/D values and measurement parameters)



- (c) Generation of a high-resolution mesh within the measured volume
- (d) Interpolation of all measured values form the measurement grid to the high-resolution grid
- (e) Extrapolation of the entire 3-D field distribution to the phantom surface over the distance from sensor to surface
- (f) Calculation of the averaged SAR within masses of 1g and 10g



# 10.2 Area & Zoom Scan Procedures

First Area Scan is used to locate the approximate location(s) of the local peak SAR value(s). The measurement grid within an Area Scan is defined by the grid extent, grid step size and grid offset. Next, in order to determine the EM field distribution in a three-dimensional spatial extension, Zoom Scan is required. The Zoom Scan measures 5x5x7 points with step size 8, 8 and 5 mm for 300 MHz to 3 GHz, and 8x8x8 points with step size 4, 4 and 2.5 mm for 3 GHz to 6 GHz. The Zoom Scan is performed around the highest E-field value to determine the averaged SAR-distribution over 10 g.

# 10.3 Volume Scan Procedures

The volume scan is used for assess overlapping SAR distributions for antennas transmitting in different frequency bands. It is equivalent to an oversized zoom scan used in standalone measurements. The measurement volume will be used to enclose all the simultaneous transmitting antennas. For antennas transmitting simultaneously in different frequency bands, the volume scan is measured separately in each frequency band. In order to sum correctly to compute the 1g aggregate SAR, the DUT remain in the same test position for all measurements and all volume scan use the same spatial resolution and grid spacing (step-size is 4, 4 and 2.5 mm). When all volume scan were completed, the software, SEMCAD postprocessor can combine and subsequently superpose these measurement data to calculating the multiband SAR.

### 10.4 SAR Averaged Methods

In DASY, the interpolation and extrapolation are both based on the modified Quadratic Shepard's method. The interpolation scheme combines a least-square fitted function method and a weighted average method which are the two basic types of computational interpolation and approximation.

Extrapolation routines are used to obtain SAR values between the lowest measurement points and the inner phantom surface. The extrapolation distance is determined by the surface detection distance and the probe sensor offset. The uncertainty increases with the extrapolation distance. To keep the uncertainty within 1% for the 1 g and 10 g cubes, the extrapolation distance should not be larger than 5 mm.


# 10.5 Power Drift Monitoring

All SAR testing is under the DUT install full charged battery and transmit maximum output power. In DASY measurement software, the power reference measurement and power drift measurement procedures are used for monitoring the power drift of DUT during SAR test. Both these procedures measure the field at a specified reference position before and after the SAR testing. The software will calculate the field difference in dB. If the power drift more than 5%, the SAR will be retested.



# 11. SAR Test Results

# 11.1 <u>SAR Test Setup</u>



Sides for SAR testing – 0mm test distance						
	Rear Face	Front Face	Тор	Bottom	Right	Left
WLAN 11b/g/n 2.4GHz	<ul> <li>Image: A second s</li></ul>	X	>	X	X	X

#### Note:

- 1. The DUT is 7" tablet with diagonal 22cm, therefore the 0 mm test distance per KDB 447498 D01.
- 2. The display can be orientated four positions, primary portrait, secondary portrait, primary landscape and secondary landscape. Above plot, the "Rear View" is secondary landscape.

The rear face, and top side (secondary landscape position) were tested 0cm due to antenna within 5cm from body, and the SAR for Right/Left/Bottom edges are not necessary, as the antenna is more than 5cm away from that edges according to KDB 447498.



Band	802.11b		Band			802.11g	
Channel	1	6	11	1	6	11	
Frequency (MHz)	2412	2437	2462	2412	2437	2462	
Power	14.86	14.38	14.61	10.71	13.73	11.07	

# 11.2 Conducted Average Power (Unit: dBm)

Band	802.11n (BW 20MHz)					
Channel	1 6 11					
Frequency (MHz)	2412	2437	2462			
Power	8.78	11.13	9.47			

#### Note:

- 1. Referring to KDB 248227, SAR tests are using 11b mode, 11g mode SAR is not necessary due to the average power is less than 1/4 dB higher than 11b mode.
- 2. Per KDB 248227, testing SAR with the highest output channel first; therefore 11b mode CH1 was tested.

# 11.3 Test Records for Tablet SAR Test

Plot No.	Band	Mode	Test Position	Test distance (cm)	Channel	SAR <sub>1g</sub> (W/kg)
1	802.11b	1Mbps	Rear Face	0	1	0.478
8	802.11b	1Mbps	Top Side	0	1	<mark>0.796</mark>
2	802.11b	1Mbps	Top Side	0	1	0.774

#### Note:

- 1. Referring to KDB 447498, if the SAR of the highest output power channel is less than 0.8W/kg, other channels SAR tests are not necessary.
- 2. The RF parts (antenna, transmitter) in three samples are electric identical, and the output power are quite close. Because the change parts (memory) are far away radio parts, the sample 1 (#1, 8) is mainly tested due to highest power, and the results are worst. The sample 2, 3 are verified, and there are no non-compliance found.

Test Engineer : Eric Huang



# 12. References

- [1] FCC 47 CFR Part 2 "Frequency Allocations and Radio Treaty Matters; General Rules and Regulations"
- [2] IEEE Std. C95.1-1991, "IEEE Standard for Safety Levels with Respect to Human Exposure to Radio Frequency Electromagnetic Fields, 3 kHz to 300 GHz", 1991
- [3] IEEE Std. 1528-2003, "Recommended Practice for Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- [4] FCC OET Bulletin 65 (Edition 97-01) Supplement C (Edition 01-01), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields", June 2001
- [5] SPEAG DASY System Handbook
- [6] FCC KDB 248227 D01 v01r02, "SAR Measurement Procedures for 802.11 a/b/g Transmitters", May 2007
- [7] FCC KDB 447498 D01 v04, "Mobile and Portable Device RF Exposure Procedures and Equipment Authorization Policies", November 2009
- [8] FCC KDB 447498 D02 v02, "SAR Measurement Procedures for USB Dongle Transmitters", November 2009
- [9] FCC KDB 616217 D01 v01r01, "SAR Evaluation Considerations for Laptop Computers with Antennas Built-in on Display Screens", November 2009
- [10] FCC KDB 616217 D03 v01, "SAR Evaluation Considerations for Laptop/Notebook/Netbook and Tablet Computers", November 2009
- [11] FCC KDB 648474 D01 v01r05, "SAR Evaluation Considerations for Handsets with Multiple Transmitters and Antennas", September 2008
- [12] FCC KDB 941225 D01 v02, "SAR Measurement Procedures for 3G Devices CDMA 2000 / Ev-Do / WCDMA / HSDPA / HSPA", October 2007
- [13] FCC KDB 941225 D03 v01, "Recommended SAR Test Reduction Procedures for GSM / GPRS / EDGE", December 2008
- [14] FCC KDB 941225 D04 v01, "Evaluating SAR for GSM/(E)GPRS Dual Transfer Mode", January 27 2010



# Appendix A. Plots of System Performance Check

The plots are shown as follows.

Date: 2011/5/25

### System Check\_Body\_2450MHz\_110525

### DUT: Dipole 2450 MHz

Communication System: CW; Frequency: 2450 MHz;Duty Cycle: 1:1 Medium: MSL\_2450\_110525 Medium parameters used: f = 2450 MHz;  $\sigma = 1.97$  mho/m;  $\epsilon_r = 53.8$ ;  $\rho$ 


=  $1000 \text{ kg/m}^3$ Ambient Temperature : 22.6 °C; Liquid Temperature : 21.5 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3071; ConvF(4, 4, 4); Calibrated: 2010/6/22
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn778; Calibrated: 2010/10/22
- Phantom: ELI 4.0\_Front; Type: QDOVA001BB; Serial: 1026
- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

**Pin=250mW/Area Scan (91x91x1):** Measurement grid: dx=10mm, dy=10mm Maximum value of SAR (interpolated) = 14.7 mW/g

Pin=250mW/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 85.9 V/m; Power Drift = 0.020 dB Peak SAR (extrapolated) = 26.1 W/kg SAR(1 g) = 12.6 mW/g; SAR(10 g) = 5.81 mW/g Maximum value of SAR (measured) = 14.4 mW/g



 $0 \, dB = 14.4 \, mW/g$ 



# Appendix B. Plots of SAR Measurement

The plots are shown as follows.

Date: 2011/5/25

### #01 802.11b\_Rear Face\_0cm\_Ch1\_Sample1

# DUT: 132346

Communication System: 802.11b ; Frequency: 2412 MHz;Duty Cycle: 1:1 Medium: MSL\_2450\_110525 Medium parameters used: f = 2412 MHz;  $\sigma$  = 1.92 mho/m;  $\epsilon_r$  = 53.9;  $\rho$ 


=  $1000 \text{ kg/m}^3$ Ambient Temperature : 22.4 °C; Liquid Temperature : 21.5 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3071; ConvF(4, 4, 4); Calibrated: 2010/6/22
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn778; Calibrated: 2010/10/22
- Phantom: ELI 4.0\_Front; Type: QDOVA001BB; Serial: 1026
- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

**Ch1/Area Scan (71x111x1):** Measurement grid: dx=20mm, dy=20mm Maximum value of SAR (interpolated) = 0.363 mW/g

### Ch1/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 1.42 V/m; Power Drift = 0.140 dB Peak SAR (extrapolated) = 1.13 W/kg SAR(1 g) = 0.478 mW/g; SAR(10 g) = 0.217 mW/g Maximum value of SAR (measured) = 0.522 mW/g



Date: 2011/5/25

# #08 802.11b\_Top Side\_0cm\_Ch1\_Sample1

# DUT: 132346

Communication System: 802.11b ; Frequency: 2412 MHz;Duty Cycle: 1:1 Medium: MSL\_2450\_110525 Medium parameters used: f = 2412 MHz;  $\sigma$  = 1.92 mho/m;  $\epsilon_r$  = 53.9;  $\rho$ 


=  $1000 \text{ kg/m}^3$ Ambient Temperature : 22.6 °C; Liquid Temperature : 21.5 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3071; ConvF(4, 4, 4); Calibrated: 2010/6/22
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn778; Calibrated: 2010/10/22
- Phantom: ELI 4.0\_Front; Type: QDOVA001BB; Serial: 1026
- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

Ch1/Area Scan (41x121x1): Measurement grid: dx=20mm, dy=20mm Maximum value of SAR (interpolated) = 0.749 mW/g

Ch1/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 14.0 V/m; Power Drift = -0.056 dB Peak SAR (extrapolated) = 1.97 W/kg SAR(1 g) = 0.796 mW/g; SAR(10 g) = 0.316 mW/g Maximum value of SAR (measured) = 1.01 mW/g



### #08 802.11b\_Top Side\_0cm\_Ch1\_2D

### DUT: 132346

Communication System: 802.11b ; Frequency: 2412 MHz;Duty Cycle: 1:1 Medium: MSL\_2450\_110525 Medium parameters used: f = 2412 MHz;  $\sigma = 1.92$  mho/m;  $\epsilon_r = 53.9$ ;


 $\rho = 1000 \text{ kg/m}^3$ Ambient Temperature : 22.6 °C; Liquid Temperature : 21.5 °C

DASY4 Configuration:

- Probe: ES3DV3 SN3071; ConvF(4, 4, 4); Calibrated: 2010/6/22
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn778; Calibrated: 2010/10/22
- Phantom: ELI 4.0\_Front; Type: QDOVA001BB; Serial: 1026
- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

**Ch1/Area Scan (41x121x1):** Measurement grid: dx=20mm, dy=20mm Maximum value of SAR (interpolated) = 0.749 mW/g

Ch1/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm Reference Value = 14.0 V/m; Power Drift = -0.056 dB Peak SAR (extrapolated) = 1.97 W/kg SAR(1 g) = 0.796 mW/g; SAR(10 g) = 0.316 mW/g Maximum value of SAR (measured) = 1.01 mW/g



### #02 802.11b\_Top Side\_0cm\_Ch1\_Sample2

### DUT: 132346

Communication System: 802.11b ; Frequency: 2412 MHz;Duty Cycle: 1:1 Medium: MSL\_2450\_110525 Medium parameters used: f = 2412 MHz;  $\sigma = 1.92$  mho/m;  $\epsilon_r = 53.9$ ;  $\rho$ 

=  $1000 \text{ kg/m}^3$ Ambient Temperature : 22.4 °C; Liquid Temperature : 21.5 °C


DASY4 Configuration:

- Probe: ES3DV3 SN3071; ConvF(4, 4, 4); Calibrated: 2010/6/22
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn778; Calibrated: 2010/10/22
- Phantom: ELI 4.0\_Front; Type: QDOVA001BB; Serial: 1026
- Measurement SW: DASY4, V4.7 Build 80; Postprocessing SW: SEMCAD, V1.8 Build 186

**Ch1/Area Scan (31x111x1):** Measurement grid: dx=20mm, dy=20mm Maximum value of SAR (interpolated) = 0.597 mW/g

### Ch1/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 19.6 V/m; Power Drift = -0.141 dB Peak SAR (extrapolated) = 2.01 W/kgSAR(1 g) = 0.774 mW/g; SAR(10 g) = 0.305 mW/g Maximum value of SAR (measured) = 0.872 mW/g





## Appendix C. DASY Calibration Certificate

The DASY calibration certificates are shown as follows.



ugnausstrasse 43, 8004 Zurie	ch, Switzerland	BC MRA	S Schweizerischer Kalibrierdiens Service suisse d'étalonnage Servizio svizzero di taratura S Swiss Calibration Service
credited by the Swiss Accredit e Swiss Accreditation Servic Iltilateral Agreement for the r	e is one of the signatorie	s to the EA	tion No.: SCS 108
lient Auden		Certificat	e No: D2450V2-735_Jun10
CALIBRATION	CERTIFICATE	E	
Object	D2450V2 - SN: 7	35	
Calibration procedure(s)	QA CAL-05:v7 Calibration proce	dure for dipole validation kits	
Calibration date.	June 17, 2010		
The measurements and the unce	artainties with confidence p	onal standards, which realize the physica robability are given on the following page ry facility: anvironment temperature (22 ±	s and are part of the certificate.
The measurements and the unce All calibrations have been condu Calibration Equipment used (M&	ertainties with confidence p cted in the closed laborator TE critical for calibration)	robability are given on the following page ry facility: anvironment temperature (22 $\pm$	s and are part of the certificate.
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards	ertainties with confidence p cted in the closed laborator TE critical for calibration)	robability are given on the following page ry facility: anvironment temperature (22 ± Cal Date (Certificate No.)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A	ertainties with confidence p cted in the closed laborator TE critical for calibration) 1D # GB37480704	robability are given on the following page ry facility: anvironment temperature (22 ± <u>Cal Date (Certificate No.)</u> 05-Oct-09 (No. 217-01086)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A	ertainties with confidence p cted in the closed laborator TE critical for calibration) ID # GB37480704 US37292783	robability are given on the following page ry facility: anvironment temperature (22 ± <u>Cal Date (Certificate No.)</u> 06-Oct-09 (No. 217-01086) 06-Oct-09 (No. 217-01086)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10 Oct-10
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator	ertainties with confidence p cted in the closed laborator TE critical for calibration) 1D # GB37480704	robability are given on the following page ry facility: anvironment temperature (22 ± <u>Cal Date (Certificate No.)</u> 05-Oct-09 (No. 217-01086)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination	ertainties with confidence p cted in the closed laborator TE critical for calibration) 1D # GB37480704 US37292783 SN: 5086 (20g)	robability are given on the following page ry facility: anvironment temperature (22 ± <u>Cal Date (Certificate No.)</u> 06-Oct-09 (No. 217-01086) 06-Oct-09 (No. 217-01086) 30-Mar-10 (No. 217-01158)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10 Oct-10 Mar-11
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination Reference Probe ES3DV3	ertainties with confidence p cted in the closed laborator TE critical for calibration) 10 # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 / 06327	robability are given on the following page ty facility: anvironment temperature (22 ± <u>Cal Date (Certificate No.)</u> 06-Oct-09 (No. 217-01086) 06-Oct-09 (No. 217-01086) 30-Mar-10 (No. 217-01158) 30-Mar-10 (No. 217-01162)	s and are part of the certificate. 3)°C and humidity <.70%. Scheduled Calibration Oct-10 Oct-10 Mar-11 Mar-11
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination Reference Probe ES3DV3 DAE4	ertainties with confidence p cted in the closed laborator TE critical for calibration) 10 # GB37480704 US37292783 SN: 5086 (20g) SN: 5047,2 / 06327 SN: 3205	Cal Date (Certificate No.)           06-Oct-09 (No. 217-01086)           30-Mar-10 (No. 217-01158)           30-Mar-10 (No. 217-01158)           30-Mar-10 (No. 217-01162)           30-Apr-10 (No. 217-01162)           30-Apr-10 (No. ES3-3205_Apr10)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10 Oct-10 Mar-11 Mar-11 Apr.11
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards Power sensor HP 8481A	ertainties with confidence p cted in the closed laborator TE critical for calibration) 1D # GB37480704 US37292783 SN: 5086 (20g) SN: 5086 (20g) SN: 5047,2 / 06327 SN: 3205 SN: 601 1D # MY41092317	Cal Date (Certificate No.)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           30-Mar-10 (No. 217-01158)           30-Mar-10 (No. 217-01162)           30-Apr-10 (No. 217-01162)           30-Apr-10 (No. ES3-3205_Apr10)           10-Jun-10 (No. DAE4-601_Jun10)           Check Date (in house)           18-Oct-02 (in house check Oct-09)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10 Oct-10 Mar-11 Mar-11 Jun-11
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards Power sensor HP 8481A RF generator R&S SMT-06	ertainties with confidence p cted in the closed laborator TE critical for calibration) 1D # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 / 06327 SN: 5047.2 / 06327 SN: 3205 SN: 601 1D # MY41092317 100005	Cal Date (Certificate No.)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           30-Mar-10 (No. 217-01158)           30-Mar-10 (No. 217-01162)           30-Apr-10 (No. 217-01162)           30-Apr-10 (No. 217-01162)           30-Apr-10 (No. ES3-3205_Apr10)           10-Jun-10 (No. DAE4-601_Jun10)           Check Date (in house)           18-Oct-02 (in house check Oct-09)           4-Aug-99 (in house check Oct-09)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10 Oct-10 Mar-11 Mar-11 Apr.11 Jun-11 Scheduled Check In house check: Oct-11 In house check: Oct-11
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards Power sensor HP 8481A RF generator R&S SMT-06	ertainties with confidence p cted in the closed laborator TE critical for calibration) 1D # GB37480704 US37292783 SN: 5086 (20g) SN: 5086 (20g) SN: 5047,2 / 06327 SN: 3205 SN: 601 1D # MY41092317	Cal Date (Certificate No.)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           30-Mar-10 (No. 217-01158)           30-Mar-10 (No. 217-01162)           30-Apr-10 (No. 217-01162)           30-Apr-10 (No. ES3-3205_Apr10)           10-Jun-10 (No. DAE4-601_Jun10)           Check Date (in house)           18-Oct-02 (in house check Oct-09)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10 Oct-10 Mar-11 Mar-11 Apr.11 Jun-11 Scheduled Check In house check: Oct-11
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer HP 8753E	ertainties with confidence p cted in the closed laborator TE critical for calibration) 1D # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 / 06327 SN: 5047.2 / 06327 SN: 3205 SN: 601 1D # MY41092317 100005	Cal Date (Certificate No.)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           30-Mar-10 (No. 217-01158)           30-Mar-10 (No. 217-01162)           30-Apr-10 (No. 217-01162)           30-Apr-10 (No. 217-01162)           30-Apr-10 (No. ES3-3205_Apr10)           10-Jun-10 (No. DAE4-601_Jun10)           Check Date (in house)           18-Oct-02 (in house check Oct-09)           4-Aug-99 (in house check Oct-09)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10 Oct-10 Mar-11 Mar-11 Apr.11 Jun-11 Scheduled Check In house check: Oct-11 In house check: Oct-11
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards Power sensor HP 8481A RF generator R&S SMT-06	ertainties with confidence p cted in the closed laborator TE critical for calibration) 1D # GB37480704 US37292783 SN: 5086 (20g) SN: 5047.2 / 06327 SN: 5047.2 / 06327 SN: 3205 SN: 601 1D # MY41092317 100005 US37390585 S4206	Cal Date (Certificate No.)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           06-Oct-09 (No. 217-01086)           30-Mar-10 (No. 217-01158)           30-Mar-10 (No. 217-01162)           30-Apr-10 (No. 217-01162)           30-Apr-10 (No. 217-01162)           30-Apr-10 (No. ES3-3205_Apr10)           10-Jun-10 (No. DAE4-601_Jun10)           Check Date (in house)           18-Oct-02 (in house check Oct-09)           4-Aug-99 (in house check Oct-09)           18-Oct-01 (in house check Oct-09)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10 Oct-10 Mar-11 Mar-11 Apr.11 Jun-11 Scheduled Check In house check: Oct-11 In house check: Oct-11 In house check: Oct-10
The measurements and the unce All calibrations have been condu Calibration Equipment used (M& Primary Standards Power meter EPM-442A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination Reference 20 dB Attenuator Type-N mismatch combination Reference 20 dB Attenuator Type-N mismatch combination Reference 20 dB Attenuator Secondary Standards Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer HP 8753E	artainties with confidence p           cted in the closed laborator           TE critical for calibration)           1D #           GB37480704           US37292783           SN: 5086 (20g)           SN: 5086 (20g)           SN: 5086 (20g)           SN: 5085 (20g)           SN: 3205           SN: 601           ID #           MY41092317           100005           US37390585 S4206           Name	Cal Date (Certificate No.) Cal Date (Certificate No.) 05-Oct-09 (No. 217-01086) 06-Oct-09 (No. 217-01086) 06-Oct-09 (No. 217-01086) 30-Mar-10 (No. 217-01162) 30-Mar-10 (No. 217-01162) 30-Apr-10 (No. 217-01162) 30-Apr-10 (No. ES3-3205_Apr10) 10-Jun-10 (No. DAE4-601_Jun10) Check Date (in house) 18-Oct-02 (in house check Oct-09) 4-Aug-99 (in house check Oct-09) 18-Oct-01 (in house check Oct-09) 18-Oct-01 (in house check Oct-09) 18-Oct-01 (in house check Oct-09) 18-Oct-01 (in house check Oct-09)	s and are part of the certificate. 3)°C and humidity < 70%. Scheduled Calibration Oct-10 Oct-10 Mar-11 Mar-11 Apr.11 Jun-11 Scheduled Check In house check: Oct-11 In house check: Oct-11 In house check: Oct-10

Certificate No: D2450V2-735\_Jun10 Page 1 of 9



### **Calibration Laboratory of** Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



BRP

S

Schweizerischer Kallbrierdienst Service suisse d'étalonnage C Servizio svizzero di taratura S Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

### Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

### Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) IEC 62209-1. "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)". February 2005
- C) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

### Additional Documentation:

d) DASY4/5 System Handbook

### Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. ..... No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters. The measured TSL parameters are used to calculate the nominal SAR result.

Certificate No: D2450V2-735\_Jun10

Page 2 of 9

SPORTON INTERNATIONAL INC.

### **Measurement Conditions**

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.2
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V4.9	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, $dy$ , $dz = 5 mm$	
Frequency	2450 MHz ± 1 MHz	

### Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0,°C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.9 ± 6 %	1.78 mho/m ± 6 %
Head TSL temperature during test	(22.0 ± 0.2) °C		544

### SAR result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.0 mW / g
SAR normalized	normalized to 1W	52.0 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	52.2 mW /g ± 17.0 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR measured	250 mW input power	6.10 mW / g
SAR normalized	normalized to 1W	24.4 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	24.4 mW /g ± 16.5 % (k=2)

Certilicate No: D2450V2-735\_Jun10

Page 3 of 9

Body TSL parameters The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52,7	1.95 mho/m
Measured Body TSL parameters	$(22.0 \pm 0.2)$ °C	52.7 ± 6 %	1.96 mho/m ± 6 %
Body TSL temperature during test	(22:0 ± 0.2) °C	Ares:	

### SAR result with Body TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	13.4 mW / g
SAR normalized	normalized to 1W	53,6 mW / g
SAR for nominal Body TSL parameters	normalized to 1W	53.5 mW / g ± 17.0 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL	condition	
the second s	condition 250 mW input power	6.28 mW / g
SAR averaged over 10 cm <sup>2</sup> (10 g) of Body TSL SAR measured SAR normalized		6.28 mW / g 25.1 mW / g

Certificate No: D2450V2-735\_Jun10

Page 4 of 9

### Appendix

### Antenna Parameters with Head TSL

Impedance, transformed to feed point	53.8 Ω + 3.4 jΩ
Return Loss	- 26.1 dB

### Antenna Parameters with Body TSL

Impedance, transformed to feed point	50.5 Ω + 3.7 jΩ
Return Loss	- 28.5 dB

### General Antenna Parameters and Design

Electrical Delay (one direction)	1.154 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

### Additional EUT Data

Manufactured by	SPEAG
Manufactured on	May 07, 2003

Certificate No: D2450V2-735\_Jun10

Page 5 of 9



### **DASY5 Validation Report for Head TSL**

Date/Time: 16.06.2010 10:56:25

Test Laboratory: SPEAG, Zurich, Switzerland

### DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:735

Communication System: CW; Frequency: 2450 MHz; Duty Cycle: 1:1 Medium: HSL U11 BB Medium parameters used: f = 2450 MHz;  $\sigma = 1.78$  mho/m;  $\varepsilon_r = 38.9$ ;  $\rho = 1000$  kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

**DASY5** Configuration:

- Probe: ES3DV3 SN3205; ConvF (4.53, 4.53, 4.53); Calibrated: 30.04.2010
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 10.06.2010
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA; Serial: 1001
- Measurement SW: DASY52, V52,2 Build 0, Version 52.2.0 (163)
- Postprocessing SW: SEMCAD X, V14.2 Build 2, Version 14.2.2 (1685)

Head/d=10mm, Pin=250 mW, dist=3.0mm (ES-Probe)/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 100.3 V/m; Power Drift = 0.046 dB Peak SAR (extrapolated) = 26.6 W/kg

SAR(1 g) = 13 mW/g; SAR(10 g) = 6.1 mW/g Maximum value of SAR (measured) = 16.6 mW/g




 $0 \, dB = 16.6 \, mW/g$ 

Certificate No: D2450V2-735\_Jun10

Page 6 of 9

### Impedance Measurement Plot for Head TSL



Certificate No: D2450V2-735\_Jun10

Page 7 of 9

### DASY5 Validation Report for Body

Date/Time: 17.06.2010 11:28:23

Test Laboratory: SPEAG, Zurich, Switzerland

### DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:735

Communication System: CW; Frequency: 2450 MHz; Duty Cycle: 1:1 Medium: MSL U11 BB Medium parameters used: f = 2450 MHz;  $\sigma = 1.96$  mho/m;  $\varepsilon_r = 52.7$ ;  $\rho = 1000$  kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

- Probe: ES3DV3 SN3205; ConvF(4.31, 4.31, 4.31); Calibrated: 30.04.2010
- Sensor-Surface; 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 10.06.2010
- Phantom: Flat Phantom 5.0 (back); Type: QD000P50AA; Serial; 1002
- Measurement SW: DASY52, V52.2 Build 0, Version 52.2.0 (163)
- Postprocessing SW: SEMCAD X, V14.2 Build 2, Version 14.2.2 (1685)

### Body/d=10mm, Pin250 mW, dist=3.0mm (ES-Probe)/Zoom Scan (7x7x7) (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mmReference Value = 94.8 V/m; Power Drift = 0.073 dB Peak SAR (extrapolated) = 27.7 W/kg SAR(1 g) = 13.4 mW/g; SAR(10 g) = 6.28 mW/g Maximum value of SAR (measured) = 16.7 mW/g




0 dB = 16.7 mW/g

Certificate No: D2450V2-735\_Jun10

Page 8 of 9

### Impedance Measurement Plot for Body TSL



Certificate No: D2450V2-735\_Jun10

Page 9 of 9



Calibration Laboratory	/ of
Schmid & Partner	
Engineering AG	
Zeughausstrasse 43, 8004 Zurich	, Switzerland



S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage Servizio svizzero di taratura

S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client Sporton (Auden)

Accreditation No.: SCS 108

Certificate No: DAE4-778 Oct10

Dbject	DAE4 - SD 000 D	04 BJ - SN: 778	and the second second
Calibration procedure(s)	QA CAL-06.v22 Calibration proces	dure for the data acquisitio	n electronics (DAE)
alibration date:	October 22, 2010		
he measurements and the unce	rtainties with confidence pre	nal standards, which realize the phy obability are given on the following p / facility: environment temperature (	bages and are part of the certificate.
rimary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
rimary Standards		Cal Date (Certificate No.) 28-Sep-10 (No:10376)	Scheduled Calibration Sep-11
nmary Standards eithley Multimeter Type 2001 econdary Standards	ID # SN: 0810278	28-Sep-10 (No:10376) Check Date (in house)	Sep-11 Scheduled Check
Calibration Equipment used (M& Primary Standards Keithley Multimeter Type 2001 Secondary Standards Calibrator Box V1.1	ID # SN: 0810278	28-Sep-10 (No:10376)	Sep-11
Primary Standards Keithley Multimeter Type 2001 Secondary Standards	ID # SN: 0810278 ID # SE UMS 006 AB 1004	28-Sep-10 (No:10376) Check Date (in house) 07-Jun-10 (in house check)	Sep-11 Scheduled Check In house check: Jun-11
nimary Standards eithley Multimeter Type 2001 econdary Standards alibrator Box V1.1	ID # SN: 0810278	28-Sep-10 (No:10376) Check Date (in house) 07-Jun-10 (in house check) Function	Sep-11 Scheduled Check
himary Standards leithley Multimeter Type 2001 econdary Standards alibrator Box V1.1	ID # SN: 0810278 ID # SE UMS 006 AB 1004	28-Sep-10 (No:10376) Check Date (in house) 07-Jun-10 (in house check)	Sep-11 Scheduled Check In house check: Jun-11 Signature
Primary Standards Geithley Multimeter Type 2001 Jecondary Standards	ID # SN: 0810278 ID # SE UMS 006 AB 1004	28-Sep-10 (No:10376) Check Date (in house) 07-Jun-10 (in house check) Function	Sep-11 Scheduled Check In house check: Jun-11

Certificate No: DAE4-778\_Oct10

Page 1 of 5



### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



SNISS BRA

Schweizerischer Kalibrierdienst S

- Service suisse d'étalonnage С
- Servizio svizzero di taratura S
- Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Glossary

DAE Connector angle

data acquisition electronics information used in DASY system to align probe sensor X to the robot coordinate system.

### Methods Applied and Interpretation of Parameters

- DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a . result from the performance test and require no uncertainty.
  - DC Voltage Measurement Linearity: Verification of the Linearity at +10% and -10% of . the nominal calibration voltage. Influence of offset voltage is included in this measurement.
  - Common mode sensitivity: Influence of a positive or negative common mode voltage on the differential measurement.
  - Channel separation: Influence of a voltage on the neighbor channels not subject to an input voltage.
  - AD Converter Values with inputs shorted: Values on the internal AD converter corresponding to zero input voltage
  - Input Offset Measurement: Output voltage and statistical results over a large number of zero voltage measurements.
  - Input Offset Current: Typical value for information; Maximum channel input offset current, not considering the input resistance.
  - Input resistance: Typical value for information: DAE input resistance at the connector, ٠ during internal auto-zeroing and during measurement.
  - Low Battery Alarm Voltage: Typical value for information. Below this voltage, a battery alarm signal is generated.
  - Power consumption: Typical value for information. Supply currents in various operating . modes

Certificate No: DAE4-778\_Oct10

Page 2 of 5

SPORTON INTERNATIONAL INC.

### **DC Voltage Measurement**

A/D - Converter Resolution nominal

Calibration Factors	X	Y	z
High Range	404.679 ± 0.1% (k=2)	403.480 ± 0.1% (k=2)	405.025 ± 0.1% (k=2)
Low Range	3.98633 ± 0.7% (k=2)	3.96375 ± 0.7% (k=2)	3.99940 ± 0.7% (k=2)

### **Connector Angle**

Connector Angle to be used in DASY system	64.5°±1°
---	----------

Certificate No: DAE4-778\_Oct10

### Appendix

### 1. DC Voltage Linearity

High Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	200004.4	1.89	0.00
Channel X + Input	20001.11	1.41	0.01
Channel X - Input	-19998.36	1.54	-0.01
Channel Y + Input	199996.1	3.42	0.00
Channel Y + Input	19999.75	0.35	0.00
Channel Y - Input	-19999.92	-0.12	0.00
Channel Z + Input	200002.7	1.29	0.00
Channel Z + Input	19996.85	-2.55	-0.01
Channel Z - Input	-20004.31	-4.61	0.02

Low Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	2000.0	0.09	0.00
Channel X + Input	200.02	0.02	0.01
Channel X - Input	-198.62	1.48	-0.74
Channel Y + Input	1999.6	-0.58	-0.03
Channel Y + Input	199.13	-0.57	-0.29
Channel Y - Input	-200.71	-0.61	0.31
Channel Z + Input	2000.1	-0.01	-0.00
Channel Z + Input	198.96	-1.14	-0.57
Channel Z - Input	-200.98	-0.98	0.49

2. Common mode sensitivity DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Common mode Input Voltage (mV)	High Range Average Reading (μV)	Low Range Average Reading (µV)
Channel X	200	-5.28	-6.07
	- 200	6.79	6.12
Channel Y	200	-1.80	-1.60
	- 200	0.97	0.35
Channel Z	200	-9.76	-9.86
	- 200	7.56	7.61

3. Channel separation DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X (µV)	Channel Y (µV)	Channel Z (µV)
Channel X	200		1.86	-0.66
Channel Y	200	2.28		2.89
Channel Z	200	1.68	-0.15	

Certificate No: DAE4-778\_Oct10

### 4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	16056	16950
Channel Y	16153	13741
Channel Z	16441	16086

### 5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec Input 10  $M\Omega$ 

	Average (µV)	min. Offset (µV)	max. Offset (μV)	Std. Deviation (µV)
Channel X	0.32	-2.35	2.08	0.55
Channel Y	-1.83	-2.96	-0.72	0.47
Channel Z	-1.93	-3.00	-0.90	0.45

### 6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

### 7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

### 8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)	
Supply (+ Vcc)	+7.9	
Supply (- Vcc)	-7.6	

## 9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+6	+14
Supply (- Vcc)	-0.01	-8	-9



credited by the Swiss Accredite the Swiss Accreditation Servi ultilateral Agreement for the	ice is one of the signator	ies to the EA	n No.: SCS 108
lient Sporton (Aud	1. 1. 1. <b>1. 1</b> . 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1.		o: ET3-1787_May10
CALIBRATION	CERTIFICAT	Ē	
Dbject	ET3DV6 - SN:1	787	
Calibration procedure(s)		QA CAL-23.v3 and QA CAL-25.v2 edure for dosimetric E-field probe	
Calibration date:	May 18, 2010		
This calibration certificate docu	ments the traceability to na	tional standards, which realize the physical un	its of measurements (SI).
The measurements and the un	certainties with confidence lucted in the closed laborat	tional standards, which realize the physical un probability are given on the following pages ar ory facility: environment temperature $(22 \pm 3)^{24}$	nd are part of the certificate.
The measurements and the uni All calibrations have been cond Calibration Equipment used (M	certainties with confidence lucted in the closed laborat &TE critical for calibration)	probability are given on the following pages ar	nd are part of the certificate.
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards	certainties with confidence lucted in the closed laborat &TE critical for calibration) ID #	probability are given on the following pages ar ory facility: environment temperature (22 ± 3) <sup>14</sup> Cal Date (Certificate No.)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration
The measurements and the unit NI calibrations have been cond Calibration Equipment used (M Primary Standards Power meter E44198	certainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874	probability are given on the following pages ar ory facility: environment temperature (22 ± 3)*4 Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter E44198 Power sensor E4412A	ertainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277	probability are given on the following pages ar ory facility: environment temperature (22 ± 3)*4 Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor E44198 Power sensor E4412A Power sensor E4412A	ertainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277 MY41496087	probability are given on the following pages an ony facility: environment temperature (22 ± 3)*4 Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Apr-11
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor E44198 Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator	ertainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277 MY41496087 SN: S5054 (3c)	probability are given on the following pages ar ory facility: environment temperature (22 ± 3)*( Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 30-Mar-10 (No. 217-01159)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Apr-11 Mar-11
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 3 dB Attenuator	ertainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277 MY41496087	probability are given on the following pages ar ory facility: environment temperature (22 ± 3)*( Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 30-Mar-10 (No. 217-01159) 30-Mar-10 (No. 217-01161)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Apr-11
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor E44198 Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 30 dB Attenuator	ertainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277 MY41496087 SN: S5054 (3c) SN: S5066 (20b)	probability are given on the following pages ar ory facility: environment temperature (22 ± 3)*( Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 30-Mar-10 (No. 217-01159)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Apr-11 Mar-11 Mar-11
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor E44198 Power sensor E4412A Reference 3 dB Attenuator Reference 3 dB Attenuator Reference 30 dB Attenuator Reference 30 dB Attenuator Reference Probe ES3DV2	ertainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277 MY41496087 SN: S5054 (3c) SN: S5056 (20b) SN: S5129 (30b)	probability are given on the following pages ar ory facility: environment temperature (22 ± 3)*( Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 30-Mar-10 (No. 217-01159) 30-Mar-10 (No. 217-01161) 30-Mar-10 (No. 217-01160)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Apr-11 Mar-11 Mar-11 Mar-11
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 30 dB Attenuator Reference 90 dB Attenuator Reference Probe ES3DV2 DAE4	LD # LD # GB41293874 MY41495277 MY41496087 SN: S5054 (3c) SN: S5086 (20b) SN: S5129 (30b) SN: 3013	probability are given on the following pages ar ory facility: environment temperature (22 ± 3)*4 Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 30-Mar-10 (No. 217-01159) 30-Mar-10 (No. 217-01161) 30-Mar-10 (No. 217-01160) 30-Dec-09 (No. ES3-3013_Dec09)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Apr-11 Mar-11 Mar-11 Mar-11 Dec-10
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator Reference 30 dB Attenuator Reference 30 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards	ertainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277 MY41496087 SN: S5054 (3c) SN: S5054 (3c) SN: S5054 (3c) SN: S5129 (30b) SN: 3013 SN: 660	probability are given on the following pages ar ory facility: environment temperature (22 ± 3) <sup>14</sup> Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 30-Mar-10 (No. 217-01161) 30-Mar-10 (No. 217-01161) 30-Mar-10 (No. 217-01161) 30-Mar-10 (No. 217-01160) 30-Dec-09 (No. ES3-3013_Dec09) 20-Apr-10 (No. DAE4-660_Apr10) Check Date (in house)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Mar-11 Mar-11 Mar-11 Dec-10 Apr-11 Scheduled Check
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter E44198 Power sensor E4412A Reference 3 dB Attenuator Reference 30 dB Attenuator Reference 30 dB Attenuator Reference 9 robe ES3DV2 DAE4 Secondary Standards RF generator HP 8648C	ertainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277 MY41496087 SN: S5056 (20b) SN: S5056 (20b) SN: S5129 (30b) SN: 3013 SN: 660 ID #	probability are given on the following pages ar ory facility: environment temperature (22 ± 3)*4 Cal Date (Certificate No.) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 1-Apr-10 (No. 217-01136) 30-Mar-10 (No. 217-01161) 30-Mar-10 (No. 217-01161) 30-Dec-09 (No. ES3-3013_Dec09) 20-Apr-10 (No. DAE4-660_Apr10)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Mar-11 Mar-11 Mar-11 Mar-11 Dec-10 Apr-11
The measurements and the un	Certainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277 MY41496087 SN: S5056 (20b) SN: S5056 (20b) SN: S5129 (30b) SN: 3013 SN: 660 ID # US3642U01700	probability are given on the following pages ar           ory facility: environment temperature (22 ± 3)*4           Cal Date (Certificate No.)           1-Apr-10 (No. 217-01136)           1-Apr-10 (No. 217-01136)           1-Apr-10 (No. 217-01136)           30-Mar-10 (No. 217-01136)           30-Mar-10 (No. 217-01161)           30-Mar-10 (No. 217-01161)           30-Dec-09 (No. ES3-3013_Dec09)           20-Apr-10 (No. DAE4-660_Apr10)           Check Date (in house)           4-Aug-99 (in house check Oct-09)	nd are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Mar-11 Mar-11 Mar-11 Dec-10 Apr-11 Scheduled Check In house check: Oct-11 In house check: Oct10
The measurements and the unit All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards RF generator HP 8648C	Certainties with confidence lucted in the closed laborat &TE critical for calibration) ID # GB41293874 MY41495277 MY41498087 SN: S5054 (3c) SN: S5054 (3c) SN: S5086 (20b) SN: S5129 (30b) SN: S5129 (30b) SN: 660 ID # US3642U01700 US37390585	Cal Date (Certificate No.)           1-Apr-10 (No. 217-01136)           30-Mar-10 (No. 217-01159)           30-Mar-10 (No. 217-01161)           30-Mar-10 (No. 217-01160)           30-Dec-09 (No. ES3-3013_Dec09)           20-Apr-10 (No. DAE4-660_Apr10)           Check Date (in house)           4-Aug-99 (in house check Oct-09)           18-Oct-01 (in house check Oct-09)	d are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-11 Apr-11 Mar-11 Mar-11 Mar-11 Dec-10 Apr-11 Scheduled Check In house check: Oct-11



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



SWISS CP D Z S Schweizerischer Kalibrierdienst

C Service suisse d'étalonnage

Servizio svizzero di taratura Surias Calibratica Service

Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

### Glossary:

eleesely.	
TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx,y,z
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C	modulation dependent linearization parameters
Polarization $\phi$	φ rotation around probe axis
Polarization 9	9 rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., 9 = 0 is normal to probe axis
	i.e., a – o is normal to probe axis

### Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

### Methods Applied and Interpretation of Parameters:

- NORMx,y,z: Assessed for E-field polarization 9 = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not effect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z \* frequency\_response (see Frequency Response Chart). This linearization is
  implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included
  in the stated uncertainty of ConvF.
- DCPx, y, z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- Ax,y,z; Bx,y,z; Cx,y,z, VRx,y,z: A, B, C are numerical linearization parameters assessed based on the data of
  power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the
  maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom
  exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

Certificate No: ET3-1787\_May10

Page 2 of 11



May 18, 2010

# Probe ET3DV6

## SN:1787

Manufactured:	May 28, 2003
Last calibrated:	May 26, 2009
Recalibrated:	May 18, 2010

Calibrated for DASY/EASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: ET3-1787\_May10

Page 3 of 11

SPORTON INTERNATIONAL INC.



May 18, 2010

## DASY/EASY - Parameters of Probe: ET3DV6 SN:1787

### **Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm (µV/(V/m) <sup>2</sup> ) <sup>A</sup>	1.60	1.79	2.10	± 10.1%
DCP (mV) <sup>B</sup>	92.4	95.5	91.0	

### **Modulation Calibration Parameters**

UID	Communication System Name	PAR		A dB	B dBuV	с	VR mV	Unc <sup>E</sup> (k=2)
10000	CW	0.00	х	0.00	0.00	1.00	300.0	± 1.5%
			Y	0.00	0.00	1.00	300.0	
			Z	0.00	0.00	1.00	300.0	

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of NormX, Y,Z do not affect the E<sup>3</sup>-field uncertainty inside TSL (see Pages 5 and 6).

<sup>8</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the maximum deviation from linear response applying recatangular distribution and is expressed for the square of the field value.

Certificate No: ET3-1787\_May10

Page 4 of 11



### May 18, 2010

## DASY/EASY - Parameters of Probe: ET3DV6 SN:1787

### Calibration Parameter Determined in Head Tissue Simulating Media

f [MHz]	Validity [MHz] <sup>C</sup>	Permittivity	Conductivity	ConvF X	ConvF Y	ConvF Z	Alpha	Depth Unc (k=2)
750	± 50 / ± 100	41.5 ± 5%	0.90 ± 5%	6.56	6.56	6.56	0.52	1.96 ± 11.0%
835	± 50 / ± 100	41.9 ± 5%	0.89 ± 5%	6.21	6.21	6.21	0.42	2.23 ± 11.0%
1750	± 50 / ± 100	40.1 ± 5%	1.37 ± 5%	5.36	5.36	5.36	0.49	1.18 ± 11.0%
1900	± 50 / ± 100	40.0 ± 5%	1.40 ± 5%	5.09	5.09	5.09	0.66	2.20 ± 11.0%
2450	± 50 / ± 100	39.2 ± 5%	1.80 ± 5%	4.50	4.50	4.50	0.99	1.63 ± 11.0%

<sup>o</sup> The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.

Certificate No: ET3-1787\_May10

Page 5 of 11



May 18, 2010

## DASY/EASY - Parameters of Probe: ET3DV6 SN:1787

Calibration Parameter Determined in Body Tissue Simulating Media

f [MHz]	Validity [MHz] <sup>C</sup>	Permittivity	Conductivity	ConvF X C	ConvF Y	ConvF Z	Alpha	Depth Unc (k=2)
750	± 50 / ± 100	55.5 ± 5%	0.96 ± 5%	6.22	6.22	6.22	0.48	2.20 ± 11.0%
835	± 50 / ± 100	55.2 ± 5%	0.97 ± 5%	6.12	6,12	6.12	0.39	2.45 ± 11.0%
1750	± 50 / ± 100	$53.4 \pm 5\%$	1.49 ± 5%	4.72	4.72	4.72	0.63	2.90 ± 11.0%
1900	± 50 / ± 100	$53.3\pm5\%$	1.52 ± 5%	4.47	4.47	4.47	0.88	2.39 ± 11.0%
2450	± 50 / ± 100	52.7 ± 5%	1.95 ± 5%	4.03	4.03	4.03	0.99	1.35 ± 11.0%

<sup>C</sup> The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


Certificate No: ET3-1787 May10

Page 6 of 11



May 18, 2010

## Frequency Response of E-Field




Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)

Certificate No: ET3-1787\_May10


Page 7 of 11



May 18, 2010



## Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$




Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

Certificate No: ET3-1787\_May10

Page 8 of 11



May 18, 2010




Uncertainty of Linearity Assessment: ± 0.6% (k=2)

Certificate No: ET3-1787\_May10

Page 9 of 11




### May 18, 2010



## **Conversion Factor Assessment**

## **Deviation from Isotropy in HSL**

Error (\$, 9), f = 900 MHz



Uncertainty of Spherical Isotropy Assessment: ± 2.6% (k=2)

Certificate No: ET3-1787\_May10

Page 10 of 11



May 18, 2010

## **Other Probe Parameters**

Sensor Arrangement	Triangular
Connector Angle (°)	Not applicable
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	10 mm
Tip Diameter	6.8 mm
Probe Tip to Sensor X Calibration Point	2.7 mm
Probe Tip to Sensor Y Calibration Point	2.7 mm
Probe Tip to Sensor Z Calibration Point	2.7 mm
Recommended Measurement Distance from Surface	4 mm

Certificate No: ET3-1787\_May10

Page 11 of 11

### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Auden

Client





S

Accreditation No.: SCS 108

S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage

- Servizio svizzero di taratura
- Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Certificate No: ES3-3071\_Jun10

CALIBRATION	CERTIFICATE
of tails to the t	O EITIT TOTTE

Object	ES3DV3 - SN:3	071	
Calibration procedure(s)		QA CAL-23.v3 and QA CAL-25.v2 edure for dosimetric E-field probes	
Calibration date:	June 22, 2010		
	물질 것 같은 것 같은 영국의 감독을 했다.	tional standards, which realize the physical uni probability are given on the following pages and	
		ory facility: environment temperature (22 ± 3)°C	and humidity < 70%.
Calibration Equipment used (M&	P.		
rimary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
ower meter E4419B	GB41293874	1-Apr-10 (No. 217-01136)	Apr-11
ower sensor E4412A	MY41495277	1-Apr-10 (No. 217-01136)	Apr-11
ower sensor E4412A	MY41498087	1-Apr-10 (No. 217-01136)	Apr-11
eference 3 dB Attenuator	SN: S5054 (3c)	30-Mar-10 (No. 217-01159)	Mar-11
eference 20 dB Attenuator	SN: S5086 (20b)	30-Mar-10 (No. 217-01161)	Mar-11
eference 30 dB Attenuator	SN: S5129 (30b)	30-Mar-10 (No. 217-01160)	Mar-11
	SN: 3013		
	011.0010	30-Dec-09 (No. ES3-3013_Dec09)	Dec-10
	SN: 660	30-Dec-09 (No. ES3-3013_Dec09) 20-Apr-10 (No. DAE4-660_Apr10)	Dec-10 Apr-11
AE4		나라는 것 같은 것 같은 것이 같은 것 같은 것 같은 것 ㅠㅠ 말 것 같아요	
AE4 secondary Standards	SN: 660	20-Apr-10 (No. DAE4-660_Apr10)	Apr-11
AE4 econdary Standards F generator HP 8648C	SN: 660	20-Apr-10 (No. DAE4-660_Apr10) Check Date (in house)	Apr-11 Scheduled Check
AE4 econdary Standards F generator HP 8648C	SN: 660	20-Apr-10 (No. DAE4-660_Apr10) Check Date (in house) 4-Aug-99 (in house check Oct-09)	Apr-11 Scheduled Check In house check: Oct-11
DAE4 Secondary Standards RF generator HP 8648C letwork Analyzer HP 8753E	SN: 660 ID # US3642U01700 US37390585	20-Apr-10 (No. DAE4-660_Apr10) Check Date (in house) 4-Aug-99 (in house check Oct-09) 18-Oct-01 (in house check Oct-09)	Apr-11 Scheduled Check In house check: Oct-11 In house check: Oct10
DAE4 Secondary Standards RF generator HP 8648C Network Analyzer HP 8753E	SN: 660 ID # US3642U01700 US37390585 Name	20-Apr-10 (No. DAE4-660_Apr10) Check Date (in house) 4-Aug-99 (in house check Oct-09) 18-Oct-01 (in house check Oct-09) Function	Apr-11 Scheduled Check In house check: Oct-11 In house check: Oct10
Reference Probe ES3DV2 DAE4 Secondary Standards RF generator HP 8648C Network Analyzer HP 8753E Calibrated by:	SN: 660 ID # US3642U01700 US37390585 Name	20-Apr-10 (No. DAE4-660_Apr10) Check Date (in house) 4-Aug-99 (in house check Oct-09) 18-Oct-01 (in house check Oct-09) Function	Apr-11 Scheduled Check In house check: Oct-11 In house check: Oct10

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst

C Service suisse d'étalonnage

Servizio svizzero di taratura

S Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

### Glossary:

TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx,y,z
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C	modulation dependent linearization parameters
Polarization $\phi$	φ rotation around probe axis
Polarization 9	9 rotation around an axis that is in the plane normal to probe axis (at measurement center),
	i.e., $\vartheta = 0$ is normal to probe axis

### Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

### Methods Applied and Interpretation of Parameters:

- NORMx, y,z: Assessed for E-field polarization θ = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx, y,z are only intermediate values, i.e., the uncertainties of NORMx, y,z does not effect the E<sup>2</sup>-field uncertainty inside TSL (see below *ConvF*).
- NORM(f)x,y,z = NORMx,y,z \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of *ConvF*.
- DCPx, y, z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- *Ax,y,z; Bx,y,z; Cx,y,z, VRx,y,z: A, B, C* are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. *VR* is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx, y, z \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

# Probe ES3DV3

## SN:3071

Manufactured: Last calibrated: Recalibrated: December 14, 2004 June 22, 2009 June 22, 2010

Calibrated for DASY/EASY Systems

(Note: non-compatible with DASY2 system!)

## DASY/EASY - Parameters of Probe: ES3DV3 SN:3071

### **Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm $(\mu V/(V/m)^2)^A$	1.24	1.22	0.97	± 10.1%
DCP (mV) <sup>B</sup>	96.6	92.2	92.6	

## **Modulation Calibration Parameters**

UID	Communication System Name	PAR		A dB	B dBuV	с	VR mV	Unc <sup>E</sup> (k=2)
10000	CW	0.00	Х	0.00	0.00	1.00	300.0	± 1.5%
			Y	0.00	0.00	1.00	300.0	
			Z	0.00	0.00	1.00	300.0	1

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>&</sup>lt;sup>A</sup> The uncertainties of NormX,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 6).

<sup>&</sup>lt;sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>&</sup>lt;sup>E</sup> Uncertainty is determined using the maximum deviation from linear response applying recatangular distribution and is expressed for the square of the field value.

## DASY/EASY - Parameters of Probe: ES3DV3 SN:3071

Validity [MHz] <sup>C</sup>	Permittivity	Conductivity	ConvF X	ConvF Y	ConvF Z	Alpha	Depth Unc (k=2)
± 50 / ± 100	41.5 ± 5%	0.90 ± 5%	5.81	5.81	5.81	0.98	1.02 ± 11.0%
± 50 / ± 100	41.5 ± 5%	0.97 ± 5%	5.67	5.67	5.67	0.75	1.15 ± 11.0%
± 50 / ± 100	40.1 ± 5%	1.37 ± 5%	4.94	4.94	4.94	0.35	1.77 ± 11.0%
± 50 / ± 100	40.0 ± 5%	1.40 ± 5%	4.73	4.73	4.73	0.57	1.35 ± 11.0%
± 50 / ± 100	40.0 ± 5%	1.40 ± 5%	4.67	4.67	4.67	0.56	1.35 ± 11.0%
± 50 / ± 100	39.2 ± 5%	1.80 ± 5%	4.20	4.20	4.20	0.38	1.93 ± 11.0%
	$\pm 50 / \pm 100$ $\pm 50 / \pm 100$	$\begin{array}{c} \pm 50 \ / \pm 100 \\ \pm 50 \ / \pm 5\% \\ \end{array}$	$\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.90 \pm 5\%$ $\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.97 \pm 5\%$ $\pm 50 / \pm 100$ $40.1 \pm 5\%$ $1.37 \pm 5\%$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$	$\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.90 \pm 5\%$ $5.81$ $\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.97 \pm 5\%$ $5.67$ $\pm 50 / \pm 100$ $40.1 \pm 5\%$ $1.37 \pm 5\%$ $4.94$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$ $4.73$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$ $4.67$	$\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.90 \pm 5\%$ $5.81$ $5.81$ $\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.97 \pm 5\%$ $5.67$ $5.67$ $\pm 50 / \pm 100$ $40.1 \pm 5\%$ $1.37 \pm 5\%$ $4.94$ $4.94$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$ $4.73$ $4.73$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$ $4.67$ $4.67$	$\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.90 \pm 5\%$ $5.81$ $5.81$ $5.81$ $\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.97 \pm 5\%$ $5.67$ $5.67$ $5.67$ $\pm 50 / \pm 100$ $40.1 \pm 5\%$ $1.37 \pm 5\%$ $4.94$ $4.94$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$ $4.73$ $4.73$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$ $4.67$ $4.67$	$\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.90 \pm 5\%$ $5.81$ $5.81$ $5.81$ $0.98$ $\pm 50 / \pm 100$ $41.5 \pm 5\%$ $0.97 \pm 5\%$ $5.67$ $5.67$ $5.67$ $0.75$ $\pm 50 / \pm 100$ $40.1 \pm 5\%$ $1.37 \pm 5\%$ $4.94$ $4.94$ $0.35$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$ $4.73$ $4.73$ $4.73$ $\pm 50 / \pm 100$ $40.0 \pm 5\%$ $1.40 \pm 5\%$ $4.67$ $4.67$ $4.67$

### Calibration Parameter Determined in Head Tissue Simulating Media


<sup>c</sup> The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.

## DASY/EASY - Parameters of Probe: ES3DV3 SN:3071

f [MHz]	Validity [MHz] <sup>C</sup>	Permittivity	Conductivity	ConvFX Co	nvFY (	ConvF Z	Alpha	Depth Unc (k=2)
835	± 50 / ± 100	55.2 ± 5%	0.97 ± 5%	5.79	5.79	5.79	0.73	1.17 ± 11.0%
900	± 50 / ± 100	55.0 ± 5%	1.05 ± 5%	5.71	5.71	5.71	0.85	1.14 ± 11.0%
1750	± 50 / ± 100	53.4 ± 5%	1.49 ± 5%	4.52	4.52	4.52	0.40	1.79 ± 11.0%
1900	± 50 / ± 100	53.3 ± 5%	1.52 ± 5%	4.30	4.30	4.30	0.38	2.04 ± 11.0%
2000	± 50 / ± 100	53.3 ± 5%	1.52 ± 5%	4.36	4.36	4.36	0.42	1.91 ± 11.0%
2450	± 50 / ± 100	52.7 ± 5%	1.95 ± 5%	4.00	4.00	4.00	0.80	1.25 ± 11.0%

## Calibration Parameter Determined in Body Tissue Simulating Media


<sup>c</sup> The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.




## **Frequency Response of E-Field**

(TEM-Cell:ifi110 EXX, Waveguide: R22)


Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)



## Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$




Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)



Dynamic Range f(SAR<sub>head</sub>) (Waveguide R22, f = 1800 MHz)


Uncertainty of Linearity Assessment: ± 0.6% (k=2)



## **Conversion Factor Assessment**

**Deviation from Isotropy in HSL** 

Error (φ, ϑ), f = 900 MHz




Uncertainty of Spherical Isotropy Assessment: ± 2.6% (k=2)

## **Other Probe Parameters**

Sensor Arrangement	Triangular				
Connector Angle (°)	Not applicable				
Mechanical Surface Detection Mode	enabled				
Optical Surface Detection Mode	disabled				
Probe Overall Length	337 mm				
Probe Body Diameter	10 mm				
Tip Length	10 mm				
Tip Diameter	4.0 mm				
Probe Tip to Sensor X Calibration Point	2 mm				
Probe Tip to Sensor Y Calibration Point	2 mm				
Probe Tip to Sensor Z Calibration Point	2 mm				
Recommended Measurement Distance from Surface	3 mm				



## Appendix D. Product Photos





## Appendix E. Test Setup Photos



Rear Face (0cm)

Top side (0cm, secondary landscape)