

Typenschildbeschriftung:

Type 5367/100

Date: MM.JJJJ

F-Nr. xxxxx

MFR I.D.: P

100-120/220-240V 4 / 2 A 50-60Hz

MEDICAL ELECTRICAL
EQUIPMENT
WITH RESPECT TO ELECTRIC
SHOCK, FIRE, AND
MECHANICAL HAZARDS ONLY
IN ACCORDANCE WITH UL2601-
1 /
CAN/CSA C22.2 NO.601.1
9F03

FCC ID: HPL5367A



AGFA HEALTHCARE NV

Made in Germany
德国制造

Hinweis: Bei Bestellung die aktuelle F-Nr. angeben.


Datecode: z.B. „01.1999“

MM = Monat, zweistellig

JJJJ = Jahr **vierstellig**1999

xxxx = Fertigungsnummer

P = Manufacturer Factory Location and Identification Peissenberg

						Masstab	Gefertigt aus:	
						2 : 1	TTR-Klebeschild	
					Datum	Name	Benennung:	Typ.
					Erstellt		Typenschild Drystar Axys	5367/ 100
					Prüfer			
					AGFA  Fototechnik		Sach-NR:	
Änd.- Zst.	Kommt vor	Mitteilung NR.	Datum	Name			Ersatz für:	Blatt von Blatt 1/1