

REV

.950

PN: Z10020-014 CONTAINS:

FCC ID: HSW-608B

MODEL: DR-10100

Operation of this equipment requires the prior coordination with a frequency coordinator designated by the FCC for the Wireless Medical Telemetry Service.

**R ONLY-US**

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

LABEL PN:  
F-2050-5592 REV A

Suitable for Use in Other Environmental Air Space in Accordance with Section 300.22(c) of the National Electric Code.



4.500

NOTES:

- 1. MYLAR LABEL, COLOR IS SILVER, LOGOS AND LETTERING ARE BLACK.
- 2. TEXT FONT IS SWITZERLAND BOLD.
- 3. UL LOGO TO COMPLY WITH UL GUIDELINES.

<b>muRata</b>		MURATA	
INNOVATOR IN ELECTRONICS		3079 PREMIERE PARKWAY, SUITE 140 DULUTH, GA 30097 USA	
Title			
Z10020-014 AP FCC Label WIT608B			
SIZE	FSCM NO.	DWG NO.	REV
A		8055927	A
SCALE	NTS	19MAR2021	SHEET 1 of 1