MEASUREMENT REPORT # 31-20

10.07.2020 Page 2 of 2

MEASUREMENT CONDITIONS

Temperature: 22.5 °C Humidity: 43 % Pressure: 101.0 kPa

MEASUREMENT EQUIPMENT

| | TVILITOUTCHVILIVI | | | | | |
|---------|------------------------|--------------|-------------------|-----------------------|-------------|--|
| Model | Model Description | Equipment ID | Cal Due Date | Certificate Number | Trace Value | |
| M 546 | Reference power meter | 163 | 24 March 2022 | 1/111-175-20 | RF Power | |
| M 534 | Reference power meter | 161 | 24 March 2022 | 1/111-173-20 | RF Power | |
| G4-186 | Signal generator | 5 | 10 July 2020 | 23-19 | RF Power | |
| RG4-14 | Signal generator | 22 | 10 July 2020 | 24-19 | RF Power | |
| V7-34 | Universal voltmeter | 0067787 | 27 September 2020 | 1994-42 | DC Voltage | |
| RCH3-72 | Frequency meter | 931200 | 13 September 2020 | 2261-43 | Frequency | |
| P6-134 | Measuring horn antenna | 14002 | 23 September 2021 | 2372-43 | Gain | |
| P6-31A | Measuring horn antenna | 35864 | 23 September 2021 | 2368-43 | Gain | |

MEASUREMENT RESULTS

Distance between the testing and generating antennas was 1.8 m at 60-75 GHz and 1.5 m at 90 GHz.

Table 1

| | | T | · |
|--|-------|-------|-------|
| Frequency, GHz | 60 | 75 | 90 |
| Power density of electromagnetic field, W/m ² | 0.063 | 0.081 | 0.121 |
| Maximum level of measured power, dBm | -16.6 | -16.9 | -16.7 |
| Gain, dBi | 22.4 | 23.0 | 23.0 |
| Expanded uncertainty, dB | 2.5 | 2.5 | 2.5 |
| Antenna Factor, dB/m | 43.4 | 44.7 | 46.3 |

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. This probability corresponds to a coverage factor of k=2 for a normal distribution.

Engineer

Quality Manager

This measurement report issued in duplicate and sent to:

^{1.} SPORTON INTERNATIONAL (SHENZHEN) INC. 1/F, 2/F, Bldg 5, Shiling Industrial Zone, Xinwei Village, Xili, Nanshan Shenzhen, 518055, People's Republic of China

^{2.} Calibration Laboratory of Microwave Measuring Equipment

Calibration Laboratory of Microwave Measuring Equipment of MWMLab





Calibration certificate



ISO 17025 ACCREDITED LABORATORY



Accreditation certificate No.

№ BY/112 5.0065

09.01.2015

Certificate number 32-20 Date when calibrated 10.07.20 Page

Item

calibrated

Antenna QWH-FPRR00 # 1011500009

Customer

SPORTON INTERNATIONAL (SHENZHEN) INC.

1/F, 2/F, Bldg 5, Shiling Industrial Zone, Xinwei Village, Xili, Nanshan Shenzhen, 518055, People's Republic of China

Method of

GOST 20271.1, MK KL 8.2-16 calibration

All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of RF. Conversion loss measurements above 178 GHz are to confirm operation functionality and traceable only to MWMLab standards and OML. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.

Authorising signature 300 A TO.

/ Technical manager

Date of issue

MWM Lab. BSUIR, 6, P. Browkistr Minsk, 220013, Belarus Phone/Fax: +375 17 293-84-96/E-mail: info@mwmlab.com

Calibration Certificate

Certificate number

32-20

Page 2

Calibration is performed by using

| Model | Model Description | Equipment ID | Cal Due Date | Certificate Number | Trace Value |
|---------|------------------------|--------------|-------------------|--------------------|-------------|
| M 523 | Reference power meter | 162 | 24 March 2022 | 1/111-172-20 | RF Power |
| M 534 | Reference power meter | 161 | 24 March 2022 | 1/111-173-20 | RF Power |
| G4-161m | Signal generator | 282 | 10 July 2020 | 25-19 | RF Power |
| RG4-14 | Signal generator | 22 | 10 July 2020 | 24-19 | RF Power |
| V7-34 | Universal voltmeter | 0067787 | 27 September 2020 | 1994-42 | DC Voltage |
| RCH3-72 | Frequency meter | 931200 | 13 September 2020 | 2261-43 | Frequency |
| P6-32 | Measuring horn antenna | 115671 | 23 September 2021 | 2369-43 | Gain |
| P6-31A | Measuring horn antenna | 35864 | 23 September 2021 | 2368-43 | Gain |

Calibration conditions

Temperature: 22.5 °C.

Humidity: 43.0 %.

Pressure: 101.0 kPa.

Calibration results are given in the measurement report # 32-20

| COURT | THE PROPERTY OF THE PARTY OF THE PROPERTY OF T | | | | | |
|-------|--|-------------------------|------------------------------------|--|--|--|
| # | Parameter | Specifications required | Specifications tested and measured | | | |
| 1 | Frequency range | 90 – 140 GHz | Corresponds | | | |
| 2 | Antenna Gain | 22.6* dBi | Corresponds (Table 1) | | | |
| 3 | Antenna Factor | 48.5 dB/m | Corresponds (Table 1) | | | |

^{* -} Expanded uncertainty of measurements 2.5 dB.

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %. This probability corresponds to a coverage factor of k=2 for a normal distribution.

Signature of the person who has performed calibration

/ Engineer

Calibration Laboratory of
Microwave Measuring Equipment

Accreditation certificate No. BY/112 5.0065

Address: 6, P. Brovki str., Minsk

220013, Belarus

Phone/Fax: +375 17 2938496

Technical Manager

mwm

Total and the state of the stat

MEASUREMENT REPORT # 32-20

July 10, 2020

| Customer: | SPORTON INTERNATIONAL (SHENZHEN) INC. 1/F, 2/F, Bldg 5, Shiling Industrial Zone, Xinwei Village, Xili, Nanshan Shenzhen, 518055, People's Republic of China |
|------------------------------|--|
| Item calibrated: | Antenna QWH-FPRR00 # 1011500009 |
| Method of calibration: | GOST 20271.1, MK KL 8.2-16 |
| Number of samples: | One |
| Delivery date of the sample: | 09.06.2020 |
| Date of calibration: | From 09.06.2020 to 10.07.2020 |

MEASUREMENT REPORT # 32-20

10.07.2020 Page 2 of 2

MEASUREMENT CONDITIONS

Temperature: 22.5 °C Humidity: 43 % Pressure: 101.0 kPa

MEASUREMENT EQUIPMENT

| | WILLIOUNDINI LYULINI | | | | | |
|---------|------------------------|--------------|-------------------|--------------------|-------------|--|
| Model | Model Description | Equipment ID | Cal Due Date | Certificate Number | Trace Value | |
| M 523 | Reference power meter | 162 | 24 March 2022 | 1/111-172-20 | RF Power | |
| M 534 | Reference power meter | 161 | 24 March 2022 | 1/111-173-20 | RF Power | |
| G4-161m | Signal generator | 282 | 10 July 2020 | 25-19 | RF Power | |
| RG4-14 | Signal generator | 22 | 10 July 2020 | 24-19 | RF Power | |
| V7-34 | Universal voltmeter | 0067787 | 27 September 2020 | 1994-42 | DC Voltage | |
| RCH3-72 | Frequency meter | 931200 | 13 September 2020 | 2261-43 | Frequency | |
| P6-32 | Measuring horn antenna | 115671 | 23 September 2021 | 2369-43 | Gain | |
| P6-31A | Measuring horn antenna | 35864 | 23 September 2021 | 2368-43 | Gain | |

MEASUREMENT RESULTS

Distance between the testing and generating antennas was 1.5 m at 90-115 GHz and 1.2 m at 140 GHz.

Table 1

| Frequency, GHz | 90 | 115 | 140 |
|--|-------|-------|-------|
| Power density of electromagnetic field, W/m ² | 0.125 | 0.174 | 0.212 |
| Maximum level of measured power, dBm | -17.2 | -17.5 | -17.9 |
| Gain, dBi | 22.4 | 22.8 | 23.2 |
| Expanded uncertainty, dB | 2.5 | 2.5 | 2.5 |
| Antenna Factor, dB/m | 46.9 | 48.6 | 50.0 |

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %. This probability corresponds to a coverage factor of k=2 for a normal distribution.

Engineer

Quality Manager

This measurement report issued in duplicate and sent to:

Duplication of Measurement report (complete or partial) must be authorized by the laboratory.

^{1.} SPORTON INTERNATIONAL (SHENZHEN) INC. 1/F, 2/F, Bldg 5, Shiling Industrial Zone, Xinwei Village, Xili, Nanshan Shenzhen, 518055, People's Republic of China

^{2.} Calibration Laboratory of Microwave Measuring Equipment

Calibration Laboratory of Microwave Measuring Equipment of MWMLab





Calibration certificate

ISO 17025

ACCREDITED LABORATORY



Accreditation certificate No.

Nº BY/112 5.0065

of

09.01.2015

Certificate number 34-20 Date when calibrated 10.07.20 Page 1 of 2

Item

calibrated

Antenna QWH-GPRR00 # 01

SPORTON INTERNATIONAL (SHENZHEN) INC.

Customer

1/F, 2/F, Bldg 5, Shiling Industrial Zone, Xinwei Village, Xili, Nanshan

Shenzhen, 518055, People's Republic of China

Method of

calibration

GOST 20271.1, MK KL 8.2-16

All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of RF. Conversion loss measurements above 178 GHz are to confirm operation functionality and traceable only to MWMLab standards and OML. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.

Authorising

/ Technical manager

Date of issue

10.07.20

MWM Lab. BSUIR, 6, P.Brovki str., Minsk, 220013, Belarus

Phone/Fax: +375 17 293-84-96/E-mail: info@mwmlab.com

Calibration Certificate

Certificate number

34-20

Page

Calibration is performed by using

| Model | Model Description | Equipment | Cal Due Date | Certificate Number | Trace Value |
|---------|------------------------|-----------|-------------------|--------------------|-------------|
| M 523 | Reference power meter | 162 | 24 March 2022 | 1/111-172-20 | RF Power |
| M 514 | Reference power meter | 165 | 24 March 2022 | 1/111-176-20 | RF Power |
| 02 | Frequency multiplier | 02 | 23 January 2020 | 02-19 | RF Power |
| RG4-14 | Signal generator | 22 | 10 July 2020 | 24-19 | RF Power |
| V7-34 | Universal voltmeter | 0067787 | 27 September 2020 | 1994-42 | DC Voltage |
| RCH3-72 | Frequency meter | 931200 | 13 September 2020 | 2261-43 | Frequency |
| P6-32 | Measuring horn antenna | 115671 | 23 September 2021 | 2369-43 | Gain |

Calibration conditions

Temperature: 22.5 °C.

Humidity: 43.0 %.

Pressure: 101.0 kPa.

Calibration results are given in the measurement report # 34-20

| | B- CILL THE BRICKS CHIEF TO THE O | | | | | |
|---|-----------------------------------|-------------------------|------------------------------------|--|--|--|
| # | Parameter | Specifications required | Specifications tested and measured | | | |
| 1 | Frequency range | 140 – 220 GHz | Corresponds | | | |
| 2 | Antenna Gain | 22.6* dBi | Corresponds (Table 1) | | | |
| 3 | Antenna Factor | 52.5 dB/m | Corresponds (Table 1) | | | |

^{* –} Expanded uncertainty of measurements 3.0 dB.

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %. This probability corresponds to a coverage factor of k=2 for a normal distribution.

Signature of the person who has performed calibration

/ Engineer

Calibration Laboratory of Microwave Measuring Equipment

Accreditation certificate No. BY/112 5.0065

Address: 6, P. Brovki str., Minsk

220013, Belarus

Phone/Fax: +375 17 2938496



MEASUREMENT REPORT # 34-20

July 10, 2020

| Customer: | SPORTON INTERNATIONAL (SHENZHEN) INC. 1/F, 2/F, Bldg 5, Shiling Industrial Zone, Xinwei Village, Xili, Nanshan Shenzhen, 518055, |
|------------------------------|---|
| Item calibrated: | People's Republic of China Antenna QWH-GPRR00 # 01 |
| Method of calibration: | GOST 20271.1, MK KL 8.2-16 |
| Number of samples: | One |
| Delivery date of the sample: | 09.06.2020 |
| Date of calibration: | From 09.06.2020 to 10.07.2020 |

MEASUREMENT REPORT # 34-20

10.07.2020

Page 2 of 2

MEASUREMENT CONDITIONS

Temperature: 22.5 °C Humidity: 43 % Pressure: 101.0 kPa

MEASUREMENT EQUIPMENT

| Model | Model Description | Equipment ID | Cal Due Date | Certificate Number | Trace Value |
|---------|------------------------|--------------|-------------------|--------------------|----------------|
| M 523 | Reference power meter | 162 | 24 March 2022 | 1/111-172-20 | RF Power |
| M 514 | Reference power meter | 165 | 24 March 2022 | 1/111-176-20 | RF Power |
| 02 | Frequency multiplier | 02 | 23 January 2020 | 02-19 | RF Power |
| RG4-14 | Signal generator | 22 | 10 July 2020 | 24-19 | RF Power |
| V7-34 | Universal voltmeter | 0067787 | 27 September 2020 | 1994-42 | DC Voltage |
| RCH3-72 | Frequency meter | 931200 | 13 September 2020 | 2261-43 | Frequency |
| P6-32 | Measuring horn antenna | 115671 | 23 September 2021 | 2369-43 | Gain |

MEASUREMENT RESULTS

Distance between the testing and generating antennas was 1 m at 140 GHz, 0.6 m at 180 GHz and 0.3 m at 220 GHz.

Table 1

| Frequency, GHz | 140 | 180 | 220 |
|--|-------|-------|-------|
| Power density of electromagnetic field, W/m ² | 0.311 | 0.369 | 0.327 |
| Maximum level of measured power, dBm | -17.1 | -18.1 | -20.2 |
| Gain, dBi | 22.3 | 22.8 | 22.9 |
| Expanded uncertainty, dB | 2.5 | 3.0 | 3.0 |
| Antenna Factor, dB/m | 50.8 | 52.5 | 54.2 |

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %. This probability corresponds to a coverage factor of k=2 for a normal distribution.

Engineer

Quality Manager

This measurement report issued in duplicate and sent to:

1. SPORTON INTERNATIONAL (SHENZHEN) INC. 1/F, 2/F, Bldg 5, Shiling Industrial Zone, Xinwei Village, Xili, Nanshan Shenzhen, 518055, People's Republic of China

2. Calibration Laboratory of Microwave Measuring Equipment

Duplication of Measurement report (complete or partial) must be authorized by the laboratory.