

System Check_Head_6500MHz

DUT:D6.5GHzV2 - SN:1026

Communication System: Custom Band; Frequency: 6500.0

Medium: HSL. Medium parameters used: $f= 6500.0$ MHz; $\sigma= 6.08$ S/m; $\epsilon_r = 34.0$

Ambient Temperature: 23.4°C; Liquid Temperature: 22.6°C

DASY6 Configuration:

- Probe: EX3DV4 - SN3935; ConvF(5.8, 5.8, 5.8); Calibrated: 2022/6/20
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1210; Calibrated: 2022/4/12
- Phantom: Twin-SAM V8.0 (Left); Serial: 2035; Section: Flat
- Measurement Software: cDASY6 V6.6.0.13926
- UID: CW, 0--

Area Scan (51.0 mm x 51.0 mm): Measurement Grid: 8.5 mm x 8.5 mm

SAR (1g) = 21.7 W/kg; SAR (10g) = 4.63 W/kg;

Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 3.4 mm x 3.4 mm x 1.4 mm

Power Drift = 0.01 dB

SAR (1g) = 27.1 W/kg; SAR (10g) = 4.88 W/kg;

psAPD (4.0cm², sq) = 119[W/m²]

