

Appendix 2

Probe Calibration Certificates

Calibration Certificate

Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1523

Place of Calibration:

Zurich

Date of Calibration:

May 11, 2001

Calibration Interval:

12 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

Alain Kohler

Approved by:

[Signature]

Probe ET3DV6

SN:1523

Manufactured:	March 21, 2000
Last calibration:	April 7, 2000
Recalibrated:	May 11, 2001

Calibrated for System DASY3

DASY3 - Parameters of Probe: ET3DV6 SN:1523**Sensitivity in Free Space**

NormX	1.57 $\mu\text{V}/(\text{V}/\text{m})^2$
NormY	1.39 $\mu\text{V}/(\text{V}/\text{m})^2$
NormZ	1.58 $\mu\text{V}/(\text{V}/\text{m})^2$

Diode Compression

DCP X	96 mV
DCP Y	96 mV
DCP Z	96 mV

Sensitivity in Tissue Simulating Liquid


Head	450 MHz	$\epsilon_r = 43.5 \pm 5\%$	$\sigma = 0.87 \pm 10\%$ mho/m
ConvF X	6.95 extrapolated		Boundary effect:
ConvF Y	6.95 extrapolated		Alpha 0.57
ConvF Z	6.95 extrapolated		Depth 1.94
Head	900 MHz	$\epsilon_r = 42 \pm 5\%$	$\sigma = 0.97 \pm 10\%$ mho/m
ConvF X	6.44 $\pm 7\%$ (k=2)		Boundary effect:
ConvF Y	6.44 $\pm 7\%$ (k=2)		Alpha 0.58
ConvF Z	6.44 $\pm 7\%$ (k=2)		Depth 1.98
Head	1500 MHz	$\epsilon_r = 40.4 \pm 5\%$	$\sigma = 1.23 \pm 10\%$ mho/m
ConvF X	5.77 interpolated		Boundary effect:
ConvF Y	5.77 interpolated		Alpha 0.60
ConvF Z	5.77 interpolated		Depth 2.03
Head	1800 MHz	$\epsilon_r = 40 \pm 5\%$	$\sigma = 1.40 \pm 10\%$ mho/m
ConvF X	5.43 $\pm 7\%$ (k=2)		Boundary effect:
ConvF Y	5.43 $\pm 7\%$ (k=2)		Alpha 0.60
ConvF Z	5.43 $\pm 7\%$ (k=2)		Depth 2.06

Sensor Offset


Probe Tip to Sensor Center	2.7	mm
Optical Surface Detection	1.3 \pm 0.2	mm

Receiving Pattern (ϕ), $\theta = 0^\circ$


f = 30 MHz, TEM cell if110




f = 100 MHz, TEM cell if110




f = 300 MHz, TEM cell if110




f = 900 MHz, TEM cell if110






Isotropy Error (ϕ), $\theta = 0^\circ$





Frequency Response of E-Field


(TEM-Cell:ifi110, Waveguide R22)



Dynamic Range $f(\text{SAR}_{\text{brain}})$ (TEM-Cell:ifi110)



Conversion Factor Assessment



Head	900 MHz	$\epsilon_r = 42 \pm 5\%$	$\sigma = 0.97 \pm 10\%$ mho/m
	ConvF X	6.44 $\pm 7\%$ (k=2)	Boundary effect:
	ConvF Y	6.44 $\pm 7\%$ (k=2)	Alpha 0.58
	ConvF Z	6.44 $\pm 7\%$ (k=2)	Depth 1.98

Head	1800 MHz	$\epsilon_r = 40 \pm 5\%$	$\sigma = 1.40 \pm 10\%$ mho/m
	ConvF X	5.43 $\pm 7\%$ (k=2)	Boundary effect:
	ConvF Y	5.43 $\pm 7\%$ (k=2)	Alpha 0.60
	ConvF Z	5.43 $\pm 7\%$ (k=2)	Depth 2.06

Schmid & Partner Engineering AG


Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Additional Conversion Factors for Dosimetric E-Field Probe

Type:	ET3DV6
Serial Number:	1523
Place of Assessment:	Zurich
Date of Assessment:	May 16, 2001
Probe Calibration Date:	May 11, 2001

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:



Dosimetric E-Field Probe ET3DV6 SN:1523

Conversion factor (\pm standard deviation)

835 MHz	ConvF	$6.51 \pm 8\%$	$\epsilon_r = 44.0$ $\sigma = 0.90 \text{ mho/m}$ (brain tissue)
900 MHz	ConvF	$6.38 \pm 8\%$	$\epsilon_r = 42.5$ $\sigma = 0.86 \text{ mho/m}$ (brain tissue)
925 MHz	ConvF	$6.32 \pm 8\%$	$\epsilon_r = 44.0$ $\sigma = 0.93 \text{ mho/m}$ (brain tissue)
835 MHz	ConvF	$6.44 \pm 8\%$	$\epsilon_r = 52.0$ $\sigma = 1.10 \text{ mho/m}$ (muscle tissue)
925 MHz	ConvF	$6.28 \pm 8\%$	$\epsilon_r = 52.0$ $\sigma = 1.20 \text{ mho/m}$ (muscle tissue)

Dosimetric E-Field Probe ET3DV6 SN:1523

Conversion factor (\pm standard deviation)

1800 MHz ConvF $5.51 \pm 8\%$

$\epsilon_r = 41.5$
 $\sigma = 1.69 \text{ mho/m}$
(brain tissue, sugar-water)

1800 MHz ConvF $5.40 \pm 8\%$

$\epsilon_r = 40.3$
 $\sigma = 1.35 \text{ mho/m}$
(brain tissue, glycol)

1900 MHz ConvF $5.27 \pm 8\%$

$\epsilon_r = 39.9$
 $\sigma = 1.42 \text{ mho/m}$
(brain tissue, glycol)

1800 MHz ConvF $5.16 \pm 8\%$

$\epsilon_r = 50.0$
 $\sigma = 1.58 \text{ mho/m}$
(muscle tissue, glycol)

1900 MHz ConvF $5.01 \pm 8\%$

$\epsilon_r = 50.0$
 $\sigma = 1.64 \text{ mho/m}$
(muscle tissue, glycol)