

## System Check\_Head\_2450MHz

### DUT: D2450V2-929

Communication System: ; Frequency: 2450.0

Medium: HSL\_2450\_220520 Medium parameters used:  $f = 2450.0$  MHz;  $\sigma = 1.82$  S/m;  $\epsilon_r = 38.9$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

#### DASY6 Configuration:

- Probe: EX3DV4 - SN7695; ConvF(8.24, 8.24, 8.24); Calibrated: 2021-11-19
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1697; Calibrated: 2021-11-09
- Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 1919; Section: Flat
- Measurement Software: cDASY6 V6.6.0.13926
- UID: , 0--
- MAIA: Area Scan: N/A; Zoom Scan: N/A

**Area Scan (40.0 mm x 96.0 mm):** Measurement Grid: 10.0 mm x 12.0 mm

SAR (1g) = 2.53 W/kg; SAR (10g) = 1.20 W/kg;

**Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm):** Measurement Grid: 5.0 mm x 5.0 mm x 1.5 mm

Power Drift = 0.11 dB

SAR (1g) = 2.52 W/kg; SAR (10g) = 1.16 W/kg;

psAPD (1.0cm<sup>2</sup>, sq) = 0 [W/m<sup>2</sup>]; psAPD (4.0cm<sup>2</sup>, sq) = 0 [W/m<sup>2</sup>]

