

## Test Equipment List with Calibration Data / Test date: November 2007

No.	Type	Inv.-No.	Model	Serial Number	Manufacturer	Calibration Organization	Date of last calibration	Date of next calibration <sup>1</sup>
1.	Test receiver	1028	ESHS 10	860043/016	Rohde & Schwarz	Rohde & Schwarz	04/2007	04/2008
2.	Test receiver	1025	ESVP	881120/024	Rohde & Schwarz	Rohde & Schwarz	10/2006	10/2007
3.	Spectrum analyzer	1078	R 3271	05050023	Advantest	Rohde & Schwarz	03/2006	03/2008
4.	Spectrum analyzer	1666	FSP 30	100063	Rohde & Schwarz	Rohde & Schwarz	02/2006	08/2007
5.	EMI test receiver	1569	ESMI	839379/013 839587/006	Rohde & Schwarz	Rohde & Schwarz	12/2005	12/2007
6.	Preamplifier	1142	R14601	13120026	Advantest	Rohde & Schwarz	11/2006	11/2007
	Matrix RELAX consists:	1941						
7.	Preamplifier	1651	CPA9231A1	3393	Schaffner	Schaffner	02/2007	02/2008
8.	Preamplifier	1684	AFS3-00100800-32-LN	847743	Miteq	Senton GmbH	06/2006	06/2007
9.	V-network	1060	ESH 3-Z5	862770/021 B1060	Rohde & Schwarz	Senton GmbH	01/2007	01/2010
10.	V-network	1218	ESH 3-Z5	830952/025 B1218	Rohde & Schwarz	Senton GmbH	05/2005	05/2008
11.	Loop antenna	1016	HFH2-Z2	882964/0001	Rohde & Schwarz	Rohde & Schwarz	12/2007	12/2008
12.	Trilog wide band antenna Chamber 2	1722	VULB 9163	9163-188 A1722	Schwarzbeck	Schwarzbeck	10/2006	10/2007
13.	Trilog wide band antenna Chamber 3	1802	VULB 9163	9163-214 A1802	Schwarzbeck	Schwarzbeck	01/2006	07/2007
14.	Biconical antenna EG 1	1518	HK 116	842204/001	Rohde & Schwarz	Rohde & Schwarz	02/2007	02/2008
15.	Log. per. antenna EG 1	1519	HL 223	841516/023	Rohde & Schwarz	Rohde & Schwarz	01/2007	01/2008
16.	Horn Antenna	1516	3115	9508-4553	Emco	ARC-Seibersdorf	05/2005	05/2008

<sup>1</sup> Accepted tolerance is  $\pm 6$  months.