

Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Additional Conversion Factors for Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1377

Place of Assessment:

Zurich

Date of Assessment:

June 29, 2001

Probe Calibration Date:

June 29, 2001

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. The evaluation is coupled with measured conversion factors. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:

Oliver Kohler

Dosimetric E-Field Probe ET3DV6 SN:1377

Conversion factor ($\pm$ standard deviation)

2400 MHz ConvF $5.47 \pm 8\%$

$$\epsilon_r = 37.0 \pm 5\%$$

$$\sigma = 2.30 \pm 5\% \text{ mho/m}$$

(brain tissue)