

Test Laboratory: Compliance Certification Services  
File Name: D2450V2\_011303-1.da4

**DUT: D2450V2 Type & Serial Number: D2450V2 SN 706**  
**Program: System Validation; D2450V2 SN 706**

Communication System: Dipole; Frequency: 2450 MHz; Duty Cycle: 1:1  
Medium: Head 2450 MHz ( $\sigma = 1.9749$  mho/m,  $\epsilon = 39.5$ ,  $\rho = 1000$  kg/m<sup>3</sup>)  
Phantom section: FlatSection

DASY4 Configuration:

- Probe: ET3DV6 - SN1578; ConvF(4.5, 4.5, 4.5); Calibrated: 2/22/2002
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn500; Calibrated: 2/26/2002
- Phantom: - TP: SAM 1
- Software: DASY4, V4.0 Build 51

**Zoom Scan (7x7x7)/Cube 0:** Measurement grid: dx=5mm, dy=5mm

Reference Value = 85.3 V/m

Peak SAR = 38.4 mW/g

SAR(1 g) = 14.3 mW/g; SAR(10 g) = 6.02 mW/g

Power Drift = 0.02 dB

**Area Scan (8x9x1):** Measurement grid: dx=10mm, dy=10mm

