

## Calibration Certificate

### Dosimetric E-Field Probe

Type:

**ET3DV6**

Serial Number:

**1578**

Place of Calibration:

**Zurich**

Date of Calibration:

**February 22, 2002**


Calibration Interval:

**12 months**


Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:



Approved by:



# Probe ET3DV6

## SN:1578

Manufactured:	April 6, 2001
Last calibration:	April 20, 2001
Recalibrated:	February 22, 2002

Calibrated for System DASY3

## DASY3 - Parameters of Probe: ET3DV6 SN:1578

### Sensitivity in Free Space

NormX	<b>1.72</b> $\mu\text{V}/(\text{V}/\text{m})^2$
NormY	<b>1.83</b> $\mu\text{V}/(\text{V}/\text{m})^2$
NormZ	<b>1.68</b> $\mu\text{V}/(\text{V}/\text{m})^2$

### Diode Compression

DCP X	<b>99</b>	mV
DCP Y	<b>99</b>	mV
DCP Z	<b>99</b>	mV

### Sensitivity in Tissue Simulating Liquid

<b>Head</b>	<b>900 MHz</b>	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.97 \pm 5\%$ mho/m
<b>Head</b>	<b>835 MHz</b>	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.90 \pm 5\%$ mho/m
ConvF X	<b>7.0</b> $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	<b>7.0</b> $\pm 9.5\%$ (k=2)	Alpha	<b>0.33</b>
ConvF Z	<b>7.0</b> $\pm 9.5\%$ (k=2)	Depth	<b>2.34</b>
<b>Head</b>	<b>1800 MHz</b>	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
<b>Head</b>	<b>1900 MHz</b>	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
ConvF X	<b>5.5</b> $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	<b>5.5</b> $\pm 9.5\%$ (k=2)	Alpha	<b>0.47</b>
ConvF Z	<b>5.5</b> $\pm 9.5\%$ (k=2)	Depth	<b>2.20</b>

### Boundary Effect


<b>Head</b>	<b>900 MHz</b>	<b>Typical SAR gradient: 5 % per mm</b>	
Probe Tip to Boundary		<b>1 mm</b>	<b>2 mm</b>
SAR <sub>be</sub> [%] Without Correction Algorithm		7.6	4.3
SAR <sub>be</sub> [%] With Correction Algorithm		0.2	0.4
<b>Head</b>	<b>1800 MHz</b>	<b>Typical SAR gradient: 10 % per mm</b>	
Probe Tip to Boundary		<b>1 mm</b>	<b>2 mm</b>
SAR <sub>be</sub> [%] Without Correction Algorithm		9.5	6.3
SAR <sub>be</sub> [%] With Correction Algorithm		0.2	0.3

### Sensor Offset


Probe Tip to Sensor Center	<b>2.7</b>	mm
Optical Surface Detection	<b>1.5 <math>\pm</math> 0.2</b>	mm

## Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$


**f = 30 MHz, TEM cell ifi110**




**f = 100 MHz, TEM cell ifi110**





**f = 300 MHz, TEM cell ifi110**




**f = 900 MHz, TEM cell ifi110**






### Isotropy Error ( $\phi$ ), $\theta = 0^\circ$





# Frequency Response of E-Field



( TEM-Cell:ifi110, Waveguide R22)



## Dynamic Range f(SAR<sub>brain</sub>) ( Waveguide R22 )




## Conversion Factor Assessment



<b>Head</b>	<b>900 MHz</b>	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.97 \pm 5\%$ mho/m
<b>Head</b>	<b>835 MHz</b>	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.90 \pm 5\%$ mho/m
	ConvF X	<b>7.0</b> $\pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	<b>7.0</b> $\pm 9.5\%$ (k=2)	Alpha <b>0.33</b>
	ConvF Z	<b>7.0</b> $\pm 9.5\%$ (k=2)	Depth <b>2.34</b>
<b>Head</b>	<b>1800 MHz</b>	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
<b>Head</b>	<b>1900 MHz</b>	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
	ConvF X	<b>5.5</b> $\pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	<b>5.5</b> $\pm 9.5\%$ (k=2)	Alpha <b>0.47</b>
	ConvF Z	<b>5.5</b> $\pm 9.5\%$ (k=2)	Depth <b>2.20</b>




## Conversion Factor Assessment



<b>Body</b>	<b>900 MHz</b>	$\epsilon_r = 55.0 \pm 5\%$	$\sigma = 1.05 \pm 5\% \text{ mho/m}$
<b>Body</b>	<b>835 MHz</b>	$\epsilon_r = 55.2 \pm 5\%$	$\sigma = 0.97 \pm 5\% \text{ mho/m}$
	ConvF X	<b>6.7</b> $\pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	<b>6.7</b> $\pm 9.5\%$ (k=2)	Alpha <b>0.29</b>
	ConvF Z	<b>6.7</b> $\pm 9.5\%$ (k=2)	Depth <b>2.76</b>
<b>Body</b>	<b>1800 MHz</b>	$\epsilon_r = 53.3 \pm 5\%$	$\sigma = 1.52 \pm 5\% \text{ mho/m}$
<b>Body</b>	<b>1900 MHz</b>	$\epsilon_r = 53.3 \pm 5\%$	$\sigma = 1.52 \pm 5\% \text{ mho/m}$
	ConvF X	<b>5.1</b> $\pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	<b>5.1</b> $\pm 9.5\%$ (k=2)	Alpha <b>0.58</b>
	ConvF Z	<b>5.1</b> $\pm 9.5\%$ (k=2)	Depth <b>2.19</b>

# Deviation from Isotropy in HSL

Error ( $\theta, \phi$ ),  $f = 900$  MHz



## Additional Conversion Factors for Dosimetric E-Field Probe

Type:

**ET3DV6**

Serial Number:

**1578**

Place of Assessment:

**Zurich**

Date of Assessment:

**February 25, 2002**

Probe Calibration Date:

**February 22, 2002**

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:

*Johnie Katz*

# Dosimetric E-Field Probe ET3DV6 SN:1578

Conversion factor ( $\pm$  standard deviation)

**450 MHz**                      ConvF                      **8.0 $\pm$  8%**

$\epsilon_r = 43.5$ $\sigma = 0.87$ mho/m (head tissue)
---

**450 MHz**                      ConvF                      **8.1  $\pm$  8%**

$\epsilon_r = 56.7$ $\sigma = 0.94$ mho/m (body tissue)
---

**2450 MHz**                      ConvF                      **4.5  $\pm$  8%**

$\epsilon_r = 39.2$ $\sigma = 1.80$ mho/m (head tissue)
---

**2450 MHz**                      ConvF                      **4.1  $\pm$  8%**

$\epsilon_r = 52.7$ $\sigma = 1.95$ mho/m (body tissue)
---

**Additional Conversion Factors**  
for Dosimetric E-Field Probe

Type:

**ET3DV6**

Serial Number:

**1578**

Place of Assessment:

**Zurich**

Date of Assessment:


**August 27, 2002**

Probe Calibration Date:

**February 22, 2002**

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:



# Dosimetric E-Field Probe ET3DV6 SN:1578

Conversion factor ( $\pm$  standard deviation)

650 MHz

ConvF

$7.3 \pm 8\%$

$\epsilon_r = 55.92 \pm 5\%$

$\sigma = 0.96 \pm 5\%$  mho/m

(body tissue)