

The device was tested using a 2 cm separation with the front facing the phantom because the user's fatty part of his arm may be in front of the antenna during transmission. It is also for consideration of bystanders. This 2 cm separation note was to account for these situations.

Thanks  
Dennis