

Date

31. July 2019



FCC ID: KWC-BSR

Pursuant to paragraph §2.925 of 47 C.F.R. the hearing aid family Phonak Bolero M-PR, subject of this filing, shall bear a permanently affixed, readily visible label listing the information as specified in §2.926 and §15.19(a) of 47 C.F.R.

Due to the risk of skin irritation related to the specific use of hearing aids in direct contact with the skin, it is not appropriate to permanently affix a label on the outer surface of the device. The only possible location for the FCC ID labeling is the serial plate located in the inside of a small compartment. However, this location is so small that it would require the FCC ID font size to be no bigger than 2 points, which would not satisfy the requirements for "readily visible" label as defined in §2.925(d)(2) and §2.925(g).

Therefore, the FCC ID and related statement specified in §15.19(a)(3) are included in the user manual and affixed on the device packaging.

Sonova AG

Phone +41 58 928 01 01

Laubisrütistrasse 28

Fax +41 58 928 20 11

8712 Stäfa

info@sonova.com