

Date 10. July 2019

To whom this may concern,

Phonak Audéo M90-RT	Base device
Phonak Audéo M70-RT	Difference to base device: SW parameters (actual SW code is identical)
Phonak Audéo M50-RT	Difference to base device: SW parameters (actual SW code is identical)
Phonak Audéo M30-RT	Difference to base device: SW parameters (actual SW code is identical)
Phonak Audéo M-RT Trial	Difference to base device: SW parameters (actual SW code is identical)

The software code for the controller/DSP is the same, compiled once for all devices. The resulting code is combined with product specific parameters (mostly flags to determine which features are enabled and which are disabled for pricing levels, M-90 RT, M-70 RT, M50-RT or M30-RT) and packaged into the "image". Because the parameters are different for each device, one such image is generated for each device, and they all have a different ID (067-1313, 067-1315 etc).

The hardware of all M-RT models is identical, (that's why the technical construction files are valid for *all* the models).

#### Regarding RF performance

Phonak Audéo M90-R	Phonak Audéo M90-RT	These devices share the same chipset, same antenna, same battery, same microphone module which is underneath the antenna.
-----------------------	------------------------	---

Based on the equivalent hardware between the Phonak Audeo M90-R and Phonak Audéo M90-RT models there is no change between the RF performance Phonak Audeo M90-R and Phonak Audéo M90-RT.

Faithfully,

Stäfa, 22 July 2019



Felix Aebersold  
Director System Integration



Glenn Borrett  
Senior Regulatory Affairs Manager

Sonova AG  
Laubisrütistrasse 28  
8712 Stäfa  
Switzerland

Phone +41 58 928 01 01  
Fax +41 58 928 20 11  
info@sonova.com  
www.sonova.com

SONOVA AG  
Laubisrütistrasse 28  
CH-8712 Stäfa  
Switzerland