

To whom it may concern

Stäfa (Switzerland), January 2022/ G.Borrett, Senior Manager Regulatory Affairs

## Product Equality Declaration

We, Sonova AG, hereby declare under our own responsibility that the products listed below as “Hardware Equivalent Products” are in all relevant parts and hardware construction identical to the corresponding product identified as “Products with basis Hardware”. The following standards and/or

Standards
Hearing Aid standards: ACOUSTIC: IEC 60118-0,-1,-2,-6 and NSH 7.0 (including Annex A); EMC: IEC 60118-13; EMC immunity: ANSI C63.19
Degrees of protection provided by enclosures (IP Code): IEC 60529:1989-11+A1:1999; EN 60529:1991-10 (incl. Corrigendum: 1993-05)+A1:2000-02
Europe: HEALTH & SAFETY: IEC/EN 60601-1, IEC/EN 60601-2-66; IEC/EN 62133; IEC/EN 62479; EMC: IEC/EN 60601-1-2; EN 301489-1,-3,-17; SPECTRUM: EN 300328; EN 300422-4; REACH; RoHS-III.
USA: 47 CFR Part 15 (B), Part 15 (C): 15.249, Part 2: 2.1091, 2.1093.
Canada: RSS-Gen, ICES-003, RSS-210, RSS-102; RSS-247.
Japan: ARIB T66, Ordinance regulating Radio Equipment (2005-08) Ar12 item 19

technical regulations and corresponding test reports fully apply accordingly:

The only difference between the listed equivalent and corresponding basis models is the model name, a different set of audiological features per performance level and new corresponding fitting software due to marketing purposes.

Products with basis Hardware (tested representative model)	Hardware Equivalent Products
<b>Phonak Audéo P90-R Fit (with Sensor M/P)</b> HW version : 050-0870-xx SW version: Target 7 FW version : 067-1466	<b>Phonak Audéo P70-R Fit (with Sensor M/P)</b> HW version : 050-0869-xx SW version: Target 7 FW version : 067-1466
	<b>Phonak Audéo P-R Fit Trial (with Sensor M/P)</b> HW version : 050-0873-P6 SW version: Target 7 FW version : 067-1466
	<b>Phonak Brio 5 Move (with Sensor M/P)</b> HW version : 050-0870-xx56 SW version: Target 8 FW version : 067-1466
	<b>Phonak Brio 5 Move Demo (with Sensor M/P)</b> HW version : 050-0870-P657 SW version: Target 8 FW version : 067-1466

where xx – means color identification

Place and date

Stäfa, 28 January 2022



B.Feldman  
Senior Director Global Quality Management

Stäfa, 28 January 2022



G.Borrett  
Senior Regulatory Affairs Manager

<b>Type:</b> PDL	<b>Status:</b> CURRENT
<b>Document No.:</b> PDL-8685	<b>Effective Date:</b> 28-Jan-2022
<b>Rev.:</b> 2	
<b>Title:</b> Product Equality Declaration Audéo Fit	
<b>Process:</b> Regulatory Affairs	
<b>Owner:</b> 11GBORRETT Glenn Borrett	

**Attributes**

<u>Attribute Type</u>	<u>Value</u>	<u>Description</u>
Affected Site	1100	Stafa (CH)
Project	Mozart Floyer Product	
Project Phase	Design Process	

**Reference**

<u>Document No.</u>	<u>Title</u>	<u>Content Type</u>	<u>Relation</u>
PDL-10004 [1]	Audeo P Fit Product Equality Declaration incl Private label	DOCUMENT	Related

**Approvals**

<u>Level</u>	<u>Actor</u>	<u>Job Title</u>	<u>Sign-off Date</u>	<u>Sign-off By</u>
1	Nataya Jaiman	TQM Regulatory Affairs Assistant	28-Jan-2022	11NJAIMAN
2	Glenn Borrett	Director Regulatory Affairs	28-Jan-2022	11GBORRETT

**Revision Notes**

<u>Access Activity</u>	<u>Note</u>	<u>Accessed By</u>	<u>Accessed Date</u>
Remark	Phonak Brio 5 Move Demo added	11NJAIMAN	28-Jan-2022