

DEKRA Testing and Certification, S.A.U.
Parque Tecnológico de Andalucía
C/ Severo Ochoa 2 & 6
29590 Campanillas
Málaga, España

Date: 14.02.2022

To Whom It May Concern,

Ref.: Label Location for FCC ID: KWC-VTP

Pursuant to paragraph §2.925 of 47 C.F.R. the FCC ID: KWC-VTP, which is subject of this filing, shall bear a permanently affixed, readily visible label listing the information as specified in §2.926 and §15.19(a) of the 47 C.F.R. Due to the specific use of the hearing instrument, its external surface is in contact with the user's skin behind the ear.

Therefore, it is not appropriate to permanently affix a label on the outer surface of the device to prevent skin irritations. In this case, the size of the FCC ID will be smaller than 2 points, which does not satisfy the requirements for "readily visible" label as defined in §2.925(d)(2) and §2.925(g).

Due to the small size of the devices and pursuant to paragraph §15.19(a)(5) of 47 C.F.R., the FCC ID and the statement specified in §15.19(a)(3) are placed in the user manual and affixed on the device packaging as shown below.



James Jasinski
Sonova USA, Inc.
444 Commerce St,
Aurora, Illinois, 60504, United States