

To whom it may concern,

Phonak AG  
Laubisrütistrasse 28  
CH-8712 Stäfa  
Switzerland  
Phone: +41 58 928 01 01  
Fax: +41 58 928 20 11

*VaS*  
Stäfa (Switzerland), 12 November 2014 / V.Shcherba, Manager Regulatory Affairs

## Product Equality Declaration

We, Phonak AG, hereby declare under our own responsibility that the products listed below as "Hardware Equivalent Products" have an identical Radio Frequency (RF) part and are in all relevant parts and hardware construction identical to the corresponding Phonak "Products with basis Hardware".

The following standards and/or technical regulations and corresponding test reports fully apply accordingly:

### Standards

Hearing Aid product standards: IEC 60118-13 (EMC), IEC 60118-0-1-2-6 and NSH 7.0 (Acoustic)  
HEALTH & SAFETY: IEC/EN 60601-1; IEC/EN 60950-1; EMC: IEC/EN 60601-1-2; EN 301489-1,-3;  
DEGREES of PROTECTION: IEC 60529; SPECTRUM: EN 300330-1,-2;  
USA: 47CFR FCC Part 15; Canada: RSS-GEN Issue 3; Japan: (EWPE) Extremely Weak Power Equipment

### Hardware Equivalent Products

Phonak Bolero V70-M  
Phonak Bolero V50-M  
Phonak Bolero V30-M  
Phonak Bolero V70-P  
Phonak Bolero V50-P  
Phonak Bolero V30-P  
Phonak Bolero V70-SP  
Phonak Bolero V50-SP  
Phonak Bolero V30-SP

### Products with basis Hardware

Phonak Bolero V90-M  
  
Phonak Bolero V90-P  
  
Phonak Bolero V90-SP

Place and Date:

Stäfa,

*12. Nov 2014*  
*A. Vonlanthen*  
A. Vonlanthen  
Group Vice President  
Research & Development

Stäfa,

*12. Nov. 2014*  
*H. Mehl*  
H. Mehl  
Group Vice President  
Operations