	<p align="center">Technical Note</p> <p align="center">Phonak Roger Earpiece Label Location</p>	<p>Vers.: V2 Page: 1 of 3 Date: 17.03.16</p>
---	---	---

ID Label / Label Location Phonak Roger Earpiece


Pursuant to Clause 2.1 of RSS-210, Clause 2.1 of RSS-Gen and Clause 3.1 of RSP-100, the product Roger Earpiece, which is subject of this filing, shall be identified with a HVIN and ISED certification number permanently indicated on the exterior of the product.

Due to the specific use of this product, its external surface is in a permanent contact with the skin within the ear canal (Figure 1). Therefore, it is not appropriate to permanently affix a label on the outer surface of the device in order to prevent any possible skin irritations. The only possible place where the PMN, HVIN and ISED certification number of the product can be put is the battery compartment. However, in this case the size of the certification number will be smaller than 2 points, which does not satisfy the requirement that "In all cases, the PMN, FVIN, HVIN and ISED certification number text shall be clearly legible." as defined in Clause 3.1 of RSP-100.

Therefore, due to the very small size of this device and pursuant to the same Clause 3.1 of RSP-100, the PMN, HVIN and ISED certification number and the statement specified in Clause 8.4 of RSS-Gen are placed in the user manual as shown on Figure 2 below:



Figure 1. Roger Earpiece

	<p align="center">Technical Note</p> <p align="center">Phonak Roger Earpiece Label Location</p>	<p>Vers.: V2 Page: 2 of 3 Date: 17.03.16</p>
---	---	---

2 Declaration of Conformity

In body worn operation, the devices have been found in compliance with the requirements of EN 62209-2, FCC rule paragraph 2.1093, the ANSI/IEEE C95.1:1999, the NCRP Report Number 86 for uncontrolled environment, according to the Health Canada's Safety Code 6 and the Industry Canada Radio Standards Specification RSS-102 for General Population/Uncontrolled exposure.

Hereby, Phonak Communications AG, Langgasse 17, CH-3280 Murten, declares that these products are in compliance with the essential requirements and other relevant provisions of the R&TTE Directive (1999/5/EC). To obtain a copy of the Declaration of conformity, please contact the manufacturer or the local Phonak Communications AG representative whose address can be taken from the list on www.phonak-communications.com/ (worldwide locations).

This device complies with Part 15 of the FCC Rules and with RSS-210 (Roger Earpiece) or RSS-247 (Roger BaseStation) of Industry Canada. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation. The device has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. The device generates uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If the device does cause harmful interference to radio or television reception, which can be determined by turning the device off and on, the user is encouraged to try to correct the interference by one of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the device and receiver.
- Connect the device into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

FCC Caution: Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the device. To comply with FCC RF exposure limits for general population/uncontrolled exposure, the antenna(s) used for Roger BaseStation must be installed to provide a separation distance of at least 20 cm from all persons and must not be collocated or operating in conjunction with any other antenna or transmitter.

This Class B digital apparatus complies with Canadian ICES-003. To comply with Industry Canada RF exposure limits for general population / uncontrolled exposure, the antenna(s) used for Roger BaseStation must be installed to provide a separation distance of at least 20 cm from all persons and must not be collocated or operating in conjunction with any other antenna or transmitter.

