 RESEARCH IN MOTION	Document Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		Page 1(27)
	Author Data <b>Daoud Attayi</b>	Dates of Test <b>June 29 – July 07, 2004</b>	Test Report No <b>RIM-0094-0407-03</b>

APPENDIX D: PROBE & DIPOLE CALIBRATION DATA



Document

Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report

Page

2(27)

Author Data

Daoud Attayi

Dates of Test

June 29 – July 07, 2004

Test Report No

RIM-0094-0407-03

FCC ID:


L6AQAP40GW

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Client

RIM

CALIBRATION CERTIFICATE			
Object(s)	ET3DV6 - SN:1643		
Calibration procedure(s)	QA CAL-01 v2 Calibration procedure for dosimetric E-field probes		
Calibration date:	October 9, 2003		
Condition of the calibrated item	In Tolerance (according to the specific calibration document)		
This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.			
All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.			
Calibration Equipment used (M&TE critical for calibration)			
Model Type	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power meter EPM E4419B	GB41293874	2-Apr-03 (METAS, No 252-0250)	Apr-04
Power sensor E4412A	MY41495277	2-Apr-03 (METAS, No 252-0250)	Apr-04
Reference 20 dB Attenuator	SN: 5086 (20b)	3-Apr-03 (METAS No. 251-0340)	Apr-04
Fluke Process Calibrator Type 702	SN: 6295803	8-Sep-03 (Sintrel SCS No. E-030020)	Sep-04
Power sensor HP 8481A	MY41092180	18-Sep-02 (Agilent, No. 20020918)	In house check: Oct 03
RF generator HP 8684C	US3642U01700	4-Aug-99 (SPEAG, in house check Aug-02)	In house check: Aug-05
Network Analyzer HP 8753E	US37390585	18-Oct-01 (Agilent, No. 24BR1033101)	In house check: Oct 03
Calibrated by:	Name Nico Vetterli	Function Technician	Signature 
Approved by:	Name Katja Pokovic	Function Laboratory Director	Signature 
Date issued: October 9, 2003			
This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.			

 <b>RESEARCH IN MOTION</b>	Document Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		Page 3(27)
	Author Data <b>Daoud Attayi</b>	Dates of Test <b>June 29 – July 07, 2004</b>	Test Report No <b>RIM-0094-0407-03</b>

Schmid & Partner Engineering AG

**s p e a g**

Zeughausstrasse 43, 8004 Zurich, Switzerland  
Phone +41 1 245 9700, Fax +41 1 245 9779  
info@speag.com, http://www.speag.com

# Probe ET3DV6

## SN:1643

**Manufactured:** November 7, 2001  
**Last calibration:** September 24, 2002  
**Recalibrated:** October 9, 2003

**Calibrated for DASY Systems**

(Note: non-compatible with DASY2 system!)



Author Data <b>Daoud Attayi</b>	Dates of Test <b>June 29 – July 07, 2004</b>	Test Report No <b>RIM-0094-0407-03</b>	FCC ID: <b>L6AQAP40GW</b>
------------------------------------	---	---	------------------------------

ET3DV6 SN:1643

October 9, 2003

**DASY - Parameters of Probe: ET3DV6 SN:1643**

Sensitivity in Free Space

Diode Compression

NormX	<b>1.73</b> $\mu\text{V}/(\text{V}/\text{m})^2$	DCP X	<b>96</b>	mV
NormY	<b>1.88</b> $\mu\text{V}/(\text{V}/\text{m})^2$	DCP Y	<b>96</b>	mV
NormZ	<b>1.81</b> $\mu\text{V}/(\text{V}/\text{m})^2$	DCP Z	<b>96</b>	mV

Sensitivity in Tissue Simulating Liquid

Head 900 MHz  $\epsilon_r = 41.5 \pm 5\%$   $\sigma = 0.97 \pm 5\%$  mho/m

Valid for f=800-1000 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

ConvF X	<b>6.5</b> $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	<b>6.5</b> $\pm 9.5\%$ (k=2)	Alpha	<b>0.37</b>
ConvF Z	<b>6.5</b> $\pm 9.5\%$ (k=2)	Depth	<b>2.72</b>

Head 1800 MHz  $\epsilon_r = 40.0 \pm 5\%$   $\sigma = 1.40 \pm 5\%$  mho/m

Valid for f=1710-1910 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

ConvF X	<b>5.2</b> $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	<b>5.2</b> $\pm 9.5\%$ (k=2)	Alpha	<b>0.47</b>
ConvF Z	<b>5.2</b> $\pm 9.5\%$ (k=2)	Depth	<b>2.87</b>

Boundary Effect

Head 900 MHz Typical SAR gradient: 5 % per mm

Probe Tip to Boundary		1 mm	2 mm
SAR <sub>be</sub> [%]	Without Correction Algorithm	10.8	6.3
SAR <sub>be</sub> [%]	With Correction Algorithm	0.4	0.6

Head 1800 MHz Typical SAR gradient: 10 % per mm

Probe Tip to Boundary		1 mm	2 mm
SAR <sub>be</sub> [%]	Without Correction Algorithm	14.5	10.1
SAR <sub>be</sub> [%]	With Correction Algorithm	0.2	0.1





Sensor Offset

Probe Tip to Sensor Center	<b>2.7</b>	mm
Optical Surface Detection	<b>1.4 <math>\pm</math> 0.2</b>	mm

ET3DV6 SN:1643



October 9, 2003

**Receiving Pattern ( $\phi$ ),  $\theta = 0^\circ$**




ET3DV6 SN:1643

October 9, 2003



**Isotropy Error ( $\phi$ ),  $\theta = 0^\circ$**




ET3DV6 SN:1643

October 9, 2003

## Frequency Response of E-Field

( TEM-Cell:if110, Waveguide R22)







ET3DV6 SN:1643

October 9, 2003

### Dynamic Range f(SARhead) ( Waveguide R22 )






ET3DV6 SN:1643

October 9, 2003

### Conversion Factor Assessment



Head                      900 MHz                       $\epsilon_r = 41.5 \pm 5\%$                        $\sigma = 0.97 \pm 5\%$  mho/m

Valid for f=800-1000 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

ConvF X	6.5 ± 9.5% (k=2)	Boundary effect:	
ConvF Y	6.5 ± 9.5% (k=2)	Alpha	<b>0.37</b>
ConvF Z	6.5 ± 9.5% (k=2)	Depth	<b>2.72</b>

Head                      1800 MHz                       $\epsilon_r = 40.0 \pm 5\%$                        $\sigma = 1.40 \pm 5\%$  mho/m


Valid for f=1710-1910 MHz with Head Tissue Simulating Liquid according to EN 50361, P1528-200X

ConvF X	5.2 ± 9.5% (k=2)	Boundary effect:	
ConvF Y	5.2 ± 9.5% (k=2)	Alpha	<b>0.47</b>
ConvF Z	5.2 ± 9.5% (k=2)	Depth	<b>2.87</b>

ET3DV6 SN:1643

October 9, 2003

### Conversion Factor Assessment



Body                      900 MHz                       $\epsilon_r = 55.0 \pm 5\%$                        $\sigma = 1.05 \pm 5\%$  mho/m

Valid for f=800-1000 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C

ConvF X	6.3 ± 9.5% (k=2)	Boundary effect:	
ConvF Y	6.3 ± 9.5% (k=2)	Alpha	0.43
ConvF Z	6.3 ± 9.5% (k=2)	Depth	2.49

Body                      1800 MHz                       $\epsilon_r = 53.3 \pm 5\%$                        $\sigma = 1.52 \pm 5\%$  mho/m

Valid for f=1710-1910 MHz with Body Tissue Simulating Liquid according to OET 65 Suppl. C


ConvF X	4.8 ± 9.5% (k=2)	Boundary effect:	
ConvF Y	4.8 ± 9.5% (k=2)	Alpha	0.57
ConvF Z	4.8 ± 9.5% (k=2)	Depth	2.74


ET3DV6 SN:1643

October 9, 2003

### Deviation from Isotropy in HSL



Error ( $\theta, \phi$ ),  $f = 900$  MHz




	Document		Page
	Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		12(27)
Author Data	Dates of Test	Test Report No	FCC ID:
<b>Daoud Attayi</b>	<b>June 29 – July 07, 2004</b>	<b>RIM-0094-0407-03</b>	<b>L6AQAP40GW</b>

**Calibration Laboratory of  
 Schmid & Partner  
 Engineering AG**  
 Zeughausstrasse 43, 8004 Zurich, Switzerland

Client **RIM**

CALIBRATION CERTIFICATE																											
Object(s)	D835V2 - SN 446																										
Calibration procedure(s)	QA CAL-05 v2 Calibration procedure for dipole validation kits																										
Calibration date:	August 21, 2003																										
Condition of the calibrated item	In Tolerance (according to the specific calibration document)																										
<p>This calibration statement documents traceability of M&amp;TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.</p> <p>All calibrations have been conducted in the closed laboratory facility; environment temperature 22 +/- 2 degrees Celsius and humidity &lt; 75%.</p> <p>Calibration Equipment used (M&amp;TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Model Type</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>RF generator R&amp;S SML-03</td> <td>100698</td> <td>27-Mar-2002 (R&amp;S, No. 20-92389)</td> <td>In house check: Mar-05</td> </tr> <tr> <td>Power sensor HP 8481A</td> <td>MY41092317</td> <td>18-Oct-02 (Agilent, No. 20021018)</td> <td>Oct-04</td> </tr> <tr> <td>Power sensor HP 8481A</td> <td>US37292783</td> <td>30-Oct-02 (METAS, No. 252-0236)</td> <td>Oct-03</td> </tr> <tr> <td>Power meter EPM E442</td> <td>GB37480704</td> <td>30-Oct-02 (METAS, No. 252-0236)</td> <td>Oct-03</td> </tr> <tr> <td>Network Analyzer HP 8753E</td> <td>US37390585</td> <td>18-Oct-01 (Agilent, No. 24BR1033101)</td> <td>In house check: Oct 03</td> </tr> </tbody> </table>				Model Type	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	RF generator R&S SML-03	100698	27-Mar-2002 (R&S, No. 20-92389)	In house check: Mar-05	Power sensor HP 8481A	MY41092317	18-Oct-02 (Agilent, No. 20021018)	Oct-04	Power sensor HP 8481A	US37292783	30-Oct-02 (METAS, No. 252-0236)	Oct-03	Power meter EPM E442	GB37480704	30-Oct-02 (METAS, No. 252-0236)	Oct-03	Network Analyzer HP 8753E	US37390585	18-Oct-01 (Agilent, No. 24BR1033101)	In house check: Oct 03
Model Type	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																								
RF generator R&S SML-03	100698	27-Mar-2002 (R&S, No. 20-92389)	In house check: Mar-05																								
Power sensor HP 8481A	MY41092317	18-Oct-02 (Agilent, No. 20021018)	Oct-04																								
Power sensor HP 8481A	US37292783	30-Oct-02 (METAS, No. 252-0236)	Oct-03																								
Power meter EPM E442	GB37480704	30-Oct-02 (METAS, No. 252-0236)	Oct-03																								
Network Analyzer HP 8753E	US37390585	18-Oct-01 (Agilent, No. 24BR1033101)	In house check: Oct 03																								
Calibrated by:	Name Judith Muefler	Function Technician	Signature 																								
Approved by:	Name Katja Pokovic	Function Laboratory Director	Signature 																								
Date issued: August 22, 2003																											
<p>This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid &amp; Partner Engineering AG is completed</p>																											

	Document		Page
	Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		13(27)
Author Data	Dates of Test	Test Report No	FCC ID:
<b>Daoud Attayi</b>	<b>June 29 – July 07, 2004</b>	<b>RIM-0094-0407-03</b>	<b>L6AQAP40GW</b>

Schmid & Partner Engineering AG

**s p e e g**

Zeughausstrasse 43, 8004 Zurich, Switzerland  
Phone +41 1 245 9700, Fax +41 1 245 9779  
info@speag.com, <http://www.speag.com>

# DASY


## Dipole Validation Kit

Type: D835V2

Serial: 446

Manufactured: October 24, 2001

Calibrated: August 21, 2003

	Document		Page
	Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		14(27)
Author Data	Dates of Test	Test Report No	FCC ID:
Daoud Attayi	June 29 – July 07, 2004	RIM-0094-0407-03	L6AQAP40GW

## 1. Measurement Conditions

The measurements were performed in the flat section of the SAM twin phantom filled with head simulating solution of the following electrical parameters at 835 MHz:

Relative Dielectricity	<b>43.3</b>	$\pm 5\%$
Conductivity	<b>0.91 mho/m</b>	$\pm 5\%$

The DASY4 System with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.7 at 835 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance spacer was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 7x7x7 fine cube was chosen for cube integration.

The dipole input power (forward power) was 250 mW  $\pm 3\%$ . The results are normalized to 1W input power.

## 2. SAR Measurement with DASY4 System


Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

averaged over 1 cm <sup>3</sup> (1 g) of tissue:	<b>9.60 mW/g <math>\pm 16.8\%</math> (k=2)<sup>1</sup></b>
averaged over 10 cm <sup>3</sup> (10 g) of tissue:	<b>6.24 mW/g <math>\pm 16.2\%</math> (k=2)<sup>1</sup></b>

<sup>1</sup> validation uncertainty





 RESEARCH IN MOTION	Document		Page
	Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		16(27)
Author Data	Dates of Test	Test Report No	FCC ID:
Daoud Attayi	June 29 – July 07, 2004	RIM-0094-0407-03	L6AQAP40GW

Page 1 of 1  
Date/Time: 08/21/03 10:03:51

Test Laboratory: SPEAG, Zurich, Switzerland  
File Name: SN446\_SN1507\_HSL835\_210803.da4

**DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN446**  
**Program: Dipole Calibration**


Communication System: CW-835; Frequency: 835 MHz; Duty Cycle: 1:1  
Medium: HSL 835 MHz ( $\sigma = 0.91$  mho/m,  $\epsilon_r = 43.28$ ,  $\rho = 1000$  kg/m<sup>3</sup>)  
Phantom section: Flat Section  
Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(6.7, 6.7, 6.7); Calibrated: 1/18/2003
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 - SN411; Calibrated: 1/16/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006
- Measurement SW: DASY4, V4.1 Build 47; Postprocessing SW: SEMCAD, V1.6 Build 115

**Pin = 250 mW; d = 15 mm/Area Scan (81x81x1); Measurement grid: dx=15mm, dy=15mm**  
Reference Value = 55.3 V/m  
Power Drift = -0.02 dB  
Maximum value of SAR = 2.55 mW/g

**Pin = 250 mW; d = 15 mm/Zoom Scan (7x7x7)/Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm**  
Peak SAR (extrapolated) = 3.52 W/kg  
SAR(1 g) = 2.4 mW/g; SAR(10 g) = 1.56 mW/g  
Reference Value = 55.3 V/m  
Power Drift = -0.02 dB  
Maximum value of SAR = 2.56 mW/g






Author Data  
**Daoud Attayi**


Dates of Test  
**June 29 – July 07, 2004**

Test Report No  
**RIM-0094-0407-03**

FCC ID:  
**L6AQAP40GW**



446




	Document		Page
	Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		18(27)
Author Data	Dates of Test	Test Report No	FCC ID:
<b>Daoud Attayi</b>	<b>June 29 – July 07, 2004</b>	<b>RIM-0094-0407-03</b>	<b>L6AQAP40GW</b>

**Calibration Laboratory of  
 Schmid & Partner  
 Engineering AG**  
 Zeughausstrasse 43, 8004 Zurich, Switzerland

Client **RIM**

CALIBRATION CERTIFICATE																											
Object(s)	D1900V2 - SN:545																										
Calibration procedure(s)	QA CAL-05.v2 Calibration procedure for dipole validation kits																										
Calibration date:	August 22, 2003																										
Condition of the calibrated item	In Tolerance (according to the specific calibration document)																										
<p>This calibration statement documents traceability of M&amp;TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity &lt; 75%.</p> <p>Calibration Equipment used (M&amp;TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Model Type</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>RF generator R&amp;S SML-03</td> <td>100698</td> <td>27-Mar-2002 (R&amp;S, No. 20-92389)</td> <td>In house check: Mar-05</td> </tr> <tr> <td>Power sensor HP 8481A</td> <td>MY41092317</td> <td>18-Oct-02 (Agilent, No. 20021018)</td> <td>Oct-04</td> </tr> <tr> <td>Power sensor HP 8481A</td> <td>US37292783</td> <td>30-Oct-02 (METAS, No. 252-0236)</td> <td>Oct-03</td> </tr> <tr> <td>Power meter EPM E442</td> <td>GB37480704</td> <td>30-Oct-02 (METAS, No. 252-0236)</td> <td>Oct-03</td> </tr> <tr> <td>Network Analyzer HP 8753E</td> <td>US37390585</td> <td>18-Oct-01 (Agilent, No. 24BR1033101)</td> <td>In house check: Oct 03</td> </tr> </tbody> </table>				Model Type	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	RF generator R&S SML-03	100698	27-Mar-2002 (R&S, No. 20-92389)	In house check: Mar-05	Power sensor HP 8481A	MY41092317	18-Oct-02 (Agilent, No. 20021018)	Oct-04	Power sensor HP 8481A	US37292783	30-Oct-02 (METAS, No. 252-0236)	Oct-03	Power meter EPM E442	GB37480704	30-Oct-02 (METAS, No. 252-0236)	Oct-03	Network Analyzer HP 8753E	US37390585	18-Oct-01 (Agilent, No. 24BR1033101)	In house check: Oct 03
Model Type	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																								
RF generator R&S SML-03	100698	27-Mar-2002 (R&S, No. 20-92389)	In house check: Mar-05																								
Power sensor HP 8481A	MY41092317	18-Oct-02 (Agilent, No. 20021018)	Oct-04																								
Power sensor HP 8481A	US37292783	30-Oct-02 (METAS, No. 252-0236)	Oct-03																								
Power meter EPM E442	GB37480704	30-Oct-02 (METAS, No. 252-0236)	Oct-03																								
Network Analyzer HP 8753E	US37390585	18-Oct-01 (Agilent, No. 24BR1033101)	In house check: Oct 03																								
Calibrated by:	Name Judith Mueller	Function Technician	Signature 																								
Approved by:	Name Katja Pokovic	Function Laboratory Director	Signature 																								
Date issued: August 24, 2003																											
<p>This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid &amp; Partner Engineering AG is completed.</p>																											

 RESEARCH IN MOTION	Document		Page
	Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		19(27)
Author Data	Dates of Test	Test Report No	FCC ID:
Daoud Attayi	June 29 – July 07, 2004	RIM-0094-0407-03	L6AQAP40GW

Schmid & Partner Engineering AG

**s p e e g**

Zeughausstrasse 43, 8004 Zurich, Switzerland  
Phone +41 1 245 9700, Fax +41 1 245 9779  
info@speag.com, http://www.speag.com

# DASY


## Dipole Validation Kit

Type: D1900V2

Serial: 545

Manufactured: November 15, 2001

Calibrated: August 22, 2003

 RESEARCH IN MOTION	Document		Page
	Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		20(27)
Author Data	Dates of Test	Test Report No	FCC ID:
Daoud Attayi	June 29 – July 07, 2004	RIM-0094-0407-03	L6AQAP40GW

## 1. Measurement Conditions

The measurements were performed in the flat section of the SAM twin phantom filled with head simulating solution of the following electrical parameters at 1900 MHz:

Relative Dielectricity	40.2	± 5%
Conductivity	1.46 mho/m	± 5%

The DASY4 System with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 5.2 at 1900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 10mm from dipole center to the solution surface. The included distance spacer was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 7x7x7 fine cube was chosen for cube integration.

The dipole input power (forward power) was 250 mW ± 3 %. The results are normalized to 1W input power.


## 2. SAR Measurement with DASY4 System

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

averaged over 1 cm <sup>3</sup> (1 g) of tissue:	41.2 mW/g ± 16.8 % (k=2) <sup>1</sup>
averaged over 10 cm <sup>3</sup> (10 g) of tissue:	21.3 mW/g ± 16.2 % (k=2) <sup>1</sup>

<sup>1</sup> validation uncertainty



 RESEARCH IN MOTION	Document		Page
	Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		22(27)
Author Data	Dates of Test	Test Report No	FCC ID:
Daoud Attayi	June 29 – July 07, 2004	RIM-0094-0407-03	L6AQAP40GW

Page 1 of 1

Date/Time: 08/22/03 15:40:53

Test Laboratory: SPEAG, Zurich, Switzerland  
File Name: SN545\_SN1507\_HSL1900\_220803.daa

**DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN545**  
**Program: Dipole Calibration**


Communication System: CW-1900; Frequency: 1900 MHz; Duty Cycle: 1:1  
Medium: HSL 1900 MHz ( $\sigma = 1.46 \text{ mho/m}$ ,  $\epsilon_r = 40.17$ ,  $\rho = 1000 \text{ kg/m}^3$ )  
Phantom section: Flat Section  
Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(5.2, 5.2, 5.2); Calibrated: 1/18/2003
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 - SN411; Calibrated: 1/16/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006
- Measurement SW: DASY4, V4.1 Build 47; Postprocessing SW: SEMCAD, V1.6 Build 115

**$P_{in} = 250 \text{ mW}$ ;  $d = 10 \text{ mm}$ /Area Scan (81x81x1); Measurement grid:  $dx=15\text{mm}$ ,  $dy=15\text{mm}$**   
Reference Value = 93.6 V/m  
Power Drift = 0.05 dB  
Maximum value of SAR = 11.5 mW/g

**$P_{in} = 250 \text{ mW}$ ;  $d = 10 \text{ mm}$ /Zoom Scan (7x7x7)/Cube 0; Measurement grid:  $dx=5\text{mm}$ ,  $dy=5\text{mm}$ ,  $dz=5\text{mm}$**   
Peak SAR (extrapolated) = 17.7 W/kg  
SAR(1 g) = 10.3 mW/g; SAR(10 g) = 5.32 mW/g  
Reference Value = 93.6 V/m  
Power Drift = 0.05 dB  
Maximum value of SAR = 11.5 mW/g








Author Data  
**Daoud Attayi**


Dates of Test  
**June 29 – July 07, 2004**

Test Report No  
**RIM-0094-0407-03**


FCC ID:  
**L6AQAP40GW**


545




 <b>RESEARCH IN MOTION</b>	Document <b>Appendices for the BlackBerry 7100t Wireless Handheld  Model No. RAQ40GW test report</b>		Page <b>24(27)</b>
	Author Data <b>Daoud Attayi</b>	Dates of Test <b>June 29 – July 07, 2004</b>	Test Report No <b>RIM-0094-0407-03</b>

APPENDIX E: SAR SET UP PHOTOS

 RESEARCH IN MOTION	Document	Appendices for the BlackBerry 7100t Wireless Handheld Model No. RAQ40GW test report		Page 25(27)
	Author Data <b>Daoud Attayi</b>	Dates of Test <b>June 29 – July 07, 2004</b>	Test Report No <b>RIM-0094-0407-03</b>	FCC ID: <b>L6AQAP40GW</b>



**Figure E1. Left ear configuration**



**Figure E2. Right ear configuration**



**Figure E3. Body worn configuration with holster and headset**