
RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 1(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RTS-0441-0611-06	L6ARBG40	GW

APPENDIX D: PROBE & DIPOLE CALIBRATION DATA

RIM Testing Services Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Author Data Dates of Test Oct. 18 – Nov. 15, 2006 Document Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Test Report No RT—0441-0611-06 L6ARBG40GW



RTS RIM Testing Services	Appendices for the Bla Model RBG41GW SA	•	ndheld	Page 3(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Calibration Laboratory of

Schmid & Pertner
Engineering AG
Zeughausstrasse 43, 8004 Zunich, Switzerland





Schweizerischer Kathrierdiere Service aulase d'étalonnage Servizio avizzaro di taratura Swiss Calibration Service

Accordingtion No.: SCS 108

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the algustories to the EA Multilatural Agreement for the recognition of calibration certificates

Glossary:

TSL tissue simulating fiquid sensitivity in free space conF sensitivity in TSL / NORMx,y,z diode compression point φ rotation around probe axis

Polarization 9 9 rotation around an axis that is in the plane normal to probe axis (at

measurement center), i.e., 9 = 0 is normal to probe axis

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001

Methods Applied and Interpretation of Parameters:

- NORMx,y,z: Assessed for E-field polarization 8 = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not effect the E²-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z * frequency_response (see Frequency Response Chart). This
 linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of
 the frequency response is included in the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a
 flat phantom exposed by a patch antenna.
- Sensor Offset. The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

Certificate No: ET3-1643_Mar06	Page 2 of 9

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 4(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

ET3DV6 SN:1643

March 16, 2006

Probe ET3DV6

SN:1643

Manufactured:

November 7, 2001 March 15, 2005

Last calibrated: Recalibrated:

March 16, 2006

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: ET3-1643_Mar05

Page 3 of 9

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	lheld	Page 5(49)
Author Data	Dates of Test			
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	$\mathbf{G}\mathbf{W}$

ET3DV6 SN:1643

March 16, 2006

DASY - Parameters of Probe: ET3DV6 SN:1643

Sensitivity in Free Space^A

Diode Compression^B

Sensitivity in Tissue Simulating Liquid (Conversion Factors)

Please see Page 8.

Boundary Effect

TSL

900 MHz Typical SAR gradient: 5 % per mm

 Sensor Center to Phantom Surface Distance
 3.7 mm
 4.7 mm

 SAR_{be} [%]
 Without Correction Algorithm
 8.3
 4.4

 SAR_{be} [%]
 With Correction Algorithm
 0.0
 0.2

TSL

1810 MHz Typical SAR gradient: 10 % per mm

 Sensor Center to Phantom Surface Distance
 3.7 mm
 4.7 mm

 SAR_{be} [%]
 Without Correction Algorithm
 7.1
 3.8

 SAR_{be} [%]
 With Correction Algorithm
 0.4
 0.3

Sensor Offset

Probe Tip to Sensor Center

2.7 mm

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.


Certificate No: ET3-1643_Mar06

Page 4 of 9

ⁿ The uncertainties of NormX,Y,Z do not affect the E²-field uncertainty inside TSL (see Page 8).

Numerical linearization parameter: uncertainty not required.

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 6(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW




Uncertainty of Frequency Response of E-field: ± 8.3% (k=2)


Cartificate No: ET3-1643_Mar06

Page 5 of 9

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAI	•	dheld	Page 7(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW



RTS RIM Testing Services	Appendices for the Bla Model RBG41GW SA	•	dheld	Page 8(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARRG40	GW





RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	lheld	Page 9(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

ET3DV6 SN:1643

March 16, 2006

Conversion Factor Assessment






f [MHz]	Avegath [MHS] _C	TSL	Permittivity	Conductivity	Alpha	Depth	CorryF Uncertainty
900	± 50 / ± 100	Head	41.5 ± 5%	0.97 ± 5%	0.59	1.80	6.42 ± 11.0% (k=2)
1610	± 50 / ± 100	Head	$40.0 \pm 5\%$	1.40 ± 5%	0.52	2.47	5.18 ± 11.0% (k=2)
900	± 50 / ± 100	Body	55.0 ± 5%	1.05 ± 5%	0.47	2.12	6.03 ± 11.0% (k=2)
1810	± 50 / ± 100	Body	53.3 ± 5%	1.62 ± 5%	0.52	2.87	4.67 ± 11.0% (k=2)

Certificate No: ET3-1843_Mar06

Page 6 of 9

⁶ The validity of 2 180 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty in the RBS of the ComrF uncertainty at calibration frequency and the uncertainty for the indicated frequency bend.

RTS RIM Testing Services	Appendices for the Blac Model RBG41GW SAR	•	dheld	10(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW



RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 11(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

RIM Testing Services Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Author Data Dates of Test Oct. 18 – Nov. 15, 2006 Document Appendices for the BlackBerry Wireless Handheld Test Report No RT—0441-0611-06 L6ARBG40GW

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurlch, Switzerland





S Schweizerischer Kalibrierdienst
C Service sulsse d'étalonnage

Issued: December 14, 2005

Servizio svizzero di taratura S Swiss Calibration Service

Accreditation No.: SCS 108

Accrecited by the Swiss Federal Office of Motrology and Accrecitation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client RIM Certificate No: EX3-3548_Dec05

CALIBRATION CERTIFICATE Object EX3DV4 - SN:3548 Calibration procedure(s) QA CAL-01.v5and QA CAL-14.v3 Calibration procedure for dosimetric E-field probes Calibration date December 12, 2005 Condition of the calibrated item In Tolerance This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI) The measurements and the uncertainties with confidence probability are given on the following pages and are part of the confidence All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Cal Date (Calibrated by, Certificate No.) Primary Standards Schoduled Calibration GB41293874 Power meter E4419B 3-May-05 (METAS, No. 251-00466) May-06 Power sensor E4412A MY41495277 3-May-95 (METAS, No. 251-00466) May-06 Power sensor E4412A MY41498087 3-May-05 (METAS, No. 251-00465). May-06 Reference 3 dB Attenuator SN: S5054 (3c) 11-Aug-05 (METAS, No. 251-00499). Aug-06 Reference 20 dB Attenuator SN: S5086 (20b) 3 May-05 (METAS, No. 251-00467) May-06 Reference 30 dB Attenuator SN: S5129 (30b) 11-Aug-05 (METAS, No. 251-00500) Aug-06 Reference Probe ES3DV2 SN: 3013 7-Jan-05 (SPEAG, No. ES3-3013_Jan05) Jan-06 SN: 654 27-Out-35 (SPFAG, No. DAE4-654_Oct05) Opt-06 Secondary Standards lip# Check Date (in house) Scheduled Check US3642U01700 RF generator HP 8648C 4-Aug 99 (SPEAG in house check Nov-05) In house check: Nov-97 Network Analyzer HP 8753E US37390585 18-Oct-0* (SPEAG, in house check Nov-05) In house check: Nov 06 Name Calibrated by: Nico Vetterli Laboratory Technician Approved by: Katja Pokovic Technical Manager

Certificate No: EX3-3548_Dec05

Page 1 of 9

This calibration cortificate shall not be reproduced except in full without written approval of the laboratory.

RTS RIM Testing Services	Appendices for the Bla Model RBG41GW SA	•	ndheld	Page 13(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kallbrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
S Swiss Callbration Service

Accreditation No.: SCS 108

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration cartificates

Glossary:

TSL tissue simulating liquid
NORMx,y,z sensitivity in free space
ConF sensitivity in TSL / NORMx,y,z
DCP diode compression point
Polarization φ rotation around probe axis
Polarization 9 protation around an axis that it

n 9 9 rotation around an axis that is in the plane normal to probe axis (at

measurement center), i.e., 9 = 0 is normal to probe axis

Calibration is Performed According to the Following Standards:

a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003

 b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001

Methods Applied and Interpretation of Parameters:

- NORMx,y,z: Assessed for E-field polarization 9 = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not effect the E²-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z * frequency_response (see Frequency Response Chart). This
 linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of
 the frequency response is included in the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

Certificate No: EX3-3548_Dec05	Page 2 of 9	-

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 14(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

EX3DV4 \$N:3548

December 12, 2005

Probe EX3DV4

SN:3548

Manufactured: November 16, 2004
Last calibrated: January 21, 2005
Recalibrated: December 12, 2005

Calibrated for DASY Systems
(Note: non-compatible with DASY2 system!)

Certificate No: EX3-3548_Dec05

Page 3 of 9

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	lheld	Page 15(49)
Author Data	Dates of Test			
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

EX3DV4 SN:3548

December 12, 2005

DASY - Parameters of Probe: EX3DV4 SN:3548

Sensitivity in Free Space^A

Diode Compression^B

Sensitivity in Tissue Simulating Liquid (Conversion Factors)

Please see Page 8.

Boundary Effect

TSL

2450 MHz

Typical SAR gradient: 12 % per mm

Sensor Center to Phantom Surface Distance 2.0 mm 3.0 mm SAR_{be} [%] Without Correction Algorithm $4.2 2.0 \text{SAR}_{be}$ [%] With Correction Algorithm $0.4 0.7 \text{SAR}_{be}$

TSL

5200 MHz

Typical SAR gradient: 26 % per mm

Sensor Center t	to Phantom Surface Distance	2.0 mm	3.0 mm
SAR _{be} [%]	Without Correction Algorithm	2.8	0.6
SAR _{be} [%]	With Correction Algorithm	0.0	0.0

Sensor Offset

Probe Tip to Sensor Center

1.0 mm

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: EX3-3548_Dec05

Page 4 of 9

^A The uncertainties of NormX,Y,Z do not affect the E²-field uncertainty inside TSL (see Page 8).

^B Numerical linearization parameter: uncertainty not required.


RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	lheld	16(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW



December 12, 2005

Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Wavegulde: R22)



Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)

Certificate No: EX3-3548_Dec05


Page 5 of 9


RTS RIM Testing Services	Appendices for the Blac Model RBG41GW SAR	•	dheld	Page 17(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW




December 12, 2005

Receiving Pattern (ϕ), θ = 0°








Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

Certificate No: EX3-3548_Dec05

Page 6 of 9

RTS RIM Testing Services	Appendices for the Blac Model RBG41GW SAI	•	dheld	Page 18(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40)GW





RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	19(49)
Author Data	Dates of Test Test Report No FCC ID:			
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

EX3DV4 SN:3548

December 12, 2005

Conversion Factor Assessment





f [MHz]	Validity [MHz] ^c	TSL.	Permittivity	Conductivity	Alpha	Depth	ConvF Uncertainty
2450	±50/±100	Head	39.2 ± 5%	1.80 ± 5%	0.60	0.77	6.96 ± 11.8% (k=2)
5200	± 50 / ± 100	Head	36.0 ± 5%	4.76 ± 5%	0.56	0.99	5.19 ± 13.1% (k=2)
5500	± 50 / ± 100	Head	35.6 ± 5%	4.96 ± 5%	0.56	0.99	4.86 ± 13.1% (k=2)
5800	± 50 / ± 100	Head	35.3 ± 5%	5.27 ± 5%	0.67	0.82	4.75 ± 13.1% (k=2)
2450	± 50 / ± 100	Body	52.7 ± 5%	1.95 ± 5%	0.86	0.60	6.94 ± 11.8% (k=2)
5200	± 50 / ± 100	Body	49.0 ± 5%	5.30 ± 5%	0.53	1.56	4.94 ± 13.1% (k=2)
5500	± 50 / ± 100	Body	48.6 ± 5%	5.65 ± 5%	0.51	1.60	4.49 ± 13.1% (k=2)
5800	± 50 / ± 100	Body	48.2 ± 5%	6.00 ± 5%	0.53	1,57	4.37 ± 13.1% (k=2)

Certificate No: EX3-3548_Dec05

Page 8 of 9

⁶ The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


RTS RIM Testing Services	Appendices for the Bla Model RBG41GW SA	•	ndheld	Page 20(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW



December 12, 2005

Deviation from Isotropy in HSL

Error (φ, θ), f = 900 MHz



Uncertainty of Spherical Isotropy Assessment: ± 2.6% (k=2)

Certificate No: EX3-3548_Dec05

Page 9 of 9

Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Author Data Dates of Test Oct. 18 – Nov. 15, 2006 Document Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Page 21(49) Page 21(49)

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurlch, Switzerland

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

S Schweizertscher Kalibrierdienst
Service sulsse d'étationnage
Servizio svizzero di taratura
S wiss Calibration Service

Accreditation No.: SCS 108

nultilateral Agreement for the recognition of calibration cert

llent RIM Certificate No: D835V2-446_Jan05

CALIBRATION CERTIFICATE Object D835V2 - SN: 446 **QA CAL-05.v6** Calibration procedure(s) Calibration procedure for dipole validation kits Calibration date: January 7, 2005 Condition of the calibrated item In Tolerance This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3) °C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Scheduled Calibration Cal Date (Calibrated by, Certificate No.) ID# Primary Standards GB37480704 Oct-05 Power meter EPM E442 12-Oct-04 (METAS, No. 251-00412) Oct-05 Power sensor HP 8481A US37292783 12-Oct-04 (METAS, No. 251-00412) Reference 20 dB Attenuator SN: 5086 (20g) 10-Aug-04 (METAS, No 251-00402) Aug-05 10-Aug-04 (METAS, No 251-00402) Aug-05 Reference 10 dB Attenuator SN: 5047.2 (10r) 26-Oct-04 (SPEAG, No. ET3-1507_Oct04) Reference Probe ET3DV6 SN 1507 SN 907 03-May-04 (SPEAG, No. DAE4-907_Mayl04) May-05 DAE4 Scheduled Check Check Date (in house) Secondary Standards MY41092317 18-Oct-02 (SPEAG, in house check Oct-03) In house check: Oct-05 Power sensor HP 8481A In house check: Dec-05 27-Mar-02 (SPEAG, in house check Dec-03) RF generator R&S SML-03 100698 In house check: Nov-05 US37390585 S4206 Oct-01 (SPEAG, in house check Nov-04) Network Analyzer HP 8753E Function Name Calibrated by: Laboratory Technician Approved by: Ketja Pokovic Issued: January 13, 2005 This calibration certificate shall not be reproduced except in full without written approval of the laboratory

Certificate No: D835V2-446_Jan05

Page 1 of 6

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 22(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurlch, Switzerland



S Schwelzerlscher Kallbrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
S Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL tissue simulating liquid
ConvF sensitivity in TSL / NOF

ConvF sensitivity in TSL / NORM x,y,z N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

d) DASY4 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
 of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
 point exactly below the center marking of the flat phantom section, with the arms oriented
 parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
 positioned under the liquid filled phantom. The impedance stated is transformed from the
 measurement at the SMA connector to the feed point. The Return Loss ensures low
 reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point.
 No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

Certificate No: D835V2-446_Jan05	Page 2 of 6	

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 23(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	$\mathbf{G}\mathbf{W}$

Measurement Conditions

DASY system configuration, as far as not given on page 1

DASY Version DASY4		V4.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V4.9	
Distance Dipole Center - TSL	15 mm	with Spacer
Area Scan resolution	dx, dy = 15 mm	
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.2 ± 6 %	0.91 mho/m ± 6 %
Head TSL temperature during test	(22.0 ± 0.2) °C		

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	condition	
SAR measured	250 mW input power	2.27 mW / g
SAR normalized	normalized to 1W	9.08 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	9.10 mW / g ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.48 mW / g
SAR normalized	normalized to 1W	5.92 mW / g
SAR for nominal Head TSL parameters 1	normalized to 1W	5.93 mW / g ± 16.5 % (k=2)

Certificate No: D835V2-446_Jan05

Page 3 of 6

¹ Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 24(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	50.1 Ω - 7.1 jΩ	
Return Loss	- 22.9 dB	

General Antenna Parameters and Design

Electrical Delay (one direction)	1.385 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG	
Manufactured on	October 24, 2001	

Certificate No: D835V2-446_Jan05

Page 4 of 6

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 25(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

DASY4 Validation Report for Head TSL

Date/Time: 01/07/05 15:08:43

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN446

Communication System: CW-835; Frequency: 835 MHz; Duty Cycle: 1:1

Medium: HSL 900 MHz;

Medium parameters used: f = 835 MHz; $\sigma = 0.91$ mho/m; $\varepsilon_r = 42.2$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

• Probe: ET3DV6 - SN1507; ConvF(6.24, 6.24, 6.24); Calibrated: 26.10.2004

Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn907; Calibrated: 03.05.2004

Phantom: Flat Phantom 4.9L; Type: QD000P50AA; Serial: SN:1001;

Measurement SW: DASY4, V4.4 Build 10; Postprocessing SW: SEMCAD, V1.8 Build 133


Pin = 250 mW; d = 15 mm/Area Scan (81x81x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 2.44 mW/g

Pin = 250 mW; d = 15 mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 54.2 V/m; Power Drift = 0.0 dB

Peak SAR (extrapolated) = 3.36 W/kg

SAR(1 g) = 2.27 mW/g; SAR(10 g) = 1.48 mW/gMaximum value of SAR (measured) = 2.45 mW/g




Certificate No: D835V2-446_Jan05

Page 5 of 6

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 26(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Impedance Measurement Plot for Head TSL



Certificate No: D835V2-446_Jan05

Page 6 of 6

RIM Testing Services Author Data Dates of Test Daoud Attayi Document Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Test Report No RT—0441-0611-06 RT—0441-0611-06 L6ARBG40GW

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurlch, Switzerland



S Schweizerischer Kalibrierdienst
C Service sulsse d'étalonnage
Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Client **RIN**

Accreditation No.: SCS 108

Certificate No: D1900V2-545_Jan05

	ERTIFICATE		***
Object	D1900V2 - SN: 5	45	
Calibration procedure(s)	QA CAL-05.v6 Calibration proces	dure for dipole validation kits	
Calibration date:	January 06, 2005	i	
Condition of the calibrated item	In Tolerance		
The measurements and the unce	ertainties with confidence po	onal standards, which realize the physical units of robability are given on the following pages and are ry facility: environment temperature (22 ± 3)°C and	part of the certificate.
Primary Standards	l 1D #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power meter EPM E442	GB37480704	12-Oct-04 (METAS, No. 251-00412)	Oct-05
Power sensor HP 8481A			
Power Sensor ar 040 IA	US37292783	12-Oct-04 (METAS, No. 251-00412)	Oct-05
	US37292783 SN: 5086 (20g)	12-Oct-04 (METAS, No. 251-00412) 10-Aug-04 (METAS, No 251-00402)	Oct-05 Aug-05
Reference 20 dB Attenuator		•	
Reference 20 dB Attenuator Reference 10 dB Attenuator	SN: 5086 (20g)	10-Aug-04 (METAS, No 251-00402)	Aug-05 Aug-05 Oct-05
Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6	SN: 5086 (20g) SN: 5047.2 (10r)	10-Aug-04 (METAS, No 251-00402) 10-Aug-04 (METAS, No 251-00402)	Aug-05 Aug-05
Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4	SN: 5086 (20g) SN: 5047.2 (10r) SN 1507	10-Aug-04 (METAS, No 251-00402) 10-Aug-04 (METAS, No 251-00402) 26-Oct-04 (SPEAG, No. ET3-1507_Oct04)	Aug-05 Aug-05 Oct-05
Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A	SN: 5086 (20g) SN: 5047.2 (10r) SN 1507 SN 907	10-Aug-04 (METAS, No 251-00402) 10-Aug-04 (METAS, No 251-00402) 26-Oct-04 (SPEAG, No. ET3-1507_Oct04) 03-May-04 (SPEAG, No. DAE4-907_May104)	Aug-05 Aug-05 Oct-05 May-05
Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A	SN: 5086 (20g) SN: 5047.2 (10r) SN 1507 SN 907	10-Aug-04 (METAS, No 251-00402) 10-Aug-04 (METAS, No 251-00402) 26-Oct-04 (SPEAG, No. ET3-1507_Oct04) 03-May-04 (SPEAG, No. DAE4-907_May104) Check Date (in house)	Aug-05 Aug-05 Oct-05 May-05 Scheduled Check
Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards	SN: 5086 (209) SN: 5047.2 (10r) SN 1507 SN 907 ID # MY41092317	10-Aug-04 (METAS, No 251-00402) 10-Aug-04 (METAS, No 251-00402) 26-Oct-04 (SPEAG, No. ET3-1507, Oct04) 03-May-04 (SPEAG, No. DAE4-907_May104) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-03)	Aug-05 Aug-05 Oct-05 May-05 Scheduled Check In house check: Oct-05
Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator R&S SML-03	SN: 5086 (209) SN: 5047.2 (10r) SN 1507 SN 907 ID # MY41092317 100698	10-Aug-04 (METAS, No 251-00402) 10-Aug-04 (METAS, No 251-00402) 26-Oct-04 (SPEAG, No. ET3-1507_Oct04) 03-May-04 (SPEAG, No. DAE4-907_Mayl04) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-03) 27-Mar-02 (SPEAG, in house check Dec-03)	Aug-05 Aug-05 Oct-05 May-05 Scheduled Check In house check: Oct-05 In house check: Dec-05
Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator R&S SML-03	SN: 5086 (209) SN: 5047.2 (10r) SN 1507 SN 907 ID # MY41092317 100698 US37390585 S4206	10-Aug-04 (METAS, No 251-00402) 10-Aug-04 (METAS, No 251-00402) 26-Oct-04 (SPEAG, No. ET3-1507, Oct04) 03-May-04 (SPEAG, No. DAE4-907_May104) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-03) 27-Mar-02 (SPEAG, in house check Nov-04) Function Laboratory Technician	Aug-05 Aug-05 Oct-05 May-05 Scheduled Check In house check: Oct-05 In house check: Dec-05 In house check: Nov 05
Reference 20 dB Attenuator Reference 10 dB Attenuator Reference Probe ET3DV6 DAE4 Secondary Standards Power sensor HP 8481A RF generator R&S SML-03 Network Analyzer HP 8753E	SN: 5086 (209) SN: 5047.2 (10r) SN 1507 SN 907 ID # MY41092317 100698 US37390585 S4206 Name	10-Aug-04 (METAS, No 251-00402) 10-Aug-04 (METAS, No 251-00402) 26-Oct-04 (SPEAG, No. ET3-1507, Oct04) 03-May-04 (SPEAG, No. DAE4-907_May104) Check Date (in house) 18-Oct-02 (SPEAG, in house check Oct-03) 27-Mar-02 (SPEAG, in house check Nov-04) Function Laboratory Technician	Aug-05 Aug-05 Oct-05 May-05 Scheduled Check In house check: Oct-05 In house check: Dec-05 In house check: Nov 05

Certificate No: D1900V2-545_Jan05

Page 1 of 6

RIM Testing Services Author Data Dates of Test Daoud Attayi Document Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Test Report No RT—0441-0611-06 L6ARBG40GW

Calibration Laboratory of

Schmid & Partner
Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
Service sulsse d'étalonnage
Servizio svizzero di taratura
S Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL

tissue simulating liquid

ConvF N/A sensitivity in TSL / NORM x,y,z not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

d) DASY4 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
 of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
 point exactly below the center marking of the flat phantom section, with the arms oriented
 parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
 positioned under the liquid filled phantom. The impedance stated is transformed from the
 measurement at the SMA connector to the feed point. The Return Loss ensures low
 reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point.
 No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

Certificate No: D1900V2-545_Jan05	Page 2 of 6	

RTS RIM Testing Services	Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report			
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Measurement Conditions

DASY Version	DASY4	V4.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V4.9	
Distance Dipole Center - TSL	10 mm	with Spacer
Area Scan resolution	dx, dy = 15 mm	
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1900 MHz ± 1 MHz	

Head TSL parameters
The following parameters and calculations were applied.

The following parameters and schodiations were	Temperature	Permittivity	Conductivity	
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m	
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.9 ± 6 %	1.45 mho/m ± 6 %	
Head TSL temperature during test	(22.0 ± 0.2) °C			

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	condition	•
SAR measured	250 mW input power	10.2 mW / g
SAR normalized	normalized to 1W	40.8 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	39.5 mW / g ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	5.34 mW / g
SAR normalized	normalized to 1W	21.4 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	20.7 mW / g ± 16.5 % (k=2)

Certificate No: D1900V2-545_Jan05

Page 3 of 6

¹ Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	lheld	^{Page} 30(49)
Author Data	Dates of Test			
Daoud Attayi	Oct. 18 – Nov. 15, 2006 RT—0441-0611-06 L6ARBG40			GW

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.7 Ω + 2.1 jΩ	
Return Loss	- 31.5 dB	

General Antenna Parameters and Design

Electrical Delay (one direction)	1.198 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG	
Manufactured on	November 15, 2001	

Certificate No: D1900V2-545_Jan05

Page 4 of 6

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	lheld	Page 31(49)
Author Data	Dates of Test Test Report No FCC ID:			
Daoud Attayi	Oct. 18 – Nov. 15, 2006 RT—0441-0611-06 L6ARBG40			GW

DASY4 Validation Report for Head TSL

Date/Time: 01/06/05 18:30:23

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN545

Communication System: CW-1900; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium: HSL 1900 MHz;

Medium parameters used: f = 1900 MHz; $\sigma = 1.45$ mho/m; $\varepsilon_r = 39.6$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

Probe: ET3DV6 - SN1507; ConvF(4.96, 4.96, 4.96); Calibrated: 26.10.2004

• Sensor-Surface: 4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn907; Calibrated: 03.05.2004

Phantom: Flat Phantom quarter size; Type: QD000P50AA; Serial: SN:1001;

Measurement SW: DASY4, V4.4 Build 10; Postprocessing SW: SEMCAD, V1.8 Build 133

Pin = 250 mW; d = 10 mm/Area Scan (81x81x1): Measurement grid: dx=15mm, dy=15mm Maximum value of SAR (interpolated) = 11.6 mW/g


Pin = 250 mW; d = 10 mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm,

dz=5mm

Reference Value = 95.2 V/m; Power Drift = 0.007 dB

Peak SAR (extrapolated) = 18 W/kg

SAR(1 g) = 10.2 mW/g; SAR(10 g) = 5.34 mW/gMaximum value of SAR (measured) = 11.7 mW/g




Certificate No: D1900V2-545 Jan05

Page 5 of 6

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 32(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW


Impedance Measurement Plot for Head TSL



Certificate No: D1900V2-545_Jan05

Page 6 of 6

RIM Testing Services Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Author Data Dates of Test Oct. 18 – Nov. 15, 2006 Document Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Test Report No RT—0441-0611-06 L6ARBG40GW



RIM Testing Services Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Author Data Dates of Test Oct. 18 – Nov. 15, 2006 ATT—0441-0611-06 ATT—0441-0611-06 ATT—0441-0611-06 ATT—0441-0611-06 ATT—0441-0611-06 ATT—0441-0611-06 ATT—0441-0611-06 ATT—0441-0611-06

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurlch, Switzerland





S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL ConvF

N/A

tissue simulating liquid

sensitivity in TSL / NORM x,y,z not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

d) DASY4 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
 of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
 point exactly below the center marking of the flat phantom section, with the arms oriented
 parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
 positioned under the liquid filled phantom. The impedance stated is transformed from the
 measurement at the SMA connector to the feed point. The Return Loss ensures low
 reflected power. No uncertainty required.
- Electrical Delay: One-way detay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

Certificate No: D2450V2-747_Dec05	Page 2 of 6

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	lheld	^{Page} 35(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Measurement Conditions

DASV system configuration, as for as not given on page 1.

DASY system configuration, as far as not	given on page 1.	
DASY Version	DASY4	V4.6
Extrapolation	Advanced Extrapolation	
Phantom	Modular Fiat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Area Scan resolution	dx, dy = 15 mm	
Zoom Scan Resolution	dx, dy = 15 mm dx, dy, dz = 5 mm	

Head TSL parameters

The following parameters and calculations were applied

The following parameters and concentrations were approxi-	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.0 ± 6 %	1.78 mho/m ± 6 %
Head TSL temperature during test	(22.0 ± 0.2) °C	<u> </u>	

SAR result with Head TSL

SAR averaged over 1 cm³ (1 g) of Head TSL	condition	
SAR measured	250 mW input power	13.6 mW / g
SAR normalized	normalized to 1W	54.4 mW / g
SAR for nominal Head TSL parameters 1	normalized to 1W	53.8 mW / g ± 17.0 % (k=2)

SAR averaged over 10 cm² (10 g) of Head TSL	condition	***
SAR measured	250 mW input power	6.29 mW / g
SAR normalized	normalized to 1W	25.2 mW / g
SAR for nominal Head TSL parameters 1	normalized to 1W	24.9 mW / g ± 16.5 % (k=2)

Certificate No: D2450V2-747_Dec05

Page 3 of 6

⁵ Correction to nominal TSL parameters according to d), chapter "SAR Sensitivities"

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 36(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Appendix

Antenna Parameters with Head TSL

impedance, transformed to feed point	51.2 Ω + 3.1 jΩ
Return Loss	-29.8 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.160 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	December 1, 2003

Certificate No: D2450V2-747_Dec05

Page 4 of 6

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 37(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

DASY4 Validation Report for Head TSL

Date/Time: 14.12.2005 17:37:31

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:747

Communication System: CW; Frequency: 2450 MHz; Duty Cycle: 1:1

Medium: HSL U10 BB;

Medium parameters used: f = 2450 MHz; $\sigma = 1.78 \text{ mho/m}$; $\varepsilon_r = 38$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ES3DV2 SN3013; ConvF(4.33, 4.33, 4.33); Calibrated: 07.01.2005
- · Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601;
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA;;
- Measurement SW: DASY4, V4.6 Build 23; Postprocessing SW: SEMCAD, V1.8 Build 160

3013_Pin = 250 mW; d = 10 mm/Area Scan (61x61x1): Measurement grid: dx=10mm, dy=10mm Maximum value of SAR (interpolated) = 16.1 mW/g


3013_Pin = 250 mW; d = 10 mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 92.4 V/m; Power Drift = 0.067 dB

Peak SAR (extrapolated) = 28.0 W/kg

SAR(1 g) = 13.6 mW/g; SAR(10 g) = 6.29 mW/g

Maximum value of SAR (measured) = 15.5 mW/g




Certificate No: D2450V2-747_Dec05

Page 5 of 6

RTS RIM Testing Services	Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report		Page 38(49)	
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Impedance Measurement Plot for Head TSL



Certificate No: D2450V2-747_Dec05

Page 6 of 6

Appendices for the BlackBerry Wireless Handheld 39(49) Model RBG41GW SAR Report **RIM Testing Services** Author Data Dates of Test FCC ID: Test Report No RT-0441-0611-06 L6ARBG40GW Daoud Attayi Oct. 18 – Nov. 15, 2006

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst S Service suisse d'étalonnage C Servizio svizzero di taratura Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

RIM

Certificate No: D5GHzV2-1033_Nov05

Accreditation No.: SCS 108

Certificate No: D5GHzV2-1033_Nov05

CALIBRATION CERTIFICATE Object D5GHzV2 - SN: 1033 QA CAL-22.v1 Calibration procedure(s) Calibration procedure for dipole validation kits between 3-6 GHz Calibration date: November 22, 2005 Condition of the calibrated item In Tolerance This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI) The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)*C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) ID# Cal Date (Calibrated by, Certificate No.) Scheduled Calibration Primary Standards GB41293874 Power meter E4419B 3-May-05 (METAS, No. 251-00466) May-06 Power sensor E4412A MY41495277 3-May-05 (METAS, No. 251-00466) May-06 Reference 20 dB Attenuator SN: S5086 (20b) 3-May-05 (METAS, No. 251-00467) May-08 SN: 5047.2 (10r) 11-Aug-05 (METAS, No 251-00498) Reference 10 dB Attenuator Aug-06 19-Mar-05 (SPEAG, No. Ex3-3503_Mar05) Mar-06 Reference Probe EX3DV4 SN 3503 07-Jan-05 (SPEAG, No. DAE4-601_Jan05) Jan-06 DAE4 SN 601 Check Date (in house) Scheduled Check Secondary Standards MY41093315 10-Aug-03 (SPEAG, in house check Oct-05) In house check: Oct-06 Power sensor HP 8481A 12-Aug-03 (SPEAG, in house check Oct-05) In house check: Oct-06 GB43310788 Power meter E4419B In house check: Dec-05 RF generator R&S SMT-06 100005 4-Aug-99 (SPEAG, in house check Dec-03) Network Analyzer HP 8753E US37390585 S4206 18-Oct-01 (SPEAG, in house check Nov-04) In house check: Nov-05 Function Signature Name Calibrated by: Katja Pokovic Technical Manager Approved by: Fin Bomholt R&D Director Issued: November 23, 2005 This calibration certificate shall not be reproduced except in full without written approval of the laboratory

Page 1 of 11

RIM Testing Services Author Data Daoud Attayi Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Test Report No RT—0441-0611-06 L6ARBG40GW

Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstresse 43, 8004 Zurich, Switzerland





S Schwelzerischer Kalibrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
Swiss Calibration Service

Accreditation No.: SCS 108

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL tissue simulating liquid
ConvF sensitivity in TSL / NORM x,y,z
N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEC Std 62209 Part 2, "Evaluation of Human Exposure to Radio Frequency Fields from Handheld and Body-Mounted Wireless Communication Devices in the Frequency Range of 30 MHz to 6 GHz: Human models, Instrumentation, and Procedures"; Part 2: "Procedure to determine the Specific Absorption Rate (SAR) for including accessories and multiple transmitters", Draft Version 0.9, December 2004
- b) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

c) DASY4 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
 of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
 point exactly below the center marking of the flat phantom section, with the arms oriented
 parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
 positioned under the liquid filled phantom. The impedance stated is transformed from the
 measurement at the SMA connector to the feed point. The Return Loss ensures low
 reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point.
 No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

Certificate No: D5GHzV2-1033_Nov05	Page 2 of 11	

RTS RIM Testing Services	Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report			Page 41(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Measurement Conditions

DASY system configuration, as far as not given on page 1

DASY Version	DASY4	V4.6
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Area Scan resolution	dx, dy = 10 mm	
Zoom Scan Resolution	dx, dy = 4.3 mm, dz = 3 mm	
Frequency	5200 MHz ± 1 MHz 5500 MHz ± 1 MHz 5800 MHz ± 1 MHz	

Head TSL parameters at 5200 MHz The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	36.0	4.66 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.6 ± 6 %	4.53 mho/m ± 6 %
Head TSL temperature during test	(22.0 ± 0.2) °C		

SAR result with Head TSL at 5200 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	condition	
SAR measured	250 mW input power	19.3 mW / g
SAR normalized	normalized to 1W	77.2 mW / g
SAR for nominal Head TSL parameters 1	normalized to 1W	76.9 mW / g ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	5.46 mW / g
SAR normalized	normalized to 1W	21.8 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	21.7 mW / g ± 19.5 % (k=2)

Certificate No: D5GHzV2-1033_Nov05

Page 3 of 11

¹ Correction to nominal TSL parameters according to c), chapter "SAR Sensitivities"

RIM Testing Services Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Author Data Dates of Test Oct. 18 – Nov. 15, 2006 Document Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Test Report No RT—0441-0611-06 L6ARBG40GW

Head TSL parameters at 5500 MHz

The following parameters and calculations were applied

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.6	4.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.1 ± 6 %	4.80 mho/m ± 6 %
Head TSL temperature during test	(22.0 ± 0.2) °C	I	

SAR result with Head TSL at 5500 MHz

SAR averaged over 1 cm ³ (1 g) of Head T\$L	condition	
SAR measured	250 mW input power	20.6 mW / g
SAR normalized	normalized to 1W	82.4 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	82.0 mW / g ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	5.79 mW/g
SAR normalized	normalized to 1W	23.2 mW / g
SAR for nominal Head TSL parameters 1	normalized to 1W	23.0 mW / g ± 19.5 % (k=2)

Head TSL parameters at 5800 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 ℃	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.6 ± 6 %	5.09 mho/m ± 6 %
Head TSL temperature during test	(22.0 ± 0.2) °C	-	

SAR result with Head TSL at 5800 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	condition	
SAR measured	250 mW input power	20.1 mW/g
SAR normalized	normalized to 1W	80.4 mW/g
SAR for nominal Head TSL parameters ¹	normalized to 1W	79.8 mW / g ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	5.58 mW/g
SAR normalized	normalized to 1W	22.3 mW / g
SAR for nominal Head TSL parameters ¹	normalized to 1W	22.1 mW / g ± 19.5 % (k=2)

Certificate No: D5GHzV2-1033_Nov05 Page 4 of 11

RTS RIM Testing Services	Appendices for the Blac Model RBG41GW SAR	•	dheld	Page 43(49)
Author Data	Dates of Test Test Report No FCC ID:			
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Body TSL parameters at 5200 MHz

The following parameters and calculations were applied

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	49.0	5.30 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	48.6 ± 6 %	5.17 mho/m ± 6 %
Body TSL temperature during test	(22.0 ± 0.2) °C		_

SAR result with Body TSL at 5200 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	condition	
SAR measured	250 mW input power	19.0 mW/g
SAR normalized	normalized to 1W	76.0 mW / g
SAR for nominal Body TSL parameters ¹	normalized to 1W	75.8 mW / g ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	250 mW input power	5,35 mW / g
SAR normalized	normalized to 1W	21.4 mW / g
SAR for nominal Body TSL parameters ¹	normalized to 1W	21.3 mW / g ± 19.5 % (k=2)

Certificate No: D5GHzV2-1033_Nov05

Page 5 of 11

RIM Testing Services Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Author Data Dates of Test Oct. 18 – Nov. 15, 2006 Document Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report Test Report No RT—0441-0611-06 L6ARBG40GW

Body TSL parameters at 5500 MHz

The following parameters and calculations were applied

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.6	5.65 mha/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	48.0 ± 6 %	5.55 mho/m ± 6 %
Body TSL temperature during test	(22.0 ± 0.2) °C	_	

SAR result with Body TSL at 5500 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	condition	
SAR measured	250 mW input power	18,7 mW / g
SAR normalized	normalized to 1W	74.8 mW/g
SAR for nominal Body TSL parameters ¹	normalized to 1W	74.5 mW / g ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	250 mW input power	5.24 mW/g
SAR normalized	normalized to 1W	21.0 mW/g
SAR for nominal Body TSL parameters ¹	normalized to 1W	20.9 mW / g ± 19.5 % (k=2)

Body TSL parameters at 5800 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.2	6.00 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	47.4 ± 6 %	5.95 mho/m ± 6 %
Body TSL temperature during test	(22.0 ± 0.2) °C	_	-

SAR result with Body TSL at 5800 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	condition	
SAR measured	250 mW input power	17.6 mW/g
SAR normalized	normalized to 1W	70.4 mW / g
SAR for nominal Body TSL parameters *	normalized to 1W	70.0 mW / g ± 19.9 % (k=2)

SAR averaged over 10 cm³ (10 g) of Body TSL	condition	
SAR measured	250 mW input power	4.91 mW / g
SAR normalized	normalized to 1W	19.6 mW/g
SAR for nominal Body TSL parameters ¹	normalized to 1W	19.5 mW / g ± 19.5 % (k=2)

Certificate No: D5GHzV2-1033_Nov05 Page 6 of 11

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	lheld	Page 45(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Appendix

Antenna Parameters with Head TSL at 5200 MHz

Impedance, transformed to feed point	48.6 Ω - 9.2jΩ	
Return Loss	-20.5 dB	

Antenna Parameters with Head TSL at 5500 MHz

Impedance, transformed to feed point	50.8 Ω - 4.3jΩ	
Return Loss	-27.4 dB	

Antenna Parameters with Head TSL at 5800 MHz

impedance, transformed to feed point	56.7 Ω - 3.2jΩ	
Return Loss	-23.1 dB	

Antenna Parameters with Body TSL at 5200 MHz

Impedance, transformed to feed point	48.6 Ω - 7.3jΩ	
Return Loss	-22.4 dB	

Antenna Parameters with Body TSL at 5500 MHz

Impedance, transformed to feed point	51.2 Ω - 3.4jΩ	
Return Loss	-29.1 dB	

Antenna Parameters with Body TSL at 5800 MHz

Impedance, transformed to feed point	57.9 Ω - 2.5jΩ	
Return Loss	-22.3 dB	

General Antenna Parameters and Design

Electrical Delay (one direction)	1.199 ns

After long term use with 40 W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	July 9, 2004

Certificate No: D5GHzV2-1033_Nov05

Page 7 of 11

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	lheld	Page 46(49)
Author Data	Dates of Test	Test Report No	FCC ID:	~
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

DASY4 Validation Report for Head TSL

Date/Time: 22.11.2005 15:46:55

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 5GHz; Type: D5GHz; Serial: D5GHzV2 - SN:1033

Communication System: CW-5GHz; Frequency: 5200 MHz Frequency: 5500 MHz Frequency: 5800 MHz;

Duty Cycle: 1:1

Medium: HSL 5800 MHz;

Medium parameters used: f = 5200 MHz; $\sigma = 4.53$ mho/m; $\varepsilon_r = 35.6$; $\rho = 1000$ kg/m³ Medium parameters used: f = 5500 MHz; $\sigma = 4.8$ mho/m; $\varepsilon_r = 35.1$; $\rho = 1000$ kg/m³ Medium parameters used: f = 5800 MHz; $\sigma = 4.8$ mho/m; $\varepsilon_r = 35.1$; $\rho = 1000$ kg/m³ Medium parameters used: $\rho = 5800$ MHz; $\rho = 1000$ kg/m³ Medium parameters used: $\rho = 5800$ MHz; $\rho = 1000$ kg/m³ Medium parameters used: $\rho = 5800$ MHz; $\rho = 1000$ kg/m³ Medium parameters used: $\rho = 1000$ kg/m³ Medium

5.09 mho/m; $\varepsilon_r = 34.6$; $\rho = 1000 \text{ kg/m}^3$ Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: EX3DV4 SN3503; ConvF(5.56, 5.56, 5.56)ConvF(5.03, 5.03, 5.03)ConvF(4.95, 4.95, 4.95); Calibrated:
 19.03.2005
- Sensor-Surface: 2mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 07.01.2005
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA
- Measurement SW: DASY4, V4.6 Build 23; Postprocessing SW: SEMCAD, V1.8 Build 160

d=10mm, Pin=250mW, f=5200 MHz/Area Scan (91x91x1): Measurement grid: dx=dy=10mm Maximum value of SAR (interpolated) = 43.6 mW/g

d=10mm, Pin=250mW, f=5200 MHz/Zoom Scan (8x8x8), dist=2mm (8x8x8)/Cube 0:

Measurement grid: dx=4.3mm, dy=4.3mm, dz=3mm Reference Vaiue = 80.8 V/m; Power Drift = 0.107 dB Peak SAR (extrapolated) = 78.5 W/kg SAR(1 g) = 20.4 mW/g; SAR(10 g) = 5.67 mW/g

Maximum value of SAR (measured) = 39.1 mW/g

d=10mm, Pin=250mW, f=5500 MHz/Zoom Scan (8x8x8), dist=2mm (8x8x8)/Cube 0:

Measurement grid: dx=4.3mm, dy=4.3mm, dz=3mm Reference Value = 81.3 V/m; Power Drift = 0.034 dB Peak SAR (extrapolated) = 80.7 W/kg SAR(1 g) = 20.6 mW/g; SAR(10 g) = 5.79 mW/g

SAR(1 g) = 20.6 mW/g; SAR(10 g) = 5.79 mW/gMaximum value of SAR (measured) = 40.1 mW/g

d=10mm, Pin=250mW, f=5800 MHz/Zoom Scan (8x8x8), dist=2mm (8x8x8)/Cube 0:

Measurement grid: dx=4.3mm, dy=4.3mm, dz=3mm Reference Value = 77.7 V/m; Power Drift = 0.070 dB Peak SAR (extrapolated) = 85.1 W/kg


SAR(1 g) = 20.1 mW/g; SAR(10 g) = 5.58 mW/g Maximum value of SAR (measured) = 40.0 mW/g

Certificate No: D5GHzV2-1033_Nov05

Page 8 of 11

RTS RIM Testing Services	Appendices for the Bla Model RBG41GW SA	•	dheld	Page 47(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attavi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40)GW

Impedance Measurement Plot for Head TSL



Certificate No: D5GHzV2-1033_Nov05

Page 9 of 11

RIM Testing Services Author Data Daoud Attavi	Dates of Test					
RTS	Appendices for the BlackBerry Wireless Handheld Model RBG41GW SAR Report					

DASY4 Validation Report for Body TSL

Date/Time: 22.11.2005 20:31:49

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 5GHz; Type: D5GHz; Serial: D5GHzV2 - SN:1033

Communication System: CW-5GHz; Frequency: 5200 MHz Frequency: 5500 MHz Frequency: 5800 MHz;

Duty Cycle: 1:1

Medium: MSL 5800 MHz;

Medium parameters used: f = 5200 MHz; $\sigma = 5.17$ mho/m; $\epsilon_r = 48.6$; $\rho = 1000$ kg/m³ Medium parameters used: f = 5500 MHz; $\sigma = 5.55$ mho/m; $\epsilon_r = 48$; $\rho = 1000$ kg/m³ Medium parameters used: f = 5800 MHz; $\sigma = 5800$ MHz; $\sigma = 5800$ MHz; $\sigma = 6800$ MHz; $\sigma =$

5.95 mho/m; ε_t = 47.4; ρ = 1000 kg/m³ Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: EX3DV4 SN3503; ConvF(5,18, 5.18, 5.18)ConvF(4.69, 4.69, 4.69)ConvF(4.78, 4.78, 4.78); Calibrated: 19.03.2005
- Sensor-Surface: 2mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 07.01.2005
- Phantom: Flat Phantom 5.0 (back); Type: QD000P50AA
- Measurement SW: DASY4, V4.6 Build 23; Postprocessing SW: SEMCAD, V1.8 Build 160

d=10mm, Pin=250mW, f=5200 MHz/Area Scan (91x91x1): Measurement grid: dx=dy=10mm Maximum value of SAR (interpolated) = 42.6 mW/g

. d=10mm, Pin=250mW, f=5200 MHz/Zoom Scan (8x8x8), dist=2mm (8x8x8)/Cube 0:

Measurement grid: dx=4.3mm, dy=4.3mm, dz=3mm

Reference Value = 79.9 V/m; Power Drift = -0.143 dB

Peak SAR (extrapolated) = 65.3 W/kg

SAR(1 g) = 19 mW/g; SAR(10 g) = 5.35 mW/g

Maximum value of SAR (measured) = 35.0 mW/g

d=10mm, Pin=250mW, f=5500 MHz/Zoom Scan (8x8x8), dist=2mm (8x8x8)/Cube 0:

Measurement grid: dx=4.3mm, dy=4.3mm, dz=3mm

Reference Value = 75.6 V/m; Power Drift = -0.066 dB

Peak SAR (extrapolated) = 70.2 W/kg

SAR(1 g) = 18.7 mW/g; SAR(10 g) = 5.24 mW/g

Maximum value of SAR (measured) = 35.3 mW/g

d=10mm, Pin=250mW, f=5800 MHz/Zoom Scan (8x8x8), dist=2mm (8x8x8)/Cube 0:

Measurement grid: dx=4.3mm, dy=4.3mm, dz=3mm

Reference Value = 70.3 V/m; Power Drift = -0.011 dB

Peak SAR (extrapolated) = 71.6 W/kg


SAR(1 g) = 17.6 mW/g; SAR(10 g) = 4.91 mW/g

Maximum value of SAR (measured) = 34.2 mW/g

Certificate No: D5GHzV	2-1033_Nov05	Page 10 of 11		
			•	
	·			

RTS RIM Testing Services	Appendices for the Black Model RBG41GW SAR	•	dheld	Page 49(49)
Author Data	Dates of Test	Test Report No	FCC ID:	
Daoud Attayi	Oct. 18 – Nov. 15, 2006	RT-0441-0611-06	L6ARBG40	GW

Impedance Measurement Plot for Body TSL



Certificate No: D5GHzV2-1033_Nov05

Page 11 of 11