| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smar |                  | 1      | Page 1(23) |
|--------------------------|---|------------------|--------|------------|
| Author Data              | Dates of Test   | Report No        | FCC ID |            |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6AF   | RBU20CW    |

# Annex B: Probe and dipole descriptions and calibration certificates

**B.1** Probe and measurement chain descriptions and specifications

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smart |                  | ı      | Page 2(23) |
|--------------------------|--|------------------|--------|------------|
| Author Data              | Dates of Test  | Report No        | FCC ID |            |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                               | RTS-0943-0801-03 | L6AI   | RBU20CW    |

DASY Dosimetric Assessment System by Schmid & Partner Engineering AG




#### **ER3DV6 ISOTROPIC E-FIELD PROBE FOR GENERAL NEAR-FIELD Applications** MEASUREMENTS Support & Downloads Download Product Flyer (PDF, 192kB) Products • DASY4 Packages Construction One dipole parallel, two dipoles normal to probe axis Built-in shielding against static charges PEEK enclosure material (resistant to organic solvents, e.g., ET3DV6 - Isotropic Dos-Probe glycolether) ES3DV3 - Isotropic Dos-Probe EX3DV4 - Isotropic Dos-Probe ET1DV3 - D-Probe Calibration In air from 100 MHz to 3.0 GHz (absolute accuracy ±6.0%, k=2) Frequency 100 MHz to > 6 GHz; Linearity: ± 0.2 dB (100 MHz to 3 GHz) EUV3 - Universal Vector E-Prol H3DV6 - Isotropic H-Probe HUV4 - Universal Vector H-Probe Directivity ± 0.2 dB in air (rotation around probe axis) ± 0.4 dB in air (rotation normal to probe axis) T1V3 - Temp-Probe DP1 - Dummy-Probe Data Acquisition System Dynamic Range 2 V/m to > 1000 V/m; Linearity: ± 0.2 dB Dimensions Overall length: 330 mm (Tip: 16 mm) Tip diameter: 8 mm (Body: 12 mm) Distance from probe tip to dipole centers: 2.5 mm Validation Kits & Calibration Dipole: General near-field measurements up to 6 GHz Application • Hearing Aid Compatibility (HAC) Ext Field component measurements • Tissue Simulating Liquids Fast automatic scanning in phantoms SPEAG Home

http://www.dasy4.com/er3.htm

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smar |                  | t      | Page 3(23) |
|--------------------------|---|------------------|--------|------------|
| Author Data              | Dates of Test   | Report No        | FCC ID |            |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6AF   | RBU20CW    |

DASY Dosimetric Assessment System by Schmid & Partner Engineering AG





http://www.dasy4.com/h3d.htm


| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smart |                  | t      | Page 4(23) |
|--------------------------|--|------------------|--------|------------|
| Author Data              | Dates of Test  | Report No        | FCC ID |            |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                               | RTS-0943-0801-03 | L6AR   | BU20CW     |

All measurements were performed to the nearest element point as per the C63.19 standard. Offset distances were entered in the DASY4 software so that the measurement was to the nearest element.


Figures 1 and 2, provided by the manufacturer, illustrate detail of the probe tip and its dimensions.

**ER3DV6** E-Field probe: The distances from the probe tip to the closest points on the dipole sensors are 1.45mm for X and Y and 1.25mm for Z. From the probe tip to the center of the sensors is 2.5mm.

**H3DV6** H-Field probe: The distance from the probe tip to the closest point of the X, Y and Z loop sensors is 1.1mm. From the probe tip to the center of the sensor is 3.00mm.



E-Field Probe (ER3DV6)



H-Field Probe (H3DV6)

The following information is from the system manufacturer user manual describing the process chain:

The first step of the evaluation is a linearization of the filtered input signal to account for the compression characteristics of the detector diode. The compensation depends on the input signal, the diode type and the DC-transmission factor from the diode to the evaluation electronics. If the exciting field is pulsed, the crest factor of the signal must be known to correctly compensate for peak power. The formula for each channel can be given as:

$$V_i = U_i + U_i^2 \cdot \frac{cf}{dcp_i}$$
(20.1)

with  $V_i$  = compensated signal of channel i (i = x, y, z)  $U_i$  = input signal of channel i (i = x, y, z) cf = crest factor of exciting field (DASY parameter)  $dcp_i$  = diode compression point (DASY parameter)

From the compensated input signals the primary field data for each channel can be evaluated:

$$\mbox{E} - \mbox{fieldprobes}: \qquad E_i = \sqrt{\frac{V_i}{Norm_i \cdot ConvF}} \label{eq:energy}$$

$${
m H-field probes}$$
 :  $H_i = \sqrt{V_i} \cdot rac{a_{i0} + a_{i1}f + a_{i2}f^2}{f}$ 

with  $V_i$  = compensated signal of channel i (i = x, y, z)  $Norm_i$  = sensor sensitivity of channel i (i = x, y, z)

= sensor sensitivity of channel (1 = x, y,  $\mu V/(V/m)^2$  for E-field Probes

ConvF = sensitivity enhancement in solution

 $a_{ij}$  = sensor sensitivity factors for H-field probes

f = carrier frequency [GHz]

 $E_i$  = electric field strength of channel i in V/m  $H_i$  = magnetic field strength of channel i in A/m

The RSS value of the field components gives the total field strength (Hermitian magnitude):

$$E_{tot} = \sqrt{E_x^2 + E_y^2 + E_z^2}$$
 (20.2)

The measurement / integration time per point is > 500 ms, as per the system manufacturer:

The time response of the field probes has been assessed by exposing the probe to a well-controlled field producing signals larger than HAC E- and H-fields of class M4. The signal response time is evaluated as the time required by the system to reach 90% of the expected final value after an on/off switch of the power source with an integration time of 500 ms and a probe response time of <5 ms. In the current implementation, DASY4 waits longer than 100 ms after having reached the grid point before starting a measurement, i.e., the response time uncertainty is negligible.

If the device under test does not emit a CW signal, the integration time applied to measure the electric field at a specific point may introduce additional uncertainties due to the discretization. The tolerances for the different systems had the worst-case of 2.6%.

| RTS RIM Testing Services | Annex B to Hearing Aid Con<br>Report for BlackBerry® Smar |                  | t      | Page 6(23) |
|--------------------------|---|------------------|--------|------------|
| Author Data              | Dates of Test   | Report No        | FCC ID |            |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6AF   | RBU20CW    |

# **B.2** Probe and dipole calibration certificates

# RTS RIM Testing Services

Annex B to Hearing Aid Compatibility RF Emissions Test Report for BlackBerry® Smartphone Model RBU21CW

Page 7(23)

Author Data

Daoud Attayi

Dates of Test

08-24 Dec, 07 and 07 Jan, 08

Report No RTS-0943-0801-03

FCC ID

L6ARBU20CW

### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client

RIM

Certificate No: ER3-2285\_Mar07

Accreditation No.: SCS 108

| Object  | ER3DV6 - SN:2  | 2285  |   |
|---|--|---|---|
| Calibration procedure(s)  | QA CAL-02.v4<br>Calibration proceevaluations in a  | cedure for E-field probes optimized for<br>air  | r close near field  |
| Calibration date:   | March 12, 2007   |   |   |
| Condition of the calibrated item  | In Tolerance   |   |   |
| All calibrations have been condu  | cted in the closed laborat   | ory facility: environment temporature /22 + 33°C and  | d humidity < 70%  |
| Calibration Equipment used (M&  | TE critical for calibration)   |   |   |
| alibration Equipment used (M& rimary Standards  | TE critical for calibration)   | Cal Date (Calibrated by, Certificate No.)   | Scheduled Calibration   |
| alibration Equipment used (M&<br>rimary Standards<br>ower meter E4419B  | TE critical for calibration)  ID #  GB41293874   | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557)   | Scheduled Calibration<br>Apr-07   |
| alibration Equipment used (M&<br>imary Standards<br>ower meter E4419B<br>ower sensor E4412A   | TE critical for calibration)   | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557)   | Scheduled Calibration<br>Apr-07<br>Apr-07   |
| alibration Equipment used (M&<br>rimary Standards<br>ower meter E4419B<br>ower sensor E4412A<br>ower sensor E4412A  | TE critical for calibration)  ID #  GB41293874  MY41495277   | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557)   | Scheduled Calibration<br>Apr-07<br>Apr-07<br>Apr-07   |
| alibration Equipment used (M&<br>rimary Standards<br>ower meter E4419B<br>ower sensor E4412A<br>ower sensor E4412A<br>eference 3 dB Attenuator  | ID # GB41293874 MY41498087   | Cal Date (Calibrated by, Certificate No.)<br>5-Apr-06 (METAS, No. 251-00557)<br>5-Apr-06 (METAS, No. 251-00557)<br>5-Apr-06 (METAS, No. 251-00557)<br>10-Aug-06 (METAS, No. 217-00592)  | Scheduled Calibration<br>Apr-07<br>Apr-07<br>Apr-07<br>Aug-07   |
| alibration Equipment used (M&<br>rimary Standards<br>ower meter E4419B<br>ower sensor E4412A<br>ower sensor E4412A<br>beference 3 dB Attenuator<br>eference 20 dB Attenuator  | TE critical for calibration)  ID # GB41293874 MY41495277 MY41498087 SN: S5054 (3c)   | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557)   | Scheduled Calibration<br>Apr-07<br>Apr-07<br>Apr-07<br>Aug-07<br>Apr-07   |
| imary Standards wer meter E4419B wer sensor E4412A wer sensor E4412A eference 3 dB Attenuator eference 30 dB Attenuator   | TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 21-00592) 4-Apr-06 (METAS, No. 21-00598) 10-Aug-06 (METAS, No. 217-00593)   | Scheduled Calibration<br>Apr-07<br>Apr-07<br>Apr-07<br>Aug-07   |
| alibration Equipment used (M&<br>rimary Standards<br>ower meter E4419B<br>ower sensor E4412A<br>ower sensor E4412A<br>eference 3 dB Attenuator<br>eference 20 dB Attenuator<br>eference 30 dB Attenuator<br>eference Probe ER3DV6   | TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)   | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-08 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 251-00558)  | Scheduled Calibration<br>Apr-07<br>Apr-07<br>Apr-07<br>Aug-07<br>Aug-07<br>Aug-07   |
| rimary Standards ower meter E4419B ower sensor E4412A ower sensor E4412A eference 3 dB Attenuator eference 20 dB Attenuator eference Probe ER3DV6 AE4 econdary Standards  | TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 2328  SN: 654                                  | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-08 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 21-00592) 4-Apr-06 (METAS, No. 217-00593) 10-Aug-06 (METAS, No. 217-00593) 2-Oct-06 (SPEAG, No. ER3-2328_Oct06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house)   | Scheduled Calibration<br>Apr-07<br>Apr-07<br>Apr-07<br>Aug-07<br>Apr-07<br>Aug-07<br>Oct-07   |
| mary Standards wer meter E4419B wer sensor E4412A wer sensor E4412A ference 3 dB Attenuator ference 30 dB Attenuator ference Probe ER3DV6 E4 condary Standards generator HP 8648C   | TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 2328  SN: 654  ID #  US3642U01700              | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 251-00558) 10-Aug-06 (METAS, No. 271-00593) 2-Oct-06 (SPEAG, No. ER3-2328_Oct06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-99 (SPEAG, in house check Nov-05)                                   | Scheduled Calibration Apr-07 Apr-07 Apr-07 Aug-07 Aug-07 Aug-07 Oct-07 Jun-07 Scheduled Check In house check: Nov-07                        |
| alibration Equipment used (M&<br>imary Standards<br>ower meter E4419B<br>ower sensor E4412A<br>ower sensor E4412A<br>eference 3 dB Attenuator<br>oference 20 dB Attenuator<br>eference 30 dB Attenuator<br>eference Probe ER3DV6<br>AE4   | TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 2328  SN: 654                                  | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-08 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 21-00592) 4-Apr-06 (METAS, No. 217-00593) 10-Aug-06 (METAS, No. 217-00593) 2-Oct-06 (SPEAG, No. ER3-2328_Oct06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house)   | Scheduled Calibration Apr-07 Apr-07 Apr-07 Aug-07 Aug-07 Aug-07 Oct-07 Jun-07 Scheduled Check In house check: Nov-07                        |
| rimary Standards ower meter E4419B ower sensor E4412A ower sensor E4412A ofference 3 dB Attenuator eference 20 dB Attenuator eference Probe ER3DV6 AE4 econdary Standards F generator HP 8648C etwork Analyzer HP 8753E   | TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c) SN: S5086 (20b) SN: S5129 (30b) SN: 2328 SN: 654  ID #  US3642U01700 US37390585  Name | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00558) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00593) 2-Oct-06 (SPEAG, No. ER3-2328_Oct06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-99 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Oct-06)                           | Scheduled Calibration Apr-07 Apr-07 Apr-07 Aug-07 Aug-07 Aug-07 Oct-07 Jun-07 Scheduled Check In house check: Nov-07                        |
| alibration Equipment used (M&<br>rimary Standards<br>ower meter E4419B<br>ower sensor E4412A<br>ower sensor E4412A<br>beference 3 dB Attenuator<br>eference 20 dB Attenuator<br>eference Probe ER3DV6<br>AE4<br>econdary Standards<br>F generator HP 8648C<br>etwork Analyzer HP 8753E  | TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c)  SN: S5086 (20b)  SN: S5129 (30b)  SN: 2328  SN: 654  ID #  US3642U01700  US37390585  | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00593) 2-Oct-06 (SPEAG, No. ER3-2328_Oct06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-99 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Oct-06)                            | Scheduled Calibration Apr-07 Apr-07 Apr-07 Aug-07 Aug-07 Oct-07 Jun-07 Scheduled Check In house check: Nov-07 In house check: Oct-07        |
| Calibration shave been conductal calibration Equipment used (M& Primary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A Reference 3 dB Attenuator Reference 20 dB Attenuator Reference Probe ER3DV6 PAE4 Recondary Standards RF generator HP 8648C Relational Standards RF generator HP 8753E Relational Standards RF generator HP 8753E | TE critical for calibration)  ID #  GB41293874  MY41495277  MY41498087  SN: S5054 (3c) SN: S5086 (20b) SN: S5129 (30b) SN: 2328 SN: 654  ID #  US3642U01700 US37390585  Name | Cal Date (Calibrated by, Certificate No.) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 5-Apr-06 (METAS, No. 251-00557) 10-Aug-06 (METAS, No. 251-00592) 4-Apr-06 (METAS, No. 217-00592) 4-Apr-06 (METAS, No. 217-00593) 2-Oct-06 (SPEAG, No. ER3-2328_Oct06) 21-Jun-06 (SPEAG, No. DAE4-654_Jun06) Check Date (in house) 4-Aug-99 (SPEAG, in house check Nov-05) 18-Oct-01 (SPEAG, in house check Oct-06) Function Technical Manager | Scheduled Calibration Apr-07 Apr-07 Apr-07 Aug-07 Aug-07 Aug-07 Oct-07 Jun-07 Scheduled Check In house check: Nov-07 In house check: Oct-07 |

Certificate No: ER3-2285\_Mar07

Page 1 of 9

#### Page Annex B to Hearing Aid Compatibility RF Emissions Test 8(23) Report for BlackBerry® Smartphone Model RBU21CW **RIM Testing Services** Author Data Report No FCC ID Dates of Test **Daoud Attayi** 08-24 Dec, 07 and 07 Jan, 08 RTS-0943-0801-03 L6ARBU20CW

## Calibration Laboratory of Schmid & Partner

Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst Service suisse d'étalonnage

C Servizio svizzero di taratura

Swiss Calibration Service Accreditation No.: SCS 108

Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

NORMx,y,z DCP

sensitivity in free space

Polarization o

diode compression point φ rotation around probe axis

Polarization 9

9 rotation around an axis that is in the plane normal to probe axis (at

measurement center), i.e.,  $\vartheta = 0$  is normal to probe axis

Connector Angle

information used in DASY system to align probe sensor X to the robot

coordinate system

### Calibration is Performed According to the Following Standards:

a) IEEE Std 1309-1996, "IEEE Standard for calibration of electromagnetic field sensors and probes, excluding antennas, from 9 kHz to 40 GHz", 1996.

### Methods Applied and Interpretation of Parameters:

- NORMx, y, z: Assessed for E-field polarization  $\vartheta = 0$  for XY sensors and  $\vartheta = 90$  for Z sensor (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide).
- NORM(f)x,y,z = NORMx,y,z \* frequency\_response (see Frequency Response Chart).
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency.
- Spherical isotropy (3D deviation from isotropy): in a locally homogeneous field realized using an open waveguide setup.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

Certificate No: ER3-2285 Mar07

Page 2 of 9

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smart |                  | t      | Page 9(23) |
|--------------------------|--|------------------|--------|------------|
| Author Data              | Dates of Test  | Report No        | FCC ID |            |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                               | RTS-0943-0801-03 | L6AR   | BU20CW     |

March 12, 2007

# Probe ER3DV6

SN:2285

Manufactured:

September 20, 2002

Last calibrated:

April 27, 2006

Recalibrated: March 12, 2007

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: ER3-2285\_Mar07

Page 3 of 9

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smart |                  |        | Page<br>10(23) |
|--------------------------|--|------------------|--------|----------------|
| Author Data              | Dates of Test  | Report No        | FCC ID |                |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                               | RTS-0943-0801-03 | L6AR   | RBU20CW        |

ER3DV6 SN:2285 March 12, 2007

### DASY - Parameters of Probe: ER3DV6 SN:2285

| Sensitivity in Free Space [ $\mu V/(V/m)^2$ ] Diode | Compression <sup>A</sup> |
|---|--------------------------|
|---|--------------------------|

 NormX
 1.24 ± 10.1 % (k=2)
 DCP X
 93 mV

 NormY
 1.40 ± 10.1 % (k=2)
 DCP Y
 93 mV

 NormZ
 1.58 ± 10.1 % (k=2)
 DCP Z
 98 mV

Frequency Correction

X 0.0 Y 0.0 Z 0.0

Sensor Offset (Probe Tip to Sensor Center)

X 2.5 mm Y 2.5 mm Z 2.5 mm

Connector Angle -99 °

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: ER3-2285\_Mar07

Page 4 of 9


A numerical linearization parameter: uncertainty not required

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smart |                  |        | Page 11(23) |
|--------------------------|--|------------------|--------|-------------|
| Author Data              | Dates of Test  | Report No        | FCC ID |             |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                               | RTS-0943-0801-03 | L6AF   | RBU20CW     |

March 12, 2007

# Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide R22)



Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)


Certificate No: ER3-2285\_Mar07

Page 5 of 9


| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smar |                  | İ      | Page 12(23) |
|--------------------------|---|------------------|--------|-------------|
| Author Data              | Dates of Test   | Report No        | FCC ID |             |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6AF   | RBU20CW     |

March 12, 2007

# Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$



# Receiving Pattern ( $\phi$ ), $\vartheta = 90^{\circ}$




Certificate No: ER3-2285\_Mar07

Page 6 of 9

| RTS RIM Testing Services | Annex B to Hearing Aid Con<br>Report for BlackBerry® Smar |                  | st     | Page 13(23) |
|--------------------------|---|------------------|--------|-------------|
| Author Data              | Dates of Test   | Report No        | FCC ID |             |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6AF   | RBU20CW     |


March 12, 2007

# Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$



Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

# Receiving Pattern ( $\phi$ ), $\vartheta$ = 90°



Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

Certificate No: ER3-2285\_Mar07

Page 7 of 9


I


| RTS RIM Testing Services | Annex B to Hearing Aid Con<br>Report for BlackBerry® Smar |                  | t      | Page<br>14(23) |
|--------------------------|---|------------------|--------|----------------|
| Author Data              | Dates of Test   | Report No        | FCC ID |                |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6AR   | BU20CW         |

March 12, 2007

# Dynamic Range f(E-field)

(Waveguide R22, f = 1800 MHz)





Uncertainty of Linearity Assessment: ± 0.6% (k=2)


Certificate No: ER3-2285\_Mar07

Page 8 of 9

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smart |                  | t      | Page<br>15(23) |
|--------------------------|--|------------------|--------|----------------|
| Author Data              | Dates of Test  | Report No        | FCC ID |                |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                               | RTS-0943-0801-03 | L6AR   | BU20CW         |

March 12, 2007

# Deviation from Isotropy in Air Error $(\phi, \vartheta)$ , f = 900 MHz



Uncertainty of Spherical Isotropy Assessment: ± 2.6% (k=2)

Certificate No: ER3-2285\_Mar07

Page 9 of 9

# RIM Testing Services Author Data Dates of Test Daoud Attayi Document Annex B to Hearing Aid Compatibility RF Emissions Test Report for BlackBerry® Smartphone Model RBU21CW Page 16(23) Page 16(23) Page 16(23) Page 16(23)

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 108

Certificate No: H3-6105\_Nov07 **CALIBRATION CERTIFICATE** H3DV6 - SN:6105 Object QA CAL-03.v5 Calibration procedure(s) Calibration procedure for H-field probes optimized for close near field evaluations in air November 9, 2007 Calibration date: Condition of the calibrated item In Tolerance This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Cal Date (Calibrated by, Certificate No.) Scheduled Calibration Primary Standards GB41293874 29-Mar-07 (METAS, No. 217-00670) Mar-08 Power meter E4419B MY41495277 29-Mar-07 (METAS, No. 217-00670) Mar-08 Power sensor E4412A Power sensor E4412A MY41498087 29-Mar-07 (METAS, No. 217-00670) Mar-08 Reference 3 dB Attenuator SN: S5054 (3c) 8-Aug-07 (METAS, No. 217-00719) Aug-08 Reference 20 dB Attenuator SN: S5086 (20b) 29-Mar-07 (METAS, No. 217-00671) Mar-08 Reference 30 dB Attenuator SN: S5129 (30b) 8-Aug-07 (METAS, No. 217-00720) Aug-08 Reference Probe H3DV6 SN: 6182 2-Oct-07 (SPEAG, No. H3-6182\_Oct07) Oct-08 DAE4 SN: 654 20-Apr-07 (SPEAG, No. DAE4-654\_Apr07) Apr-08 Secondary Standards ID# Check Date (in house) Scheduled Check 4-Aug-99 (SPEAG, in house check Oct-07) RF generator HP 8648C US3642U01700 In house check: Oct-09 Network Analyzer HP 8753E US37390585 18-Oct-01 (SPEAG, in house check Oct-07) In house check: Oct-08 Function Signature Calibrated by: Katja Pokovic Technical Manager Niels Kuster Quality Manager Approved by: Issued: November 12, 2007 This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: H3-6105\_Nov07

Page 1 of 8

#### Page Annex B to Hearing Aid Compatibility RF Emissions Test 17(23) Report for BlackBerry® Smartphone Model RBU21CW **RIM Testing Services** Author Data Report No FCC ID Dates of Test **Daoud Attayi** 08-24 Dec, 07 and 07 Jan, 08 RTS-0943-0801-03 L6ARBU20CW

### Calibration Laboratory of

Schmid & Partner **Engineering AG** Zeughausstrasse 43, 8004 Zurich, Switzerland





- Schweizerischer Kalibrierdienst S C
  - Service suisse d'étalonnage
- Servizio svizzero di taratura **Swiss Calibration Service**

Accreditation No.: SCS 108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

NORMx,y,z sensitivity in free space DCP diode compression point Polarization φ φ rotation around probe axis

Polarization 9 9 rotation around an axis that is in the plane normal to probe axis (at

measurement center), i.e.,  $\vartheta = 0$  is normal to probe axis

Connector Angle information used in DASY system to align probe sensor X to the robot

coordinate system

Calibration is Performed According to the Following Standards:

a) IEEE Std 1309-2005, "IEEE Standard for calibration of electromagnetic field sensors and probes, excluding antennas, from 9 kHz to 40 GHz", December 2005.

### Methods Applied and Interpretation of Parameters:

- X,Y,Z\_a0a1a2: Assessed for E-field polarization  $\vartheta = 90$  for XY sensors and  $\vartheta = 0$  for Z sensor (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide).
- X,Y,Z(f)\_a0a1a2= X,Y,Z\_a0a1a2\* frequency\_response (see Frequency Response Chart).
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency.
- Spherical isotropy (3D deviation from isotropy): in a locally homogeneous field realized using an open waveguide setup.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the X\_a0a1a2 (no uncertainty required).

Certificate No: H3-6105 Nov07 Page 2 of 8

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smart |                  | 1      | Page 18(23) |
|--------------------------|--|------------------|--------|-------------|
| Author Data              | Dates of Test  | Report No        | FCC ID |             |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                               | RTS-0943-0801-03 | L6AR   | BU20CW      |

H3DV6 SN:6105 November 9, 2007

# Probe H3DV6

SN:6105

Manufactured: January 4, 2002 Last calibrated: November 15, 2006 Recalibrated: November 9, 2007

Calibrated for DASY Systems
(Note: non-compatible with DASY2 system!)

Certificate No: H3-6105\_Nov07 Page 3 of 8

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smar |                  | t      | Page<br>19(23) |
|--------------------------|---|------------------|--------|----------------|
| Author Data              | Dates of Test   | Report No        | FCC ID |                |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6AF   | RBU20CW        |

H3DV6 SN:6105 November 9, 2007

### DASY - Parameters of Probe: H3DV6 SN:6105

Sensitivity in Free Space [A/m /  $\sqrt{(\mu V)}$ ]

a0 a1 a2 X 2.879E-03 7.363E-5 -1.925E-5 ± 5.1 % (k=2) Y 2.588E-03 1.222E-4 -1.861E-5 ± 5.1 % (k=2) Z 2.929E-03 -6.729E-6 -1.558E-5 ± 5.1 % (k=2)

Diode Compression<sup>1</sup>

DCP X **87** mV DCP Y **87** mV DCP Z **87** mV

Sensor Offset (Probe Tip to Sensor Center)

X 3.0 mm Y 3.0 mm Z 3.0 mm

Connector Angle -234 °

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: H3-6105\_Nov07

Page 4 of 8


<sup>&</sup>lt;sup>1</sup> numerical linearization parameter: uncertainty not required

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smart |                  | :      | Page 20(23) |
|--------------------------|--|------------------|--------|-------------|
| Author Data              | Dates of Test  | Report No        | FCC ID |             |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                               | RTS-0943-0801-03 | L6AF   | RBU20CW     |

November 9, 2007

# Frequency Response of H-Field

(TEM-Cell:ifi110, Waveguide R22)



Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)


Certificate No: H3-6105\_Nov07

Page 5 of 8


| RTS RIM Testing Services | Annex B to Hearing Aid Con<br>Report for BlackBerry® Smar |                  | t      | Page 21(23) |
|--------------------------|---|------------------|--------|-------------|
| Author Data              | Dates of Test   | Report No        | FCC ID |             |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6AR   | BU20CW      |

November 9, 2007

# Receiving Pattern ( $\phi$ ), $\theta$ = 90°



Receiving Pattern ( $\phi$ ),  $\vartheta$  = 0°




Certificate No: H3-6105\_Nov07

Page 6 of 8

| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smar |                  | Page 22(23) |
|--------------------------|---|------------------|-------------|
| Author Data              | Dates of Test   | Report No        | FCC ID      |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6ARBU20CW  |


November 9, 2007

# Receiving Pattern ( $\phi$ ), $\vartheta$ = 90°



Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

# Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$



Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)

Certificate No: H3-6105\_Nov07


Page 7 of 8


| RTS RIM Testing Services | Annex B to Hearing Aid Com<br>Report for BlackBerry® Smar |                  | t      | Page 23(23) |
|--------------------------|---|------------------|--------|-------------|
| Author Data              | Dates of Test   | Report No        | FCC ID |             |
| Daoud Attayi             | 08-24 Dec, 07 and 07 Jan, 08                              | RTS-0943-0801-03 | L6AF   | RBU20CW     |

November 9, 2007

# Dynamic Range f(H-field)

(Waveguide R22, f = 1800 MHz)





Uncertainty of Linearity Assessment: ± 0.6% (k=2)

Certificate No: H3-6105\_Nov07

Page 8 of 8