
	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 1(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

APPENDIX D: PROBE & DIPOLE CALIBRATION DATA

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 2(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Calibration Laboratory of Schmid & Partner Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
 Service suisse d'étalonnage
 Servizio svizzero di taratura
 Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client: **RTS (RIM Testing Services)**

Certificate No.: **ES3-3225_Jan13**

CALIBRATION CERTIFICATE

Object: **ES3DV3 - SN:3225**

Calibration procedure(s): **QA CAL-01.v8, QA CAL-23.v4, QA CAL-25.v4
 Calibration procedure for dosimetric E-field probes**

Calibration date: **January 10, 2013**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 2°C and humidity < 70%).

Calibration Equipment used (MATE critical for calibration):


Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter E4419B	084129074	29-Mar-12 (No. 217-01008)	Apr-13
Power sensor E4413A	8711498087	29-Mar-12 (No. 217-01008)	Apr-13
Reference 3-dB Attenuator	SN: 80084 (20)	20-Mar-12 (No. 217-01001)	Apr-13
Reference 20-dB Attenuator	SN: 80086 (20)	27-Mar-12 (No. 217-01029)	Apr-13
Reference 30-dB Attenuator	SN: 80426 (20)	27-Mar-12 (No. 217-01032)	Apr-13
Reference Probe E5387V2	SN: 3013	28-Dec-12 (No. 832-3013, Dec12)	Dec-13
DAT4	SN: 660	20-Jun-12 (No. DAT4-660, Jun12)	Jun-13
Secondary Standards	ID	Check Date (in house)	Scheduled Check
RF generator HP 8463C	US264801700	4-Aug-09 (in house check Apr-11)	In house check: Apr-15
Network Analyzer HP 8710B	US37300401	18-Oct-01 (in house check Oct-12)	In house check: Oct-13

Calibrated by: **Name: Jovan Kostov, Function: Laboratory Technician, Signature: [Signature]**

Approved by: **Name: Rajko Petrovic, Function: Technical Manager, Signature: [Signature]**

Issued: January 14, 2013

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 3(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
S Service suisse de taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

Accreditation No.: **SCS 108**

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:


TSL	issue simulating liquid
NORM _{x,y,z}	sensitivity in free space
ConvF	sensitivity in TSL / NORM _{x,y,z}
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization ψ	ψ rotation around probe axis
Polarization θ	θ rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., $\theta = 0$ is normal to probe axis

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

Methods Applied and Interpretation of Parameters:

- NORM_{x,y,z}: Assessed for E-field polarization $\theta = 0$ ($f < 900$ MHz in TEM-cell; $f > 1800$ MHz: R22 waveguide). NORM_{x,y,z} are only intermediate values, i.e., the uncertainties of NORM_{x,y,z} does not affect the E²-field uncertainty inside TSL (see below ConvF).
- NORM_f(_{x,y,z}) = NORM_{x,y,z} * frequency_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP_{x,y,z}: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A_{k,y,z}; B_{k,y,z}; C_{k,y,z}; D_{k,y,z}; VR_{k,y,z}: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f < 800$ MHz) and inside waveguide using analytical field distributions based on power measurements for $f > 800$ MHz. The same setups are used for assessment of the parameters applied for boundary-compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM_{x,y,z} * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 4(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

ES3DV3 - SN3225

January 10, 2013

Probe ES3DV3

SN:3225

Manufactured: September 1, 2009
Calibrated: January 10, 2013

Calibrated for DASY/EASY Systems
(Note: non-compatible with DASY2 system!)

Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

ES3DV3- SN:3225

January 10, 2013

DASY/EASY - Parameters of Probe: ES3DV3 - SN:3225

Basic Calibration Parameters

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm $(\mu V/(V/m)^2)^{\#}$	1.28	1.18	1.31	$\pm 10.1\%$
DCP (mV) [†]	100.5	101.5	99.9	

Modulation Calibration Parameters

USB	Communication System Name		A dB	B dB/√μV	C	D dB	VR mV	Unc [‡] (k=2)
0	Cir	X	0.0	0.0	1.0	0.00	107.5	$\pm 7.7\%$
		Y	0.0	0.0	1.0		108.4	
		Z	0.0	0.0	1.0		100.0	

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

[#] The uncertainties of NormX,Y,Z do not affect the E² field uncertainty inside TSI. (see Pages 8 and 9).

[†] Numerical linearization parameter; uncertainty not required.

[‡] Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.



Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

ES3DV3- SN:3225

January 10, 2013

DASY/EASY - Parameters of Probe: ES3DV3 - SN:3225

Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) ¹	Relative Permittivity ²	Conductivity (S/m) ²	Coeff X	Coeff Y	Coeff Z	Alpha	Depth (mm)	Unc. (k=2)
750	41.9	0.89	6.56	6.56	6.56	0.42	1.54	± 12.0 %
900	41.5	0.97	6.19	6.19	6.19	0.43	1.52	± 12.0 %
1810	40.0	1.40	5.35	5.35	5.35	0.63	1.39	± 12.0 %
1990	40.0	1.40	5.09	5.09	5.09	0.60	1.23	± 12.0 %
2450	39.2	1.80	4.65	4.65	4.65	0.61	1.63	± 12.0 %
2600	39.0	1.96	4.43	4.43	4.43	0.60	1.32	± 12.0 %

¹ Frequency validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the Coeff uncertainty at calibration frequency and the uncertainty for the indicated frequency band.

² At frequencies below 1 GHz, the validity of tissue parameters (ρ and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 1 GHz, the validity of tissue parameters (ρ and σ) is restricted to ± 5%. The uncertainty is the RSS of the Coeff uncertainty for indicated target tissue parameters.



Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

ES3DV3- SN:3225

January 10, 2013


DASY/EASY - Parameters of Probe: ES3DV3 - SN:3225

Calibration Parameter Determined in Body Tissue Simulating Media

f (MHz) ¹	Relative Permittivity ²	Conductivity (S/m) ²	CorrF X	CorrF Y	CorrF Z	Alpha	Depth (mm)	Unc. (k=2)
750	55.5	0.96	6.27	6.27	6.27	0.48	1.51	± 12.0 %
900	55.0	1.05	6.12	6.12	6.12	0.73	1.25	± 12.0 %
1810	53.3	1.52	5.04	5.04	5.04	0.57	1.47	± 12.0 %
1950	53.3	1.52	4.94	4.94	4.94	0.58	1.50	± 12.0 %
2450	52.7	1.95	4.35	4.35	4.35	0.70	1.16	± 12.0 %
2600	52.5	2.16	4.11	4.11	4.11	0.67	0.99	± 12.0 %

¹ Frequency validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the CorrF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.


² At frequencies below 1 GHz, the validity of tissue parameters (ρ and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 1 GHz, the validity of tissue parameters (ρ and σ) is restricted to ± 5%. The uncertainty is the RSS of the CorrF uncertainty for indicated target tissue parameters.

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 8(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW


ESD-3225- 04-3225

January 10, 2013

Frequency Response of E-Field (TEM-Cell: iR1110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field: $\pm 0.3\%$ ($k=2$)

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 9(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

ES3203-04-3225

January 10, 2013

Receiving Pattern (ϕ), $\theta = 0^\circ$





Uncertainty of Axial Isotropy Assessment: $\pm 0.5\%$ ($k=2$)


ES3043- 043225

January 10, 2013

Dynamic Range f(SAR_{head})
 (TEM cell, f = 900 MHz)




Uncertainty of Linearity Assessment: ± 0.6% (k=2)

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 11(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW



ESD-3225-34-3225

January 10, 2013

Conversion Factor Assessment



Deviation from Isotropy in Liquid Error (δ, θ), f = 900 MHz





Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

ES3DV3 - SN:3225

January 10, 2013

DASY/EASY - Parameters of Probe: ES3DV3 - SN:3225

Other Probe Parameters

Sensor Arrangement	Triangular
Connector Angle (°)	6.3
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	10 mm
Tip Diameter	4 mm
Probe Tip to Sensor X Calibration Point	2 mm
Probe Tip to Sensor Y Calibration Point	2 mm
Probe Tip to Sensor Z Calibration Point	2 mm
Recommended Measurement Distance from Surface	3 mm

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 42, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
 C Service suisse d'étalonnage
 S Servizio svizzero di taratura
 S Swiss Calibration Service


Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **RTS (RIM Testing Services)**

Certificate No: **D750V3-1021_Jan13**

CALIBRATION CERTIFICATE			
Object	D750V3 - SN: 1021		
Calibration procedure(s)	OA CAL-05.v9 Calibration procedure for dipole validation kits above 700 MHz		
Calibration date:	January 07, 2013		
<p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closest laboratory facility: environment temperature (22 ± 3)°C and humidity = 70%.</p> <p>Calibration Equipment used (MPE critical for calibration)</p>			
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter CPM-442A	04827480704	01-Nov-12 (No. 217-01040)	Oct-13
Power sensor HP 8481A	US3250785	01-Nov-12 (No. 217-01040)	Oct-13
Reference 20-dB Attenuator	SN: 3058 (20x)	27-Mar-12 (No. 217-01030)	Apr-13
Type-N mismatch combination	SN: 6047.3 / 06327	27-Mar-12 (No. 217-01030)	Apr-13
Reference Probe E5307x3	SN: 3705	28-Dec-12 (No. E53-3205_Dec12)	Dec-13
DAU4	SN: 601	27-Jun-12 (No. DAU4-601_Jun12)	Jun-13
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	M741002317	18-Oct-09 (in house check Oct-11)	In house check: Oct-13
RF generator RLS (SMT-06)	100005	04-Aug-09 (in house check Oct-11)	In house check: Oct-13
Network Analyzer HP 8713E	US32700585 (4206)	18-Oct-09 (in house check Oct-12)	In house check: Oct-13
Calibrated by:	Name Leif Klyener	Function Laboratory Technician	Signature 
Approved by:	Name Katja Polowik	Function Technical Manager	Signature 
			Issued: January 8, 2013
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 14(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Calibration Laboratory of
 Schmid & Partner
 Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
 S Service suisse d'étalonnage
 C Servizio svizzero di taratura
 S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 108

Glossary:

TSL tissue simulating liquid
 ConvF sensitivity in TSL / NORM x,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- d) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- *Measurement Conditions:* Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- *Antenna Parameters with TSL:* The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- *Feed Point Impedance and Return Loss:* These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- *Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- *SAR measured:* SAR measured at the stated antenna input power.
- *SAR normalized:* SAR as measured, normalized to an input power of 1 W at the antenna connector.
- *SAR for nominal TSL parameters:* The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52 a.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	750 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.9	0.89 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 6 %	0.89 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.12 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	8.48 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.58 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	5.91 W/kg ± 14.5 % (k=2)



Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW	IC 2503A-RFR100LW
-------------------------------------	---	--	-------------------------------	-----------------------------

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	55.7 Ω - 0.2 j Ω
Return Loss	- 25.4 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.033 ns
----------------------------------	----------


After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	December 01, 2010

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 17(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

DASYS Validation Report for Head TSL

Date: 07.01.2013

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 750 MHz; Type: D750V3; Serial: D750V3 - SN: 1021

Communication System: CW; Frequency: 750 MHz

Medium parameters used: $f = 750 \text{ MHz}$; $\sigma = 0.89 \text{ S/m}$; $\epsilon_r = 41.4$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASYS (IEEE/IEC/ANSI C63.19-2007)

DASYS2 Configuration:

- Probe: ES3DV3 - SN:205; CorvF(6.28, 6.28, 6.28); Calibrated: 28.12.2012;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sa601; Calibrated: 27.06.2012
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASYS2 52.8.4(1052); SEMCAD X 14.6.8(7028)

Dipole Calibration for Head Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: $d_x=5\text{mm}$, $d_y=5\text{mm}$, $d_z=5\text{mm}$


Reference Value = 54.107 V/m; Power Drift = 0.01 dB

Peak SAR (extrapolated) = 3.23 W/kg


SAR(1 g) = 2.12 W/kg; SAR(10 g) = 1.38 W/kg

Maximum value of SAR (measured) = 2.47 W/kg



	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 18(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Impedance Measurement Plot for Head TSL



Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW	IC 2503A-RFR100LW
-------------------------------------	---	--	-------------------------------	-----------------------------

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zengenhofstrasse 43, 8004 Zurich, Switzerland



S: Schweizerischer Kalibrierdienst
S: Service suisse d'étalonnage
C: Servizio svizzero di taratura
S: Swiss Calibration Service


Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SGS 108**

Client **RTS (RIM Testing Services)**

Certificate No: **D835V2-4d043_Apr11**

CALIBRATION CERTIFICATE			
Object	D835V2 - SN: 4d043		
Calibration procedure(s)	QA CAL-05.v8 Calibration procedure for dipole validation kits		
Calibration date:	April 07, 2011		
<p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the client laboratory facility, environment temperature (23 ± 3)°C and humidity < 75%.</p> <p>Calibration Equipment used (NISTE official for calibration):</p>			
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter EPM-4424	D837460104	06-Oct-10 (No. 217-01266)	Oct-11
Power sensor HP 8481A	US37292780	06-Oct-10 (No. 217-01266)	Oct-11
Reference 20 dB Attenuator	SN: 3085 (20g)	29-Mar-11 (No. 217-01069)	Apr-12
Type N Network combination	SN: 3041.2 / 00007	29-Mar-11 (No. 217-01071)	Apr-12
Reference Probe ESC040	SN: 3205	30-Apr-10 (No. 233-3205_Apr10)	Apr-11
CalE	SN: 601	16-Jun-10 (No. 042-4-601_Jun10)	Jun-11
Secondary Standards	ID #	Check Date (in hours)	Scheduled Check
Power sensor HP 8481A	MY11020117	16-Oct-09 (in house check Oct-09)	in house check Oct-11
RF generator PXS SMT-05	170005	4-Aug-09 (in house check Oct-09)	in house check Oct-11
Network Analyzer HP 8710B	US37290580 (4206)	16-Oct-09 (in house check Oct-10)	in house check Oct-11
Calibrated by:	Name Jesse Kaelmel	Function Laboratory Technician	Signature   Issued: April 7, 2011
Approved by:	Name Kaja Polovic	Technical Manager	
<p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p>			

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 20(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Bruggstrasse 42, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
 C Service suisse d'étalonnage
 S Servizio svizzero di taratura
 S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL tissue simulating liquid
 ConvF sensitivity in TSL / NORM x,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields: Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW	IC 2503A-RFR100LW
-------------------------------------	---	--	-------------------------------	-----------------------------

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V5.2.6.2
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V4.9	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Size Resolution	dx, dy, dz = 5 mm	
Frequency	935 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.8 ± 6 %	0.90 mho/m ± 6 %
Head TSL temperature during test	(22.0 ± 0.2) °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.33 mW / g
SAR normalized	normalized to 1W	9.32 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	9.43 mW / g ± 17.6 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.52 mW / g
SAR normalized	normalized to 1W	6.08 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	6.14 mW / g ± 16.9 % (k=2)



Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW	IC 2503A-RFR100LW
-------------------------------------	---	--	-------------------------------	-----------------------------

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	52.9 Ω ± 3.4 jΩ
Return Loss	-27.2 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.291 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.


The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. No excessive force must be applied to the dipole arms, because they might bend or the soldered connectors near the feedpoint may be damaged.

Design Modification by End User

The dipole has been modified with Teflon Rings (TR) placed within identified markings close to the end of each dipole arm. Calibration has been performed with TR attached to the dipole.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	April 07, 2006

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 23(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

DASY5 Validation Report for Head TSL

Date/Time: 07.04.2011 09:28:21

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN:4d043

Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1

Medium: HSI900

Medium parameters used: $f = 835 \text{ MHz}$; $\sigma = 0.88 \text{ nho/m}$; $\epsilon_r = 40.6$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63, 19-2007)

DASY5 Configuration:

- Probe: ES3DV3 - SN3205; ConsP6.03, 6.03, 6.03; Calibrated: 30.04.2010
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Set601; Calibrated: 10.06.2010
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- Measurement SW: DASY52, V52A.2 Build (424)
- Postprocessing SW: SEMCAD X, V14.4.4 Build (2829)

Pin=250 mW /d=15mm/Cube 0:


Measurement grid: $d_x=5\text{mm}$, $d_y=5\text{mm}$, $d_z=5\text{mm}$


Reference Value = 57.201 V/m; Power Drift = 0.07 dB

Peak SAR (extrapolated) = 3.504 W/kg


SAR(1g) = 2.33 mW/g; SAR(10g) = 1.52 mW/g


Maximum value of SAR (measured) = 2.730 mW/g



	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 24(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Impedance Measurement Plot for Head T3L



	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 25(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 45, 8004 Zurich, Switzerland



S Schweizerischer Kantonaleinst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAC)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client: **RTS (RIM Testing Services)**

Certificate No: **D835V2-446_Jan13**

CALIBRATION CERTIFICATE

Object: **D835V2 - SN: 446**
 Calibration procedure(s): **QA CAL-05.v9
 Calibration procedure for dipole validation kits above 700 MHz**
 Calibration date: **January 07, 2013**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (20 ± 3)°C and humidity < 70%.

Calibration Equipment used (MATE critical for calibration)


Primary Standards	ID #	Cal. Date / Certificate No.	Scheduled Calibration
Power meter EPM-412A	0820480704	01-Nov-12 (No. 217-01640)	Oct-13
Power sensor HP 8481A	US37292783	01-Nov-12 (No. 217-01640)	Oct-13
Reference 20 dB Attenuator	SN: 5058 (304)	27-Mar-12 (No. 217-01530)	Apr-13
Type-N mismatch combination	SN: 5047.3 / 06327	27-Mar-12 (No. 217-01530)	Apr-13
Reference Probe ESGDV0	SN: 3205	28-Dec-12 (No. 853-3205_Dec12)	Dec-13
DAE4	SN: 601	27-Jun-12 (No. DAE4-601_Jun12)	Jun-13

Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	MY41092317	18-Oct-02 (in house check Oct-11)	In house check: Oct-13
HP generator R&S (SM7-09)	100005	01-Aug-09 (in house check Oct-11)	In house check: Oct-13
Network Analyzer HP 8713E	US37390385 54206	18-Oct-01 (in house check Oct-12)	In house check: Oct-13

Calibrated by:	Name Leif Ryner	Function Laboratory Technician	Signature 
Approved by:	Name Katja Poliwski	Function Technical Manager	Signature 

Issued: **January 8, 2013**

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 26(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL tissue simulating liquid
 ConvF sensitivity in TSL / NORM x,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- d) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- *Measurement Conditions:* Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- *Antenna Parameters with TSL:* The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- *Feed Point Impedance and Return Loss:* These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- *Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- *SAR measured:* SAR measured at the stated antenna input power.
- *SAR normalized:* SAR as measured, normalized to an input power of 1 W at the antenna connector.
- *SAR for nominal TSL parameters:* The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASYS	V92.8.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

Head TSL parameters


The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.0 ± 6 %	0.92 mho/m ± 8 %
Head TSL temperature change during test	< 0.5 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.38 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	9.39 W/kg ± 17.6 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	1.55 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	6.13 W/kg ± 16.4 % (k=2)

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 28(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	50.1 Ω - 6.5 jΩ
Return Loss	- 23.7 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.395 ns
----------------------------------	----------


After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR1 data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	October 24, 2001

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 29(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

DASY5 Validation Report for Head TSL

Date: 07.01.2013

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 446

Communication System: CW; Frequency: 835 MHz

Medium parameters used: $f = 835 \text{ MHz}$; $\sigma = 0.92 \text{ S/m}$; $\epsilon_r = 42$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(6.05, 6.05, 6.05); Calibrated: 28.12.2012;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 27.06.2012
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASY52 52.8.4(1052); SEMCAD X 14.6.8(7028)

Dipole Calibration for Head Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0;


Measurement grid: $d_x=5\text{mm}$, $d_y=5\text{mm}$, $d_z=5\text{mm}$


Reference Value = 56.650 V/m; Power Drift = 0.01 dB

Peak SAR (extrapolated) = 3.61 W/kg


SAR(1 g) = 2.38 W/kg; SAR(10 g) = 1.55 W/kg


Maximum value of SAR (measured) = 2.79 W/kg



	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 30(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Impedance Measurement Plot for Head TSL



	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 31(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**


Class **RTS (RIM Testing Services)**

Certificate No: **D1800V2-2d020_Jan13**

CALIBRATION CERTIFICATE			
Object	D1800V2 - SN: 2d020		
Calibration procedure(s)	QA CAL-05.v9) Calibration procedure for dipole validation kits above 700 MHz		
Calibration date:	January 09, 2013		
<p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.</p> <p>Calibration Equipment used (MATE critical for calibration):</p>			
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter EPM-442A	0837480704	01-Nov-12 (No. 217-01640)	Oct-13
Power sensor HP 8481A	US37262763	01-Nov-12 (No. 217-01640)	Oct-13
Reference 20 dB Attenuator	SN: 5058 (204)	27-Mar-12 (No. 217-01630)	Apr-13
Type-N mismatch combination	SN: 5047 3.1 06327	27-Mar-12 (No. 217-01630)	Apr-13
Reference Probe F330V3	SN: 3025	28-Dec-12 (No. F33-3205_Dec12)	Dec-13
DAE4	SN: 601	27-Jan-12 (No. DAE4-601_Jan12)	Jan-13
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	MFV1062217	18-Oct-09 (in house check Oct-11)	In house check: Oct-13
RF generator RLS SMT-06	100205	04-Aug-09 (in house check Oct-11)	In house check: Oct-13
Network Analyzer HP 8713E	US37390585 04206	18-Oct-01 (in house check Oct-12)	In house check: Oct-13
Calibrated by:	Name Irene El-Masry	Function Laboratory Technician	Signature 
Approved by:	Name Rajja Pokovic	Function Technical Manager	Signature 
Issued: January 9, 2013			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			

Certificate No: D1800V2-2d020_Jan13

Page 1 of 6

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 32(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Calibration Laboratory of
Schmid & Partner
 Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 108

Glossary:

TSL tissue simulating liquid
 ConvF sensitivity in TSL / NORM x,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.8.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1800 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.9 ± 6 %	1.38 mho/m ± 6 %
Head TSL temperature change during test	± 0.5 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	9.61 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	38.5 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	5.06 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	20.3 W/kg ± 16.5 % (k=2)



Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	40.2 (Ω) - 8.3 (jΩ)
Return Loss	- 20.5 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.216 ns
----------------------------------	----------


After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end-caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	September 07, 2001

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 35(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

DASYS Validation Report for Head TSL

Date: 09.01.2013

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 1800 MHz; Type: D1800V2; Serial: D1800V2 - SN: 20020

Communication System: CW; Frequency: 1800 MHz

Medium parameters used: $f = 1800$ MHz; $\sigma = 1.38$ S/m; $\epsilon_r = 38.9$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASYS (IEEE/IEC/ANSI C63.19-2007)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; CorrF(5.04, 5.04, 5.04); Calibrated: 28.12.2012;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DA84 Sn601; Calibrated: 27.06.2012
- Phantom: Flat Phantom 3.0 (front); Type: QD000P50AA; Serial: 1001
- DASY52 52.8.4(1052); SEMCAD X 14.6.8(7028)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm


Reference Value = 95.870 V/m; Power Drift = 0.04 dB

Peak SAR (extrapolated) = 17.5 W/kg


SAR(1 g) = 9.61 W/kg; SAR(10 g) = 5.06 W/kg

Maximum value of SAR (measured) = 11.8 W/kg



	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 36(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Impedance Measurement Plot for Head TSL



Calibration Laboratory of Schmid & Partner Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
 S Service suisse d'étalonnage
 C Servizio svizzero di taratura
 S Swiss Calibration Service


Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **RTS (RIM Testing Services)**

Certificate No: **D1900V2-5d075_Apr11**

CALIBRATION CERTIFICATE			
Object	D1900V2 - SN: 5d075		
Calibration procedure(s)	GA_CAL-05.v8 Calibration procedure for dipole validation kits		
Calibration date	April 5, 2011		
<p>This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 0.1°C and humidity < 70%).</p> <p>Calibration Equipment used (MPE) critical for calibration:</p>			
Primary Standards	ID #	Cal Date (Cert Issue No.)	Scheduled Calibration
Power meter EPM 443A	6257460704	08-Oct-10 (No. 217-01260)	Oct-11
Power sensor HP 8461A	L527292783	08-Oct-10 (No. 217-01260)	Oct-11
Reference 20-dB Attenuator	SN: 5084 (J02)	29-Mar-11 (No. 217-01260)	Apr-12
Type-N mismatch combination	SN: 5047.2 / 08027	29-Mar-11 (No. 217-01271)	Apr-12
Reference Probe ES30V3	SN: 3205	30-Apr-10 (No. 853-3205_Apr10)	Apr-11
SAB4	SN: 601	10-Jun-10 (No. DAB446A_Jun10)	Jun-11
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8461A	M141002317	18-Oct-07 (in house check Oct-09)	In house check Oct-11
Ref generator RAG SMT-60	105005	4-Aug-99 (in house check Oct-09)	In house check Oct-11
Network Analyzer HP 8713C	L527290245-54000	18-Oct-01 (in house check Oct-10)	In house check Oct-11
Calibrated by:	Name Mike Mall	Function Laboratory Technician	Signature 
Approved by:	Name Kolja Pollock	Function Technical Manager	Signature 
Issued: April 5, 2011			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 38(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL tissue simulating liquid
 ConvF sensitivity in TSL / NORM x,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW	IC 2503A-RFR100LW
-------------------------------------	---	--	-------------------------------	-----------------------------

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASYS	V50.0.2
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V1.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1900 MHz ± 1 MHz	

Head TSL parameters


The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 6 %	1.41 mho/m ± 6 %
Head TSL temperature during test	(21.3 ± 0.2) °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	10.2 mW / g
SAR normalized	normalized to 1W	40.8 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	40.4 mW / g ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	5.29 mW / g
SAR normalized	normalized to 1W	21.2 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	21.0 mW / g ± 16.5 % (k=2)

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 40(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	$53.5 \Omega + 6.1 j\Omega$
Return Loss	-23.3 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.197 ns
----------------------------------	----------


After long term use with 100W radiated power, only a slight warping of the dipole near the feedpoint can be measured.

The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	January 24, 2006

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 41(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

DASY5 Validation Report for Head TSL

Date/Time: 05.04.2011 12:41:39

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN:54075

Communication System: CW; Frequency: 1900 MHz; Duty Cycle: 1:1

Medium: HSL U12 BB

Medium parameters used: $f = 1900$ MHz; $\sigma = 1.41$ mho/m; $\epsilon_r = 39$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

- Probe: ESCHAC - SNT200; Cal: HSL09, 1.09, 5.09; Calibrated: 30.04.2010
- Sensor-Surface: 3mm (Mechanical Surface Detectors)
- Electronics: DAB S600; Calibrated: 10.06.2010
- Phantom: Flat Phantom SNT (anti); Type: QD000P95AA; Serial: 1001
- Measurement SW: DASY52, V12.6.2 Build (424)
- Postprocessing SW: SIMCAD-X, V14.4.4 Build (2029)

Head / $d=10$ mm, $P_{in}=250$ mW / Cube 0:


Measurement grid: $d_x=5$ mm, $d_y=5$ mm, $d_z=5$ mm


Reference Value = 97.376 V/m; Power Drift = 0.04 dB

Peak SAR (extrapolated) = 18.796 W/kg


SAR(1 g) = 10.2 mW/kg; SAR(10 g) = 5.29 mW/kg

Maximum value of SAR (measured) = 12.476 mW/kg



	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 42(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Impedance Measurement Plot for Head T&L



Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
C Service suisse d'Etalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the I.A.
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **RTS (RIM Testing Services)**

Certificate No: **D1900V2-545_Jan13**

CALIBRATION CERTIFICATE

Object **D1900V2 - SN: 545**

Calibration procedure(s) **QA CAL-05.v9
 Calibration procedure for dipole validation kits above 700 MHz**

Calibration date: **January 09, 2013**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 0.5)°C and humidity < 70%.

Calibration Equipment used (NMTE critical for calibration)


Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter EPM-40A	0807480704	01-Nov-12 (No. 217-01940)	Oct-13
Power sensor HP 8481A	US37292703	01-Nov-12 (No. 217-01940)	Oct-13
Reference 20-dB Attenuator	SR: 5258 (P04)	27-Mar-12 (No. 217-01530)	Apr-13
Type-N mismatch combination	SR: 5047.3 / 90307	27-Mar-12 (No. 217-01530)	Apr-13
Reference Probe S33DV3	SR: 3095	28-Dec-12 (No. E53-3095_Dec12)	Dec-13
DAE4	SR: 601	27-Jun-12 (No. D464-601_Jun12)	Jun-13
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	HP11082317	18-Oct-07 (in house check Oct-11)	In house check: Oct-13
RF generator R&S SM7-06	100005	04-Aug-09 (in house check Oct-11)	In house check: Oct-13
Network Analyzer HP 8752E	US37380585 54208	18-Oct-07 (in house check Oct-12)	In house check: Oct-13

Calibrated by: **Ismael El-Harouq** (Name) / **Laboratory Technician** (Function) / *Ismael El-Harouq* (Signature)

Approved by: **Kolja Pokovic** (Name) / **Technical Manager** (Function) / *Kolja Pokovic* (Signature)

Issued: January 9, 2013

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 44(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Calibration Laboratory of Schmid & Partner Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst
S Service suisse d'étalonnage
C Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary:

TSL tissue simulating liquid
 ConwF sensitivity in TSL / NORM x,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V02.8.4
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1900 MHz ± 1 MHz	

Head TSL parameters


The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.4 ± 6 %	1.38 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	10.6 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	40.2 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	5.26 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	21.1 W/kg ± 16.5 % (k=2)

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 46(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.0 Ω + 1.7 jΩ
Return Loss	- 34.3 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.198 ns
----------------------------------	----------


After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	November 15, 2001

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 47(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

DASY5 Validation Report for Head TSL

Date: 09.01.2013

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 545

Communication System: CW; Frequency: 1900 MHz

Medium parameters used: $f = 1900 \text{ MHz}$; $\sigma = 1.38 \text{ S/m}$; $\epsilon_r = 39.4$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; CorvP(4.98, 4.98, 4.98); Calibrated: 28.12.2012;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAB4 Sa601; Calibrated: 27.06.2012
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA; Serial: 1001
- DASY52 52.8.4(1052); SEMCAD X 14.6.8(7028)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: $d_x=5\text{mm}$, $d_y=5\text{mm}$, $d_z=5\text{mm}$

Reference Value = 95.493 V/m; Power Drift = 0.05 dB


Peak SAR (extrapolated) = 18.1 W/kg

SAR(1 g) = 10 W/kg; SAR(10 g) = 5.26 W/kg


Maximum value of SAR (measured) = 12.2 W/kg




0 dB = 12.2 W/kg = 10.86 dBW/kg

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 48(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Impedance Measurement Plot for Head TSL



	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 49(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

Calibration Laboratory of Schmid & Partner Engineering AG
Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
S Servizio svizzero di taratura
S Swiss Calibration Service


Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **RTS (RIM Testing Services)**

Certificate No: **D2450V2-791_Apr11**

CALIBRATION CERTIFICATE			
Object	D2450V2 - SN: 791		
Calibration procedure(s)	QA CAL-05.v8 Calibration procedure for dipole validation kits		
Calibration date:	April 5, 2011		
This calibration certificate documents the (traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter EPM442A	GB37480704	06-Oct-10 (No. 217-01286)	Oct-11
Power sensor HP 8481A	US37292783	06-Oct-10 (No. 217-01286)	Oct-11
Reference 20 dB Attenuator	SN: 5086 (20g)	29-Mar-11 (No. 217-01368)	Apr-12
Type-N mismatch combination	SN: 5047.2 / 06327	29-Mar-11 (No. 217-01371)	Apr-12
Reference Probe ES3DV3	SN: 3205	30-Apr-10 (No. ES3-3205_Apr10)	Apr-11
DAE4	SN: 601	10-Jun-10 (No. DAE4-601_Jun10)	Jun-11
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power sensor HP 8481A	MY41092317	18-Oct-02 (in house check Oct-09)	In house check: Oct-11
RF generator R&S SMT-06	100005	4-Aug-99 (in house check Oct-09)	In house check: Oct-11
Network Analyzer HP 8753E	US37390585 S4206	18-Oct-01 (in house check Oct-10)	In house check: Oct-11
Calibrated by:	Name Mike Meli	Function Laboratory Technician	Signature 
Approved by:	Name Katja Pokovic	Technical Manager	
Issued: April 6, 2011			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 50(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

**Calibration Laboratory of
 Schmid & Partner
 Engineering AG**
 Dreuhausstrasse 63, 8004 Zurich, Switzerland



S Schweizerischer Kantonverband
 Service Suisse d'Attestation
C Servizio svizzero di attesta
 Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
 The Swiss Accreditation Service is one of the signatories to the EA
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.1 SCS 108

Glossary:

TSL tissue simulating liquid
 CornF sensitivity in TSL / NORM x,y,z
 N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- Federal Communications Commission Office of Engineering & Technology (FCC OET), "Evaluating Compliance with FCC Guidelines for Human Exposure to Radiofrequency Electromagnetic Fields; Additional Information for Evaluating Compliance of Mobile and Portable Devices with FCC Limits for Human Exposure to Radiofrequency Emissions", Supplement C (Edition 01-01) to Bulletin 65

Additional Documentation:

- GASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW	IC 2503A-RFR100LW
-------------------------------------	---	--	-------------------------------	-----------------------------

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.6.2
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.60 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 6 %	1.72 mho/m ± 6 %
Head TSL temperature during test	(21.6 ± 0.2) °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.3 mW / g
SAR normalized	normalized to 1W	53.2 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	54.1 mW / g ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	6.21 mW / g
SAR normalized	normalized to 1W	24.8 mW / g
SAR for nominal Head TSL parameters	normalized to 1W	25.0 mW / g ± 16.5 % (k=2)



Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

Appendix

Antenna Parameters with Head TSL

Impedance, transformed to feed point	55.5 (± 3.8) jΩ
Return Loss	-24.1 dB

General Antenna Parameters and Design


Electrical Delay (one direction)	1.152 ns
----------------------------------	----------

After long term use with 100mW radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	January 24, 2008

	Document Appendix D for the BlackBerry® Smartphone Model RFR101LW SAR Report			Page 53(55)
	Author Data Andrew Becker	Dates of Test Mar 04 – May 30, 2013	Test Report No RTS-6036-1305-06B	FCC ID: L6ARFR100LW

DASY5 Validation Report for Head TSL

Date/Time: 05.04.2011 15:06:24

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:791

Communication System: CW; Frequency: 2450 MHz; Duty Cycle: 1:1

Medium: HSL U12 BB

Medium parameters used; $f = 2450$ MHz; $\sigma = 1.74$ mho/m; $\epsilon_r = 38.6$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IECIANSI C63.19-2007)

DASY5 Configuration:

- Probe: ES-DW1 - SN3209, C-co-R(4.5), 4.53, 4.53; Calibrated: 30.04.2010
- Sensor (Surface): 3mm (Mechanical Surface Disposition)
- Electronics: DASA S600; Calibrated: 10.06.2010
- Phantom: Flat Phantom 5.0 (Front); Type: QD000FN0AA; Serial: 1001
- Measurement SW: DASY52, v12.A.2 (Build 414)
- Postprocessing SW: SEMCAD X, v14.4.4 (Build 17879)

Head / $d = 10$ mm, Pin=250 mW / Cube 0:


Measurement grid: $d_x = 5$ mm, $d_y = 5$ mm, $d_z = 5$ mm

Reference Value = 102.4 W/m; Power Drift = 0.04 dB

Peak SAR (extrapolated) = 27.237 W/kg

SAR(1 g) = 13.3 mW/g; SAR(10 g) = 6.21 mW/g

Maximum value of SAR (measured) = 16.899 mW/g



Author Data
Andrew Becker


Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

Impedance Measurement Plot for Head T&L



Author Data
Andrew Becker

Dates of Test
Mar 04 – May 30, 2013

Test Report No
RTS-6036-1305-06B

FCC ID:
L6ARFR100LW

IC
2503A-RFR100LW

Impedance Measurement Plot for Head T&L

