

Information about the Applicant	
Grantee (Company Name)	Medtronic Inc.
FCC Contact Person	Charles Farlow
Address	8200 Coral Sea St.
City, State, Zip	Mounds View, MN 55112
Job Number	MDTR0202
Model	Acute Human Monitor
FCC ID	LF519061
Agent	
Approval Type	Original
Equipment Class	DCD Part 15, low power Tx below 1.705 MHz
Rule Part	unlicensed 15.209

Overview

The model 19061 is an Acute Human Monitor used in clinical studies only and not available to the public. The Acute Human Monitor module is a component being used as part of a Clinical Study System in the Electrophysiology Lab. It uses inductive telemetry to communicate with an FCC approved implant, FCC ID: LF5MICSIMPLANT3

Confidentiality requested for allowed files. Request is properly addressed and referenced. Special confidentiality requested and granted by the FCC for internal photos exhibit.

No short term confidentiality requested.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

Signature



Dave Tolman, TCB Committee
5/29/2013