

Information about the Applicant	
Grantee (Company Name)	Medtronic, Inc.
FCC Contact Person	Charles S. Farlow
Address	8200 Coral Seas St. NE
City, State, Zip	Mounds View, MN 55112
Job Number	MDTR0294
Model	MyCareLink Patient Monitor Model 24950
FCC ID	LF524950
Agent	Janet Gavidia
Approval Type	Class II Permissive Change
Equipment Class	DSS Direct Sequence Spread Spectrum
Rule Part	15.247

Overview

Metronic, Inc. is seeking approval of a class II permissive change to FCC ID: LF525950, which contains a Bluetooth radio. The change is to reduce the output power of the Bluetooth radio. A KDB inquiry was submitted regarding the specific changes. The inquiry and FCC response, allowing the change, are documented in the Operational Description exhibit. The change is in accordance with the change request policies KDB 178919.

There were no other changes to the device.

Confidentiality requested for allowed files. Request is properly addressed and referenced.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

Signature

Dave Tolman, TCB Committee

1/23/2014