

Information about the Applicant	
Grantee (Company Name)	Medtronic, Inc.
FCC Contact Person	Charles S. Farlow
Address	8200 Coral Seas St. NE
City, State, Zip	Mounds View, MN 55112
Job Number	MDTR0293
Model	MyCareLink Patient Monitor
FCC ID	LF524955
Agent	Janet Gavidia
Approval Type	Class II Permissive Change
Equipment Class	DCD Part 15, low power Tx below 1.705 MHz
Rule Part	15.209

**Overview**


Metronic, Inc. is seeking approval of a class II permissive change to FCC ID: LF525955, which is a Medical Implant Reader. The class II change is sought in order to reduce component parts count in the inductive transmitter. A KDB inquiry was submitted regarding the specific changes. The inquiry and FCC response, allowing the changes, are documented in the Operational Description exhibit.

Confidentiality requested for allowed files. Request is properly addressed and referenced.

**Recommendation**

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

**Signature**



Dave Tolman, TCB Committee  
1/22/2014