

Information about the Applicant	
<b>Grantee (Company Name)</b>	Medtronic Inc.
<b>FCC Contact Person</b>	Charles S. Farlow
<b>Address</b>	8200 Coral Sea Street N. E.
<b>City, State, Zip</b>	Moundsview, MN 55112
<b>Job Number</b>	MDTR0119
<b>Model</b>	CareLink Encore Programmer Model 29901
<b>FCC ID</b>	LF529901
<b>Agent</b>	Yogendra Shaw, Janet Gavidia, Jacko Chen
<b>Approval Type</b>	Original
<b>Equipment Class</b>	DSS Part 15 Spread Spectrum Transmitter DTS Digital Transmission System NII National Information Infrastructure DCD Low Power Transmitter below 1.75 MHz DXX Low Power Communications Device Transmitter
<b>Rule Part</b>	15.247, 15.407, 15.225, 15.209

## Overview

Medtronic, Inc. is seeking original approval for their CareLink Encore Programmer. The programmer is used to interrogate and program selected Medtronic implantable devices. It contains an 802.11 a/b/g/n, Bluetooth, 13.56 MHz RFID, and a 175 kHz inductive telemetry radios.

Confidentiality requested for allowed files. Request is properly addressed and referenced.

## Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

## Signature

Dave Tolman, TCB Committee  
9/5/2012