Medtronic

Cardiac Rhythm and Heart Failure 8200 Coral Sea Street NE, MVS11 Mounds View, MN 55112 USA www.medtronic.com

tel 763.526.0581 fax 651.367.0603

May 29, 2016

Federal Communications Commission Laboratory Division 7435 Oakland Mills Rd Columbia, MD 21045

Ref: Medtronic, Inc.

Request for processing application for Carelink EncoreTM 26901

FCC ID: LF529901B

I hereby apply to seek original grant for Carelink Encore™ 26901- Programming head for Medtronic and Vitatron devices. The Programmer and the RF head is used to interrogate and program selected Medtronic implantable devices. It contains only inductive radio for near field communication. The Programmer RF head 26901 is compliant to the applicable FCC rule parts of 15.207 and 15.209.

The following report is submitted to support this filing. The results are representative of the RF heads utilized with the Carelink EncoreTM Programmer. The radio for the RF head as tested is identical electrically and mechanically for all units in the field.

Inductive Telemetry	Report	No.	MDTR0	451.2	Model 26901
	(FCC	15.20	07:2016	and	
	15.209:2016)				

If you have any questions or need additional information, please feel free to contact me at (763) 526-0581 or email at yogendra.shah@medtronic.com.

Sincerely,

(Yogendra A. Shah)

Yogend A. Shelm

Principal Telecom/Radio Regulatory Affairs Specialist