

May 31st, 2012

Federal Communications Commission Authorization and Evaluation Division 7435 Oakland Mills Road Columbia, Maryland 21046

Medtronic, Inc. 7000 Central Avenue Northeast Minneapolis, MN 55432

Model No: 3537

Description: Interstim Patient Therapy Manager (PTM3)

FCC ID: LF53537 FRN: 0008250839

Sir/Madam,

This is an application for FCC approval of the Interstim PTM3 (Model 3537) with a Bluetooth transceiver for original certification.

Test data and applicable documents required for certification are included in the accompanying documentation. The test data provided shows the PTM3 meets FCC Part 15.247, 15.209, and 2.1093 requirements. Medtronic, Inc therefore requests FCC certification for the PTM3.

Sincerely,

Anne Coldwell

Director, Quality and Reliability

u Calchall